

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/03/2019 19:37

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 015

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 79960	Download date/time: 02/07/2019	07:43	Record ID ADV8
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012323 Accuracy check date/time: 02/07/2019 07:08

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3065
Check sum 58236	Download date/time: 02/07/2019	07:43	Record ID AE30
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/07/2019 07:12

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 023

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57582	Download date/time: 02/07/2019	07:43	Record ID ADW3
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012323 Accuracy check date/time: 02/07/2019 07:14

First calibration check: .126 Interferent detected: No  
Second calibration check: .128 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.01  
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 023

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 19037	Download date/time: 02/07/2019	07:43	Record ID ADX6
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/14/2019 09:52

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12351	Download date/time: 02/21/2019	10:29	Record ID AF86
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012323 Accuracy check date/time: 02/14/2019 09:56

First calibration check: .124 Interferent detected: No  
Second calibration check: .124 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.02  
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 014

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 31612	Download date/time: 02/21/2019	10:29	Record ID AF91
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/21/2019 09:56

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3077
Check sum	43160	Download date/time:	02/21/2019	10:29		Record ID	AFB1
Record imported (Y/N)				N	Source Diagnostic		

Instrument serial No: 68-012323 Accuracy check date/time: 02/21/2019 10:01

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 010

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	54570	Download date/time:	02/21/2019	10:29		Record ID	AFA8
Record imported (Y/N)				N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/21/2019 10:04

First calibration check: .121 Interferent detected: No  
Second calibration check: .123 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 33.99  
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 010

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 10900	Download date/time: 02/21/2019	10:31	Record ID AFC2
Record imported (Y/N) N		Source Calibration	

---

Instrument serial No: 68-012323 Accuracy check date/time: 02/26/2019 01:35

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 62263	Download date/time: 03/07/2019	09:03	Record ID AFE2
Record imported (Y/N) N		Source Exception	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/28/2019 08:39

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3058	
Check sum	20370	Download date/time:	03/07/2019	09:03		Record ID	AFC2	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 02/28/2019 08:41

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 013

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	18166	Download date/time:	03/07/2019	09:03		Record ID	AF86	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 02/28/2019 08:43

First calibration check: .126 Interferent detected: No  
Second calibration check: .126 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.03  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 013

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51520	Download date/time: 03/07/2019	09:03	Record ID AF91
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.