

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 04:16

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12380	Download date/time: 02/07/2019	07:41	Record ID ADI0
Record imported (Y/N) N	Source	Exception	

---

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 19:35

First calibration check: .STD

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 92730	Download date/time: 02/07/2019	07:40	Record ID AD66
Record imported (Y/N) N	Source	Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 19:35

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 13500	Download date/time: 02/07/2019	07:41	Record ID ADJ1
Record imported (Y/N) N	Source Exception		

---

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 19:37

First calibration check: .STD

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 42563	Download date/time: 02/07/2019	07:40	Record ID AD75
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 19:37

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 35538	Download date/time: 02/07/2019	07:41	Record ID ADK6
Record imported (Y/N) N	Source	Exception	

---

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 19:38

First calibration check: .STD

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 22651	Download date/time: 02/07/2019	07:40	Record ID AD88
Record imported (Y/N) N	Source	Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/03/2019 19:38

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 56972	Download date/time: 02/07/2019	07:41	Record ID ADL7
Record imported (Y/N) N	Source Exception		

Instrument serial No: 68-012335 Accuracy check date/time: 02/07/2019 07:08

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Fail
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 2805
Check sum 44941	Download date/time: 02/07/2019	07:40	Record ID ADF7
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/07/2019 07:10

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 017

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 46172	Download date/time: 02/07/2019	07:40	Record ID AD97
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012335 Accuracy check date/time: 02/07/2019 07:16

First calibration check: .128 Interferent detected: No  
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.02  
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 017

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49574	Download date/time: 02/07/2019	07:45	Record ID AE83
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/14/2019 09:51

First calibration check: .STD Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 011

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 29163	Download date/time: 03/07/2019	09:08	Record ID AFF5
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012335 Accuracy check date/time: 02/14/2019 09:51

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 59370	Download date/time: 03/07/2019	09:08	Record ID AFM8
Record imported (Y/N) N		Source Exception	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/14/2019 09:54

First calibration check: .119 Interferent detected: No  
Second calibration check: .121 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.00  
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 011

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 79910	Download date/time: 03/07/2019	09:08	Record ID AFG6
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012335 Accuracy check date/time: 02/18/2019 00:19

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 46875	Download date/time: 03/07/2019	09:08	Record ID AFN1
Record imported (Y/N) N	Source Exception		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/21/2019 09:55

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2796	
Check sum	52610	Download date/time:	03/07/2019	09:08		Record ID	AFJ7	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012335 Accuracy check date/time: 02/21/2019 10:00

First calibration check: .STD Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 021

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	23923	Download date/time:	03/07/2019	09:08		Record ID	AFH3	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.



**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/21/2019 10:00

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 47333	Download date/time: 03/07/2019	09:08	Record ID AFO2
Record imported (Y/N) N	Source	Exception	

---

Instrument serial No: 68-012335 Accuracy check date/time: 02/21/2019 10:06

First calibration check: .123

Interferent detected: No

Second calibration check: .123

Interferent detected: No

Concentration (%W/V) .129

Solution lot#: IRS013019

Soln temp C: 34.00

Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 021

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 97430	Download date/time: 03/07/2019	09:08	Record ID AFI4
Record imported (Y/N) N	Source	Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012335 Accuracy check date/time: 02/28/2019 08:38

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2736
Check sum	48160	Download date/time:	03/07/2019 09:08			Record ID	AFK8
		Record imported (Y/N)	N	Source	Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.