

LONG BEACH POLICE DEPARTMENT
CSULB PD
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012321
01/17/2019

SIM CONC (%W/V) = .135
SIM TEMP (°C) = 34.00
LOT # = IRS041118
SOLN EXP DATE = 04/11/19
SIM SERIAL # = DR4116
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS = PLACED INTO SERVICE

SUB TESTS SINCE LAST CAL CHECK = 002

TEST	%BAC	TIME
AIR BLANK	.000	10:17
CAL. CHECK	.135	10:17
AIR BLANK	.000	10:17
CAL. CHECK	.136	10:17
AIR BLANK	.000	10:18

LONG BEACH POLICE DEPARTMENT
CSULB PD
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012321
01/24/2019

SIM CONC (%W/V) = .135
SIM TEMP (°C) = 34.01
LOT # = IRS041118
SOLN EXP DATE = 04/11/19
SIM SERIAL # = DR4116
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS =

SUB TESTS SINCE LAST CAL CHECK = 000

TEST	%BAC	TIME
AIR BLANK	.000	14:41
CAL. CHECK	.120	14:41
AIR BLANK	.000	14:41
CAL. CHECK	.125	14:42
AIR BLANK	.000	14:42

TIME	%BAC	TEST
10:15		AIR BLANK
10:15	.498	INTERNAL 5
10:15	.398	INTERNAL 4
10:15	.298	INTERNAL 3
10:15	.100	INTERNAL 2
10:15	.100	INTERNAL 1
10:15	.000	AIR BLANK

ADDITIONAL INFORMATION AND/OR REMARKS

01/17/2019
CA MODEL 5000EN SN 68-012321
INTOXILYZER - ALCOHOL ANALYZER
CSULB PD
LONG BEACH POLICE DEPARTMENT

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

TIME	%BAC	TEST
14:37		AIR BLANK
14:36	.502	INTERNAL 5
14:36	.400	INTERNAL 4
14:36	.300	INTERNAL 3
14:35	.200	INTERNAL 2
14:36	.099	INTERNAL 1
14:36	.000	AIR BLANK

ADDITIONAL INFORMATION AND/OR REMARKS

01/24/2019
CA MODEL 5000EN SN 68-012321
INTOXILYZER - ALCOHOL ANALYZER
CSULB PD
LONG BEACH POLICE DEPARTMENT

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

LONG BEACH POLICE DEPARTMENT
CSULB PD
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012321
01/24/2019

SIM CONC (%W/V) = .135
SIM TEMP (°C) = 34.01
LOT # = IRS041118
SOLN EXP DATE = 04/11/19
SIM SERIAL # = DR4116
NAME/PHOTO # =

G GOSSAGE 7885
REMARKS = LAST CAL CHECK - COLD SIM
HOSE
SUB TESTS SINCE LAST CAL CHECK = 000

TEST	%BAC	TIME
AIR BLANK	.000	14:44
CAL. CHECK	.128	14:44
AIR BLANK	.000	14:45
CAL. CHECK	.129	14:45
AIR BLANK	.000	14:45

LONG BEACH POLICE DEPARTMENT
CSULB PD
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012321
01/31/2019

SIM CONC (%W/V) = .135
SIM TEMP (°C) = 34.00
LOT # = IRS041118
SOLN EXP DATE = 04/11/19
SIM SERIAL # = DR4116
NAME/PHOTO # =

G GOSSAGE 7885
REMARKS =
SUB TESTS SINCE LAST CAL CHECK = 000

TEST	%BAC	TIME
AIR BLANK	.000	10:07
CAL. CHECK	.129	10:07
AIR BLANK	.000	10:07
CAL. CHECK	.129	10:08
AIR BLANK	.000	10:08

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

AIR BLANK

INTERNAL 5

INTERNAL 4

INTERNAL 3

INTERNAL 2

INTERNAL 1

AIR BLANK

TEST

TIME FIRST OBSERVED

ADDITIONAL INFORMATION AND/OR REMARKS

01/31/2019
CA MODEL 5000EN SN 68-012321
INTOXILYZER - ALCOHOL ANALYZER
CSULB PD
LONG BEACH POLICE DEPARTMENT

