

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/01/2019 02:30

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3035
Check sum	31290	Download date/time:	01/03/2019 09:30	Record ID	AC05		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012323 Accuracy check date/time: 01/03/2019 08:52

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3063
Check sum	67400	Download date/time:	01/03/2019 09:30	Record ID	AC10		
Record imported (Y/N) N				Source Diagnostic			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/05/2019 17:13

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 25497	Download date/time: 02/07/2019	07:43	Record ID AE47
Record imported (Y/N) N	Source Exception		

Instrument serial No: 68-012323 Accuracy check date/time: 01/05/2019 17:16

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 20563	Download date/time: 02/07/2019	07:43	Record ID AE58
Record imported (Y/N) N	Source Exception		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/10/2019 09:06

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3059	
Check sum	48479	Download date/time:	02/07/2019	07:43		Record ID	ADY1	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 01/10/2019 09:11

First calibration check: .128
Second calibration check: .129

Interferent detected: No
Interferent detected: No

Concentration (%W/V) .135

Solution lot#: IRS041118

Soln temp C: 34.00
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 011

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	26783	Download date/time:	02/07/2019	07:43		Record ID	ADM2	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/12/2019 01:56

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3049	
Check sum	64236	Download date/time:	02/07/2019	07:43		Record ID	ADZ2	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 01/13/2019 18:47

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	48743	Download date/time:	02/07/2019	07:43		Record ID	AE67	
Record imported (Y/N)				N				Source Exception

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/17/2019 08:22

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3043
Check sum	23440	Download date/time:	02/07/2019	07:43		Record ID	AE09
Record imported (Y/N)				Source Diagnostic			

Instrument serial No: 68-012323 Accuracy check date/time: 01/17/2019 08:25

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 006

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	23237	Download date/time:	02/07/2019	07:43		Record ID	ADN3
Record imported (Y/N)				Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/17/2019 08:28

First calibration check: .124 Interferent detected: No
Second calibration check: .128 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 33.97
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 006

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 17832	Download date/time: 02/07/2019	07:43	Record ID ADO8
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012323 Accuracy check date/time: 01/17/2019 08:31

First calibration check: .130 Interferent detected: No
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.00
Simulator Ser No. DR4116

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 43278	Download date/time: 02/07/2019	07:43	Record ID ADP0
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/24/2019 11:56

First calibration check: Interferent detected:
Second calibration check: Interferent detected:
Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass
Prom Pass Real Time clock Pass Motor Pass Printer Pass
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 3063
Check sum 20974 Download date/time: 02/07/2019 07:43 Record ID AE10
Record imported (Y/N) N Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 01/24/2019 11:57

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No
Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 006

Operator _____

Diagnostics check

Eeprom Serial No. Temp Internal standards
Prom Real Time clock Motor Printer
Ram Auto calibration Range Digital voltmeter
Check sum 17778 Download date/time: 02/07/2019 07:43 Record ID ADQ5
Record imported (Y/N) N Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/24/2019 12:00

First calibration check: .130 Interferent detected: No
Second calibration check: .130 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.05
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 006

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 24704	Download date/time: 02/07/2019	07:43	Record ID ADR6
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012323 Accuracy check date/time: 01/30/2019 15:28

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Ambient air condition

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 56977	Download date/time: 02/07/2019	07:43	Record ID AE78
Record imported (Y/N) N		Source Exception	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/31/2019 09:26

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3045	
Check sum	36442	Download date/time:	02/07/2019	07:43		Record ID	AE29	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 01/31/2019 09:28

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 010

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	33568	Download date/time:	02/07/2019	07:43		Record ID	ADS1	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 01/31/2019 09:29

First calibration check: .124 Interferent detected: No
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.00
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 010

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 28917	Download date/time: 02/07/2019	07:43	Record ID ADT2
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012323 Accuracy check date/time: 01/31/2019 09:36

First calibration check: .130 Interferent detected: No
Second calibration check: .130 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.01
Simulator Ser No. DR4116

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 62400	Download date/time: 02/07/2019	07:43	Record ID ADU7
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.