

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 06/02/2019 19:16

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-----------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 48093 | Download date/time: 06/06/2019 | 08:19 | Record ID AKC8 |
| Record imported (Y/N) N | Source | Exception | |

Instrument serial No: 68-012323 Accuracy check date/time: 06/06/2019 07:41

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|------------|-------------------------|
| Eeprom Pass | Serial No. Pass | Temp Pass | Internal standards Pass |
| Prom Pass | Real Time clock Pass | Motor Pass | Printer Pass |
| Ram Pass | Auto calibration Pass | Range Pass | Digital voltmeter 2985 |
| Check sum 20102 | Download date/time: 06/06/2019 | 08:19 | Record ID AKB5 |
| Record imported (Y/N) N | Source | Diagnostic | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 06/06/2019 07:43

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 012

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 50346 | Download date/time: 06/06/2019 | 08:19 | Record ID AK99 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012323 Accuracy check date/time: 06/06/2019 07:45

First calibration check: .130 Interferent detected: No
Second calibration check: .131 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 012

Operator G GOSSAGE 7885 _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 32927 | Download date/time: 06/06/2019 | 08:19 | Record ID AKA8 |
| Record imported (Y/N) N | Source Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 06/07/2019 21:08

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 53080 | Download date/time: 06/14/2019 | 21:13 | Record ID AKS3 |
| Record imported (Y/N) N | Source Exception | | |

Instrument serial No: 68-012323 Accuracy check date/time: 06/13/2019 08:55

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|------------|-------------------------|
| Eeprom Pass | Serial No. Pass | Temp Pass | Internal standards Pass |
| Prom Pass | Real Time clock Pass | Motor Pass | Printer Pass |
| Ram Pass | Auto calibration Pass | Range Pass | Digital voltmeter 2972 |
| Check sum 45610 | Download date/time: 06/14/2019 | 21:13 | Record ID AKP6 |
| Record imported (Y/N) N | Source Diagnostic | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 06/13/2019 08:59

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 013

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 58067 | Download date/time: 06/14/2019 | 21:13 | Record ID AKN7 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012323 Accuracy check date/time: 06/13/2019 09:04

First calibration check: .131 Interferent detected: No
Second calibration check: .131 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.02
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 013

Operator G GOSSAGE 7885 _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 42128 | Download date/time: 06/14/2019 | 21:13 | Record ID AKO0 |
| Record imported (Y/N) N | Source Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 06/14/2019 02:41

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | |
|-----------|-------|-----------------------|------------------|--------|------------|-----------------------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards |
| Prom | Pass | Real Time clock | | Motor | Pass | Printer |
| Ram | Pass | Auto calibration | Pass | Range | Fail | Digital voltmeter ??? |
| Check sum | 18698 | Download date/time: | 06/14/2019 21:13 | | | Record ID AKQ3 |
| | | Record imported (Y/N) | N | Source | Diagnostic | |

Instrument serial No: 68-012323 Accuracy check date/time: 06/14/2019 19:15

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | |
|-----------|-------|-----------------------|------------------|--------|------------|-----------------------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards |
| Prom | Pass | Real Time clock | | Motor | Pass | Printer |
| Ram | Pass | Auto calibration | Pass | Range | Fail | Digital voltmeter ??? |
| Check sum | 54724 | Download date/time: | 06/14/2019 21:13 | | | Record ID AKR6 |
| | | Record imported (Y/N) | N | Source | Diagnostic | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 06/14/2019 20:16

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

TAKEN OUT OF SERVICE - SOURCE OUT

Number of people tested since last periodic determination of accuracy:

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 22357 | Download date/time: 06/14/2019 | 21:13 | Record ID AKU3 |
| Record imported (Y/N) N | | Source Memo | |

Instrument serial No: 68-012323 Accuracy check date/time: 06/14/2019 20:17

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 53669 | Download date/time: 06/14/2019 | 21:13 | Record ID AKT6 |
| Record imported (Y/N) N | | Source Exception | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.