

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 06/06/2019 09:56

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3442	
Check sum	32630	Download date/time:	06/06/2019	10:30		Record ID	AKM0	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012333 Accuracy check date/time: 06/06/2019 09:57

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .129

Solution lot#: IRS013019

Soln temp C: 34.01

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	25854	Download date/time:	06/06/2019	10:30		Record ID	AKK0	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 06/06/2019 10:00

First calibration check: .133 Interferent detected: No
Second calibration check: .134 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 33899	Download date/time: 06/06/2019	10:30	Record ID AKL7
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012333 Accuracy check date/time: 06/13/2019 10:05

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3410
Check sum 47936	Download date/time: 07/03/2019	11:13	Record ID AMK6
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 06/13/2019 10:06

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 10831	Download date/time: 07/03/2019	11:13	Record ID AMCO
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012333 Accuracy check date/time: 06/13/2019 10:08

First calibration check: .132 Interferent detected: No
Second calibration check: .135 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 27570	Download date/time: 07/03/2019	11:13	Record ID AMD9
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 06/20/2019 12:00

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3408	
Check sum	36259	Download date/time:	07/03/2019	11:13		Record ID	AML5	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012333 Accuracy check date/time: 06/20/2019 12:01

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .129

Solution lot#: IRS013019

Soln temp C: 34.01

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	55112	Download date/time:	07/03/2019	11:13		Record ID	AME6	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 06/20/2019 12:03

First calibration check: .131 Interferent detected: No
Second calibration check: .132 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 41641	Download date/time: 07/03/2019	11:13	Record ID AMF1
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012333 Accuracy check date/time: 06/27/2019 13:43

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3391
Check sum 50213	Download date/time: 07/03/2019	11:13	Record ID AMM2
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 06/27/2019 13:44

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 29851	Download date/time: 07/03/2019	11:13	Record ID AMG8
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012333 Accuracy check date/time: 06/27/2019 13:46

First calibration check: .134 Interferent detected: No
Second calibration check: .133 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57365	Download date/time: 07/03/2019	11:13	Record ID AMH7
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.