

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/02/2019 08:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3036	
Check sum	23618	Download date/time:	05/10/2019	10:43		Record ID	AIL7	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 05/02/2019 09:00

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	16189	Download date/time:	05/10/2019	10:43		Record ID	AID5	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/02/2019 09:02

First calibration check: .128 Interferent detected: No
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 014

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 54350	Download date/time: 05/10/2019	10:43	Record ID AIE2
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012323 Accuracy check date/time: 05/09/2019 08:40

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3036
Check sum 85860	Download date/time: 05/10/2019	10:43	Record ID AIM4
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/09/2019 08:42

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 015

Operator _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 22200	Download date/time: 05/10/2019	10:43	Record ID AIF5
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012323 Accuracy check date/time: 05/09/2019 08:45

First calibration check: .128 Interferent detected: No
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 015

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 59427	Download date/time: 05/10/2019	10:43	Record ID AIG2
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/16/2019 08:31

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 47053	Download date/time: 05/16/2019	09:12	Record ID AIQ7
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012323 Accuracy check date/time: 05/16/2019 08:35

First calibration check: .130 Interferent detected: No
Second calibration check: .130 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 014

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 19400	Download date/time: 05/16/2019	09:12	Record ID AIR4
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/16/2019 08:45

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3008
Check sum	72710	Download date/time:	05/30/2019	09:30		Record ID	AJR9
Record imported (Y/N)				Source Diagnostic			

Instrument serial No: 68-012323 Accuracy check date/time: 05/17/2019 22:19

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	60170	Download date/time:	05/30/2019	09:30		Record ID	AJU6
Record imported (Y/N)				Source Exception			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/18/2019 01:18

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 20944	Download date/time: 05/30/2019	09:30	Record ID AJV9
Record imported (Y/N) N	Source Exception		

Instrument serial No: 68-012323 Accuracy check date/time: 05/23/2019 08:16

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No. Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2987
Check sum 26757	Download date/time: 05/30/2019	09:30	Record ID	AJS2			
Record imported (Y/N) N	Source Diagnostic						

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/23/2019 08:22

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 032

Operator _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 64940	Download date/time: 05/30/2019	09:30	Record ID AJN0
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012323 Accuracy check date/time: 05/23/2019 08:25

First calibration check: .129 Interferent detected: No
Second calibration check: .130 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.03
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 032

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 54074	Download date/time: 05/30/2019	09:30	Record ID AJO7
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/30/2019 08:43

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2983
Check sum	56570	Download date/time:	05/30/2019	09:30		Record ID	AJT3
Record imported (Y/N)				N	Source Diagnostic		

Instrument serial No: 68-012323 Accuracy check date/time: 05/30/2019 08:49

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 022

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	57017	Download date/time:	05/30/2019	09:30		Record ID	AJP3
Record imported (Y/N)				N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 05/30/2019 08:51

First calibration check: .127 Interferent detected: No
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.02
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 022

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57240	Download date/time: 05/30/2019	09:30	Record ID AJQ6
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.