

LONG BEACH POLICE DEPARTMENT
BOOKING
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012335
05/02/2019

SIM CONC (%W/V) = .129
SIM TEMP (°C) = 34.03
LOT # = IRS013019
SOLN EXP DATE = 01/30/20
SIM SERIAL # = MP5021
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS =

SUB TESTS SINCE LAST CAL CHECK = 011

TEST	%BAC	TIME
AIR BLANK	.000	09:01
CAL. CHECK	.128	09:02
AIR BLANK	.000	09:02
CAL. CHECK	.128	09:02
AIR BLANK	.000	09:03

LONG BEACH POLICE DEPARTMENT
BOOKING
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012335
05/09/2019

SIM CONC (%W/V) = .129
SIM TEMP (°C) = ~~34.03~~ 34.00c
LOT # = IRS013019
SOLN EXP DATE = 01/30/20
SIM SERIAL # = MP5021
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS =

SUB TESTS SINCE LAST CAL CHECK = 016

TEST	%BAC	TIME
AIR BLANK	.000	08:44
CAL. CHECK	.127	08:44
AIR BLANK	.000	08:44
CAL. CHECK	.128	08:45
AIR BLANK	.000	08:45

SUBJECT'S NAME

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

TIME FIRST OBSERVED

INSTRUMENT LOCATION

OPERATOR

OPERATOR

ADDITIONAL INFORMATION AND / OR REMARKS

ADDITIONAL INFORMATION AND / OR REMARKS

INTERNAL STD OUT OF TOLERANCE
INVALID TEST
DATE 05/02/2019
SLAVE 75 2242
SN 68-012335

INTERNAL STD OUT OF TOLERANCE
INVALID TEST
DATE 05/09/2019
SLAVE 75 2242
SN 68-012335

INTOXILYZER® INSTRUMENT PRINTER CARD



INTOXILYZER® INSTRUMENT PRINTER CARD



LONG BEACH POLICE DEPARTMENT
BOOKING
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012335
05/16/2019

SIM CONC (%W/V) = .129
SIM TEMP (°C) = 34.00
LOT # = IRS013019
SOLN EXP DATE = 01/30/20
SIM SERIAL # = MP5021
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS =

SUB TESTS SINCE LAST CAL CHECK = 011

TEST	%BAC	TIME
AIR BLANK	.000	08:37
CAL. CHECK	.129	08:37
AIR BLANK	.000	08:38
CAL. CHECK	.130	08:38
AIR BLANK	.000	08:38

LONG BEACH POLICE DEPARTMENT
BOOKING
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012335
05/23/2019

SIM CONC (%W/V) = .129
SIM TEMP (°C) = 34.02
LOT # = IRS013019
SOLN EXP DATE = 01/30/20
SIM SERIAL # = MP5021
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS =

SUB TESTS SINCE LAST CAL CHECK = 000

TEST	%BAC	TIME
AIR BLANK	.000	08:26
CAL. CHECK	.130	08:27
AIR BLANK	.000	08:27
CAL. CHECK	.132	08:27
AIR BLANK	.000	08:28

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

100 188 280 282 423

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

INTERNAL STD OUT OF TOLERANCE
INVALID TEST
DATE 05/16/2019
TIME 08:30
SN 68-012335
SLAVE 75 2242
VERSION 02106.12

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

*INTERNAL 5
INTERNAL 4
INTERNAL 3
INTERNAL 2
INTERNAL 1
AIR BLANK*

OPERATOR

ADDITIONAL INFORMATION AND/OR REMARKS

05/23/2019
CA MODEL 5000EN SN 68-012335
INTOXILYZER - ALCOHOL ANALYZER
BOOKING
LONG BEACH POLICE DEPARTMENT

INTOXILYZER® INSTRUMENT PRINTER CARD



© 1986 by CMI INC.

LONG BEACH POLICE DEPARTMENT
BOOKING
INTOXILYZER - ALCOHOL ANALYZER
CA MODEL 5000EN SN 68-012335
05/30/2019

SIM CONC (%W/V) = .129
SIM TEMP (°C) = 34.02
LOT # = IRS013019
SOLN EXP DATE = 01/30/20
SIM SERIAL # = MP5021
NAME/PHOTO # =
G GOSSAGE 7885
REMARKS =

SUB TESTS SINCE LAST CAL CHECK = 000

TEST	%BAC	TIME
AIR BLANK	.000	08:51
CAL. CHECK	.128	08:52
AIR BLANK	.000	08:52
CAL. CHECK	.127	08:52
AIR BLANK	.000	08:52

SUBJECT'S NAME

TIME FIRST OBSERVED

INSTRUMENT LOCATION

0100 081 080 033 044

OPERATOR

ADDITIONAL INFORMATION AND/ OR REMARKS

INTERNAL STD OUT OF TOLERANCE

INVALID TEST

DATE 05/30/2019

TIME 08:49

SLAVE 25.2242

VERSION 02105.12

SN 68-012335

INTOXILYZER® INSTRUMENT-PRINTER CARD

