

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/01/2018 08:23

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3040	
Check sum	52088	Download date/time:	11/01/2018	09:06		Record ID	A8Q1	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 11/01/2018 08:25

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 019

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	20699	Download date/time:	11/01/2018	09:06		Record ID	A8H7	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/01/2018 08:36

First calibration check: .131 Interferent detected: No  
Second calibration check: .132 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.04  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 019

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 30146	Download date/time: 11/01/2018	09:06	Record ID A818
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012323 Accuracy check date/time: 11/04/2018 01:12

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

CHANGED TIME FROM 0212 TO 0112

Number of people tested since last periodic determination of accuracy:

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 43812	Download date/time: 12/06/2018	08:52	Record ID AAH4
Record imported (Y/N) N	Source Memo		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/08/2018 08:44

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3013	
Check sum	27600	Download date/time:	12/06/2018	08:52		Record ID	AA98	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 11/08/2018 08:45

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	57537	Download date/time:	12/06/2018	08:52		Record ID	A9Y6	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/08/2018 08:49

First calibration check: .124 Interferent detected: No  
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.04  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 018

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 28127	Download date/time: 12/06/2018	08:52	Record ID A9Z7
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012323 Accuracy check date/time: 11/08/2018 08:52

First calibration check: .126 Interferent detected: No  
Second calibration check: .127 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.21  
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SOLN

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 42994	Download date/time: 12/06/2018	08:52	Record ID AA09
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/15/2018 08:02

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3026	
Check sum	82030	Download date/time:	12/06/2018	08:52		Record ID	AAA1	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 11/15/2018 08:04

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	30353	Download date/time:	12/06/2018	08:52		Record ID	AA10	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/15/2018 08:08

First calibration check: .131 Interferent detected: No  
Second calibration check: .130 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.03  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 018

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57652	Download date/time: 12/06/2018	08:52	Record ID AA27
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012323 Accuracy check date/time: 11/21/2018 10:20

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3035
Check sum 36598	Download date/time: 12/06/2018	08:52	Record ID AAB6
Record imported (Y/N) N	Source Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/21/2018 10:21

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 006

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 53360	Download date/time: 12/06/2018	08:52	Record ID AA38
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012323 Accuracy check date/time: 11/21/2018 10:27

First calibration check: .127 Interferent detected: No  
Second calibration check: .128 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.04  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 006

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 44102	Download date/time: 12/06/2018	08:52	Record ID AA41
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/29/2018 08:41

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3021	
Check sum	86930	Download date/time:	12/06/2018	08:52		Record ID	AAC1	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 11/29/2018 08:43

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 022

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	20016	Download date/time:	12/06/2018	08:52		Record ID	AA52	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.



**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012323 Accuracy check date/time: 11/29/2018 08:45

First calibration check: .132 Interferent detected: No  
Second calibration check: .132 Interferent detected: No

Concentration (%W/V) .135 Solution lot#: IRS041118 Soln temp C: 34.03  
Simulator Ser No. DR4116

Number of people tested since last periodic determination of accuracy: 022

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 26689	Download date/time: 12/06/2018	08:52	Record ID AA69
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.