

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/03/2019 10:00

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3446	
Check sum	37265	Download date/time:	10/03/2019	10:38		Record ID	AS99	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012321 Accuracy check date/time: 10/03/2019 10:01

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	58976	Download date/time:	10/03/2019	10:38		Record ID	AS40	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/03/2019 10:03

First calibration check: .125 Interferent detected: No
Second calibration check: .132 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 33.74
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 28072	Download date/time: 10/03/2019	10:38	Record ID AS53
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012321 Accuracy check date/time: 10/10/2019 08:53

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3437
Check sum 37763	Download date/time: 11/07/2019	12:21	Record ID AUO8
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/10/2019 08:54

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 33.74
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 41528	Download date/time: 11/07/2019	12:21	Record ID AUD1
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012321 Accuracy check date/time: 10/10/2019 08:56

First calibration check: .131 Interferent detected: No
Second calibration check: .131 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.02
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 21505	Download date/time: 11/07/2019	12:21	Record ID AUE8
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/17/2019 12:40

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards
Prom	Pass	Real Time clock		Motor	Pass	Printer
Ram	Pass	Auto calibration	Pass	Range	Fail	Digital voltmeter ???
Check sum	27060	Download date/time:	11/07/2019	12:21		Record ID AUP2
		Record imported (Y/N)	N	Source	Diagnostic	

Instrument serial No: 68-012321 Accuracy check date/time: 10/17/2019 12:41

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards
Prom		Real Time clock		Motor		Printer
Ram		Auto calibration		Range		Digital voltmeter
Check sum	45768	Download date/time:	11/07/2019	12:21		Record ID AUF5
		Record imported (Y/N)	N	Source	Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/17/2019 12:45

First calibration check: .128 Interferent detected: No
Second calibration check: .131 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 33.97
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 41762	Download date/time: 11/07/2019	12:21	Record ID AUG2
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012321 Accuracy check date/time: 10/24/2019 09:25

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3438
Check sum 27400	Download date/time: 11/07/2019	12:21	Record ID AUQ5
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/24/2019 09:26

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 33.97
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12488	Download date/time: 11/07/2019	12:21	Record ID AUH5
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012321 Accuracy check date/time: 10/24/2019 09:28

First calibration check: .130 Interferent detected: No
Second calibration check: .131 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 30898	Download date/time: 11/07/2019	12:21	Record ID AUI2
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/31/2019 09:44

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3448
Check sum	43773	Download date/time:	11/07/2019 12:21	Record ID	AUR6		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012321 Accuracy check date/time: 10/31/2019 09:46

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards			
Prom	Real Time clock	Motor	Printer			
Ram	Auto calibration	Range	Digital voltmeter			
Check sum	63808	Download date/time:	11/07/2019 12:21	Record ID	AUJ9	
Record imported (Y/N) N				Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012321 Accuracy check date/time: 10/31/2019 09:47

First calibration check: .130 Interferent detected: No
Second calibration check: .132 Interferent detected: No

Concentration (%W/V) .129 Solution lot#: IRS013019 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 32016	Download date/time: 11/07/2019	12:21	Record ID AUK6
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.