

CHP South LA PAS Usage Logs 10/8/19-10/18/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
96	0.087		5:30 AM	10/09/2019	7:46 AM	04/12/2019	0.095	1:13 PM	10/08/2019	16040140
97	0.088		5:33 AM	10/09/2019	7:46 AM	04/12/2019	0.095	1:13 PM	10/08/2019	16040140
98	0.212		2:36 AM	10/18/2019	7:46 AM	04/12/2019	0.095	1:13 PM	10/08/2019	16040140
99	0.218		2:39 AM	10/18/2019	7:46 AM	04/12/2019	0.095	1:13 PM	10/08/2019	16040140
56	0		5:43 PM	10/09/2019	3:41 PM	03/28/2019	0.104	10:53 AM	10/08/2019	16040141
57	0.194		9:21 PM	10/10/2019	3:41 PM	03/28/2019	0.104	10:53 AM	10/08/2019	16040141
58	0.21		9:24 PM	10/10/2019	3:41 PM	03/28/2019	0.104	10:53 AM	10/08/2019	16040141
279	0.143		3:03 AM	10/11/2019	8:56 AM	04/25/2017	0.103	11:02 AM	10/08/2019	16040161
280	0.139		3:07 AM	10/11/2019	8:56 AM	04/25/2017	0.103	11:02 AM	10/08/2019	16040161
349	0.111		11:25 PM	10/12/2019	4:09 PM	06/25/2019	0.1	1:14 PM	10/08/2019	16040162
350	0.112		11:26 PM	10/12/2019	4:09 PM	06/25/2019	0.1	1:14 PM	10/08/2019	16040162
351	0.149		12:41 AM	10/13/2019	4:09 PM	06/25/2019	0.1	1:14 PM	10/08/2019	16040162
352	0.141		12:43 AM	10/13/2019	4:09 PM	06/25/2019	0.1	1:14 PM	10/08/2019	16040162
353	0.109		5:02 AM	10/13/2019	4:09 PM	06/25/2019	0.1	1:14 PM	10/08/2019	16040162
354	0.114		5:04 AM	10/13/2019	4:09 PM	06/25/2019	0.1	1:14 PM	10/08/2019	16040162
421	0.327		2:57 PM	10/08/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
422	0.301		2:59 PM	10/08/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
423	0.278		3:01 PM	10/08/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
424	0.031		11:17 PM	10/17/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
425	0.038		11:19 PM	10/17/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
426	0.038		11:21 PM	10/17/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
427	0.032		11:22 PM	10/17/2019	7:30 AM	04/12/2019	0.099	10:55 AM	10/08/2019	16040163
104	0.314		3:07 AM	10/11/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
105	0.307		3:09 AM	10/11/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
106	0.091		8:17 PM	10/12/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
107	0.091		8:19 PM	10/12/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
108	0		8:23 PM	10/12/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
109	0.22		8:03 PM	10/13/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
110	0.18		8:06 PM	10/13/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
111	0.181		8:08 PM	10/13/2019	3:58 PM	03/28/2019	0.102	11:05 AM	10/08/2019	16080027
39	0.16		10:55 PM	10/12/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
40	0.158		10:58 PM	10/12/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
41	0.157		2:12 AM	10/13/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
42	0.164		2:15 AM	10/13/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
43	0.014		3:43 AM	10/13/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
44	0.013		3:44 AM	10/13/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
45	0.08		3:38 AM	10/14/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
46	0.081		3:41 AM	10/14/2019	4:07 PM	03/28/2019	0.103	11:14 AM	10/08/2019	16080030
75	0.083		11:22 PM	10/11/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
76	0.087		11:25 PM	10/11/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
77	0.071		1:45 AM	10/12/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
78	0.066		1:46 AM	10/12/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
79	0.12		4:51 AM	10/12/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
80	0.122		4:53 AM	10/12/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
81	0		11:54 PM	10/12/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
82	0.135		2:36 AM	10/13/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
83	0.112		2:38 AM	10/13/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
84	0.122		2:40 AM	10/13/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
85	0.043		2:05 AM	10/17/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
86	0.042		2:07 AM	10/17/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
87	0.032		3:07 PM	10/17/2019	4:12 PM	03/28/2019	0.102	11:15 AM	10/08/2019	16080064
429	0.101		12:46 AM	10/11/2019	10:41 AM	08/14/2017	0.101	11:21 AM	10/08/2019	16080071
430	0.099		12:48 AM	10/11/2019	10:41 AM	08/14/2017	0.101	11:21 AM	10/08/2019	16080071
431	0.269		11:49 PM	10/14/2019	10:41 AM	08/14/2017	0.101	11:21 AM	10/08/2019	16080071
127	0.179		2:45 AM	10/13/2019	9:50 AM	04/25/2017	0.105	11:39 AM	10/08/2019	16120177
128	0.179		2:48 AM	10/13/2019	9:50 AM	04/25/2017	0.105	11:39 AM	10/08/2019	16120177
97	0.102		2:21 AM	10/13/2019	8:16 AM	10/21/2016	0.102	11:37 AM	10/08/2019	16120188
98	0.102		2:23 AM	10/13/2019	8:16 AM	10/21/2016	0.102	11:37 AM	10/08/2019	16120188
99	0.126		3:10 AM	10/13/2019	8:16 AM	10/21/2016	0.102	11:37 AM	10/08/2019	16120188
100	0.125		3:13 AM	10/13/2019	8:16 AM	10/21/2016	0.102	11:37 AM	10/08/2019	16120188
375	0		10:50 PM	10/08/2019	12:48 PM	06/09/2019	0.106	11:44 AM	10/08/2019	16120195
376	0.144		2:54 AM	10/12/2019	12:48 PM	06/09/2019	0.106	11:44 AM	10/08/2019	16120195
377	0.151		2:58 AM	10/12/2019	12:48 PM	06/09/2019	0.106	11:44 AM	10/08/2019	16120195
378	0.107		1:10 AM	10/17/2019	12:48 PM	06/09/2019	0.106	11:44 AM	10/08/2019	16120195
379	0.107		1:13 AM	10/17/2019	12:48 PM	06/09/2019	0.106	11:44 AM	10/08/2019	16120195
380	0.01		10:27 PM	10/17/2019	12:48 PM	06/09/2019	0.106	11:44 AM	10/08/2019	16120195

CHP South LA PAS Usage Logs 10/8/19-10/18/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
100	0.163		10:37 PM	10/08/2019	5:05 PM	03/28/2019	0.105	11:54 AM	10/08/2019	17170064
101	0.156		10:40 PM	10/08/2019	5:05 PM	03/28/2019	0.105	11:54 AM	10/08/2019	17170064

#1

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16040162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	1119	34	19DP41092	.100	23	.091					M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:09 PM	06/25/2019	.091	.100	11:20 AM	10/18/2019	BAC	16040162	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16040119

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
10/18/2019	1127	34	19DDP41092	100	27		.094				X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	3:25 PM	03/28/2019	.094	.100	11:28 AM	10/18/2019	BAC	16040119	8.7.24	

**Lifeloc Technologies**

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16040140

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1	TEST #2 (Optional)	YES	NO					
10/18/2019	1119	34	19DP41092	.100	23		.094				X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:46 AM	04/12/2019	.094	.100	11:19 AM	10/18/2019	BAC	16040140	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16040141

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
10/18/2019	1124	34	19DP41092	.100	22	.101				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:41 PM	03/28/2019	.101	.100	11:24 AM	10/18/2019	BAC	16040141	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	120134	19DP41092		.100	23		.099				X	M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	:099	.100	12:01 PM	10/18/2019	BAC	16040163	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#7

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	1217 34	19DP41092	.100	22	.103			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	12:59 PM	01/29/2018	.103	.100	12:17 PM	10/18/2019	BAC	16080013	8.7.24

**Lifeloc Technologies**

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 8

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifoloc FC20    PAS SERIAL NO.: 16080014

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					YES
10/18/2019	12:12	34	19DP41092	.100	23		.103				X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:51 AM	07/05/2019	.103	.100	12:12 PM	10/18/2019	BAC	16080014	8.7.24	Lifoloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080015

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/18/2019	12/11	34	19DP41092	.100	23	.101						M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.101	.100	12:12 PM	10/18/2019	BAC	16080015	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080027

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	1211 34	19DP41092	.100	22	.102		X		M. Tippett	20137		

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.102	.100	12:12 PM	10/18/2019	BAC	16080027	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 11

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16080028

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2	YES	NO				
10/18/2019	12/6 34	19DP41092		.100	22		.101			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	4:05 PM	03/28/2019	.101	.100	12:17 PM	10/18/2019	BAC	16080028	8.7.24

**Lifeloc Technologies**

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080030

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2 (Optional)	YES	NO			
10/18/2019	12:16:34	19DP41092	.100	22		.103						M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:07 PM	03/28/2019	.103	.100	12:17 PM	10/18/2019	BAC	16080030	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	1217	34	19DP41092	.100	22	.103	X			M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:12 PM	03/28/2019	.103	.100	12:17 PM	10/18/2019	BAC	16080064	8.7.24	

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.1100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	12:19	34	19DP41092	.100	22	.104			X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	1:54 PM	12/19/2017	.104	.100	12:20 PM	10/18/2019	BAC	16080065	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO: 16080069

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	122434	19DP41092	.100	22	.100			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
	.100	4:18 PM	03/28/2019	.100	.100	12:25 PM	10/18/2019	BAC	16080069	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lieloc FC20      PAS SERIAL NO.: 16080071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	1223	34	19DP41092	.100	22	.102				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.102	.100	12:23 PM	10/18/2019	BAC	16080071	8.7.24	

**Lieloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
10/18/2019	1241 34	19DP41092		.100	22	.105			X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.105	.100	12:41 PM	10/18/2019	BAC	16080072	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	1241	34	19DP41092	.100	22	.102				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.102	.100	12:41 PM	10/18/2019	BAC	16080075	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
								TEST #1 (Optional)	TEST #2 (Optional)						
10/18/2019	12:34	34	19DP41092		.100	22		.104		NO	X	M. Tippett	20137		
Solution prepared by: Cesar Munoz															
Date: 04-22-19															
Expiration date: 04-22-21															
		Cal Standard				Time: 4:30 PM		Last Cal		03/28/2019		Result: .104			
		Cal Standard				Time: 12:36 PM		Last Check		10/18/2019		UOM: BAC		Serial No.: 16080077	
		FC20				I.D. NO.: 8.7.24								Calibration / Check Printout	

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles      AREA NO: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO: 16120175

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	1230	34	19DP41092	.100	22	.100		X		M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:36 PM	03/28/2019	.100	.100	12:30 PM	10/18/2019	BAC	16120175	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16120177

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	12:34:34	19DP41092	.100	22	105		X		M. Tippett	20137		

Solution prepared by:		Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
Cesar Munoz	Date: 04-22-19												
Expiration date: 04-22-21		.100	09:50 AM	04/25/2017	.105	.100	12:35 PM	10/18/2019	BAC	16120177	8.7.24		
Calibration / Check Printout													
<b>Lifeloc Technologies</b>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16046161

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	160834			19DP41092	.100	24		.101			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.101	.100	4:09 PM	10/18/2019	BAC	16040161	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#22

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16120188

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	1605 34	19DP41092	.100	23	.099			X		M. Tippett	20137		

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.099	.100	4:06 PM	10/18/2019	BAC	16120188	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16120189

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	161034		19DP41092	.100	22		.103			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	08:50 AM	04/12/2019	.103	.100	4:11 PM	10/18/2019	BAC	16120189	8.7.24	Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	1606 34		19DP41092	.100	22		.101			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:20 PM	08/15/2019	.101	.100	4:07 PM	10/18/2019	BAC	16120194	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS/ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
								TEST #1	TEST #2 (Optional)					
10/18/2019	6/12	34	19DP41092	.100	22			107				M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.107	.100	4:13 PM	10/18/2019	BAC	16120195	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/18/2019	162034		19DP41092	.100	72		.101			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	3:25 PM	06/25/2019	.101	.100	4:21 PM	10/18/2019	BAC	16120197	8.7.24

**Lifeloc Technologies**

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/18/2019	1624 34	19DP41092	.100	72		.105		X				M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:42 PM	03/28/2019	.105	.100	4:24 PM	10/18/2019	BAC	17170066	8.7.24	

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**#28**

AREA NAME: **South Los Angeles** AREA NO: **530** COORDINATOR NAME: **Michael Tippett** I.D. NUMBER: **20137** PAS MAKE/MODEL: **Lifeloc FC20** PAS SERIAL NO.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
10/18/2019			19DP41092							M. Tippett	20137	<i>oo service</i>

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 17170155

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/18/2019	1626 39		19DP41092		.100	22	.104	X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:46 PM	03/28/2019	.104	.100	4:26 PM	10/18/2019	BAC	17170155	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 30

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.:

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/18/2019		19DP41092		.100								M. Tippett	20137	no service	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 51

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 17170064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		LOT NO.	VALUE			TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
10/18/2019	1624/34	19DP41092	.100	21		.107				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	5:05 PM	03/28/2019	.107	.100	4:24 PM	10/18/2019	BAC	17170064	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.