

CHP South LA PAS Usage Logs 10/18/19-10/29/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
66	0.141		8:24 PM	10/20/2019	3:25 PM	03/28/2019	0.094	11:28 AM	10/18/2019	16040119
67	0.14		8:27 PM	10/20/2019	3:25 PM	03/28/2019	0.094	11:28 AM	10/18/2019	16040119
68	0.156		7:19 PM	10/26/2019	3:25 PM	03/28/2019	0.094	11:28 AM	10/18/2019	16040119
69	0.165		7:21 PM	10/26/2019	3:25 PM	03/28/2019	0.094	11:28 AM	10/18/2019	16040119
98	0.212		2:36 AM	10/18/2019	7:46 AM	04/12/2019	0.095	1:13 PM	10/08/2019	16040140
99	0.218		2:39 AM	10/18/2019	7:46 AM	04/12/2019	0.095	1:13 PM	10/08/2019	16040140
100	0.079		8:31 AM	10/20/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
101	0.08		8:33 AM	10/20/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
102	0.078		8:35 AM	10/20/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
103	0		10:33 PM	10/20/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
104	0.044		11:13 PM	10/20/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
105	0		4:43 AM	10/21/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
106	0.178		1:10 AM	10/25/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
107	0.18		1:13 AM	10/25/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
108	0		9:10 PM	10/26/2019	7:46 AM	04/12/2019	0.094	11:19 AM	10/18/2019	16040140
59	0.228		2:05 AM	10/19/2019	3:41 PM	03/28/2019	0.101	11:24 AM	10/18/2019	16040141
60	0.23		2:07 AM	10/19/2019	3:41 PM	03/28/2019	0.101	11:24 AM	10/18/2019	16040141
61	0.158		12:48 AM	10/24/2019	3:41 PM	03/28/2019	0.101	11:24 AM	10/18/2019	16040141
62	0.154		12:51 AM	10/24/2019	3:41 PM	03/28/2019	0.101	11:24 AM	10/18/2019	16040141
281	0.159		3:14 AM	10/20/2019	8:56 AM	04/25/2017	0.101	4:09 PM	10/18/2019	16040161
282	0.166		3:16 AM	10/20/2019	8:56 AM	04/25/2017	0.101	4:09 PM	10/18/2019	16040161
283	NEG		8:44 PM	10/24/2019	8:56 AM	04/25/2017	0.101	4:09 PM	10/18/2019	16040161
284	0.027		1:48 AM	10/25/2019	8:56 AM	04/25/2017	0.101	4:09 PM	10/18/2019	16040161
355	0.048		3:07 AM	10/20/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
356	0.083		8:06 AM	10/20/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
357	0.082		8:08 AM	10/20/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
358	0.128		8:08 PM	10/20/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
359	0.129		8:11 PM	10/20/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
360	0.111		3:58 AM	10/21/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
361	0.113		4:01 AM	10/21/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
362	0.178		3:18 AM	10/27/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
363	0.178		3:20 AM	10/27/2019	4:09 PM	06/25/2019	0.091	11:20 AM	10/18/2019	16040162
428	0.072		4:09 PM	10/19/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163

CHP South LA PAS Usage Logs 10/18/19-10/29/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
429	0.087		4:11 PM	10/19/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
430	0.093		4:13 PM	10/19/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
431	0.037		8:29 AM	10/21/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
432	0.014		8:20 PM	10/27/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
433	0.015		8:23 PM	10/27/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
434	0.027		1:41 AM	10/29/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
112	0.141		2:56 AM	10/19/2019	3:58 PM	03/28/2019	0.102	12:12 PM	10/18/2019	16080027
113	0.134		2:59 AM	10/19/2019	3:58 PM	03/28/2019	0.102	12:12 PM	10/18/2019	16080027
114	NEG		8:12 PM	10/28/2019	3:58 PM	03/28/2019	0.102	12:12 PM	10/18/2019	16080027
58	0.062		1:56 AM	10/19/2019	4:05 PM	03/28/2019	0.101	12:17 PM	10/18/2019	16080028
59	0.064		1:56 AM	10/19/2019	4:05 PM	03/28/2019	0.101	12:17 PM	10/18/2019	16080028
47	0.208		12:03 AM	10/22/2019	4:07 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080030
48	0.211		12:08 AM	10/22/2019	4:07 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080030
49	0.197		1:55 AM	10/26/2019	4:07 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080030
50	0.2		1:58 AM	10/26/2019	4:07 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080030
88	0.122		11:08 PM	10/19/2019	4:12 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080064
89	0.116		11:10 PM	10/19/2019	4:12 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080064
90	0.088		2:12 AM	10/20/2019	4:12 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080064
91	0.082		2:15 AM	10/20/2019	4:12 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080064
92	0.182		5:00 AM	10/20/2019	4:12 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080064
93	0.166		5:03 AM	10/20/2019	4:12 PM	03/28/2019	0.103	12:17 PM	10/18/2019	16080064
286	0		1:06 AM	10/19/2019	1:54 PM	12/19/2017	0.104	12:20 PM	10/18/2019	16080065
432	0.105		1:29 AM	10/20/2019	10:41 AM	08/14/2017	0.102	12:23 PM	10/18/2019	16080071
433	0.11		1:31 AM	10/20/2019	10:41 AM	08/14/2017	0.102	12:23 PM	10/18/2019	16080071
434	0.157		10:39 PM	10/27/2019	10:41 AM	08/14/2017	0.102	12:23 PM	10/18/2019	16080071
435	0.162		10:41 PM	10/27/2019	10:41 AM	08/14/2017	0.102	12:23 PM	10/18/2019	16080071
436	0.275		2:35 AM	10/28/2019	10:41 AM	08/14/2017	0.102	12:23 PM	10/18/2019	16080071
437	0.256		2:38 AM	10/28/2019	10:41 AM	08/14/2017	0.102	12:23 PM	10/18/2019	16080071
33	0		11:50 PM	10/19/2019	2:05 PM	12/19/2017	0.105	12:41 PM	10/18/2019	16080072
34	0		11:53 PM	10/19/2019	2:05 PM	12/19/2017	0.105	12:41 PM	10/18/2019	16080072
25	0.045		2:51 AM	10/20/2019	4:24 PM	03/28/2019	0.102	12:41 PM	10/18/2019	16080075
26	0.035		3:44 AM	10/20/2019	4:24 PM	03/28/2019	0.102	12:41 PM	10/18/2019	16080075
27	0.035		3:47 AM	10/20/2019	4:24 PM	03/28/2019	0.102	12:41 PM	10/18/2019	16080075

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
75	0.197		11:46 PM	10/25/2019	4:36 PM	03/28/2019	0.1	12:30 PM	10/18/2019	16120175
76	0.205		11:48 PM	10/25/2019	4:36 PM	03/28/2019	0.1	12:30 PM	10/18/2019	16120175
129	0.224		9:42 PM	10/19/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
130	0.218		9:44 PM	10/19/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
131	0.16		12:27 AM	10/20/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
132	0.149		12:30 AM	10/20/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
133	0.145		3:34 AM	10/27/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
134	0.198		3:31 AM	10/28/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
135	0.198		3:33 AM	10/28/2019	9:50 AM	04/25/2017	0.105	12:35 PM	10/18/2019	16120177
101	0.141		2:46 AM	10/26/2019	8:16 AM	10/21/2016	0.099	4:06 PM	10/18/2019	16120188
102	0.135		2:48 AM	10/26/2019	8:16 AM	10/21/2016	0.099	4:06 PM	10/18/2019	16120188
381	0.015		2:54 AM	10/19/2019	12:48 PM	06/09/2019	0.107	4:13 PM	10/18/2019	16120195
382	0.04		3:28 AM	10/19/2019	12:48 PM	06/09/2019	0.107	4:13 PM	10/18/2019	16120195
383	0.125		3:52 AM	10/21/2019	12:48 PM	06/09/2019	0.107	4:13 PM	10/18/2019	16120195
384	0.132		3:55 AM	10/21/2019	12:48 PM	06/09/2019	0.107	4:13 PM	10/18/2019	16120195
337	0		12:24 AM	10/19/2019	3:25 PM	06/25/2019	0.101	4:21 PM	10/18/2019	16120197
338	0.171		3:41 AM	10/19/2019	3:25 PM	06/25/2019	0.101	4:21 PM	10/18/2019	16120197
339	0.171		3:43 AM	10/19/2019	3:25 PM	06/25/2019	0.101	4:21 PM	10/18/2019	16120197
340	0.145		3:58 AM	10/28/2019	3:25 PM	06/25/2019	0.101	4:21 PM	10/18/2019	16120197
341	0.145		4:01 AM	10/28/2019	3:25 PM	06/25/2019	0.101	4:21 PM	10/18/2019	16120197
102	0.105		2:07 AM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
103	0.096		2:08 AM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
104	0.099		4:15 AM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
105	0.096		4:17 AM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
106	0.097		4:20 AM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
107	0.098		5:43 PM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
108	0.099		5:46 PM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
109	0		7:39 PM	10/20/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
110	0.144		1:55 AM	10/21/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
111	0.14		1:57 AM	10/21/2019	5:05 PM	03/28/2019	0.107	4:24 PM	10/18/2019	17170064
12	0.013		2:21 AM	10/21/2019	4:42 PM	03/28/2019	0.105	4:24 PM	10/18/2019	17170066
13	0		2:23 AM	10/25/2019	4:42 PM	03/28/2019	0.105	4:24 PM	10/18/2019	17170066
14	0.11		2:56 AM	10/25/2019	4:42 PM	03/28/2019	0.105	4:24 PM	10/18/2019	17170066

CHP South LA PAS Usage Logs 10/18/19-10/29/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
15	0.1		2:58 AM	10/25/2019	4:42 PM	03/28/2019	0.105	4:24 PM	10/18/2019	17170066

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS		
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
10/29/2019	1510 ³⁴		19DPR41092	.100		27		.099				X		M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	4:09 PM	06/25/2019	.099	.100	3:11 PM	10/29/2019	BAC	16040162	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1 (Optional)	TEST #2 (Optional)			
10/29/2019	1519	34	19DP41092	.100	24	.098						M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	03/28/2019	.098	.100	3:19 PM	10/29/2019	BAC	16040119	8.7.24	

Lifoloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/29/2019	1148	19DP41092	.100	21	.092			X		M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:46 AM	04/12/2019	.092	.100	11:48 AM	10/29/2019	BAC	16040140	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1153	34	19DP41092	.100	21	.098				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:41 PM	03/28/2019	.098	.100	11:53 AM	10/29/2019	BAC	16040141	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1204 34	19DP41092		100	21	.101		X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:56 AM	04/25/2017	.101	.100	12:05 PM	10/29/2019	BAC	16040161	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	11:58:34		19DP41092	.100	23		.096				M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.096	.100	11:58 AM	10/29/2019	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

47

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1219 34		19DP41092	.100	22							M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	12:59 PM	01/29/2018	.101	.100	12:19 PM	10/29/2019	BAC	16080013	8.7.24

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

Handwritten signature

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1214	74	19DP41092	100	73	.102					X	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:51 AM	07/05/2019	.102	.100	12:14 PM	10/29/2019	BAC	16080014	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/29/2019	1212 34	19DP41092	.100	22	.102			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	3:58 PM	03/28/2019	.102	.100	12:13 PM	10/29/2019	BAC	16080027	8.7.24	
Calibration / Check Printout										
Lifeloc Technologies										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	12/18 34		19DP41092	.100		21		.100				M. Tippett	20137	

Solution prepared by:
 Cesar Minoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Call Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:05 PM	03/28/2019	.100	.100	12:19 PM	10/29/2019	BAC	16080028	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
10/29/2019	1218	34	19DP41092	.100	21		.102		NO	X	M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:07 PM	03/28/2019	.102	.100	12:18 PM	10/29/2019	BAC	16080030	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 160 800 64

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
10/29/2019	12:19	34	19DP41092	.100	21	.100					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:12 PM	03/28/2019	.100	.100	12:20 PM	10/29/2019	BAC	16080064	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/29/2019	15:26:34	19DP41092		.100	24-						M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	1:54 PM	12/19/2017	.102	.100	3:26 PM	10/29/2019	BAC	16080065	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2	YES	NO				
10/29/2019	123634	19DP41092		.100	21		.105				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.105	.100	12:36 PM	10/29/2019	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/29/2019	12:29	34	19DDP41092		.100	.102				M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	4:24 PM	03/28/2019	.102	.100	12:29 PM	10/29/2019	BAC	16080075	8.7.24

Lifoloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifloc FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/29/2019	122834	19DDP41092	100	22		.104			X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:30 PM	03/28/2019	.104	.100	12:29 PM	10/29/2019	BAC	16080077	8.7.24	Lifloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 167 20175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	12:27	34	19DP41092	.100	21	.104				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:36 PM	03/28/2019	.104	.100	12:27 PM	10/29/2019	BAC	16120175	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	12:36:34	19DP41092		.100	22		104					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	09:50 AM	04/25/2017	.104	.100	12:36 PM	10/29/2019	BAC	16120177	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1233	34	19DP41092	.100	22	.100	X			M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	08:16 AM	10/21/2016	.100	.100	12:34 PM	10/29/2019	BAC	16120188	8.7.24	
Calibration / Check Printout										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/29/2019	123834	19DP41092	100	22		.103			X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.103	.100	12:38 PM	10/29/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1234 24	19DP41092	.100	22		.100					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:20 PM	08/15/2019	.100	.100	12:34 PM	10/29/2019	BAC	16120194	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
10/29/2019	124034	19DPA1092	.100	22	.105				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:48 PM	06/09/2019	.105	.100	12:40 PM	10/29/2019	BAC	16120195	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1247	34	19DP41092							M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:25 PM	06/25/2019	.101	.100	12:48 PM	10/29/2019	BAC	16120197	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/29/2019	1253	34	19DPR41092	.100	21	.105		X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
4:42 PM	03/28/2019	.105	.100	12:54 PM	10/29/2019	BAC	17170066	8.7.24		
.100										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#229

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 177 70155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/29/2019	12:57:34		19DP41092	.100									M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
	.100	4:46 PM	03/28/2019	.103	.100	12:51 PM	10/29/2019	BAC	17170155	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

