

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
64	0.197		1:59 AM	09/27/2019	3:25 PM	03/28/2019	0.104	1:23 PM	09/17/2019	16040119
65	0.192		2:02 AM	09/27/2019	3:25 PM	03/28/2019	0.104	1:23 PM	09/17/2019	16040119
55	0		2:17 AM	10/04/2019	3:41 PM	03/28/2019	0.103	2:40 PM	09/27/2019	16040141
266	0.181		2:09 AM	09/28/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
267	0.171		2:12 AM	09/28/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
268	0		11:51 PM	09/28/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
269	0		11:52 PM	09/28/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
270	0		12:21 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
271	0.05		12:42 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
272	0.048		12:44 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
273	0.266		1:01 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
274	0.267		1:03 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
275	0.087		3:45 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
276	0.085		3:47 AM	09/29/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
277	0.146		1:11 AM	10/04/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
278	0.145		1:13 AM	10/04/2019	8:56 AM	04/25/2017	0.103	2:49 PM	09/27/2019	16040161
344	0.121		1:59 AM	10/04/2019	4:09 PM	06/25/2019	0.099	2:36 PM	09/27/2019	16040162
345	0.137		2:01 AM	10/04/2019	4:09 PM	06/25/2019	0.099	2:36 PM	09/27/2019	16040162
346	0.185		10:42 PM	10/06/2019	4:09 PM	06/25/2019	0.099	2:36 PM	09/27/2019	16040162
347	0.184		10:44 PM	10/06/2019	4:09 PM	06/25/2019	0.099	2:36 PM	09/27/2019	16040162
348	0		1:57 AM	10/07/2019	4:09 PM	06/25/2019	0.099	2:36 PM	09/27/2019	16040162
417	0.145		2:42 AM	10/04/2019	7:30 AM	04/12/2019	0.097	2:43 PM	09/27/2019	16040163
418	0.14		2:44 AM	10/04/2019	7:30 AM	04/12/2019	0.097	2:43 PM	09/27/2019	16040163
419	0.025		1:48 AM	10/05/2019	7:30 AM	04/12/2019	0.097	2:43 PM	09/27/2019	16040163
420	0.022		1:50 AM	10/05/2019	7:30 AM	04/12/2019	0.097	2:43 PM	09/27/2019	16040163
94	0.067		11:40 PM	10/05/2019	3:58 PM	03/28/2019	0.103	3:23 PM	09/27/2019	16080015
95	0.109		11:41 PM	10/05/2019	3:58 PM	03/28/2019	0.103	3:23 PM	09/27/2019	16080015
96	0.044		12:19 AM	10/06/2019	3:58 PM	03/28/2019	0.103	3:23 PM	09/27/2019	16080015
97	0.042		12:20 AM	10/06/2019	3:58 PM	03/28/2019	0.103	3:23 PM	09/27/2019	16080015
103	0.073		1:49 AM	09/29/2019	3:58 PM	03/28/2019	0.103	3:23 PM	09/27/2019	16080027
50	0.094		2:13 AM	09/29/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
51	0.096		2:17 AM	09/29/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
52	0.164		4:31 AM	09/29/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
53	0.17		4:33 AM	09/29/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
54	0.086		8:53 AM	09/29/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
55	0.084		8:56 AM	09/29/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
56	0.316		11:45 PM	10/04/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
57	0.324		11:48 PM	10/04/2019	4:05 PM	03/28/2019	0.1	3:28 PM	09/27/2019	16080028
73	0.015		1:00 PM	10/05/2019	4:12 PM	03/28/2019	0.101	3:30 PM	09/27/2019	16080064
74	0.014		1:05 PM	10/05/2019	4:12 PM	03/28/2019	0.101	3:30 PM	09/27/2019	16080064
280	0.277		10:40 AM	09/29/2019	1:54 PM	12/19/2017	0.103	3:31 PM	09/27/2019	16080065
281	0.286		10:43 AM	09/29/2019	1:54 PM	12/19/2017	0.103	3:31 PM	09/27/2019	16080065
282	0.065		3:40 AM	10/06/2019	1:54 PM	12/19/2017	0.103	3:31 PM	09/27/2019	16080065
283	0.065		3:43 AM	10/06/2019	1:54 PM	12/19/2017	0.103	3:31 PM	09/27/2019	16080065
284	0.11		4:18 AM	10/06/2019	1:54 PM	12/19/2017	0.103	3:31 PM	09/27/2019	16080065
285	0.115		4:21 AM	10/06/2019	1:54 PM	12/19/2017	0.103	3:31 PM	09/27/2019	16080065
158	0.079		10:50 PM	09/28/2019	4:18 PM	03/28/2019	0.1	3:36 PM	09/27/2019	16080069
427	0		2:42 AM	09/28/2019	10:41 AM	08/14/2017	0.101	3:36 PM	09/27/2019	16080071
428	0		2:43 AM	09/28/2019	10:41 AM	08/14/2017	0.101	3:36 PM	09/27/2019	16080071
23	0.135		10:03 PM	10/02/2019	4:24 PM	03/28/2019	0.101	3:45 PM	09/27/2019	16080075
24	0.136		10:06 PM	10/02/2019	4:24 PM	03/28/2019	0.101	3:45 PM	09/27/2019	16080075
68	0.159		8:58 PM	09/27/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
69	0.166		9:01 PM	09/27/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
70	0.148		1:08 AM	10/05/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
71	0.153		1:11 AM	10/05/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
72	0.118		4:50 AM	10/06/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
73	0.161		4:52 AM	10/06/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
74	0.161		4:53 AM	10/06/2019	4:36 PM	03/28/2019	0.107	3:45 PM	09/27/2019	16120175
126	0.017		9:37 PM	09/28/2019	9:50 AM	04/25/2017	0.102	3:53 PM	09/27/2019	16120177
330	0		9:39 PM	09/30/2019	12:20 PM	08/15/2019	0.097	4:15 PM	09/27/2019	16120194
331	0		9:41 PM	09/30/2019	12:20 PM	08/15/2019	0.097	4:15 PM	09/27/2019	16120194
332	0.149		2:32 AM	10/06/2019	12:20 PM	08/15/2019	0.097	4:15 PM	09/27/2019	16120194
333	0.146		2:35 AM	10/06/2019	12:20 PM	08/15/2019	0.097	4:15 PM	09/27/2019	16120194
373	0.102		12:51 AM	09/27/2019	12:48 PM	06/09/2019	0.108	2:15 PM	09/17/2019	16120195
374	0.107		12:53 AM	09/27/2019	12:48 PM	06/09/2019	0.108	2:15 PM	09/17/2019	16120195
330	0.222		12:10 AM	10/04/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
331	0.168		12:12 AM	10/04/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197
332	0.218		12:14 AM	10/04/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197
333	0.202		12:25 AM	10/04/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197
334	0.201		12:27 AM	10/04/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197
335	0.129		2:34 AM	10/07/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197
336	0.125		2:36 AM	10/07/2019	3:25 PM	06/25/2019	0.099	4:11 PM	09/27/2019	16120197
84	0.072		10:04 PM	09/27/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
85	0.058		10:05 PM	09/27/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
86	0.171		1:14 AM	09/28/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
87	0.187		1:17 AM	09/28/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
88	0.019		9:45 PM	09/28/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
89	0.017		9:48 PM	09/28/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
90	0.157		2:32 AM	09/29/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
91	0.147		2:34 AM	09/29/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
92	0.088		10:39 PM	10/04/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
93	0.091		10:42 PM	10/04/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
94	0.084		10:40 PM	10/05/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
95	0.075		10:44 PM	10/05/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
96	0.215		1:41 AM	10/06/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
97	0.213		1:43 AM	10/06/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
98	0.2		12:10 AM	10/07/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064
99	0.213		12:11 AM	10/07/2019	5:05 PM	03/28/2019	0.104	4:23 PM	09/27/2019	17170064

#1

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS			
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO							
10/08/2019	1314 34	19DP41092		.100		27		.100		.100		X		M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	4:09 PM	06/25/2019	.100	.100	1:14 PM	10/08/2019	BAC	16040162	8.7.24

Lifeloc Technologies

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 2

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16040119

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES				
10/08/2019	105834		19DP41092	.100	23								M. Tippett	20137	

Solution prepared by  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:25 PM	03/28/2019	.100	.100	10:58 AM	10/08/2019	BAC	16040119	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	171354	19DP41092	.100		23		095		X		M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:46 AM	04/12/2019	.095	.100	1:13 PM	10/08/2019	BAC	16040140	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16040141

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1052 54	19DP41092	.100	22	.104			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	3:41 PM	03/28/2019	.104	.100	10:53 AM	10/08/2019	BAC	16040141	8.7.24	
Calibration / Check Printout										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16040161

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	11/01/34		19DP41092	.100		22		.103			X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	08:56 AM	04/25/2017	.103	.100	11:02 AM	10/08/2019	BAC	16040161	8.7.24	Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#6

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16040163

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1055	34	19DP41092	.100	23		.099				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.099	.100	10:55 AM	10/08/2019	BAC	16040163	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
10/08/2019	11:34	19DP41092	.100	23		.102						M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.102	.100	11:11 AM	10/08/2019	BAC	16080013	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1106 24	19DP41092	.100	23		.103				X	M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:51 AM	07/05/2019	.103	.100	11:06 AM	10/08/2019	BAC	16080014	8.7.24	

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16080015

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	11:06:34		19DP41092	.100	-22		.102				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.102	.100	11:06 AM	10/08/2019	BAC	16080015	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16080027

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1105	34	19DP41092	.100	22	.102				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.102	.100	11:05 AM	10/08/2019	BAC	16080027	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080028

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1114	34	19DP41092	.100	22		.100				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.100	.100	11:14 AM	10/08/2019	BAC	16080028	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 ID NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080030

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2	TEST #1 (Optional)	TEST #2	YES	NO				
10/08/2019	11/4	34	19DP41092	.100	22		.103		X		M. Tippett	20137		

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:07 PM	03/28/2019	.103	.100	11:14 AM	10/08/2019	BAC	16080030	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #	(Optional)	YES	NO				
10/08/2019	11:53	34	19DP41092	.100	22	102				X		M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	4:12 PM	03/28/2019	.102	.100	11:15 AM	10/08/2019	BAC	16080064	8.7.24	Calibration / Check Printout

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#14

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	11:16	34	19DP41092	.100	21	.103					M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	1:54 PM	12/19/2017	.103	.100	11:17 AM	10/08/2019	BAC	16080065	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080069

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1/22 34		19DP41092	.100		22		.099			X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:18 PM	03/28/2019	.099	.100	11:23 AM	10/08/2019	BAC	16080069	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#17

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1139	34	19DP41092	.100	21		.104				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.104	.100	11:39 AM	10/08/2019	BAC	16080072	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080075

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	11:31	34	19DP41092	.100	22		.103			X		M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.103	.100	11:32 AM	10/08/2019	BAC	16080075	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	PAS TEMP (°C)				TEST #1 (Optional)	TEST #2 (Optional)					
10/08/2019	1130	34	19DP41092		.100	22				X		M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:30 PM	03/28/2019	.105	.100	11:31 AM	10/08/2019	BAC	16080077	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 20

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      ID NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1/29 34		19DP41092	.100		22		.107			X	M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:36 PM	03/28/2019	.107	.100	11:30 AM	10/08/2019	BAC	16120175	8.7.24	

Calibration / Check Printout  
**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME: South Los Angeles  
AREA NO: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	11:39	34	19DP41092	.100	22		.105				X	M. Tippett	20137	

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.105	.100	11:39 AM	10/08/2019	BAC	16120177	8.7.24	

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



# 22

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Michael Tippett  
 ID NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO: 16120188

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO.				
10/08/2019	11:37	34	19DP41092	.100	22		.102					M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.102	.100	11:37 AM	10/08/2019	BAC	16120188	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifeloc FC20  
PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1140	34	19DP41092	.100	22	.104					M. Tippett	20137	

Solution prepared by:  
Cesar Munoz  
Date: 04-22-19  
Expiration date:  
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.104	.100	11:41 AM	10/08/2019	BAC	16120189	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    ID NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1137	34	19DP41092	.100								M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	12:20 PM	08/15/2019	.099	.100	11:37 AM	10/08/2019	BAC	16120194	8.7.24

Lifeloc Technologies  
 Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 I.D. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16/20195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
10/08/2019	1143	34	19DP41092	100	22	-106				X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:48 PM	06/09/2019	.106	.100	11:44 AM	10/08/2019	BAC	16120195	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#27

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 17170066

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2	TEST #1 (Optional)	TEST #2	YES	NO				
10/08/2019	1159	34	19DP41092	.100		20		.105			X	M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:42 PM	03/28/2019	.105	.100	11:55 AM	10/08/2019	BAC	17170066	8.7.24	
Calibration / Check Printout										

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#29

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    ID. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 17170155

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
10/08/2019	1152 34	19DP41092	.100	21		.104			X			M. Tippett	20137	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:46 PM	03/28/2019	.104	.100	11:56 AM	10/08/2019	BAC	17170155	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



# 30

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Michael Tippett  
I.D. NUMBER: 20137  
PAS MAKE/MODEL: Lifefoc FC20  
PAS SERIAL NO.:

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2	TEST #1 (Optional)	TEST #2	YES	NO				
10/08/2019			19DP41092	100								M. Tippett	20137	DEVICE IS OUT OF SERVICE

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 31

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    ID. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 17170064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2019	11:54	34	19DP41092	.100	21		.105			X		M. Tippett	20137	
Solution prepared by: Cesar Munoz														
Date: 04-22-19														
Expiration date: 04-22-21														
Cal Standard														
Time: 5:05 PM														
Last Cal														
03/28/2019														
Result														
.105														
Cal Standard														
Time: 11:54 AM														
Last Check														
10/08/2019														
UOM														
BAC														
Serial No.														
17170064														
FC20														
8.7.24														
Calibration / Check Printout														
<b>Lifeloc Technologies</b>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.