

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
70	0.213		11:59 PM	11/02/2019	3:25 PM	03/28/2019	0.098	3:19 PM	10/29/2019	16040119
71	0.213		12:01 AM	11/03/2019	3:25 PM	03/28/2019	0.098	3:19 PM	10/29/2019	16040119
109	NEG		9:23 PM	10/31/2019	7:46 AM	04/12/2019	0.092	11:48 AM	10/29/2019	16040140
110	0.135		3:59 AM	11/04/2019	7:46 AM	04/12/2019	0.092	11:48 AM	10/29/2019	16040140
111	0.133		4:01 AM	11/04/2019	7:46 AM	04/12/2019	0.092	11:48 AM	10/29/2019	16040140
63	0.169		12:11 AM	10/31/2019	3:41 PM	03/28/2019	0.098	11:53 AM	10/29/2019	16040141
64	0.114		12:14 AM	10/31/2019	3:41 PM	03/28/2019	0.098	11:53 AM	10/29/2019	16040141
65	0.149		12:17 AM	10/31/2019	3:41 PM	03/28/2019	0.098	11:53 AM	10/29/2019	16040141
66	0.119		11:00 PM	11/03/2019	3:41 PM	03/28/2019	0.098	11:53 AM	10/29/2019	16040141
67	0.124		11:02 PM	11/03/2019	3:41 PM	03/28/2019	0.098	11:53 AM	10/29/2019	16040141
285	0.171		10:58 PM	11/02/2019	8:56 AM	04/25/2017	0.101	12:05 PM	10/29/2019	16040161
286	0.168		11:00 PM	11/02/2019	8:56 AM	04/25/2017	0.101	12:05 PM	10/29/2019	16040161
364	0.226		10:50 PM	10/30/2019	4:09 PM	06/25/2019	0.099	3:11 PM	10/29/2019	16040162
365	0.232		10:53 PM	10/30/2019	4:09 PM	06/25/2019	0.099	3:11 PM	10/29/2019	16040162
434	0.027		1:41 AM	10/29/2019	7:30 AM	04/12/2019	0.099	12:01 PM	10/18/2019	16040163
435	NEG		9:28 PM	10/29/2019	7:30 AM	04/12/2019	0.096	11:58 AM	10/29/2019	16040163
436	0.114		6:34 AM	11/02/2019	7:30 AM	04/12/2019	0.096	11:58 AM	10/29/2019	16040163
437	0.117		6:37 AM	11/02/2019	7:30 AM	04/12/2019	0.096	11:58 AM	10/29/2019	16040163
438	0.188		5:01 AM	11/03/2019	7:30 AM	04/12/2019	0.096	11:58 AM	10/29/2019	16040163
439	0.183		5:04 AM	11/03/2019	7:30 AM	04/12/2019	0.096	11:58 AM	10/29/2019	16040163
115	0.171		1:35 AM	11/02/2019	3:58 PM	03/28/2019	0.102	12:13 PM	10/29/2019	16080027
116	0.175		1:37 AM	11/02/2019	3:58 PM	03/28/2019	0.102	12:13 PM	10/29/2019	16080027
60	0		10:11 PM	10/29/2019	4:05 PM	03/28/2019	0.1	12:19 PM	10/29/2019	16080028
61	0.156		12:20 AM	11/02/2019	4:05 PM	03/28/2019	0.1	12:19 PM	10/29/2019	16080028
62	0.166		12:24 AM	11/02/2019	4:05 PM	03/28/2019	0.1	12:19 PM	10/29/2019	16080028
51	0.133		7:38 PM	11/02/2019	4:07 PM	03/28/2019	0.102	12:18 PM	10/29/2019	16080030
52	0.134		7:41 PM	11/02/2019	4:07 PM	03/28/2019	0.102	12:18 PM	10/29/2019	16080030
53	0.063		5:38 PM	11/03/2019	4:07 PM	03/28/2019	0.102	12:18 PM	10/29/2019	16080030
54	0.059		5:39 PM	11/03/2019	4:07 PM	03/28/2019	0.102	12:18 PM	10/29/2019	16080030
94	0.096		2:32 AM	11/03/2019	4:12 PM	03/28/2019	0.1	12:20 PM	10/29/2019	16080064
95	0.099		2:34 AM	11/03/2019	4:12 PM	03/28/2019	0.1	12:20 PM	10/29/2019	16080064
159	0.147		3:07 AM	11/02/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
160	0.146		3:10 AM	11/02/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069

CHP South LA PAS Usage Logs 10/29/19-11/6/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
161	0.068		1:08 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
162	0.066		1:10 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
163	0.082		1:51 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
164	0.08		1:53 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
165	0.163		3:33 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
166	0.153		3:36 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
167	0.078		4:10 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
168	0.14		4:11 AM	11/03/2019	4:18 PM	03/28/2019	0.099	3:15 PM	10/29/2019	16080069
438	0.007		1:30 AM	11/03/2019	10:41 AM	08/14/2017	0.1	3:13 PM	10/29/2019	16080071
77	0.085		4:12 AM	11/02/2019	4:36 PM	03/28/2019	0.104	12:27 PM	10/29/2019	16120175
78	0.085		4:15 AM	11/02/2019	4:36 PM	03/28/2019	0.104	12:27 PM	10/29/2019	16120175
136	0.08		2:42 AM	11/02/2019	9:50 AM	04/25/2017	0.104	12:36 PM	10/29/2019	16120177
137	0.082		2:44 AM	11/02/2019	9:50 AM	04/25/2017	0.104	12:36 PM	10/29/2019	16120177
138	0.075		2:47 AM	11/02/2019	9:50 AM	04/25/2017	0.104	12:36 PM	10/29/2019	16120177
385	0		11:58 PM	11/02/2019	12:48 PM	06/09/2019	0.105	12:40 PM	10/29/2019	16120195
112	0		10:24 PM	10/30/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
113	0.095		1:55 AM	11/01/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
114	0.091		1:57 AM	11/01/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
115	0.092		11:32 PM	11/02/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
116	0.094		11:34 PM	11/02/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
117	0.05		11:37 PM	11/02/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
118	0.188		11:38 PM	11/02/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
119	0.107		12:35 AM	11/03/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
120	0.109		12:38 AM	11/03/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
121	0.141		11:33 PM	11/04/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
122	0.144		11:35 PM	11/04/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
123	0.161		10:33 PM	11/05/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064
124	0.161		10:36 PM	11/05/2019	5:05 PM	03/28/2019	0.105	12:51 PM	10/29/2019	17170064

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO						
11/06/2019	1259 34	19DP41092	.100	22	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	X	M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:09 PM	06/25/2019	.088	.100	1:59 PM	11/06/2019	BAC	16040162	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
11/06/2019	1:09 PM	19DP41092		.100	22		.093		X				M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	3:25 PM	03/28/2019	.093	.100	1:09 PM	11/06/2019	BAC	16040119	8.7.24

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1239 34	19DP41092	.100	22		.095					M Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:46 AM	04/12/2019	.095	.100	1:39 PM	11/06/2019	BAC	16040140	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: **16040141**

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #2 (Optional)	TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	14:05 1344		19DP41092	.100	-1								M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	3:41 PM	03/28/2019	.097	.100	1:44 PM	11/06/2019	BAC	16040141	8.7.24	

Calibration / Check Printout

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1:55	34	19DP41092	.100	20	.098		X			M. Tippett	20137	AS

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.098	.100	1:55 PM	11/06/2019	BAC	16040161	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	134834		19DP41092		.100	22		.094		NO	YES	X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	07:30 AM	04/12/2019	.094	.100	1:48 PM	11/06/2019	BAC	16040163	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1405	34	19DP41092	.100	22	.098			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.098	.100	2:05 PM	11/06/2019	BAC	16080013	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: *16080014*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1:59:34		19DP41092	.100	23		.099					X	20137	<i>DST</i>

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard		Last Cal		Cal Standard		Last Check		UOM		Serial No.		FC20	
Time:	Result	Time:	Result	Time:	Result	Time:	Result						
11:51 AM	.100	07/05/2019	.099	1:59 PM	.100	11/06/2019	.099	BAC		16080014		8.7.24	

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
11/06/2019	1:59 34	19DP41092		.100	22	.098		X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	3:58 PM	03/28/2019	.098	.100	1:59 PM	11/06/2019	BAC	16080015	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS TEMP #1 (Optional)	PAS TEMP #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2	YES	NO				
11/06/2019	1359	34	19DP41092	.100	21		.097				X	M. Tippett	20137	
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21														
		Cal Standard												
		Time:												
		Last Cal												
		Time:												
		Result												
		Cal Standard												
		Time:												
		Last Check												
		Time:												
		UOM												
		Serial No.												
		FC20												
		I.D. NO.												
Calibration / Check Printout														
Lifeloc Technologies														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1504	54	19DP41092		.100	21	.097		X	M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:05 PM	03/28/2019	.097	.100	2:05 PM	11/06/2019	BAC	16080028	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
11/06/2019	140434	19DP41092	.100	21	.097			X			M. Tippett	20137	AST	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:07 PM	03/28/2019	.097	.100	2:04 PM	11/06/2019	BAC	16080030	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS							
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO											
11/06/2019	1405 34		19DP41092	.100	23		.100				X	M. Tippett	20137	DST							
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21																					
Cal Standard					Time:	Last Cal					Result	Cal Standard		Time:	Last Check		UOM	Serial No.	FC20	Calibration / Check Printout	
					4:12 PM	03/28/2019					.100			2:05 PM	11/06/2019		BAC	16080064	8.7.24		
Lifeloc Technologies																					

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

4/14

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Michael Tippet	20137	Lifeloc FC20	16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
11/06/2019		19DP41092		.100							M. Tippet	20137	DEVICE WAS NOT IN LOCKER FOR TESTING
		Solution prepared by: Cesar Munoz											
		Date: 04-22-19											
		Expiration date: 04-22-21											
													DEVICE WAS PIKED OO SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1410	34	19DP41092	.100	21	.094			X	M. Tippett	20137	AST

Solution prepared by:												
Cesar Munoz												
Date: 04-22-19												
Expiration date: 04-22-21												
Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies		
.100	4:18 PM	03/28/2019	.094	.100	2:10 PM	11/06/2019	BAC	16080069	8.7.24	Calibration / Check Printout		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1409 34	19DP41092	.100	22	097		X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.097	.100	2:09 PM	11/06/2019	BAC	16080071	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1427 34	19DP41092	.100	21		.101				X	M. Tippett	20137	AS
Solution prepared by: Cesar Munoz													
Date: 04-22-19													
Expiration date: 04-22-21													
		Cal Standard		Time: 2:05 PM		Last Cal		Result: .101		Cal Standard: .100		Time: 2:27 PM	
										Last Check		11/06/2019	
												UOM	
												Serial No.	
												FC20	
												8.7.24	
Calibration / Check Printout													
Lifeloc Technologies													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
									TEST #1	TEST #2	YES	NO				
11/06/2019	1421	34	19DP41092		.100	21			.097				X	M. Tippett	20137	BST
			Solution prepared by: Cesar Munoz		Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20		
			Date: 04-22-19		.100	4:24 PM	03/28/2019	.097	.100	2:21 PM	11/06/2019	BAC	16080075	8.7.24		
			Expiration date: 04-22-21													

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
11/06/2019	1426 34	19DP41092	.100	21		.097		X				M. Tippett	20137	DSY

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:30 PM	03/28/2019	.097	.100	2:26 PM	11/06/2019	BAC	16080077	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
11/06/2019	1424 34		19DDP41092	.100									M. Tippett	20137	Δ 57

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.099	.100	2:24 PM	11/06/2019	BAC	16120175	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1433	34	19DP41092	.100	22		.099			X		M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.099	.100	2:33 PM	11/06/2019	BAC	16120177	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16720188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2 (Optional)	YES	NO				
11/06/2019	1431	34	19DP41092	.100	21						X	M. Tippett	20137	DSY

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.094	.100	2:31 PM	11/06/2019	BAC	16120188	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16/20189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	14:24	19DP41092	.100	21	.095			X		M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	08:50 AM	04/12/2019	.095	.100	2:36 PM	11/06/2019	BAC	16120189	8.7.24

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1431	34	19DP41092	.100	21		.093				X	M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:20 PM	08/15/2019	.093	.100	2:31 PM	11/06/2019	BAC	16120194	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REIEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
11/06/2019	1439	34	19DP41092	.100	21	.100		X		M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.100	.100	2:39 PM	11/06/2019	BAC	16120195	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
11/06/2019	144734		19DP41092	.100	21		.094					X		M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	06/25/2019	.094	.100	2:47 PM	11/06/2019	BAC	16120197	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1451 34		19DP41092	.100	20	.098					M. Tippett	20137	DST

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:42 PM	03/28/2019	.098	.100	2:54 PM	11/06/2019	BAC	17170066	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

428

AREA NAME
South Los Angeles

AREA NO.
530

COORDINATOR NAME
Michael Tippett

I.D. NUMBER
20137

PAS MAKE/MODEL
Lifeloc FC20

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/06/2019			19DP41092	.100								M. Tippett	20137	OO SERVICE
Solution prepared by: Cesar Munoz														
Date: 04-22-19														
Expiration date: 04-22-21														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	145734	19DPR41092	.100	21		.097			X		M. Tippett	20137	DST

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifoloc Technologies
.100	4:46 PM	03/28/2019	.097	.100	2:51 PM	11/06/2019	BAC	17170155	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

30

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
11/06/2019		19DP41092		.100							M. Tippett	20137	<i>OO SERVICE</i>

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
11/06/2019	1457	34	19DP41092	.100	20	.098					M. Tippett	20137	Δ 57

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	5:05 PM	03/28/2019	.098	.100	2:51 PM	11/06/2019	BAC	17170064	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.