

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS ACCURACY CHECK RESULTS* | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|-------|-----------|-----------|--------------|------------|--------------------|-----------------------------|----------------------|--------------------|--------------|----------|----------------------------|
| | | | | | | | | | | | | |
| 05/02/2019 | 12:59 | 24 | 17GP41049 | .100 | 30 | | .098 | X | | B. Marzorini | 14413 | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | |
| | | | | FC20 | 8.7.24 | | | | | | | |
| | | | | Serial No. | 16040162 | | | | | | | |
| | | | | UOM | BAC | | | | | | | |
| | | | | Last Check | 05/02/2019 | | | | | | | |
| | | | | Time: | 1:17 PM | | | | | | | |
| | | | | Cal Standard | .100 | | | | | | | |
| | | | | Result | .098 | | | | | | | |
| | | | | Last Cal | 01/17/2019 | | | | | | | |
| | | | | Time: | 4:50 PM | | | | | | | |
| | | | | Cal Standard | .100 | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/ FC20
 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|---|------|-----------|-----------|--------|--------------------|--------------------|-----------------------------|--------------------|----------------------|--------------|--------------------|--------------|----------|----------------------------|--|
| | | | | | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | | |
| 05/02/2019 | 7:34 | 34 | 17GP41049 | .100 | | 22 | | 101 | | | | B. Marzorini | 14413 | | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 16040119 | | UOM | | BAC | | Last Check | |
| | | | | | | Time: | | 07:34 AM | | Cal Standard | | .100 | | Result | |
| | | | | | | | | | | | | | | Last Cal | |
| | | | | | | | | | | | | | | Time: | |
| | | | | | | | | | | | | | | Cal Standard | |
| | | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzolini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------|-------|---------|--------------------|---------------|---------|-----------------------------|---------|----------------------|----|--------------------|--------------|----------|----------------------------|
| | | | | | | | TEST #1 | TEST #2 | TEST #1 | TEST #2 | YES | NO | | | | |
| 05/02/2019 | 7:45 | 34 | 17GP41049 | .100 | 22 | | | | | | | | | B. Marzolini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

Lifeloc Technologies

Calibration / Check Printout

| | |
|--------------|------------|
| FC20 | 8.7.24 |
| Serial No. | 16040141 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 07:45 AM |
| Cal Standard | .100 |
| Result | .104 |
| Last Cal | 03/28/2019 |
| Time: | 3:41 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/ FC20
 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | LOT NO. | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS** | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-----------|-------|--------------------|--------------------|------------------------------|--------------------|----------------------|----|--------------------|--------------|----------|----------------------------|
| | | | | | | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 05/02/2019 | 7:49 | 34 | | 17GR41049 | .100 | | 23 | | | | | | B. Marzorini | 14413 | |

| | |
|------------------------------|------------|
| Lifeloc Technologies | |
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16040161 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 07:50 AM |
| Cal Standard | .100 |
| Result | .104 |
| Last Cal | 04/25/2017 |
| Time: | 08:56 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 ID NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/ FC20
 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|--------------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|--------------|----------|---------------------------|
| | | | | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 05/02/2019 | 7:48 | 34 | 17GP41049 | .100 | 24 | .099 | | X | | | B. Marzorini | 14413 | |

Prepared by:
 Cesar A. Mejia

7/24/19

Expire Date:
 7/24/19

Date: 7-24-17

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16040163 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 07:48 AM |
| Cal Standard | .100 |
| Result | .099 |
| Last Cal | 04/12/2019 |
| Time: | 07:30 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-----------|-------|---------------|-----------------------------|--------------------|----------------------|--------------------|--------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | | TEST #1 (Optional) | TEST #2 (Optional) | | | | | |
| 05/02/2019 | 7:59 | 34 | 17GP41049 | .100 | 23 | | | | | B. Marzorini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16080014 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 07:59 AM |
| Cal Standard | .100 |
| Result | .102 |
| Last Cal | 03/28/2019 |
| Time: | 3:52 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzolini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/ FC20
 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|-----------------------------------|--------------------|---------------|-----------------------------|--------------------|----------------------|--------------------|--------------|----------|---------------------------|
| | | | | VALUE | TEST #1 (Optional) | | TEST #2 (Optional) | TEST #1 (Optional) | | | | | |
| 05/02/2019 | 8:01 | 39 | 17GP41049 | .100 | 23 | | 102 | | X | | B. Marzolini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16080027 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 08:01 AM |
| Cal Standard | .100 |
| Result | .102 |
| Last Cal | 03/28/2019 |
| Time: | 3:58 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** . If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzolini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16086030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|-------|---------|--------------------|-----------------------------|--------------------|----------------------|--------------------|--------------|----------|---------------------------|
| | | | | | | | TEST #1 (Optional) | TEST #2 (Optional) | | | | | |
| 05/02/2019 | 8:05 | 34 | 17GP41049 | .100 | 22 | | 100 | | NO | | B. Marzolini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16080030 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 08:06 AM |
| Cal Standard | .100 |
| Result | .100 |
| Last Cal | 03/28/2019 |
| Time: | 4:07 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|--------------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|--------------|----------|---------------------------|
| | | | | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 05/02/2019 | 8:09 | 34 | 17GP41049 | .100 | 22 | 102 | | | | | B. Marzorini | 14413 | |

Lifeloc Technologies
 Calibration / Check Printout

| | |
|--------------|------------|
| FC20 | 8.7.24 |
| Serial No. | 16080064 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 08:09 AM |
| Cal Standard | .100 |
| Result | .102 |
| Last Cal | 03/28/2019 |
| Time: | 4:12 PM |
| Cal Standard | .100 |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#14

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|---|------|--------------|-----------|------------|---------|--------------------|-----------------------------|--------------------|----------------------|--------------------|--------------|----------|---------------------------|
| | | | | | | | TEST #1 (Optional) | TEST #2 (Optional) | | | | | |
| 05/02/2019 | 8:13 | 34 | 17GP41049 | .100 | 22 | | 105 | | NO | ✓ | B. Marzorini | 14413 | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 16080065 | | UOM | | BAC | |
| | | Last Check | | 05/02/2019 | | Time: | | 08:14 AM | | Cal Standard | | .100 | |
| | | Result | | .105 | | Last Cal | | 12/19/2017 | | Time: | | 1:54 PM | |
| | | Cal Standard | | .100 | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: B. Marzorini | I.D. NUMBER: 14413 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS | |
|---|------|-----------|-----------|----------|---------|--------------------|---------------|----------|-----------------------------|--------|----------------------|------|--------------------|--------------|----------|---------------------------|--|
| | | | | | | | TEST #1 | TEST #2 | YES | NO | YES | NO | | | | | |
| 05/02/2019 | 8:18 | 34 | 17GP41049 | .100 | 22 | | | | | 099 | | | X | B. Marzorini | 14413 | | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 16080069 | | UOM | | BAC | | Last Check | | 05/02/2019 | |
| | | Time: | | 08:18 AM | | Cal Standard | | .100 | | Result | | .099 | | Last Cal | | 03/28/2019 | |
| | | Time: | | 4:18 PM | | Cal Standard | | .100 | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|-----------------------------------|---------|--------------------|---------|----------------------------|-----|----------------------|-----|--------------------|--|--------------|----------|---------------------------|
| | | | | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | YES | NO | | | | |
| 05/02/2019 | 8:19 | 34 | 17GP41049 | .100 | .22 | | | .101 | | | | | | B. Marzorini | 14413 | |

Lifeloc Technologies

Calibration / Check Printout

| | |
|--------------|------------|
| FC20 | 8.7.24 |
| Serial No. | 16080071 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 08:19 AM |
| Cal Standard | .100 |
| Result | .101 |
| Last Cal | 08/14/2017 |
| Time: | 10:41 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: B. Marzotini | I.D. NUMBER: 14413 | PAS MAKE/MODEL: Lifeloc/ FC20 | PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | TEST #1 | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|-----------|-------|---------|--------------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|--------------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | | | TEST #2 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 05/02/2019 | 8:54 | 34 | 17GP41049 | 100 | 22 | | | | | | | | B. Marzotini | 14413 | |

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16080072 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 08:54 AM |
| Cal Standard | .100 |
| Result | .104 |
| Last Cal | 12/19/2017 |
| Time: | 2:05 PM |
| Cal Standard | .100 |

Date: 7-24-17
Expire Date: 7/24/19
Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS | |
|---|------|-----------|-----------|----------|---------|--------------------|---------------|----------|-----------------------------|---------|----------------------|------|--------------------|--------------|----------|---------------------------|--|
| | | | | | | | TEST #1 | TEST #2 | TEST #1 | TEST #2 | YES | NO | | | | | |
| 05/02/2019 | 8:17 | 31 | 17GP41049 | 100 | 22 | | | | | | | | | B. Marzorini | 14413 | | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 16120175 | | UOM | | BAC | | Last Check | | 04/22/2019 | |
| | | Time: | | 08:50 AM | | Cal Standard | | .100 | | Result | | .100 | | Last Cal | | 03/28/2019 | |
| | | Time: | | 4:36 PM | | Cal Standard | | .100 | | | | | | | | | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS | |
|---|------|-----------|-----------|----------|---------|--------------------|---------------|----------|-----------------------------|---------|----------------------|------|--------------------|--------------|----------|---------------------------|-----|
| | | | | | | | TEST #1 | TEST #2 | TEST #1 | TEST #2 | YES | NO | | | | | YES |
| 05/02/2019 | 8:52 | 34 | 17GP41049 | .100 | 23 | | | | | 103 | | | | B. Marzorini | 14413 | | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 16080077 | | UOM | | BAC | | Last Check | | 05/02/2019 | |
| | | Time: | | 08:52 AM | | Cal Standard | | .100 | | Result | | .103 | | Last Cal | | 03/28/2019 | |
| | | Time: | | 4:30 PM | | Cal Standard | | .100 | | | | | | | | | |

#19

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

20

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: B. Marzorini
I.D. NUMBER: 14413
PAS MAKE/MODEL: Lifeloc/FC20
PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------------------------------|--|-------|---------|--------------------|---------------|---------|-----------------------------|----|----------------------|----|--------------------|--------------|----------|---------------------------|
| | | | LOT NO. | | | | | TEST #1 | TEST #2 | YES | NO | YES | NO | | | | |
| 05/02/2019 | 8:54 | 34 | 17GR41049 | | .100 | 22 | | | | | | | | X | B. Marzorini | 14413 | |

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120175 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 08:55 AM |
| Cal Standard | .100 |
| Result | .105 |
| Last Cal | 03/28/2019 |
| Time: | 4:36 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------|-------|---------|--------------------|----------------------|----|--------------------|--------------|----------|----------------------------|
| | | | | | | | YES | NO | | | | |
| 05/02/2019 | 9:02 | 34 | 17GP41049 | .100 | 23 | | | | | B. Marzorini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

Lifeloc Technologies

Calibration / Check Printout

| | |
|--------------|------------|
| FC20 | 8.7.24 |
| Serial No. | 16120188 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 09:02 AM |
| Cal Standard | .100 |
| Result | .100 |
| Last Cal | 10/21/2016 |
| Time: | 08:16 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | LOT NO. | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | ID. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------------------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|--------------------|--------------|---------|---------------------------|
| | | | | | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | | | | | |
| 05/02/2019 | 9:06 | 34 | | 17GP41049 | .100 | 23 | | 101 | | NO | | B. Marzorini | 14413 | |

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120189 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 09:06 AM |
| Cal Standard | .100 |
| Result | .101 |
| Last Cal | 04/12/2019 |
| Time: | 08:50 AM |
| Cal Standard | .100 |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

24

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------------------------------|------|-------|---------|--------------------|---------------|-----------------------------|--------------------|----------------------|--------------------|--------------|----------|---------------------------|
| | | | LOT NO. | DATE | | | | | TEST #1 (Optional) | TEST #2 (Optional) | | | | | |
| 05/02/2019 | 9:03 | 34 | 17GR41049 | | .100 | 22 | | | | | NO | | B. Marzorini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120194 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 09:04 AM |
| Cal Standard | .100 |
| Result | .095 |
| Last Cal | 01/29/2018 |
| Time: | 08:50 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

25

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzotini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-----------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|--------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 05/02/2019 | 9:18 | 39 | 17GP41049 | .100 | 23 | .103 | | | X | B. Marzotini | 14413 | |

Lifeloc Technologies

Calibration / Check Printout

| | |
|--------------|------------|
| FC20 | 8.7.24 |
| Serial No. | 16120195 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 09:18 AM |
| Cal Standard | .100 |
| Result | .103 |
| Last Cal | 04/25/2017 |
| Time: | 08:47 AM |
| Cal Standard | .100 |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|-------|--------------------|--------------------|--------------------|--------------------|----------------------|----|--------------------|--------------|----------|---------------------------|
| | | | | | | | | | YES | NO | | | | |
| 05/02/2019 | 9:22 | 34 | 17GP41049 | .100 | 22 | | .097 | | | X | | B. Marzorini | 14413 | |

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120197 |
| UOM | BAC |
| Last Check | 05/02/2019 |
| Time: | 09:22 AM |
| Cal Standard | .100 |
| Result | .097 |
| Last Cal | 10/21/2016 |
| Time: | 10:22 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

26

27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 (Optional) | TEST #2 (Optional) | PAS ACCURACY CHECK RESULTS: | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|------|-----------|-----------|--------|--------------------|--------------------|-----------------------------|----------|----------------------|-----|--------------------|--------------|----------|---------------------------|--|------------|--|-------|--|----------|--|--------------|--|------|--|--------|--|------|--|----------|--|------------|--|-------|--|---------|--|--------------|--|------|--|
| | | | | | | | TEST #1 | TEST #2 | YES | NO | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 05/02/2019 | 9:24 | 34 | 17GP41049 | .100 | 22 | | 103 | | | | X | B. Marzorini | 14413 | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Lifeloc Technologies Calibration / Check Printout | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 17170066 | | UOM | | BAC | | Last Check | | 05/02/2019 | | Time: | | 09:25 AM | | Cal Standard | | .100 | | Result | | .103 | | Last Cal | | 03/28/2019 | | Time: | | 4:42 PM | | Cal Standard | | .100 | |
| Prepared by: Cesar A. Mejia Date: 7-24-17 Expire Date: 7/24/19 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | TEST #1 | TEST #2 (Optional) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|------|-----------|-----------|-----------------------------------|---------|------------|---------|--------------------|-----------------------------|-----|----------------------|--------------------|--------------|------------|----------------------------|------------|--|-------|--|----------|--|--------------|--|------|--|--------|--|------|--|----------|--|------------|--|-------|--|---------|--|--------------|--|------|--|
| | | | | TEMP (°C) | LOT NO. | | | | YES | NO | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 05/02/2019 | 9:28 | 34 | 17GP41049 | | | .100 | 22 | | | | X | | B. Marzorini | 14413 | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Lifeloc Technologies | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Calibration / Check Printout | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | FC20 | | 8.7.24 | | Serial No. | | 17170155 | | UOM | | BAC | | Last Check | | 05/02/2019 | | Time: | | 09:28 AM | | Cal Standard | | .100 | | Result | | .101 | | Last Cal | | 03/28/2019 | | Time: | | 4:46 PM | | Cal Standard | | .100 | |
| Date: 7-24-17 Expire Date: 7/24/19 Prepared by: Cesar A. Mejia | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

| Test Number | Result | ID | Test Time | Test Date | Cal Time | Cal Date | Check Result | Check Time | Check Date | Unit Serial Number |
|-------------|--------|----|-----------|------------|----------|------------|--------------|------------|------------|--------------------|
| 27 | 0 | | 5:55 PM | 04/23/2019 | 3:25 PM | 03/28/2019 | 0.096 | 7:43 AM | 04/22/2019 | 16040119 |
| 28 | NEG | | 9:15 PM | 04/25/2019 | 3:25 PM | 03/28/2019 | 0.096 | 7:43 AM | 04/22/2019 | 16040119 |
| 24 | NEG | | 4:03 AM | 04/27/2019 | 7:46 AM | 04/12/2019 | 0.094 | 7:42 AM | 04/22/2019 | 16040140 |
| 21 | 0.16 | | 11:42 PM | 04/26/2019 | 3:41 PM | 03/28/2019 | 0.1 | 7:48 AM | 04/22/2019 | 16040141 |
| 22 | 0.151 | | 11:46 PM | 04/26/2019 | 3:41 PM | 03/28/2019 | 0.1 | 7:48 AM | 04/22/2019 | 16040141 |
| 204 | 0.106 | | 2:57 AM | 04/30/2019 | 8:56 AM | 04/25/2017 | 0.099 | 7:52 AM | 04/22/2019 | 16040161 |
| 205 | 0.111 | | 3:00 AM | 04/30/2019 | 8:56 AM | 04/25/2017 | 0.099 | 7:52 AM | 04/22/2019 | 16040161 |
| 206 | 0.011 | | 3:04 AM | 04/30/2019 | 8:56 AM | 04/25/2017 | 0.099 | 7:52 AM | 04/22/2019 | 16040161 |
| 207 | 0.062 | | 3:06 AM | 04/30/2019 | 8:56 AM | 04/25/2017 | 0.099 | 7:52 AM | 04/22/2019 | 16040161 |
| 280 | 0 | | 7:46 AM | 04/28/2019 | 4:50 PM | 01/17/2019 | 0.098 | 7:35 AM | 04/22/2019 | 16040162 |
| 281 | 0 | | 7:47 AM | 04/28/2019 | 4:50 PM | 01/17/2019 | 0.098 | 7:35 AM | 04/22/2019 | 16040162 |
| 282 | 0 | | 7:47 AM | 04/28/2019 | 4:50 PM | 01/17/2019 | 0.098 | 7:35 AM | 04/22/2019 | 16040162 |
| 365 | 0.097 | | 2:18 AM | 04/29/2019 | 7:30 AM | 04/12/2019 | 0.093 | 7:52 AM | 04/22/2019 | 16040163 |
| 366 | 0.096 | | 2:21 AM | 04/29/2019 | 7:30 AM | 04/12/2019 | 0.093 | 7:52 AM | 04/22/2019 | 16040163 |
| 67 | NEG | | 9:03 PM | 04/22/2019 | 12:59 PM | 01/29/2018 | 0.092 | 12:38 PM | 04/22/2019 | 16080013 |
| 68 | 0 | | 11:02 PM | 04/27/2019 | 12:59 PM | 01/29/2018 | 0.092 | 12:38 PM | 04/22/2019 | 16080013 |
| 69 | 0.155 | | 2:11 AM | 04/29/2019 | 12:59 PM | 01/29/2018 | 0.092 | 12:38 PM | 04/22/2019 | 16080013 |
| 70 | 0.155 | | 2:13 AM | 04/29/2019 | 12:59 PM | 01/29/2018 | 0.092 | 12:38 PM | 04/22/2019 | 16080013 |
| 71 | 0.141 | | 2:10 AM | 05/02/2019 | 12:59 PM | 01/29/2018 | 0.092 | 12:38 PM | 04/22/2019 | 16080013 |
| 72 | 0.141 | | 2:13 AM | 05/02/2019 | 12:59 PM | 01/29/2018 | 0.092 | 12:38 PM | 04/22/2019 | 16080013 |
| 24 | 0.151 | | 10:14 PM | 04/23/2019 | 3:52 PM | 03/28/2019 | 0.096 | 8:01 AM | 04/22/2019 | 16080014 |
| 25 | 0.151 | | 10:18 PM | 04/23/2019 | 3:52 PM | 03/28/2019 | 0.096 | 8:01 AM | 04/22/2019 | 16080014 |
| 26 | 0.091 | | 3:17 AM | 04/28/2019 | 3:52 PM | 03/28/2019 | 0.096 | 8:01 AM | 04/22/2019 | 16080014 |
| 27 | 0.09 | | 3:20 AM | 04/28/2019 | 3:52 PM | 03/28/2019 | 0.096 | 8:01 AM | 04/22/2019 | 16080014 |
| 28 | NEG | | 2:28 AM | 04/29/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 29 | NEG | | 2:28 AM | 04/29/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 30 | 0.106 | | 2:30 AM | 04/29/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 31 | 0.112 | | 2:31 AM | 04/29/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 32 | 0.152 | | 2:38 AM | 04/29/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 33 | 0.101 | | 1:23 AM | 05/01/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 34 | 0.103 | | 1:25 AM | 05/01/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:04 AM | 04/22/2019 | 16080015 |
| 26 | 0.059 | | 4:06 AM | 04/28/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:02 AM | 04/22/2019 | 16080027 |
| 27 | 0.056 | | 4:08 AM | 04/28/2019 | 3:58 PM | 03/28/2019 | 0.096 | 8:02 AM | 04/22/2019 | 16080027 |

CHP SLA LIFELOC USAGE LOG 5/2/2019

| | | | | | | | | | |
|-----|-------|----------|------------|---------|------------|-------|---------|------------|----------|
| 27 | 0 | 5:55 PM | 04/23/2019 | 3:25 PM | 03/28/2019 | 0.096 | 7:43 AM | 04/22/2019 | 16040119 |
| 11 | 0.139 | 3:24 AM | 04/28/2019 | 4:07 PM | 03/28/2019 | 0.096 | 8:06 AM | 04/22/2019 | 16080030 |
| 12 | 0.13 | 3:26 AM | 04/28/2019 | 4:07 PM | 03/28/2019 | 0.096 | 8:06 AM | 04/22/2019 | 16080030 |
| 13 | 0.012 | 9:06 PM | 04/24/2019 | 4:12 PM | 03/28/2019 | 0.096 | 8:09 AM | 04/22/2019 | 16080064 |
| 14 | 0.012 | 9:08 PM | 04/24/2019 | 4:12 PM | 03/28/2019 | 0.096 | 8:09 AM | 04/22/2019 | 16080064 |
| 15 | 0.028 | 9:49 PM | 04/24/2019 | 4:12 PM | 03/28/2019 | 0.096 | 8:09 AM | 04/22/2019 | 16080064 |
| 16 | 0.029 | 1:55 AM | 04/28/2019 | 4:12 PM | 03/28/2019 | 0.096 | 8:09 AM | 04/22/2019 | 16080064 |
| 17 | 0.03 | 1:56 AM | 04/28/2019 | 4:12 PM | 03/28/2019 | 0.096 | 8:09 AM | 04/22/2019 | 16080064 |
| 85 | NEG | 9:21 PM | 04/23/2019 | 4:18 PM | 03/28/2019 | 0.095 | 8:14 AM | 04/22/2019 | 16080069 |
| 6 | NEG | 8:42 AM | 04/22/2019 | 4:24 PM | 03/28/2019 | 0.097 | 8:40 AM | 04/22/2019 | 16080075 |
| 7 | 0 | 12:55 AM | 04/28/2019 | 4:24 PM | 03/28/2019 | 0.097 | 8:40 AM | 04/22/2019 | 16080075 |
| 8 | 0 | 5:04 AM | 04/28/2019 | 4:24 PM | 03/28/2019 | 0.097 | 8:40 AM | 04/22/2019 | 16080075 |
| 9 | 0 | 5:04 AM | 04/28/2019 | 4:24 PM | 03/28/2019 | 0.097 | 8:40 AM | 04/22/2019 | 16080075 |
| 10 | NEG | 8:48 AM | 05/02/2019 | 4:24 PM | 03/28/2019 | 0.101 | 8:48 AM | 05/02/2019 | 16080075 |
| 44 | 0.075 | 2:52 AM | 04/28/2019 | 4:30 PM | 03/28/2019 | 0.098 | 8:47 AM | 04/22/2019 | 16080077 |
| 45 | 0.086 | 2:55 AM | 04/28/2019 | 4:30 PM | 03/28/2019 | 0.098 | 8:47 AM | 04/22/2019 | 16080077 |
| 46 | 0.081 | 1:08 AM | 04/29/2019 | 4:30 PM | 03/28/2019 | 0.098 | 8:47 AM | 04/22/2019 | 16080077 |
| 47 | 0.082 | 1:10 AM | 04/29/2019 | 4:30 PM | 03/28/2019 | 0.098 | 8:47 AM | 04/22/2019 | 16080077 |
| 48 | 0.078 | 2:42 AM | 04/29/2019 | 4:30 PM | 03/28/2019 | 0.098 | 8:47 AM | 04/22/2019 | 16080077 |
| 49 | 0.079 | 2:45 AM | 04/29/2019 | 4:30 PM | 03/28/2019 | 0.098 | 8:47 AM | 04/22/2019 | 16080077 |
| 18 | 0.181 | 1:51 AM | 04/24/2019 | 4:36 PM | 03/28/2019 | 0.1 | 8:50 AM | 04/22/2019 | 16120175 |
| 19 | 0.187 | 1:54 AM | 04/24/2019 | 4:36 PM | 03/28/2019 | 0.1 | 8:50 AM | 04/22/2019 | 16120175 |
| 20 | 0.18 | 10:46 PM | 04/24/2019 | 4:36 PM | 03/28/2019 | 0.1 | 8:50 AM | 04/22/2019 | 16120175 |
| 21 | 0.186 | 10:48 PM | 04/24/2019 | 4:36 PM | 03/28/2019 | 0.1 | 8:50 AM | 04/22/2019 | 16120175 |
| 306 | 0.121 | 2:34 AM | 04/29/2019 | 8:50 AM | 01/29/2018 | 0.093 | 9:00 AM | 04/22/2019 | 16120194 |
| 307 | 0.119 | 2:37 AM | 04/29/2019 | 8:50 AM | 01/29/2018 | 0.093 | 9:00 AM | 04/22/2019 | 16120194 |
| 308 | 0.219 | 2:40 AM | 04/29/2019 | 8:50 AM | 01/29/2018 | 0.093 | 9:00 AM | 04/22/2019 | 16120194 |