

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzolini ID. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/30/2019	054034	17GR41049	100	23	089			X	100	B. Marzolini	14413	NEEDED CALIBRATION

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040162
UOM	BAC
Last Check	05/30/2019
Time:	05:43 AM
Cal Standard	100
Result	100
Last Cal	05/30/2019
Time:	05:42 AM
Cal Standard	100

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/ FC20
 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							YES	NO				
05/30/2019	5:52	34	17GP41049	.100	22				X	B. Marzorini	14413	
Lifeloc Technologies												
Calibration / Check Printout												
				FC20		8.7.24						
				Serial No.		16040119						
				UOM		BAC						
				Last Check		05/30/2019						
				Time:		05:52 AM						
				Cal Standard		.100						
				Result		.094						
				Last Cal		03/28/2019						
				Time:		3:25 PM						
				Cal Standard		.100						

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)					
05/30/2019	5:49	34	17GP41049	.100	22		.091		X		B. Marzorini	14413	

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040140
UOM	BAC
Last Check	05/30/2019
Time:	05:50 AM
Cal Standard	.100
Result	.091
Last Cal	04/12/2019
Time:	07:46 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
05/30/2019	5:55	34	17GP41049	.100	25		.098		X		B. Marzorini	14413	
Lifeloc Technologies													
Calibration / Check Printout													
		FC20		8.7.24		Serial No.		16040141		UOM		BAC	
		Last-Check		05/30/2019		Time:		05:55 AM		Cal Standard		.100	
		Result		.098		Last Cal		03/28/2019		Time:		3:41 PM	
		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
05/30/2019	0:01	34	17GP41049	.100	21		.096		X		B. Marzorini	14413	
Lifeloc Technologies													
Calibration / Check Printout													
		FC20		8.7.24		Serial No.		16040161		UOM		BAC	
		Last Check		05/30/2019		Time:		06:01 AM		Cal Standard		.100	
		Result		.096		Last Cal		04/25/2017		Time:		08:56 AM	
		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 ID. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
05/30/2019	6:15	34	17GP41049	100	28		.091	X		B. Marzorini	14413	

Date: 7-24-17
 Expire Date:
 7/24/19
 Prepared by:
 Cesar A. Mejia

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040163
UOM	BAC
Last Check	05/30/2019
Time:	06:16 AM
Cal Standard	100
Result	.091
Last Cal	04/12/2019
Time:	07:30 AM
Cal Standard	100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									YES	NO				
05/30/2019	6:10	34	17GP41049	.100	22		.095			X		B. Marzorini	14413	

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080013
UOM	BAC
Last Check	05/30/2019
Time:	06:11 AM
Cal Standard	.100
Result	.095
Last Cal	01/29/2018
Time:	12:59 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 ID. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
									YES	NO				
05/30/2019	6:08	34	17GP41049	.100	23		104			X		B. Marzorini	14413	
Lifeloc Technologies														
Calibration / Check Printout														
				FC20	8.7.24		Serial No.		16080014		UOM		BAC	
				Last Check	05/30/2019		Time:		06:08 AM		Cal Standard		.100	
				Result	.104		Last Cal		05/20/2019		Time:		1:28 PM	
				Cal Standard	.100									

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

9

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini ID. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									YES	NO				
05/30/2019	6:10	34	17GP41049	.100	22		.093			X		B. Marzorini	14413	
Lifeloc Technologies														
Calibration / Check Printout														
					FC20	8.7.24								
					Serial No.	16080015								
					UOM	BAC								
					Last Check	05/30/2019								
					Time:	06:11 AM								
					Cal Standard	.100								
					Result	.093								
					Last Cal	03/28/2019								
					Time:	3:58 PM								
					Cal Standard	.100								

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
05/30/2019	6:08	34	17GP41049	.100	23		.095			B. Marzorini	14413	

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080027
UOM	BAC
Last Check	05/30/2019
Time:	06:09 AM
Cal Standard	.100
Result	.095
Last Cal	03/28/2019
Time:	3:58 PM
Cal Standard	.100

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES	NO			
05/30/2019	06:13	34		17GP41049	.100		21		.093			X		B. Marzorini	14413	

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080028
UOM	BAC
Last Check	05/30/2019
Time:	06:13 AM
Cal Standard	.100
Result	.093
Last Cal	03/28/2019
Time:	4:05 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									YES	NO				
05/30/2019	6:14	34	17GP41049	.100	22		.094			X		B. Marzorini	14413	
Lifeloc Technologies														
Calibration / Check Printout														
					FC20	8.7.24								
					Serial No.	16080030								
					UOM	BAC								
					Last Check	05/30/2019								
					Time:	06:14 AM								
					Cal Standard	.100								
					Result	.094								
					Last Cal	03/28/2019								
					Time:	4:07 PM								
					Cal Standard	.100								

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
									YES	NO					
05/30/2019	06:25	34	17GP41049	.100	23		.093			X		B. Marzorini	14413		
Lifeloc Technologies															
Calibration / Check Printout															
		FC20		8.7.24		Serial No.		16080064		UOM		BAC		Last Check	
				05/30/2019		Time:		06:25 AM		Cal Standard		.100		Result	
														.093	
														Last Cal	
														03/28/2019	
														Time:	
														4:12 PM	
														Cal Standard	
														.100	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#114

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					YES	NO				
05/30/2019	02:34		17GP41049	.100	21		.096		X		B. Marzorini	14413	
Lifeloc Technologies													
Calibration / Check Printout													
				FC20	8.7.24						Serial No.		16080065
				UOM	BAC						Last Check		05/30/2019
				Time:	06:27 AM						Cal Standard		.100
				Result	.096						Last Cal		12/19/2017
				Time:	1:54 PM						Cal Standard		.100

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzotini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																										
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO																														
05/30/2019	06:31	34	17GP41049		.100	21		.092					B. Marzotini	14413																											
Lifeloc Technologies																																									
Calibration / Check Printout																																									
		FC20		8.7.24		Serial No.		16080069		UOM		BAC		Last Check		05/30/2019		Time:		06:31 AM		Cal Standard		.100		Result		.092		Last Cal		03/28/2019		Time:		4:18 PM		Cal Standard		.100	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/30/2019	6:31	34	17GP41049	.100	.21		.094					B. Marzorini	14413	

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16080071
UOM	BAC
Last Check	05/30/2019
Time:	06:32 AM
Cal Standard	.100
Result	.094
Last Cal	08/14/2017
Time:	10:41 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#117

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
05/30/2019	06:47:34		17GP41049	100	21						B. Marzorini	14413	

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080072
UOM	BAC
Last Check	05/30/2019
Time:	06:48 AM
Cal Standard	100
Result	.094
Last Cal	12/19/2017
Time:	2:05 PM
Cal Standard	100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
05/30/2019	6:41	34	17GP41049	.100	21	-	.093		X		B. Marzorini	14413	
Lifeloc Technologies													
Calibration / Check Printout													
		FC20		8.7.24		Serial No.		16080075		UOM		BAC	
		Last Check		05/30/2019		Time:		06:41 AM		Cal Standard		.100	
		Result		.093		Last Cal		03/28/2019		Time:		4:24 PM	
		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 ID. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.							YES	NO				
05/30/2019	6:41	37	17GP41049	100	21		.095			X			B. Marzorini	14413	

Date: 7-24-17
 Expire Date: 7/24/19
 Prepared by: Cesar A. Mejia

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080077
UOM	BAC
Last Check	05/30/2019
Time:	06:42 AM
Cal Standard	.100
Result	.095
Last Cal	03/28/2019
Time:	4:30 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 16/20175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/30/2019	6:44	34	17GPA1049	.100	21		.093			X		B. Marzorini	14413	
Lifeloc Technologies Calibration / Check Printout														
				FC20	8.7.24		Serial No.		16120175		UOM		BAC	
				Last Check	05/30/2019		Time:		06:45 AM		Cal Standard		.100	
				Result	.093		Last Cal		03/28/2019		Time:		4:36 PM	
				Cal Standard	.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 ID. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/ FC20
 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
							YES	NO				
05/30/2019	6:51	34	17GP41049	.100	21				X	B. Marzorini	14413	
Lifeloc Technologies												
Calibration / Check Printout												
		FC20		8.7.24								
		Serial No.		16120177								
		UOM		BAC								
		Last Check		05/30/2019								
		Time:		06:51 AM								
		Cal Standard		.100								
		Result		.095								
		Last Cal		04/25/2017								
		Time:		09:50 AM								
		Cal Standard		.100								

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.0990 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: B. Marzorini | I.D. NUMBER: 14413 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#22

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
										YES	NO				
05/30/2019	6:56	3Y	17GP41049	.100	21		.091						B. Marzorini	14413	
Lifeloc Technologies															
Calibration / Check Printout															
				8.7.24		16120188		BAC		05/30/2019		06:51 AM		.100	
				FC20		Serial No.		UOM		Last Check		Time:		Cal Standard	
														.100	
														.091	
														10/21/2016	
														08:16 AM	
														.100	
														Cal Standard	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: B. Marzorini
I.D. NUMBER: 14413
PAS MAKE/MODEL: Lifeloc/ FC20
PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#23

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									YES	NO				
05/30/2019	6:54	34	17GP41049	.100	21						X	B. Marzorini	14413	
Lifeloc Technologies Calibration / Check Printout														
				FC20	8.7.24									
				Serial No.	16120189									
				UOM	BAC									
				Last Check	05/30/2019									
				Time:	06:55 AM									
				Cal Standard	.100									
				Result	.092									
				Last Cal	04/12/2019									
				Time:	08:50 AM									
				Cal Standard	.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

H24

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzotini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							YES	NO					
05/30/2019	6:52	34	17GP41049	.100	21		X			B. Marzotini	14413	NEEDED CAL BRATTOW	
Lifeloc Technologies													
Calibration / Check Printout													
		FC20		8.7.24		Serial No.		16120194		UOM		BAC	
		Last Check		05/30/2019		Time:		06:53 AM		Cal Standard		.100	
		Result		.098		Last Cal		05/30/2019		Time:		06:53 AM	
		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/ FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							YES	NO					
05/30/2019	6:59	34	17GP41049	.100	21		X		101	B. Marzorini	14413	NEED CALIBRATION	
Lifeloc Technologies													
Calibration / Check Printout													
		FC20		8.7.24		Serial No.		16120195		UOM		BAC	
		Last Check		05/30/2019		Time:		07:02 AM		Cal Standard		.100	
		Result		.101		Last Cal		05/30/2019		Time:		07:01 AM	
		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: B. Marzotini | ID. NUMBER: 14413 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO					YES
05/30/2019	7:06	37	17GP41049	.100	2	1								B. Marzotini	14413	NEEDED CALIBRATION	
Lifeloc Technologies Calibration / Check Printout																	
		FC20		8.7.24		Serial No.		16120197		UOM		BAC		Last Check		05/30/2019	
		Time:		07:08 AM		Cal Standard		.100		Result		.100		Last Cal		05/30/2019	
		Time:		07:07 AM		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzorini ID. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
05/30/2019	7:10 3Y		17GP41049	.100	20		.095		X		B. Marzorini	14413	
Lifeloc Technologies													
Calibration / Check Printout													
						8.7.24							
						Serial No.	17170066						
						UOM	BAC						
						Last Check	05/30/2019						
						Time:	07:10 AM						
						Cal Standard	.100						
						Result	.095						
						Last Cal	03/28/2019						
						Time:	4:42 PM						
						Cal Standard	.100						

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 17170150

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2		YES	NO			
05/30/2019			17GP41049	.100										B. Marzorini	14413	OUT OF SERVICE

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

29

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
05/30/2019	7:12	34	17GP41049	.100	20		.093			B. Marzorini	14413	
Lifeloc Technologies Calibration / Check Printout												
FC20		8.7.24		Serial No.		17170155		UOM		BAC		
Last Check		05/30/2019		Time:		07:13 AM		Cal Standard		.100		
Result		.093		Last Cal		03/28/2019		Time:		4:46 PM		
Cal Standard		.100		Cal Standard		.100						

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: B. Marzotini I.D. NUMBER: 14413 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 90102844-82

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
05/30/2019	7:15	34	17GP41049	.100					X		B. Marzotini	14413	
Lifeloc Technologies													
Calibration / Check Printout													
		FC20		6.24c		Serial No.		90102844		UOM		BAC	
		Last Check		05/30/2019		Time:		07:19 AM		Cal Standard		.100	
		Result		.096		Last Cal		03/28/2019		Time:		5:06 PM	
		Cal Standard		.100									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions Chp202k_1108.pat

#31

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: B. Marzorini
 I.D. NUMBER: 14413
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
									YES	NO					
05/30/2019	7:15	34	17GP41049	.100	20		.096			X		B. Marzorini	14413		
Lifeloc Technologies															
Calibration / Check Printout															
		FC20		8.7.24		Serial No.		17170064		UOM		BAC		Last Check	
		Time:		07:15 AM		Cal Standard		.100		Result		.096		Last Cal	
		Time:		5:05 PM		Cal Standard		.100							
Date: 7-24-17 Expire Date: 7/24/19 Prepared by: Cesar A. Mejia															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Num	UOM
37	0.128		5:20 AM	05/25/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140	
38	0.125		5:22 AM	05/25/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140	
39	0.161		8:52 PM	05/26/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140	
40	0.153		8:53 PM	05/26/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140	
41	0.053		1:42 AM	05/30/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140	
42	0.053		1:44 AM	05/30/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140	
23	0.067		10:16 PM	05/26/2019	3:41 PM	03/28/2019	0.104	9:20 AM	05/20/2019	16040141	
24	0.07		10:18 PM	05/26/2019	3:41 PM	03/28/2019	0.104	9:20 AM	05/20/2019	16040141	
25	0.191		3:14 PM	05/28/2019	3:41 PM	03/28/2019	0.104	9:20 AM	05/20/2019	16040141	
26	0.187		3:15 PM	05/28/2019	3:41 PM	03/28/2019	0.104	9:20 AM	05/20/2019	16040141	
214	0.108		1:02 AM	05/24/2019	8:56 AM	04/25/2017	0.104	9:27 AM	05/20/2019	16040161	
215	0.114		1:04 AM	05/24/2019	8:56 AM	04/25/2017	0.104	9:27 AM	05/20/2019	16040161	
371	0.1		11:37 PM	05/22/2019	7:30 AM	04/12/2019	0.099	9:34 AM	05/20/2019	16040163	
372	0.1		11:40 PM	05/22/2019	7:30 AM	04/12/2019	0.099	9:34 AM	05/20/2019	16040163	
373	0		3:57 AM	05/25/2019	7:30 AM	04/12/2019	0.099	9:34 AM	05/20/2019	16040163	
374	0		11:54 PM	05/27/2019	7:30 AM	04/12/2019	0.099	9:34 AM	05/20/2019	16040163	
375	0.206		1:26 AM	05/30/2019	7:30 AM	04/12/2019	0.099	9:34 AM	05/20/2019	16040163	
78	0.131		11:40 PM	05/22/2019	12:59 PM	01/29/2018	0.102	9:39 AM	05/20/2019	16080013	
79	0.127		11:51 PM	05/22/2019	12:59 PM	01/29/2018	0.102	9:39 AM	05/20/2019	16080013	
80	0.126		11:54 PM	05/22/2019	12:59 PM	01/29/2018	0.102	9:39 AM	05/20/2019	16080013	
81	0.009		5:24 PM	05/27/2019	12:59 PM	01/29/2018	0.102	9:39 AM	05/20/2019	16080013	
82	0		5:27 PM	05/27/2019	12:59 PM	01/29/2018	0.102	9:39 AM	05/20/2019	16080013	
34	NEG		1:22 PM	05/20/2019	3:52 PM	03/28/2019	0.097	10:55 AM	05/10/2019	16080014	
35	0		1:23 PM	05/20/2019	3:52 PM	03/28/2019	0.097	10:55 AM	05/10/2019	16080014	
36	0		1:23 PM	05/20/2019	3:52 PM	03/28/2019	0.097	10:55 AM	05/10/2019	16080014	
37	0.083		2:13 AM	05/24/2019	1:28 PM	05/20/2019	0.101	1:28 PM	05/20/2019	16080014	
38	0.083		2:17 AM	05/24/2019	1:28 PM	05/20/2019	0.101	1:28 PM	05/20/2019	16080014	
39	0.078		2:34 AM	05/25/2019	1:28 PM	05/20/2019	0.101	1:28 PM	05/20/2019	16080014	
40	0.072		2:36 AM	05/25/2019	1:28 PM	05/20/2019	0.101	1:28 PM	05/20/2019	16080014	
41	0		10:40 PM	05/27/2019	1:28 PM	05/20/2019	0.101	1:28 PM	05/20/2019	16080014	
42	0		11:34 PM	05/27/2019	1:28 PM	05/20/2019	0.101	1:28 PM	05/20/2019	16080014	
48	0.212		12:58 AM	05/20/2019	3:58 PM	03/28/2019	0.095	10:57 AM	05/10/2019	16080015	
49	0.213		1:01 AM	05/20/2019	3:58 PM	03/28/2019	0.095	10:57 AM	05/10/2019	16080015	

37	0.128	5:20 AM	05/25/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140
50	0.083	12:11 AM	05/27/2019	3:58 PM	03/28/2019	0.097	9:43 AM	05/20/2019	16080015
51	0.088	12:14 AM	05/27/2019	3:58 PM	03/28/2019	0.097	9:43 AM	05/20/2019	16080015
52	0.169	3:17 AM	05/27/2019	3:58 PM	03/28/2019	0.097	9:43 AM	05/20/2019	16080015
53	0.163	3:20 AM	05/27/2019	3:58 PM	03/28/2019	0.097	9:43 AM	05/20/2019	16080015
37	0.01	10:36 PM	05/22/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
38	0.165	1:30 AM	05/25/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
39	0.171	1:33 AM	05/25/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
40	0.169	3:11 AM	05/25/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
41	0.166	3:13 AM	05/25/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
42	0.146	4:07 AM	05/26/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
43	0.142	4:10 AM	05/26/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
44	0	3:43 AM	05/27/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
45	0.163	11:32 PM	05/27/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
46	0.174	11:34 PM	05/27/2019	3:58 PM	03/28/2019	0.102	9:42 AM	05/20/2019	16080027
16	0.102	11:31 PM	05/26/2019	4:07 PM	03/28/2019	0.101	9:55 AM	05/20/2019	16080030
17	0.106	11:34 PM	05/26/2019	4:07 PM	03/28/2019	0.101	9:55 AM	05/20/2019	16080030
20	0.147	9:40 PM	05/27/2019	4:12 PM	03/28/2019	0.101	9:58 AM	05/20/2019	16080064
21	0.146	9:43 PM	05/27/2019	4:12 PM	03/28/2019	0.101	9:58 AM	05/20/2019	16080064
22	0.09	4:36 AM	05/29/2019	4:12 PM	03/28/2019	0.101	9:58 AM	05/20/2019	16080064
23	0.092	4:39 AM	05/29/2019	4:12 PM	03/28/2019	0.101	9:58 AM	05/20/2019	16080064
218	0.151	12:34 AM	05/27/2019	1:54 PM	12/19/2017	0.107	10:00 AM	05/20/2019	16080065
219	0.143	12:38 AM	05/27/2019	1:54 PM	12/19/2017	0.107	10:00 AM	05/20/2019	16080065
89	0.041	12:06 AM	05/26/2019	4:18 PM	03/28/2019	0.098	10:05 AM	05/20/2019	16080069
90	0.284	12:57 AM	05/26/2019	4:18 PM	03/28/2019	0.098	10:05 AM	05/20/2019	16080069
91	0.26	1:00 AM	05/26/2019	4:18 PM	03/28/2019	0.098	10:05 AM	05/20/2019	16080069
92	0.252	1:02 AM	05/26/2019	4:18 PM	03/28/2019	0.098	10:05 AM	05/20/2019	16080069
93	0.089	2:10 AM	05/27/2019	4:18 PM	03/28/2019	0.098	10:05 AM	05/20/2019	16080069
94	0.082	2:13 AM	05/27/2019	4:18 PM	03/28/2019	0.098	10:05 AM	05/20/2019	16080069
378	0.094	1:21 AM	05/25/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
379	0.09	1:25 AM	05/25/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
380	0.071	2:59 AM	05/26/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
381	0.068	3:02 AM	05/26/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
382	0.192	3:53 AM	05/26/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071

37	0.128	5:20 AM	05/25/2019	7:46 AM	04/12/2019	0.098	10:16 AM	05/20/2019	16040140
383	0.172	3:55 AM	05/26/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
384	0.176	3:58 AM	05/26/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
385	0	2:18 AM	05/28/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
386	0.073	2:22 AM	05/28/2019	10:41 AM	08/14/2017	0.101	10:06 AM	05/20/2019	16080071
12	0	9:47 PM	05/25/2019	2:05 PM	12/19/2017	0.104	10:25 AM	05/20/2019	16080072
13	0.171	5:16 AM	05/26/2019	2:05 PM	12/19/2017	0.104	10:25 AM	05/20/2019	16080072
14	0.178	5:20 AM	05/26/2019	2:05 PM	12/19/2017	0.104	10:25 AM	05/20/2019	16080072
55	0.159	10:49 PM	05/24/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
56	0.164	10:51 PM	05/24/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
57	0.161	2:08 AM	05/25/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
58	0.156	2:11 AM	05/25/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
59	0	11:13 PM	05/25/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
60	0.147	1:46 AM	05/26/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
61	0.146	1:49 AM	05/26/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
62	0.071	3:41 AM	05/26/2019	4:30 PM	03/28/2019	0.104	10:25 AM	05/20/2019	16080077
84	0.181	11:03 PM	05/27/2019	8:16 AM	10/21/2016	0.1	10:47 AM	05/20/2019	16120188
309	0.166	2:31 AM	05/20/2019	8:50 AM	01/29/2018	0.09	1:09 PM	05/10/2019	16120194
310	0.166	2:34 AM	05/20/2019	8:50 AM	01/29/2018	0.09	1:09 PM	05/10/2019	16120194
311	0	2:17 AM	05/21/2019	8:50 AM	01/29/2018	0.093	10:49 AM	05/20/2019	16120194
312	0.129	1:54 AM	05/26/2019	8:50 AM	01/29/2018	0.093	10:49 AM	05/20/2019	16120194
313	0.122	1:57 AM	05/26/2019	8:50 AM	01/29/2018	0.093	10:49 AM	05/20/2019	16120194
314	0.119	1:42 AM	05/27/2019	8:47 AM	04/25/2017	0.103	11:01 AM	05/20/2019	16120195
315	0.251	1:43 AM	05/27/2019	8:47 AM	04/25/2017	0.103	11:01 AM	05/20/2019	16120195
316	0.122	1:45 AM	05/27/2019	8:47 AM	04/25/2017	0.103	11:01 AM	05/20/2019	16120195
307	0.261	6:52 PM	05/25/2019	10:22 AM	10/21/2016	0.097	11:08 AM	05/20/2019	16120197
308	0.29	6:55 PM	05/25/2019	10:22 AM	10/21/2016	0.097	11:08 AM	05/20/2019	16120197
309	0.28	6:57 PM	05/25/2019	10:22 AM	10/21/2016	0.097	11:08 AM	05/20/2019	16120197
310	0.171	2:32 AM	05/26/2019	10:22 AM	10/21/2016	0.097	11:08 AM	05/20/2019	16120197
311	0.16	2:35 AM	05/26/2019	10:22 AM	10/21/2016	0.097	11:08 AM	05/20/2019	16120197
312	0	5:08 AM	05/26/2019	10:22 AM	10/21/2016	0.097	11:08 AM	05/20/2019	16120197