

CHP South LA PAS Usage Logs 6-19-19

Test #	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
38	0.163	7:03 AM	06/16/2019	3:25 PM	03/28/2019	0.094	5:52 AM	05/30/2019	16040119	FC20
39	0.159	7:05 AM	06/16/2019	3:25 PM	03/28/2019	0.094	5:52 AM	05/30/2019	16040119	FC20
27	0.136	2:32 AM	06/17/2019	3:41 PM	03/28/2019	0.1	11:47 AM	06/09/2019	16040141	FC20
28	0.135	2:34 AM	06/17/2019	3:41 PM	03/28/2019	0.1	11:47 AM	06/09/2019	16040141	FC20
220	0.025	2:46 AM	06/11/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
221	0.079	11:07 PM	06/15/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
222	0.075	11:09 PM	06/15/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
223	0.111	12:13 AM	06/16/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
224	0.124	12:15 AM	06/16/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
225	0.155	3:15 AM	06/16/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
226	0.152	3:18 AM	06/16/2019	8:56 AM	04/25/2017	0.102	11:54 AM	06/09/2019	16040161	FC20
301	0	8:45 PM	06/15/2019	5:42 AM	05/30/2019	0.108	11:36 AM	06/09/2019	16040162	FC20
302	0.25	5:52 PM	06/16/2019	5:42 AM	05/30/2019	0.108	11:36 AM	06/09/2019	16040162	FC20
303	0.249	5:55 PM	06/16/2019	5:42 AM	05/30/2019	0.108	11:36 AM	06/09/2019	16040162	FC20
304	0.01	12:07 AM	06/17/2019	5:42 AM	05/30/2019	0.108	11:36 AM	06/09/2019	16040162	FC20
380	0.216	11:56 PM	06/15/2019	7:30 AM	04/12/2019	0.095	11:47 AM	06/09/2019	16040163	FC20
381	0.213	11:58 PM	06/15/2019	7:30 AM	04/12/2019	0.095	11:47 AM	06/09/2019	16040163	FC20
85	0	12:05 AM	06/10/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
86	0.219	11:09 PM	06/11/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
87	0.232	11:11 PM	06/11/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
88	0	12:10 AM	06/14/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
89	0.162	1:16 AM	06/16/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
90	0.143	1:19 AM	06/16/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
91	0	2:55 AM	06/16/2019	12:59 PM	01/29/2018	0.099	12:03 PM	06/09/2019	16080013	FC20
59	0.1	12:01 PM	06/09/2019	3:58 PM	03/28/2019	0.093	6:11 AM	05/30/2019	16080015	FC20
60	0.122	10:32 PM	06/15/2019	3:58 PM	03/28/2019	0.097	12:05 PM	06/09/2019	16080015	FC20
61	0.119	10:35 PM	06/15/2019	3:58 PM	03/28/2019	0.097	12:05 PM	06/09/2019	16080015	FC20
62	0.181	2:39 AM	06/17/2019	3:58 PM	03/28/2019	0.097	12:05 PM	06/09/2019	16080015	FC20
63	0.178	2:41 AM	06/17/2019	3:58 PM	03/28/2019	0.097	12:05 PM	06/09/2019	16080015	FC20
51	0.106	1:37 AM	06/09/2019	3:58 PM	03/28/2019	0.095	6:09 AM	05/30/2019	16080027	FC20
52	0.108	1:41 AM	06/09/2019	3:58 PM	03/28/2019	0.095	6:09 AM	05/30/2019	16080027	FC20
53	0.111	1:43 AM	06/09/2019	3:58 PM	03/28/2019	0.095	6:09 AM	05/30/2019	16080027	FC20
24	0	1:55 AM	06/15/2019	4:05 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080028	FC20

CHP South LA PAS Usage Logs 6-19-19

Test #	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
25	0.148	3:07 AM	06/15/2019	4:05 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080028	FC20
26	0.137	3:11 AM	06/15/2019	4:05 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080028	FC20
24	0.356	5:39 PM	06/09/2019	4:07 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080030	FC20
25	0.344	5:42 PM	06/09/2019	4:07 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080030	FC20
26	0.159	2:01 AM	06/15/2019	4:07 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080030	FC20
27	0.163	2:04 AM	06/15/2019	4:07 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080030	FC20
28	0	10:33 PM	06/15/2019	4:07 PM	03/28/2019	0.098	12:08 PM	06/09/2019	16080030	FC20
28	0.055	2:14 AM	06/09/2019	4:12 PM	03/28/2019	0.093	6:25 AM	05/30/2019	16080064	FC20
29	0.052	2:16 AM	06/09/2019	4:12 PM	03/28/2019	0.093	6:25 AM	05/30/2019	16080064	FC20
30	0.095	3:24 AM	06/09/2019	4:12 PM	03/28/2019	0.093	6:25 AM	05/30/2019	16080064	FC20
31	0.096	3:28 AM	06/09/2019	4:12 PM	03/28/2019	0.093	6:25 AM	05/30/2019	16080064	FC20
230	0.113	3:18 AM	06/09/2019	1:54 PM	12/19/2017	0.096	6:27 AM	05/30/2019	16080065	FC20
231	0.111	3:21 AM	06/09/2019	1:54 PM	12/19/2017	0.096	6:27 AM	05/30/2019	16080065	FC20
232	0.155	2:57 AM	06/16/2019	1:54 PM	12/19/2017	0.1	12:12 PM	06/09/2019	16080065	FC20
233	0.141	3:00 AM	06/16/2019	1:54 PM	12/19/2017	0.1	12:12 PM	06/09/2019	16080065	FC20
234	0.119	1:20 AM	06/18/2019	1:54 PM	12/19/2017	0.1	12:12 PM	06/09/2019	16080065	FC20
235	0.114	1:23 AM	06/18/2019	1:54 PM	12/19/2017	0.1	12:12 PM	06/09/2019	16080065	FC20
98	0.171	2:45 AM	06/17/2019	4:18 PM	03/28/2019	0.098	12:16 PM	06/09/2019	16080069	FC20
99	0.159	2:49 AM	06/17/2019	4:18 PM	03/28/2019	0.098	12:16 PM	06/09/2019	16080069	FC20
387	0.126	1:57 AM	06/17/2019	10:41 AM	08/14/2017	0.097	12:24 PM	06/09/2019	16080071	FC20
388	0.128	1:59 AM	06/17/2019	10:41 AM	08/14/2017	0.097	12:24 PM	06/09/2019	16080071	FC20
72	0	2:53 AM	06/17/2019	4:30 PM	03/28/2019	0.096	12:32 PM	06/09/2019	16080077	FC20
30	0.119	11:50 PM	06/11/2019	4:36 PM	03/28/2019	0.102	12:33 PM	06/09/2019	16120175	FC20
31	0.116	11:53 PM	06/11/2019	4:36 PM	03/28/2019	0.102	12:33 PM	06/09/2019	16120175	FC20
32	0.09	1:50 AM	06/16/2019	4:36 PM	03/28/2019	0.102	12:33 PM	06/09/2019	16120175	FC20
33	0.089	1:52 AM	06/16/2019	4:36 PM	03/28/2019	0.102	12:33 PM	06/09/2019	16120175	FC20
34	0.088	2:02 AM	06/16/2019	4:36 PM	03/28/2019	0.102	12:33 PM	06/09/2019	16120175	FC20
35	0.09	2:04 AM	06/16/2019	4:36 PM	03/28/2019	0.102	12:33 PM	06/09/2019	16120175	FC20
321	0.135	2:14 AM	06/16/2019	12:48 PM	06/09/2019	0.099	12:48 PM	06/09/2019	16120195	FC20
322	0.146	2:17 AM	06/16/2019	12:48 PM	06/09/2019	0.099	12:48 PM	06/09/2019	16120195	FC20
31	0.125	11:47 PM	06/16/2019	5:05 PM	03/28/2019	0.101	12:59 PM	06/09/2019	17170064	FC20
32	0.127	11:50 PM	06/16/2019	5:05 PM	03/28/2019	0.101	12:59 PM	06/09/2019	17170064	FC20
33	0	10:13 PM	06/18/2019	5:05 PM	03/28/2019	0.101	12:59 PM	06/09/2019	17170064	FC20

#1

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1454 34	17GP41049	.100	23	2/105						M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	05:42 AM	05/30/2019	.105	.100	2:55 PM	06/19/2019	BAC	16040162	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1505	34	17GP41049	.100	22	.095	X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	3:25 PM	03/28/2019	.095	.100	3:06 PM	06/19/2019	BAC	16040119	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	153234	17GP41049	.100	28	.101			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:41 PM	03/28/2019	.101	.100	3:33 PM	06/19/2019	BAC	16040141	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS WAKEMODEL: Lifeloc FC20 PAS SERIAL NO: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1	TEST #2	YES	NO					
06/19/2019	1555	34	17GP41049	.100	22		.098					X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:56 AM	04/25/2017	.098	.100	3:16 PM	06/19/2019	BAC	16040161	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1572	34	17GP41049	.100	23		093				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date: 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.093	.100	3:13 PM	06/19/2019	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1506	54	17GP41049	.100	27	.092	X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.092	.100	3:06 PM	06/19/2019	BAC	16080013	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

49

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2019	1541	34	17GP41049	.100	27	.094		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100		3:58 PM	03/28/2019	.094	.100	3:44 PM	06/19/2019	BAC	16080015	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)					
06/19/2019	15/16	39	17GP41049	.100	22		.098		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.098	.100	3:16 PM	06/19/2019	BAC	16080027	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-09) OPI 051

11

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	152034	17GP41049	.100	22	.096			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.096	.100	3:20 PM	06/19/2019	BAC	16080028	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1521	54	17GR41049	.100	21		.094			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:07 PM	03/28/2019	.094	.100	3:22 PM	06/19/2019	BAC	16080030	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#13

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2019	152337		17GP41049	.100	22		.097				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date: 07-24-19

Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100		4:12 PM	03/28/2019	.097	.100	3:23 PM	06/19/2019	BAC	16080064	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2019	1555	34	17GP41049	.100	22	.099					M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	2:05 PM	12/19/2017	.099	.100	3:56 PM	06/19/2019	BAC	16080072	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1 (Optional)	TEST #2 (Optional)			
06/19/2019	155034	17GP41049		.100	21	.095						M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:24 PM	03/28/2019	.095	.100	3:50 PM	06/19/2019	BAC	16080075	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1558	34	17GP41049	.100	22		.099					M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:30 PM	03/28/2019	.099	.100	3:58 PM	06/19/2019	BAC	16080077	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	155834		17GP41049	.100	22		.101			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.101	.100	3:59 PM	06/19/2019	BAC	16120175	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEMP (°C)				TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1605 34	17	17GP41049		.100	22		.099			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	09:50 AM	04/25/2017	.099	.100	4:05 PM	06/19/2019	BAC	16120177	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)						
06/19/2019	160334		17GPA1049	.100	22		.095		YES	NO	X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.095	.100	4:04 PM	06/19/2019	BAC	16120188	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	16/0	24	17GP41049	.100	22	.096	X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.096	.100	4:10 PM	06/19/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2019	1607	34	17GP41049		22	.100	.099				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	06:53 AM	05/30/2019	.099	.100	4:07 PM	06/19/2019	BAC	16120194	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2019	1612 34		17GP41049	.100	22		.099				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:48 PM	06/09/2019	.099	.100	4:12 PM	06/19/2019	BAC	16120195	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

26

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2019	6:16	34	I7GP41049	.100	22		.106				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	07:07 AM	05/30/2019	.106	.100	4:17 PM	06/19/2019	BAC	16120197	8.7.24	

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
06/19/2019	1619 34	17	GP41049	100	21	099					X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date: 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:42 PM	03/28/2019	.099	.100	4:19 PM	06/19/2019	BAC	17170066	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2019	1626 Jy	17	17GP41049		.100		21		.096			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:46 PM	03/28/2019	.096	.100	4:21 PM	06/19/2019	BAC	17170155	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

