

Test Number	Result	Test Time	Test Date	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
305	0.058	8:54 AM	06/23/2019	05/30/2019	0.105	2:55 PM	06/19/2019	16040162	FC20
306	0.126	5:21 PM	06/23/2019	05/30/2019	0.105	2:55 PM	06/19/2019	16040162	FC20
307	0.127	5:23 PM	06/23/2019	05/30/2019	0.105	2:55 PM	06/19/2019	16040162	FC20
382	0.264	6:34 PM	06/23/2019	04/12/2019	0.093	3:13 PM	06/19/2019	16040163	FC20
383	0.238	6:35 PM	06/23/2019	04/12/2019	0.093	3:13 PM	06/19/2019	16040163	FC20
27	0.079	1:37 AM	06/23/2019	03/28/2019	0.096	3:20 PM	06/19/2019	16080028	FC20
28	0.081	1:39 AM	06/23/2019	03/28/2019	0.096	3:20 PM	06/19/2019	16080028	FC20
29	0.126	12:26 AM	06/23/2019	03/28/2019	0.094	3:22 PM	06/19/2019	16080030	FC20
30	0.126	12:29 AM	06/23/2019	03/28/2019	0.094	3:22 PM	06/19/2019	16080030	FC20
31	0	9:07 PM	06/23/2019	03/28/2019	0.094	3:22 PM	06/19/2019	16080030	FC20
32	0.175	1:43 AM	06/23/2019	03/28/2019	0.097	3:23 PM	06/19/2019	16080064	FC20
33	0.168	1:45 AM	06/23/2019	03/28/2019	0.097	3:23 PM	06/19/2019	16080064	FC20
236	0.125	12:35 AM	06/25/2019	12/19/2017	0.099	3:24 PM	06/19/2019	16080065	FC20
237	0.119	12:38 AM	06/25/2019	12/19/2017	0.099	3:24 PM	06/19/2019	16080065	FC20
100	0.185	2:03 AM	06/22/2019	03/28/2019	0.094	3:29 PM	06/19/2019	16080069	FC20
101	0.184	2:07 AM	06/22/2019	03/28/2019	0.094	3:29 PM	06/19/2019	16080069	FC20
15	0.087	10:29 PM	06/23/2019	12/19/2017	0.099	3:56 PM	06/19/2019	16080072	FC20
16	0.091	10:32 PM	06/23/2019	12/19/2017	0.099	3:56 PM	06/19/2019	16080072	FC20
11	0	12:05 AM	06/22/2019	03/28/2019	0.095	3:50 PM	06/19/2019	16080075	FC20
12	0	12:27 AM	06/22/2019	03/28/2019	0.095	3:50 PM	06/19/2019	16080075	FC20
13	0.037	2:53 AM	06/22/2019	03/28/2019	0.095	3:50 PM	06/19/2019	16080075	FC20
73	0.161	11:22 PM	06/22/2019	03/28/2019	0.099	3:58 PM	06/19/2019	16080077	FC20
74	0.159	11:26 PM	06/22/2019	03/28/2019	0.099	3:58 PM	06/19/2019	16080077	FC20
75	0.036	1:52 AM	06/23/2019	03/28/2019	0.099	3:58 PM	06/19/2019	16080077	FC20
76	0.036	1:54 AM	06/23/2019	03/28/2019	0.099	3:58 PM	06/19/2019	16080077	FC20
77	0.028	3:55 AM	06/23/2019	03/28/2019	0.099	3:58 PM	06/19/2019	16080077	FC20
78	0.042	4:25 AM	06/23/2019	03/28/2019	0.099	3:58 PM	06/19/2019	16080077	FC20
318	0.213	1:43 AM	06/22/2019	05/30/2019	0.099	4:07 PM	06/19/2019	16120194	FC20
319	0.203	1:45 AM	06/22/2019	05/30/2019	0.099	4:07 PM	06/19/2019	16120194	FC20
323	0.184	3:29 AM	06/22/2019	06/09/2019	0.099	4:12 PM	06/19/2019	16120195	FC20
324	0.173	3:32 AM	06/22/2019	06/09/2019	0.099	4:12 PM	06/19/2019	16120195	FC20
325	0.146	9:45 PM	06/23/2019	06/09/2019	0.099	4:12 PM	06/19/2019	16120195	FC20
326	0.134	9:47 PM	06/23/2019	06/09/2019	0.099	4:12 PM	06/19/2019	16120195	FC20

CHP South LA PAS usage logs 6-25-19

Test Number	Result	Test Time	Test Date	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
34	0.214	10:58 PM	06/21/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
35	0.209	11:00 PM	06/21/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
36	0.078	3:31 AM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
37	0.077	3:34 AM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
38	0.077	3:40 AM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
39	0.171	6:32 AM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
40	0.182	6:34 AM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
41	0.107	11:05 PM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
42	0.121	11:08 PM	06/23/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20
43	0.232	10:01 PM	06/24/2019	03/28/2019	0.101	4:07 PM	06/19/2019	17170064	FC20

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2019	1608	34	17GP41049	.100	22	.111		X		.099	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date: 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:09 PM	06/25/2019	.099	.100	4:11 PM	06/25/2019	BAC	16040162	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 081

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2019	1614	34	17GP41049	.100	22	.096		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	07:46 AM	04/12/2019	.096	.100	4:14 PM	06/25/2019	BAC	16040140	8.7.24	
Calibration / Check Printout										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/25/2019	16:22	34	17GP41049	.100	22	.100				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.100	.100	4:23 PM	06/25/2019	BAC	16040161	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/25/2019	16/19	34	17GP41049		.100	23	.095	X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.095	.100	4:19 PM	06/25/2019	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES				
06/25/2019	1601	34	17GP41049	.100	22	.099	X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19												
			Cal Standard		Time:							
				.100	12:59 PM							
					Last Cal							
				.099	01/29/2018							
			Cal Standard		Time:							
				.100	4:02 PM							
					Last Check							
					06/25/2019							
			UOM		BAC							
			Serial No.		16080013							
			FC20		8.7.24							
Calibration / Check Printout												Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 160 800 14

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)					TEST #2 (Optional)	YES
06/25/2019	1557 34		17GP41049		.100	23	.108			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	1:28 PM	05/20/2019	.108	.100	3:58 PM	06/25/2019	BAC	16080014	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1			
06/25/2019	1603 34		17GP41049	.100	22		.999					M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.099	.100	4:03 PM	06/25/2019	BAC	16080027	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

4/11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: **16080028**

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES				
06/25/2019	<i>160734</i>		17GP41049		.100					X		

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.098	.100	4:07 PM	06/25/2019	BAC	16080028	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				TEST #1 (Optional)	TEST #2 (Optional)					
06/25/2019	160734		17GP41049	.100	21		.099		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:07 PM	03/28/2019	.099	.100	4:07 PM	06/25/2019	BAC	16080030	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
06/25/2019	1540	34	17GPA41049	.100	22		101		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	1:54 PM	12/19/2017	.101	.100	3:40 PM	06/25/2019	BAC	16080065	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2019	1544 34	17GP41049	.100	22	.098		X		M. Tippett	20137		
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19												
			Cal Standard	Time:								
				4:18 PM								
			Last Cal	03/28/2019								
			Result		.098							
			Cal Standard	Time:		.100						
				3:44 PM								
			Last Check	06/25/2019								
			UOM			BAC						
			Serial No.			16080069						
			FC20			8.7.24						
Calibration / Check Printout												Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2019	1545	34	17GP41049	.100	22		.098			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.098	.100	3:45 PM	06/25/2019	BAC	16080071	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#117

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2019	1559	34	17GP41049	.100	22	.103			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.103	.100	4:00 PM	06/25/2019	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
06/25/2019	1552	34	17GP41049	100	22								X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard		Last Cal		Cal Standard		Last Check		UOM		Serial No.		FC20		Calibration / Check Printout	Lifeloc Technologies
Result	Time:	Result	Time:	Result	Time:	Result	Time:								
.100	4:24 PM	.100	03/28/2019	.100	3:52 PM		06/25/2019	BAC		16080075		8.7.24			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

HTA

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2019	1507	34	17GPA1049	.100	22			.099	.100	YES	NO	X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	4:30 PM	03/28/2019	.099	.100	3:07 PM	06/25/2019	BAC	16080077	8.7.24	Calibration / Check Printout
										Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: *16120175*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/25/2019	<i>1508 34</i>		17GP41049	.100	<i>22</i>	<i>.104</i>					<i>X</i>			

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.104	.100	3:09 PM	06/25/2019	BAC	16120175	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2019	1514	34	17GP41049	.100	23		.103			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	09:50 AM	04/25/2017	.103	.100	3:15 PM	06/25/2019	BAC	16120177	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/25/2019	1513 34		17GP41049	.100		22						X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:16 AM	10/21/2016	.101	.100	3:13 PM	06/25/2019	BAC	16120188	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2019	151734		17GP41049	.100	22		.102			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:50 AM	04/12/2019	.102	.100	3:17 PM	06/25/2019	BAC	16120189	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1 (Optional)	TEST #2 (Optional)					
06/25/2019	1517	34	17GPA1049	.100	22								X		M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	06:53 AM	05/30/2019	.110	.100	3:14 PM	06/25/2019	BAC	16120194	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2019	1524	34	17GP41049	.100	22		.113		X		.099	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	06/25/2019	.099	.100	3:26 PM	06/25/2019	BAC	16120197	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/25/2019	153034		17GP41049	.100	21		.101			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:46 PM	03/28/2019	.101	.100	3:30 PM	06/25/2019	BAC	17170155	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

431

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/25/2019	1529	34	17GP41049	.100	21	.103			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifoloc Technologies
.100	5:05 PM	03/28/2019	.103	.100	3:30 PM	06/25/2019	BAC	17170064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.