

CHP South Los Angeles PAS Usage logs 7-5-2019 to 7-15-2019

Test Number	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
42	0.033	2:41 AM	07/14/2019	3:25 PM	03/28/2019	0.1	11:23 AM	07/05/2019	16040119	
43	0.073	4:13 AM	07/14/2019	3:25 PM	03/28/2019	0.1	11:23 AM	07/05/2019	16040119	
44	0.068	4:15 AM	07/14/2019	3:25 PM	03/28/2019	0.1	11:23 AM	07/05/2019	16040119	
52	0	11:18 AM	07/06/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
53	0.076	1:31 AM	07/07/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
54	0.075	1:33 AM	07/07/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
55	0.079	2:20 AM	07/07/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
56	0.078	2:21 AM	07/07/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
57	0.114	3:20 AM	07/14/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
58	0.106	3:23 AM	07/14/2019	7:46 AM	04/12/2019	0.098	11:18 AM	07/05/2019	16040140	
31	0.19	2:58 AM	07/07/2019	3:41 PM	03/28/2019	0.103	11:23 AM	07/05/2019	16040141	
32	0.191	3:01 AM	07/07/2019	3:41 PM	03/28/2019	0.103	11:23 AM	07/05/2019	16040141	
227	0.138	3:56 AM	07/06/2019	8:56 AM	04/25/2017	0.102	11:29 AM	07/05/2019	16040161	
228	0.117	3:58 AM	07/06/2019	8:56 AM	04/25/2017	0.102	11:29 AM	07/05/2019	16040161	
229	0.147	4:32 AM	07/06/2019	8:56 AM	04/25/2017	0.102	11:29 AM	07/05/2019	16040161	
230	0.158	4:33 AM	07/06/2019	8:56 AM	04/25/2017	0.102	11:29 AM	07/05/2019	16040161	
317	0.122	2:35 AM	07/15/2019	4:09 PM	06/25/2019	0.103	1:55 PM	07/05/2019	16040162	
318	0.116	2:38 AM	07/15/2019	4:09 PM	06/25/2019	0.103	1:55 PM	07/05/2019	16040162	
319	0	4:35 AM	07/15/2019	4:09 PM	06/25/2019	0.103	1:55 PM	07/05/2019	16040162	
74	0.015	2:57 AM	07/06/2019	3:58 PM	03/28/2019	0.102	12:48 PM	07/05/2019	16080015	
75	0.122	3:40 AM	07/06/2019	3:58 PM	03/28/2019	0.102	12:48 PM	07/05/2019	16080015	
54	0	2:30 AM	07/07/2019	3:58 PM	03/28/2019	0.103	12:48 PM	07/05/2019	16080027	
55	0	2:38 AM	07/07/2019	3:58 PM	03/28/2019	0.103	12:48 PM	07/05/2019	16080027	
56	0.12	3:16 AM	07/11/2019	3:58 PM	03/28/2019	0.103	12:48 PM	07/05/2019	16080027	
57	0.117	3:18 AM	07/11/2019	3:58 PM	03/28/2019	0.103	12:48 PM	07/05/2019	16080027	
58	0	3:45 AM	07/13/2019	3:58 PM	03/28/2019	0.103	12:48 PM	07/05/2019	16080027	
32	0.102	2:28 AM	07/14/2019	4:07 PM	03/28/2019	0.102	12:55 PM	07/05/2019	16080030	
33	0.108	2:31 AM	07/14/2019	4:07 PM	03/28/2019	0.102	12:55 PM	07/05/2019	16080030	
34	0	2:57 AM	07/07/2019	4:12 PM	03/28/2019	0.102	1:01 PM	07/05/2019	16080064	
35	0	5:51 PM	07/10/2019	4:12 PM	03/28/2019	0.102	1:01 PM	07/05/2019	16080064	
36	0.095	11:23 PM	07/10/2019	4:12 PM	03/28/2019	0.102	1:01 PM	07/05/2019	16080064	
37	0.09	11:26 PM	07/10/2019	4:12 PM	03/28/2019	0.102	1:01 PM	07/05/2019	16080064	
38	0.245	6:46 AM	07/13/2019	4:12 PM	03/28/2019	0.102	1:01 PM	07/05/2019	16080064	

Test Number	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
39	0.243	6:48 AM	07/13/2019	4:12 PM	03/28/2019	0.102	1:01 PM	07/05/2019	16080064	
242	0.141	8:05 PM	07/10/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
243	0.146	8:08 PM	07/10/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
244	0.177	4:24 AM	07/12/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
245	0.172	4:28 AM	07/12/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
246	0.201	4:01 AM	07/13/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
247	0.195	4:03 AM	07/13/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
248	0.181	11:34 PM	07/13/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
249	0.182	11:37 PM	07/13/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
250	0.163	3:00 AM	07/14/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
251	0.151	3:03 AM	07/14/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
252	NEG	3:06 AM	07/14/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
253	0.148	3:07 AM	07/14/2019	1:54 PM	12/19/2017	0.103	1:02 PM	07/05/2019	16080065	
104	0	2:32 AM	07/06/2019	4:18 PM	03/28/2019	0.099	1:42 PM	07/05/2019	16080069	
105	0.149	3:33 AM	07/13/2019	4:18 PM	03/28/2019	0.099	1:42 PM	07/05/2019	16080069	
106	0.151	3:37 AM	07/13/2019	4:18 PM	03/28/2019	0.099	1:42 PM	07/05/2019	16080069	
392	0.015	12:42 AM	07/06/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071	
393	0.013	12:46 AM	07/06/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071	
394	0.084	4:01 AM	07/06/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071	
395	0.083	4:04 AM	07/06/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071	
396	0.026	12:14 AM	07/15/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071	
397	0.025	12:17 AM	07/15/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071	
86	0.061	2:03 AM	07/07/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
87	0.062	2:06 AM	07/07/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
88	0.166	2:34 AM	07/07/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
89	0.172	2:38 AM	07/07/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
90	0.15	4:32 AM	07/07/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
91	0.142	4:35 AM	07/07/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
92	0.073	12:28 AM	07/08/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
93	0.069	12:30 AM	07/08/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
94	0.145	10:15 PM	07/13/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
95	0.138	10:17 PM	07/13/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
96	0.166	2:32 AM	07/14/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	

CHP South Los Angeles PAS Usage logs 7-5-2019 to 7-15-2019

Test Number	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number	Unit Model
97	0.173	2:34 AM	07/14/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
98	0.14	5:32 AM	07/14/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
99	0.138	5:35 AM	07/14/2019	4:30 PM	03/28/2019	0.102	1:15 PM	07/05/2019	16080077	
44	0.089	10:19 PM	07/05/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
45	0.099	10:22 PM	07/05/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
46	0	2:41 AM	07/06/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
47	0.125	5:02 AM	07/06/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
48	0.117	5:05 AM	07/06/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
49	0.074	4:09 AM	07/07/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
50	0.053	2:25 AM	07/08/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
51	0	1:01 AM	07/13/2019	4:36 PM	03/28/2019	0.105	1:16 PM	07/05/2019	16120175	
320	0.155	1:31 AM	07/14/2019	6:53 AM	05/30/2019	0.11	1:22 PM	07/05/2019	16120194	
321	0.153	1:34 AM	07/14/2019	6:53 AM	05/30/2019	0.11	1:22 PM	07/05/2019	16120194	
333	0.141	2:30 AM	07/06/2019	12:48 PM	06/09/2019	0.103	1:44 PM	07/05/2019	16120195	
334	0.142	2:32 AM	07/06/2019	12:48 PM	06/09/2019	0.103	1:44 PM	07/05/2019	16120195	
335	0.052	2:40 AM	07/06/2019	12:48 PM	06/09/2019	0.103	1:44 PM	07/05/2019	16120195	
336	0.258	4:44 AM	07/14/2019	12:48 PM	06/09/2019	0.103	1:44 PM	07/05/2019	16120195	
337	0.263	4:47 AM	07/14/2019	12:48 PM	06/09/2019	0.103	1:44 PM	07/05/2019	16120195	
58	0.192	3:25 AM	07/06/2019	5:05 PM	03/28/2019	0.104	1:58 PM	07/05/2019	17170064	
59	0.177	3:28 AM	07/06/2019	5:05 PM	03/28/2019	0.104	1:58 PM	07/05/2019	17170064	
60	0	1:45 AM	07/07/2019	5:05 PM	03/28/2019	0.104	1:58 PM	07/05/2019	17170064	
61	0.055	10:56 PM	07/14/2019	5:05 PM	03/28/2019	0.104	1:58 PM	07/05/2019	17170064	
62	0.053	10:59 PM	07/14/2019	5:05 PM	03/28/2019	0.104	1:58 PM	07/05/2019	17170064	

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	0954	34	17GP41049	.100	23		.093			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	3:25 PM	03/28/2019	.093	.100	09:54 AM	07/15/2019	BAC	16040119	8.7.24

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
ID. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	0950	34	17GP41049	.100	23		.092					M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	07:46 AM	04/12/2019	.092	.100	09:51 AM	07/15/2019	BAC	16040140	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			YES	NO				
07/15/2019	0958 34		17GP41049						M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:41 PM	03/28/2019	.099	.100	09:56 AM	07/15/2019	BAC	16040141	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

[Signature]

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1206 54		17GP41049	.100	23	.099			X		M. Tippett	20137	

Solution prepared by:													
Cesar A. Mejia													
Date: 07-24-17													
Expiration date: 07-24-19													
Cal Standard													
Time:		08:56 AM		Last Cal		04/25/2017		Result		.099		Cal Standard	
Time:		10:07 AM		Last Check		07/15/2019		UOM		BAC		Serial No. 16040161	
Time:		8:7.24		FC20									
Calibration / Check Printout													
Lifeloc Technologies													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

#6

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1063	34	17GP41049	.100	23		.093		X		M. Tippett	20137		

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	07:30 AM	04/12/2019	.093	.100	10:04 AM	07/15/2019	BAC	16040163	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	10/834	17GP41049			.100	23		.097			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.097	.100	10:18 AM	07/15/2019	BAC	16080013	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	10:23:34		17GP41049	.100	25			.097			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	11:51 AM	07/05/2019	.097	.100	10:23 AM	07/15/2019	BAC	16080014	8.7.24

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
07/13/2019	1024	34	17GP41049	100	24	.098				X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.098	.100	10:25 AM	07/15/2019	BAC	16080015	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)				
07/15/2019	1028	34	17GP41049	100	25	.096	X		M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.096	.100	10:28 AM	07/15/2019	BAC	16080028	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1028 34			17GP41049	.100	23		.096				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:07 PM	03/28/2019	.096	.100	10:29 AM	07/15/2019	BAC	16080030	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

Handwritten initials

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: *16080064*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
07/15/2019	1031	34	17GP41049	100	23									M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:12 PM	03/28/2019	.096	.100	10:31 AM	07/15/2019	BAC	16080064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#14

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
										YES	NO				
07/15/2019	1035	34	17GP41049	.100	23		.096			X		M. Tippett	20137		

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	1:54 PM	12/19/2017	.096	.100	10:35 AM	07/15/2019	BAC	16080065	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	103934	17GP41049	.100	23		.096			X		M. Tippett	20137		

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:18 PM	03/28/2019	.096	.100	10:40 AM	07/15/2019	BAC	16080069	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	104034	17GP41049		100	25	097			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	10:41 AM	08/14/2017	.097	.100	10:40 AM	07/15/2019	BAC	16080071	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
07/15/2019	1055	34	17GP41049	.100	23		.100		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies Calibration / Check Printout
.100	2:05 PM	12/19/2017	.100	.100	10:56 AM	07/15/2019	BAC	16080072	8.7.24	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1048 34		17GP41049	100	22		.097		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.097	.100	10:49 AM	07/15/2019	BAC	16080075	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	105734		17GP41049	100	23		.096				X	M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia
Date: 07-24-17
Expiration date: 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:30 PM	03/28/2019	.096	.100	10:57 AM	07/15/2019	BAC	16080077	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16/20175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/15/2019	1057	34	17GP41049	100	23		.099					M. Tippett	20137		
Solution prepared by: Cesar A. Mejia															
Date: 07-24-17															
Expiration date: 07-24-19															
Cal Standard															
Time:		Last Cal		Result		Cal Standard		Time:		Last Check		UOM		Serial No.	
4:36 PM		03/28/2019		.099		.100		10:58 AM		07/15/2019		BAC		16120175	
														FC20	
Calibration / Check Printout															
Lifeloc Technologies															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	11:03	34	17GP41049	.100	23	.095			X		M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia
Date: 07-24-17
Expiration date: 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.095	.100	11:03 AM	07/15/2019	BAC	16120188	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
07/15/2019	1107	34	17GP41049	.100	23		.096		NO	X	M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.096	.100	11:08 AM	07/15/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
07/15/2019	1104	34	17GPA1049		.100		23						X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	06:53 AM	05/30/2019	.104	.100	11:04 AM	07/15/2019	BAC	16120194	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1111	34	17GP41049	.100	23		.099		X		M. Tippett	20137		

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.099	.100	11:11 AM	07/15/2019	BAC	16120195	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2019	11:16	39	17GP41049	.100	22		.095				X	M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	06/25/2019	.095	.100	11:17 AM	07/15/2019	BAC	16120197	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1119	34	17GP41049	.100	21	.099		X			M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:42 PM	03/28/2019	.099	.100	11:19 AM	07/15/2019	BAC	17170066	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	11:21	34	17GP41049	.100	22		.098		X		M. Tippett	20137		

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:46 PM	03/28/2019	.098	.100	11:21 AM	07/15/2019	BAC	17170155	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

#31

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2019	1121	34	17GP41049	.100	71	.101		X			M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	5:05 PM	03/28/2019	.101	.100	11:21 AM	07/15/2019	BAC	17170064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.