

Test Number	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
45	0.143	12:07 AM	07/21/2019	3:25 PM	03/28/2019	0.093	9:54 AM	07/15/2019	16040119
46	0.167	12:10 AM	07/21/2019	3:25 PM	03/28/2019	0.093	9:54 AM	07/15/2019	16040119
59	0.115	11:26 PM	07/20/2019	7:46 AM	04/12/2019	0.092	9:51 AM	07/15/2019	16040140
60	0.114	11:28 PM	07/20/2019	7:46 AM	04/12/2019	0.092	9:51 AM	07/15/2019	16040140
61	0	1:01 AM	07/21/2019	7:46 AM	04/12/2019	0.092	9:51 AM	07/15/2019	16040140
62	0.119	11:12 PM	07/24/2019	7:46 AM	04/12/2019	0.092	9:51 AM	07/15/2019	16040140
63	0.121	11:14 PM	07/24/2019	7:46 AM	04/12/2019	0.092	9:51 AM	07/15/2019	16040140
33	0	5:20 AM	07/19/2019	3:41 PM	03/28/2019	0.099	9:56 AM	07/15/2019	16040141
34	0.099	4:06 AM	07/21/2019	3:41 PM	03/28/2019	0.099	9:56 AM	07/15/2019	16040141
35	0.096	4:09 AM	07/21/2019	3:41 PM	03/28/2019	0.099	9:56 AM	07/15/2019	16040141
231	0.142	1:03 AM	07/21/2019	8:56 AM	04/25/2017	0.099	10:07 AM	07/15/2019	16040161
232	0.145	1:05 AM	07/21/2019	8:56 AM	04/25/2017	0.099	10:07 AM	07/15/2019	16040161
233	0.21	9:55 PM	07/21/2019	8:56 AM	04/25/2017	0.099	10:07 AM	07/15/2019	16040161
234	0.197	9:59 PM	07/21/2019	8:56 AM	04/25/2017	0.099	10:07 AM	07/15/2019	16040161
317	0.122	2:35 AM	07/15/2019	4:09 PM	06/25/2019	0.103	1:55 PM	07/05/2019	16040162
318	0.116	2:38 AM	07/15/2019	4:09 PM	06/25/2019	0.103	1:55 PM	07/05/2019	16040162
319	0	4:35 AM	07/15/2019	4:09 PM	06/25/2019	0.103	1:55 PM	07/05/2019	16040162
392	0.063	9:21 PM	07/17/2019	7:30 AM	04/12/2019	0.093	10:04 AM	07/15/2019	16040163
393	0.068	9:24 PM	07/17/2019	7:30 AM	04/12/2019	0.093	10:04 AM	07/15/2019	16040163
394	0	2:37 AM	07/23/2019	7:30 AM	04/12/2019	0.093	10:04 AM	07/15/2019	16040163
395	0.022	3:35 AM	07/23/2019	7:30 AM	04/12/2019	0.093	10:04 AM	07/15/2019	16040163
53	0	6:32 PM	07/20/2019	11:51 AM	07/05/2019	0.097	10:23 AM	07/15/2019	16080014
54	0	6:34 PM	07/20/2019	11:51 AM	07/05/2019	0.097	10:23 AM	07/15/2019	16080014
55	0.205	11:56 PM	07/20/2019	11:51 AM	07/05/2019	0.097	10:23 AM	07/15/2019	16080014
56	0.207	11:59 PM	07/20/2019	11:51 AM	07/05/2019	0.097	10:23 AM	07/15/2019	16080014
59	0.161	1:20 AM	07/19/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
60	0.16	1:23 AM	07/19/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
61	0.119	1:15 AM	07/21/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
62	0.125	1:18 AM	07/21/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
63	0.038	7:00 PM	07/22/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
64	0.155	1:53 AM	07/23/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
65	0.104	1:56 AM	07/23/2019	3:58 PM	03/28/2019	0.098	10:24 AM	07/15/2019	16080027
34	0.084	2:35 AM	07/21/2019	4:07 PM	03/28/2019	0.096	10:29 AM	07/15/2019	16080030

CHP South LA PAS Usage Log 7-15-2019 to 7-25-2019

Test Number	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
35	0.081	2:37 AM	07/21/2019	4:07 PM	03/28/2019	0.096	10:29 AM	07/15/2019	16080030
254	0.223	3:31 AM	07/20/2019	1:54 PM	12/19/2017	0.096	10:35 AM	07/15/2019	16080065
255	0.228	3:34 AM	07/20/2019	1:54 PM	12/19/2017	0.096	10:35 AM	07/15/2019	16080065
107	0.192	1:39 AM	07/21/2019	4:18 PM	03/28/2019	0.096	10:40 AM	07/15/2019	16080069
108	0.213	1:42 AM	07/21/2019	4:18 PM	03/28/2019	0.096	10:40 AM	07/15/2019	16080069
109	0.14	3:29 AM	07/21/2019	4:18 PM	03/28/2019	0.096	10:40 AM	07/15/2019	16080069
110	0.144	3:32 AM	07/21/2019	4:18 PM	03/28/2019	0.096	10:40 AM	07/15/2019	16080069
111	0.094	12:49 AM	07/24/2019	4:18 PM	03/28/2019	0.096	10:40 AM	07/15/2019	16080069
112	0.1	12:52 AM	07/24/2019	4:18 PM	03/28/2019	0.096	10:40 AM	07/15/2019	16080069
396	0.026	12:14 AM	07/15/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071
397	0.025	12:17 AM	07/15/2019	10:41 AM	08/14/2017	0.101	1:05 PM	07/05/2019	16080071
398	0.087	12:40 AM	07/21/2019	10:41 AM	08/14/2017	0.097	10:40 AM	07/15/2019	16080071
399	0.092	12:42 AM	07/21/2019	10:41 AM	08/14/2017	0.097	10:40 AM	07/15/2019	16080071
400	0.016	3:38 AM	07/21/2019	10:41 AM	08/14/2017	0.097	10:40 AM	07/15/2019	16080071
401	0.139	12:58 AM	07/22/2019	10:41 AM	08/14/2017	0.097	10:40 AM	07/15/2019	16080071
402	0.145	1:01 AM	07/22/2019	10:41 AM	08/14/2017	0.097	10:40 AM	07/15/2019	16080071
22	0.174	9:56 PM	07/18/2019	2:05 PM	12/19/2017	0.1	10:56 AM	07/15/2019	16080072
23	0.2	9:59 PM	07/18/2019	2:05 PM	12/19/2017	0.1	10:56 AM	07/15/2019	16080072
24	0.197	10:02 PM	07/18/2019	2:05 PM	12/19/2017	0.1	10:56 AM	07/15/2019	16080072
100	0.013	12:44 AM	07/17/2019	4:30 PM	03/28/2019	0.096	10:57 AM	07/15/2019	16080077
101	0.013	12:47 AM	07/17/2019	4:30 PM	03/28/2019	0.096	10:57 AM	07/15/2019	16080077
102	0	1:17 AM	07/22/2019	4:30 PM	03/28/2019	0.096	10:57 AM	07/15/2019	16080077
88	0.149	10:07 PM	07/17/2019	8:16 AM	10/21/2016	0.095	11:03 AM	07/15/2019	16120188
89	0.146	10:09 PM	07/17/2019	8:16 AM	10/21/2016	0.095	11:03 AM	07/15/2019	16120188
322	NEG	12:33 PM	07/25/2019	6:53 AM	05/30/2019	0.104	11:04 AM	07/15/2019	16120194
338	0.145	3:18 AM	07/20/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
339	0.155	3:21 AM	07/20/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
340	0.065	12:31 AM	07/21/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
341	0	2:37 AM	07/21/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
342	0	11:38 PM	07/24/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
343	0.12	1:08 AM	07/25/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
344	0.123	1:11 AM	07/25/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
315	0	11:02 PM	07/24/2019	3:25 PM	06/25/2019	0.095	11:17 AM	07/15/2019	16120197

CHP South LA PAS Usage Log 7-15-2019 to 7-25-2019

Test Number	Result	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
63	0.016	12:16 AM	07/17/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
64	0.136	2:48 AM	07/20/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
65	0.134	2:51 AM	07/20/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
66	0	9:34 PM	07/20/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
67	0	2:12 AM	07/21/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
68	0.176	3:09 AM	07/21/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
69	0.172	3:11 AM	07/21/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
70	0.015	2:27 AM	07/23/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
71	0.013	2:31 AM	07/23/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
72	0.079	2:39 AM	07/24/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064
73	0.08	2:42 AM	07/24/2019	5:05 PM	03/28/2019	0.101	11:21 AM	07/15/2019	17170064

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
07/25/2019	1:26:34			181PA1076	.100		25						X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia																	
Date: 09-10-18																	
Expiration date: 09-10-20																	
Cal Standard																	
Time: 4:09 PM																	
Last Cal																	
06/25/2019																	
Result																	
.095																	
Cal Standard																	
Time: 11:37 AM																	
Last Check																	
07/25/2019																	
UOM																	
BAC																	
Serial No.																	
16040162																	
FC20																	
8.7.24																	
Calibration / Check Printout																	
Lifeloc Technologies																	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/25/2019	1148 34	181P41076		.100	24	.096					M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:25 PM	03/28/2019	.096	.100	11:48 AM	07/25/2019	BAC	16040119	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	114234	181P41076	.100	25	.095			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:46 AM	04/12/2019	.095	.100	11:42 AM	07/25/2019	BAC	16040140	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1147 34		181P41076			.102				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:41 PM	03/28/2019	.102	.100	11:47 AM	07/25/2019	BAC	16040141	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/25/2019	115934		181P41076		.100	TEST #1	TEST #2 (Optional)			M. Tippett	20137	

Solution prepared by:		Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
Cesar A. Mejia												
Date: 09-10-18												
Expiration date: 09-10-20												
		.100	08:56 AM	04/25/2017	.100	.100	11:55 AM	07/25/2019	BAC	16040161	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	115034		181P41076			.096				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.096	.100	11:51 AM	07/25/2019	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1204	34	181P41076	.100	25	.100	.100	NO	X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.100	.100	12:04 PM	07/25/2019	BAC	16080013	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

8

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/25/2019	1200 34	181P41076	.100	25	.102			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	11:51 AM	07/05/2019	.102	.100	12:00 PM	07/25/2019	BAC	16080014	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/25/2019	12:01:34	181P41076			.100	25	101	X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.101	.100	12:02 PM	07/25/2019	BAC	16080015	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/25/2019	12:01 PM	181P41076		100	25	103		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	3:58 PM	03/28/2019	.103	.100	12:01 PM	07/25/2019	BAC	16080027	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	12:34		181P41076							M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.102	.100	12:06 PM	07/25/2019	BAC	16080028	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#112

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAC VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1205	34	181P41076	100	24			.102			X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia															
Date: 09-10-18															
Expiration date: 09-10-20															
Cal Standard															
Time:		Last Cal		Result		Cal Standard		Time:		Last Check		UOM		Serial No.	
4:07 PM		03/28/2019		.102		.100		12:05 PM		07/25/2019		BAC		16080030	
FC20															
8.7.24															
Calibration / Check Printout															
Lifeloc Technologies															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		REPAIRS/PROBLEMS/COMMENTS			
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS	OPERATOR		ID. NO.		
07/25/2019	12:07:34		181P41076	.100		24		.101			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	4:12 PM	03/28/2019	.101	.100	12:07 PM	07/25/2019	BAC	16080064	8.7.24	
Calibration / Check Printout										
Lifeloc Technologies										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080055

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	12:08	34	181P41076	.100	24	.101		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	1:54 PM	12/19/2017	.101	.100	12:09 PM	07/25/2019	BAC	16080065	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
07/25/2019	12:15:34	181P41076	.100	24		.099						M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:18 PM	03/28/2019	.099	.100	12:13 PM	07/25/2019	BAC	16080069	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

4/16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	12:16	34	181P41076	.100	25	.102			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.102	.100	12:17 PM	07/25/2019	BAC	16080071	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#117

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1232 34	181P41076	.100	24	.106			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.106	.100	12:33 PM	07/25/2019	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#118

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			YES	NO				
07/25/2019	12:25	34	181P41076	.100	2-4	.103	X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.103	.100	12:26 PM	07/25/2019	BAC	16080075	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1225	34	181P41076	.100	25	.104			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:30 PM	03/28/2019	.104	.100	12:26 PM	07/25/2019	BAC	16080077	8.7.24	
Calibration / Check Printout										
Lifeloc Technologies										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
07/25/2019	1226	34	181P41076	.100	24	.105			X		M. Tippett	20137		
Solution prepared by:														
Cesar A. Mejia														
Date: 09-10-18														
Expiration date: 09-10-20														

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:36 PM	03/28/2019	.105	.100	12:26 PM	07/25/2019	BAC	16120175	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES	NO			
07/25/2019	1233 34	181P41076	.100	24		102				X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.102	.100	12:34 PM	07/25/2019	BAC	16120177	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	YES	NO				
07/25/2019	1258	34	181P41076	.100	25		.099			X		M. Tippett	20137	
				Cal Standard			.100							
				Time:	08:16 AM									
				Last Cal	10/21/2016									
				Result			.099							
				Cal Standard			.100							
				Time:	12:32 PM									
				Last Check	07/25/2019									
				UOM	BAC									
				Serial No.	16120188									
				FC20	8.7.24									
				Calibration / Check Printout										
				Lifeloc Technologies										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 161 20189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1236	34	181P41076	.100	24						M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.103	.100	12:36 PM	07/25/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1232	34	181P41076	.100	24	109	X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	06:53 AM	05/30/2019	.109	.100	12:33 PM	07/25/2019	BAC	16120194	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1238	34	181P41076	.100	24	.105				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.105	.100	12:39 PM	07/25/2019	BAC	16120195	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	124634	181P41076		.100		24		.098			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	06/25/2019	.098	.100	12:46 PM	07/25/2019	BAC	16120197	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	1249	34	181P41076			102				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:42 PM	03/28/2019	.102	.100	12:49 PM	07/25/2019	BAC	17170066	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME: South Los Angeles | AREA NO: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifaloc FC20 | PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/25/2019		181P41076		.100							M. Tippett	20137	DEVICE IS OUT OF SERVICE
		Solution prepared by: Cesar A. Mejia Date: 09-10-18 Expiration date: 09-10-20											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 CHP202K_1109.pdf

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/25/2019	1251	34	181P41076							M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:46 PM	03/28/2019	.101	.100	12:52 PM	07/25/2019	BAC	17170155	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

30

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/25/2019			181P41076	.100							M. Tippett	20137	DEVICE IS OUT OF SERVICE
		Solution prepared by: Cesar A. Mejia											
		Date: 09-10-18											
		Expiration date: 09-10-20											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#31

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 17170664

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/25/2019	12:57	34	181P41076	.100	23	.107	X			M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	5:05 PM	03/28/2019	.107	.100	12:51 PM	07/25/2019	BAC	17170064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.