

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
40	0		12:48 AM	07/04/2019	3:25 PM	03/28/2019	0.095	3:06 PM	06/19/2019	16040119
41	0.062		1:36 AM	07/04/2019	3:25 PM	03/28/2019	0.095	3:06 PM	06/19/2019	16040119
50	0.224		12:33 PM	06/30/2019	7:46 AM	04/12/2019	0.096	4:14 PM	06/25/2019	16040140
51	0.214		12:35 PM	06/30/2019	7:46 AM	04/12/2019	0.096	4:14 PM	06/25/2019	16040140
308	0.103		7:15 PM	06/27/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
309	0.104		7:18 PM	06/27/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
310	0		10:08 PM	06/30/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
311	0.15		4:52 AM	07/01/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
312	0.142		4:55 AM	07/01/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
313	0.211		5:14 PM	07/03/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
314	0.203		5:17 PM	07/03/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
315	0.16		6:30 PM	07/04/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
316	0.163		6:34 PM	07/04/2019	4:09 PM	06/25/2019	0.099	4:11 PM	06/25/2019	16040162
384	0.203		5:58 PM	06/28/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
385	0.204		6:00 PM	06/28/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
386	0		11:25 PM	07/02/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
387	0		12:45 PM	07/04/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
388	0.011		4:24 PM	07/04/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
389	0.011		4:26 PM	07/04/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
390	0.141		7:47 PM	07/04/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
391	0.13		7:50 PM	07/04/2019	7:30 AM	04/12/2019	0.095	4:19 PM	06/25/2019	16040163
51	0.101		3:07 AM	06/29/2019	1:28 PM	05/20/2019	0.108	3:58 PM	06/25/2019	16080014
52	0.094		3:10 AM	06/29/2019	1:28 PM	05/20/2019	0.108	3:58 PM	06/25/2019	16080014
72	0.224		9:34 PM	06/27/2019	3:58 PM	03/28/2019	0.094	3:44 PM	06/19/2019	16080015
73	0.213		9:36 PM	06/27/2019	3:58 PM	03/28/2019	0.094	3:44 PM	06/19/2019	16080015
29	0.085		2:10 AM	07/02/2019	4:05 PM	03/28/2019	0.098	4:07 PM	06/25/2019	16080028
30	0.088		2:12 AM	07/02/2019	4:05 PM	03/28/2019	0.098	4:07 PM	06/25/2019	16080028
236	0.125		12:35 AM	06/25/2019	1:54 PM	12/19/2017	0.099	3:24 PM	06/19/2019	16080065
237	0.119		12:38 AM	06/25/2019	1:54 PM	12/19/2017	0.099	3:24 PM	06/19/2019	16080065
238	0.128		3:25 AM	06/28/2019	1:54 PM	12/19/2017	0.101	3:40 PM	06/25/2019	16080065
239	0.126		3:28 AM	06/28/2019	1:54 PM	12/19/2017	0.101	3:40 PM	06/25/2019	16080065
240	0.184		12:44 AM	06/29/2019	1:54 PM	12/19/2017	0.101	3:40 PM	06/25/2019	16080065
241	0.199		12:46 AM	06/29/2019	1:54 PM	12/19/2017	0.101	3:40 PM	06/25/2019	16080065

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
102	0.185		11:51 PM	06/27/2019	4:18 PM	03/28/2019	0.098	3:44 PM	06/25/2019	16080069
103	0.181		11:54 PM	06/27/2019	4:18 PM	03/28/2019	0.098	3:44 PM	06/25/2019	16080069
389	0		12:45 AM	06/28/2019	10:41 AM	08/14/2017	0.098	3:45 PM	06/25/2019	16080071
390	0		12:47 AM	06/28/2019	10:41 AM	08/14/2017	0.098	3:45 PM	06/25/2019	16080071
391	0		2:29 AM	07/02/2019	10:41 AM	08/14/2017	0.098	3:45 PM	06/25/2019	16080071
17	0.071		1:18 AM	06/30/2019	2:05 PM	12/19/2017	0.103	4:00 PM	06/25/2019	16080072
18	0.068		1:21 AM	06/30/2019	2:05 PM	12/19/2017	0.103	4:00 PM	06/25/2019	16080072
19	0.094		2:00 AM	06/30/2019	2:05 PM	12/19/2017	0.103	4:00 PM	06/25/2019	16080072
20	0.091		2:02 AM	06/30/2019	2:05 PM	12/19/2017	0.103	4:00 PM	06/25/2019	16080072
21	0		1:19 AM	07/04/2019	2:05 PM	12/19/2017	0.103	4:00 PM	06/25/2019	16080072
79	0.126		10:04 PM	06/27/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
80	0.135		10:06 PM	06/27/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
81	0		5:01 AM	06/29/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
82	0.013		5:33 PM	06/29/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
83	0		5:35 PM	06/29/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
84	0.074		10:53 PM	07/01/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
85	0.081		10:56 PM	07/01/2019	4:30 PM	03/28/2019	0.099	3:07 PM	06/25/2019	16080077
36	0.175		12:33 AM	06/28/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
37	0.171		12:35 AM	06/28/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
38	0.075		8:03 PM	06/29/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
39	0.069		8:04 PM	06/29/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
40	0.129		9:25 PM	06/29/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
41	0.126		9:27 PM	06/29/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
42	0.214		10:44 PM	07/04/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
43	0.226		10:47 PM	07/04/2019	4:36 PM	03/28/2019	0.104	3:09 PM	06/25/2019	16120175
113	0.241		11:03 PM	06/30/2019	9:50 AM	04/25/2017	0.103	3:15 PM	06/25/2019	16120177
114	0.237		11:05 PM	06/30/2019	9:50 AM	04/25/2017	0.103	3:15 PM	06/25/2019	16120177
85	0.113		11:47 PM	06/30/2019	8:16 AM	10/21/2016	0.101	3:13 PM	06/25/2019	16120188
86	0.113		11:50 PM	06/30/2019	8:16 AM	10/21/2016	0.101	3:13 PM	06/25/2019	16120188
87	0.124		11:57 PM	06/30/2019	8:16 AM	10/21/2016	0.101	3:13 PM	06/25/2019	16120188
327	0.079		3:10 AM	06/29/2019	12:48 PM	06/09/2019	0.105	3:19 PM	06/25/2019	16120195
328	0.077		3:13 AM	06/29/2019	12:48 PM	06/09/2019	0.105	3:19 PM	06/25/2019	16120195
329	0.133		2:07 AM	07/04/2019	12:48 PM	06/09/2019	0.105	3:19 PM	06/25/2019	16120195

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
330	0.137		2:09 AM	07/04/2019	12:48 PM	06/09/2019	0.105	3:19 PM	06/25/2019	16120195
331	0.139		11:10 PM	07/04/2019	12:48 PM	06/09/2019	0.105	3:19 PM	06/25/2019	16120195
332	0.134		11:13 PM	07/04/2019	12:48 PM	06/09/2019	0.105	3:19 PM	06/25/2019	16120195
44	0.099		3:35 AM	06/28/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
45	0.096		3:38 AM	06/28/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
46	0.027		1:32 AM	06/29/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
47	0.024		1:34 AM	06/29/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
48	0.105		2:36 AM	06/29/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
49	0.101		2:38 AM	06/29/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
50	0		11:38 PM	06/29/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
51	0.034		3:53 AM	06/30/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
52	0.033		3:55 AM	06/30/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
53	0.14		12:20 AM	07/05/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
54	0.146		12:21 AM	07/05/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
55	0.196		2:44 AM	07/05/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
56	0.194		2:47 AM	07/05/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
57	0		10:49 AM	07/05/2019	5:05 PM	03/28/2019	0.103	3:30 PM	06/25/2019	17170064
6	0.141		3:06 AM	06/30/2019	4:42 PM	03/28/2019	0.103	3:28 PM	06/25/2019	17170066
7	0.139		3:08 AM	06/30/2019	4:42 PM	03/28/2019	0.103	3:28 PM	06/25/2019	17170066
8	0.017		3:12 AM	06/30/2019	4:42 PM	03/28/2019	0.103	3:28 PM	06/25/2019	17170066

#1

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16040162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1755	34	17GP41049	.100	27		.103				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:09 PM	06/25/2019	.103	.100	1:55 PM	07/05/2019	BAC	16040162	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16040119

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/05/2019	1122 34		17GP41049	.100	22		.100					X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	3:25 PM	03/28/2019	.100	.100	11:23 AM	07/05/2019	BAC	16040119	8.7.24

Lifeloc Technologies

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Michael Tippett    I.D. NUMBER: 20137    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16040140

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/05/2019	11:17	39	17GP41049	.100	22	.098			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:46 AM	04/12/2019	.098	.100	11:18 AM	07/05/2019	BAC	16040140	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16040141

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/05/2019	11:25	34	17GPA1049	.100	21	.103		X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:41 PM	03/28/2019	.103	.100	11:23 AM	07/05/2019	BAC	16040141	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Michael Tippett	I.D. NUMBER 20137	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16046161
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
07/05/2019	1129 34		17GP41049		.100	22	.102		X		M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
			.100			BAC	16040161	8.7.24	
	08:56 AM				07/05/2019				

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**Lifoloc Technologies**



*46*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: *16040163*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS															
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO																			
07/05/2019	<i>11:25</i>	<i>24</i>	17GP41049	100		<i>23</i>		<i>.097</i>				<i>X</i>																	
		Solution prepared by:		Cal Standard		Time:		Result		Last Cal		Cal Standard		Time:		Last Check		UOM		Serial No.		FC20		Calibration / Check Printout		Lifeloc Technologies			
		Cesar A. Mejia				07:30 AM		.097		04/12/2019		.100		11:26 AM		07/05/2019		BAC		16040163		8.7.24							
		Date: 07-24-17																											
		Expiration date:																											
		07-24-19																											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Michael Tippett  
 ID. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	11:53	34	17GP41049	.100	23	.101			X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
Cal Standard													
					Time:		Last Cal		Result		Cal Standard		Time:
				.100	12:59 PM		01/29/2018		.101				11:54 AM
											Last Check	07/05/2019	
											UOM	BAC	
											Serial No.	16080013	
											FC20	8.7.24	
Calibration / Check Printout													
<b>Lifeloc Technologies</b>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080014

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	115034		17GP41049	.100	23		///		X		.100	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:51 AM	07/05/2019	.100	.100	11:51 AM	07/05/2019	BAC	16080014	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080015

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/05/2019	1247	34	17GP41049	.100	25	.102				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.102	.100	12:48 PM	07/05/2019	BAC	16080015	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080027

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
07/05/2019	124734		17GP41049	.100	22		103			X		M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.103	.100	12:48 PM	07/05/2019	BAC	16080027	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080028

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	12:53:34	17	GP41049	.100	22		101			X		M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.101	.100	12:54 PM	07/05/2019	BAC	16080028	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080030

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1254	34	17GP41049	.100	22		.102				X	M. Tippett	20137	

Solution prepared by:														
Cesar A. Mejia														
Date: 07-24-17														
Expiration date: 07-24-19														
Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout			
.100		4:07 PM	03/28/2019	.102	.100	12:55 PM	07/05/2019	BAC	16080030	8.7.24	Lifeloc Technologies			

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1301	34	17GP41049	.100	23		.102				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:12 PM	03/28/2019	.102	.100	1:01 PM	07/05/2019	BAC	16080064	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*Handwritten:* #14

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/05/2019	130234	17	GP41049	.100	23			103					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	1:54 PM	12/19/2017	.103	.100	1:02 PM	07/05/2019	BAC	16080065	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080069

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
07/05/2019	1341	34	17GP41049	.100	55		0.99				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:18 PM	03/28/2019	.099	.100	1:42 PM	07/05/2019	BAC	16080069	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*416*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: *16080071*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	<i>130434</i>		17GPA1049	.100	<i>23</i>		<i>.101</i>			<i>X</i>		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.101	.100	1:05 PM	07/05/2019	BAC	16080071	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*417*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: *16080072*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
07/05/2019	<i>1319 34</i>		17GP41049	.100	<i>22</i>		<i>.105</i>		NO		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies Calibration / Check Printout
.100	2:05 PM	12/19/2017	.105	.100	1:20 PM	07/05/2019	BAC	16080072	8.7.24	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080075

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/05/2019	12:23	34	17GP41049	.100	23		.102				X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.102	.100	1:13 PM	07/05/2019	BAC	16080075	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*#19*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080077

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1:31	34	17GP41049	.100		23		.102			X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:30 PM	03/28/2019	.102	.100	1:15 PM	07/05/2019	BAC	16080077	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*Handwritten initials/signature*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: *16120175*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1:55	34	17GP41049	.100	22		105				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:36 PM	03/28/2019	.105	.100	1:16 PM	07/05/2019	BAC	16120175	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120177

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1:22:34	17GP41049	.100	22	103			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.103	.100	1:23 PM	07/05/2019	BAC	16120177	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*#22*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: *16120188*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
07/05/2019	<i>1321</i>	<i>34</i>	17GP41049	.100	<i>23</i>		<i>.098</i>		<i>X</i>		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.098	.100	1:22 PM	07/05/2019	BAC	16120188	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120189

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
07/05/2019	1525	34	17GP41049	.100	23		.102	NO	X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.102	.100	1:25 PM	07/05/2019	BAC	16120189	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

*#24*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	132234		17GP41049	.100	22		.110				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	06:53 AM	05/30/2019	.110	.100	1:22 PM	07/05/2019	BAC	16120194	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*Handwritten signature*

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1343	34	17GP41049	.100	23		103				X		20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:48 PM	06/09/2019	.103	.100	1:44 PM	07/05/2019	BAC	16120195	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1:55:34	17GP41049	.100	22		.100			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:25 PM	06/25/2019	.100	.100	1:54 PM	07/05/2019	BAC	16120197	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 171700666

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
07/05/2019	1356	34	17GP41049	.100	22		.103		X	M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Cal Standard		Last Cal		Cal Standard		Last Check		UOM		FC20		Calibration / Check Printout	
Result	Time:	Result	Time:	Result	Time:	Result	Time:		Serial No.			Lifeloc Technologies	
.100	4:42 PM	.103	03/28/2019	.100	1:57 PM		07/05/2019	BAC	17170066		8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Michael Tippett	I.D. NUMBER 20137	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170152
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019		17GP41049		.100							M. Tippett	20137	OUT OF SERVICE

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Michael Tippett  
 ID. NUMBER: 20137  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO: 17170155

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				VALUE	TEST #1	TEST #2	YES				
07/05/2019	1:58	34	17GP41049	.100	22		.105		X			M. Tippett	20137	

Solution prepared by:		Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
		.100	4:46 PM	03/28/2019	.105	.100	1:59 PM	07/05/2019	BAC	17170155	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*#50*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippert      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 90102844-82

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/05/2019			17GP41049	.100							M. Tippert	20137	OUT OF SERVICE

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Michael Tippett      I.D. NUMBER: 20137      PAS MAKE/MODEL: Lifloc FC20      PAS SERIAL NO.: 17170064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/05/2019	1:35	87.4	17GP41049	.100	22		.104			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:12 PM	03/28/2019	.102	.100	1:01 PM	07/05/2019	BAC	16080064	8.7.24	

**Lifloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.