

CHP South Los Angeles PAS Usage Logs 8-5-2019 to 8-15-2019

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
52	0.148		7:09 AM	08/10/2019	3:25 PM	03/28/2019	0.094	8:10 AM	08/05/2019	16040119
53	0.146		7:11 AM	08/10/2019	3:25 PM	03/28/2019	0.094	8:10 AM	08/05/2019	16040119
64	0.248		2:44 AM	08/07/2019	7:46 AM	04/12/2019	0.096	8:04 AM	08/05/2019	16040140
65	0.261		2:47 AM	08/07/2019	7:46 AM	04/12/2019	0.096	8:04 AM	08/05/2019	16040140
66	0		3:37 AM	08/07/2019	7:46 AM	04/12/2019	0.096	8:04 AM	08/05/2019	16040140
237	0		10:50 PM	08/06/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
238	0.131		2:59 AM	08/07/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
239	0.122		3:03 AM	08/07/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
240	0.074		3:13 AM	08/07/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
241	0.142		2:45 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
242	0.137		2:47 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
243	0.076		2:51 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
244	0.058		2:54 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
245	0.086		2:58 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
246	0.08		3:01 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
247	0.063		3:13 AM	08/11/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
248	0		12:11 AM	08/12/2019	8:56 AM	04/25/2017	0.101	9:16 AM	08/05/2019	16040161
327	0.162		3:12 AM	08/09/2019	4:09 PM	06/25/2019	0.093	9:40 AM	08/05/2019	16040162
328	0.158		3:15 AM	08/09/2019	4:09 PM	06/25/2019	0.093	9:40 AM	08/05/2019	16040162
329	0		1:16 PM	08/11/2019	4:09 PM	06/25/2019	0.093	9:40 AM	08/05/2019	16040162
330	0		1:17 PM	08/11/2019	4:09 PM	06/25/2019	0.093	9:40 AM	08/05/2019	16040162
331	0.106		12:21 AM	08/12/2019	4:09 PM	06/25/2019	0.093	9:40 AM	08/05/2019	16040162
332	0.108		12:23 AM	08/12/2019	4:09 PM	06/25/2019	0.093	9:40 AM	08/05/2019	16040162
59	0.114		2:27 AM	08/09/2019	11:51 AM	07/05/2019	0.101	8:20 AM	08/05/2019	16080014
60	0.115		2:31 AM	08/09/2019	11:51 AM	07/05/2019	0.101	8:20 AM	08/05/2019	16080014
61	0.185		2:20 AM	08/11/2019	11:51 AM	07/05/2019	0.101	8:20 AM	08/05/2019	16080014
62	0.182		2:22 AM	08/11/2019	11:51 AM	07/05/2019	0.101	8:20 AM	08/05/2019	16080014
89	0.18		2:06 AM	08/08/2019	3:58 PM	03/28/2019	0.1	8:21 AM	08/05/2019	16080015
90	0.186		2:09 AM	08/08/2019	3:58 PM	03/28/2019	0.1	8:21 AM	08/05/2019	16080015
91	0.184		1:48 AM	08/10/2019	3:58 PM	03/28/2019	0.1	8:21 AM	08/05/2019	16080015
92	0.195		1:50 AM	08/10/2019	3:58 PM	03/28/2019	0.1	8:21 AM	08/05/2019	16080015
93	0.155		4:27 AM	08/10/2019	3:58 PM	03/28/2019	0.1	8:21 AM	08/05/2019	16080015
73	0.059		2:53 AM	08/11/2019	3:58 PM	03/28/2019	0.1	8:20 AM	08/05/2019	16080027

CHP South Los Angeles PAS Usage Logs 8-5-2019 to 8-15-2019

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
74	0.084		7:12 PM	08/11/2019	3:58 PM	03/28/2019	0.1	8:20 AM	08/05/2019	16080027
75	0.084		7:14 PM	08/11/2019	3:58 PM	03/28/2019	0.1	8:20 AM	08/05/2019	16080027
76	0.208		7:26 PM	08/11/2019	3:58 PM	03/28/2019	0.1	8:20 AM	08/05/2019	16080027
77	0.211		7:27 PM	08/11/2019	3:58 PM	03/28/2019	0.1	8:20 AM	08/05/2019	16080027
45	NEG		7:07 PM	08/05/2019	4:12 PM	03/28/2019	0.1	8:28 AM	08/05/2019	16080064
46	0		12:18 AM	08/08/2019	4:12 PM	03/28/2019	0.1	8:28 AM	08/05/2019	16080064
47	0.126		1:24 AM	08/08/2019	4:12 PM	03/28/2019	0.1	8:28 AM	08/05/2019	16080064
48	0.131		1:27 AM	08/08/2019	4:12 PM	03/28/2019	0.1	8:28 AM	08/05/2019	16080064
49	0		3:21 AM	08/09/2019	4:12 PM	03/28/2019	0.1	8:28 AM	08/05/2019	16080064
50	0.06		12:10 AM	08/14/2019	4:12 PM	03/28/2019	0.1	8:28 AM	08/05/2019	16080064
261	0		3:23 AM	08/07/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
262	0.124		3:35 AM	08/08/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
263	0.126		3:38 AM	08/08/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
264	0.094		2:34 AM	08/10/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
265	0.101		2:36 AM	08/10/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
266	0.24		2:44 AM	08/11/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
267	0.21		2:48 AM	08/11/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
268	0.214		2:51 AM	08/11/2019	1:54 PM	12/19/2017	0.103	8:30 AM	08/05/2019	16080065
115	0.153		12:09 AM	08/10/2019	4:18 PM	03/28/2019	0.098	8:35 AM	08/05/2019	16080069
116	0.135		11:04 PM	08/10/2019	4:18 PM	03/28/2019	0.098	8:35 AM	08/05/2019	16080069
117	0.157		11:07 PM	08/10/2019	4:18 PM	03/28/2019	0.098	8:35 AM	08/05/2019	16080069
118	0.149		11:09 PM	08/10/2019	4:18 PM	03/28/2019	0.098	8:35 AM	08/05/2019	16080069
119	0.074		2:15 AM	08/11/2019	4:18 PM	03/28/2019	0.098	8:35 AM	08/05/2019	16080069
120	0.071		2:17 AM	08/11/2019	4:18 PM	03/28/2019	0.098	8:35 AM	08/05/2019	16080069
407	0.132		3:21 AM	08/10/2019	10:41 AM	08/14/2017	0.102	8:35 AM	08/05/2019	16080071
408	0.131		3:24 AM	08/10/2019	10:41 AM	08/14/2017	0.102	8:35 AM	08/05/2019	16080071
55	0		10:31 PM	08/14/2019	4:36 PM	03/28/2019	0.107	8:46 AM	08/05/2019	16120175
115	0.141		2:36 AM	08/12/2019	9:50 AM	04/25/2017	0.104	8:53 AM	08/05/2019	16120177
116	0.14		2:39 AM	08/12/2019	9:50 AM	04/25/2017	0.104	8:53 AM	08/05/2019	16120177
324	0		3:04 AM	08/11/2019	6:53 AM	05/30/2019	0.109	8:53 AM	08/05/2019	16120194
325	0.187		5:43 AM	08/11/2019	6:53 AM	05/30/2019	0.109	8:53 AM	08/05/2019	16120194
326	0.201		5:46 AM	08/11/2019	6:53 AM	05/30/2019	0.109	8:53 AM	08/05/2019	16120194
352	0.162		2:54 AM	08/12/2019	12:48 PM	06/09/2019	0.105	8:59 AM	08/05/2019	16120195

CHP South Los Angeles PAS Usage Logs 8-5-2019 to 8-15-2019

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
353	0.124		2:55 AM	08/12/2019	12:48 PM	06/09/2019	0.105	8:59 AM	08/05/2019	16120195
9	0.006		11:57 PM	08/05/2019	4:42 PM	03/28/2019	0.103	9:06 AM	08/05/2019	17170066

#1

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 160 40162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #	TEST #2 (Optional)	YES	NO				
08/15/2019	1105	34	181P41076	.100	22		.100		X		M. Tippett	20137		

Solution prepared by: Cesar A. Mejia
Date: 09-10-18
Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:09 PM	06/25/2019	.100	.100	11:06 AM	08/15/2019	BAC	16040162	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1337	34	181P41076	.100	31		.100			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard		Last Cal		Cal Standard		Last Check		UOM		Serial No.		FC20		Calibration / Check Printout	Lifeloc Technologies
Value	Time	Value	Time	Value	Time	Value	Time	Value	Time	Value	Time	Value	Time		
.100	3:25 PM	.100	03/28/2019	.100	1:37 PM	.100	08/15/2019	BAC		16040119		8.7.24			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1124 34		181P41076							M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:46 AM	04/12/2019	.100	.100	11:24 AM	08/15/2019	BAC	16040140	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	11:30	34	181P41076	.100	22		105				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:41 PM	03/28/2019	.105	.100	11:30 AM	08/15/2019	BAC	16040141	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	11:37	34	181P41076	.100	22		.105				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.105	.100	11:38 AM	08/15/2019	BAC	16040161	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
08/15/2019	11:33	34	181P41076	.100	23		.098					X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.098	.100	11:33 AM	08/15/2019	BAC	16040163	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

7

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							YES	NO				
08/15/2019	1148	39	181P41076	.100	25			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.102	.100	11:48 AM	08/15/2019	BAC	16080013	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.					TEMP #1	TEMP #2	TEST #1	TEST #2	YES	NO				
08/15/2019	1149	34	181P41076		.100	24								X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:51 AM	07/05/2019	.103	.100	11:44 AM	08/15/2019	BAC	16080014	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1144	34	181P41076	.100	23		.103				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.103	.100	11:45 AM	08/15/2019	BAC	16080015	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID NUMBER: 20137 PAS MAKE/MODEL: Liteloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1143-34	181P41076		100	23		104				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 09-10-18														
Expiration date: 09-10-20														
Cal Standard														
Time: 3:58 PM														
Last Cal														
03/28/2019														
Result														
.104														
Cal Standard														
.100														
Time: 11:44 AM														
Last Check														
08/15/2019														
UOM														
BAC														
Serial No.														
16080027														
FC20														
8.7.24														
Calibration / Check Printout														

Liteloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID NUMBER: 20137
 PAS MAKE/MODEL: Lifoloc FC20
 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1148 34		181P41076	.100		22		.103			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifoloc Technologies
.100	4:05 PM	03/28/2019	.103	.100	11:48 AM	08/15/2019	BAC	16080028	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
08/15/2019	1148	34	181P41076	.100	22								X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:07 PM	03/28/2019	.103	.100	11:48 AM	08/15/2019	BAC	16080030	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	11:51 AM	34	181P41076	.100	22		.101				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:12 PM	03/28/2019	.101	.100	11:51 AM	08/15/2019	BAC	16080064	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#114

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1152 34		181P41076	100	22							M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	1:54 PM	12/19/2017	.104	.100	11:53 AM	08/15/2019	BAC	16080065	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	11:57 AM	18.1	P41076	.100	2.3		.098		X		M. Tippett	20137		

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard		Last Cal		Cal Standard		Last Check		UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
Time:	Result	Time:	Result	Time:	Result	Time:	Result					
4:18 PM	.100	03/28/2019	.098	11:57 AM	.100	08/15/2019	.098	BAC	16080069	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	11:58	54	181P41076	.100	23		100					M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifloc Technologies
.100	10:41 AM	08/14/2017	.100	.100	11:58 AM	08/15/2019	BAC	16080071	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1215	34	181P41076	.100	22			104				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.104	.100	12:15 PM	08/15/2019	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1208	34	181P41076	.100		22		.103				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 09-10-18 Expiration date: 09-10-20														
Cal Standard														
Time: 4:24 PM														
Last Cal														
03/28/2019														
Result														
.103														
Cal Standard														
Time: 12:08 PM														
Last Check														
08/15/2019														
UOM														
BAC														
Serial No.														
16080075														
FC20														
8.7.24														
Calibration / Check Printout														

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: *16080077*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)					
08/15/2019	<i>12:09</i>	<i>34</i>	181P41076	.100	<i>23</i>		<i>104</i>		NO	X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 09-10-18 Expiration date: 09-10-20													
				Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
				.100	4:30 PM	03/28/2019	.104	.100	12:10 PM	08/15/2019	BAC	16080077	8.7.24
Calibration / Check Printout													
Lifeloc Technologies													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16/20175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2					
08/15/2019	12:10	34	181P41076	.100	22		.108		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.108	.100	12:10 PM	08/15/2019	BAC	16120175	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1219	34	181P41076	.100	27						X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	09:50 AM	04/25/2017	.104	.100	12:19 PM	08/15/2019	BAC	16120177	8.7.24

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	12:18	34	181P41076	.100	23		.101				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:16 AM	10/21/2016	.101	.100	12:18 PM	08/15/2019	BAC	16120188	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
08/15/2019	122234	181P41076		.100	23	.103			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.103	.100	12:23 PM	08/15/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16190194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)					
08/15/2019	1219	34	181P41076	.100	22		111		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	12:20 PM	08/15/2019	.099	.100	12:20 PM	08/15/2019	BAC	16120194	8.7.24

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)					
08/15/2019	12:29	34	181P41076	.100	22		.104		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.104	.100	12:29 PM	08/15/2019	BAC	16120195	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	1235	34	181P41076	.100		22		.099			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 09-10-18														
Expiration date: 09-10-20														
Cal Standard														
Time: 3:25 PM														
Last Cal														
06/25/2019														
Result														
.099														
Cal Standard														
Time: 12:36 PM														
Last Check														
08/15/2019														
UOM														
BAC														
Serial No.														
16120197														
FC20														
8.7.24														
Calibration / Check Printout														
Lifeloc Technologies														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019	12:38	34	181P41076	.100	21		.104			NO		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:42 PM	03/28/2019	.104	.100	12:38 PM	08/15/2019	BAC	17170066	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 17170150

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/15/2019			181P41076	.100								M. Tippett	20137	00 SERVICE
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 09-10-18											
			Expiration date:											
			09-10-20											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

#29

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	YES	NO				
08/15/2019	1240	34	181P41076	.100	22	102					X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	4:46 PM	03/28/2019	.102	.100	12:40 PM	08/15/2019	BAC	17170155	8.7.24	
Calibration / Check Printout										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

50

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 90102844-82

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/15/2019		181P41076		.100							M. Tippett	20137	DO SERVICE
		Solution prepared by: Cesar A. Mejia											
		Date: 09-10-18											
		Expiration date: 09-10-20											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
08/15/2019	1239	34	181P41076	.100	21								X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date: 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	5:05 PM	03/28/2019	.106	.100	12:39 PM	08/15/2019	BAC	17170064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.