

CHP SLA PAS Usage logs 8-15-19 to 8-26-19

| Test Number | Result | ID | Test Time | Test Date | Cal Time | Cal Date | Check Result | Check Time | Check Date | Unit Serial Number |
|-------------|--------|----|-----------|------------|----------|------------|--------------|------------|------------|--------------------|
| 54 | 0.189 | | 11:43 PM | 08/18/2019 | 3:25 PM | 03/28/2019 | 0.1 | 1:37 PM | 08/15/2019 | 16040119 |
| 55 | 0.183 | | 11:45 PM | 08/18/2019 | 3:25 PM | 03/28/2019 | 0.1 | 1:37 PM | 08/15/2019 | 16040119 |
| 56 | 0 | | 11:07 PM | 08/22/2019 | 3:25 PM | 03/28/2019 | 0.1 | 1:37 PM | 08/15/2019 | 16040119 |
| 57 | NEG | | 8:13 PM | 08/24/2019 | 3:25 PM | 03/28/2019 | 0.1 | 1:37 PM | 08/15/2019 | 16040119 |
| 67 | 0 | | 6:29 PM | 08/17/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 68 | 0 | | 6:29 PM | 08/17/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 69 | 0.077 | | 11:36 PM | 08/17/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 70 | 0.096 | | 2:11 AM | 08/18/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 71 | 0.092 | | 2:13 AM | 08/18/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 72 | 0.091 | | 11:20 PM | 08/23/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 73 | 0.087 | | 11:23 PM | 08/23/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 74 | 0.102 | | 1:32 AM | 08/24/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 75 | 0.103 | | 1:36 AM | 08/24/2019 | 7:46 AM | 04/12/2019 | 0.1 | 11:24 AM | 08/15/2019 | 16040140 |
| 44 | 0.214 | | 12:07 AM | 08/25/2019 | 3:41 PM | 03/28/2019 | 0.105 | 11:30 AM | 08/15/2019 | 16040141 |
| 45 | 0.212 | | 12:10 AM | 08/25/2019 | 3:41 PM | 03/28/2019 | 0.105 | 11:30 AM | 08/15/2019 | 16040141 |
| 249 | 0.086 | | 12:09 AM | 08/17/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 250 | 0.081 | | 12:12 AM | 08/17/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 251 | 0.133 | | 11:16 PM | 08/17/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 252 | 0.131 | | 11:19 PM | 08/17/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 253 | 0.104 | | 11:26 PM | 08/24/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 254 | 0.11 | | 11:29 PM | 08/24/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 255 | 0.223 | | 6:13 PM | 08/25/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 256 | 0.23 | | 6:17 PM | 08/25/2019 | 8:56 AM | 04/25/2017 | 0.105 | 11:38 AM | 08/15/2019 | 16040161 |
| 333 | 0 | | 1:43 PM | 08/17/2019 | 4:09 PM | 06/25/2019 | 0.1 | 11:06 AM | 08/15/2019 | 16040162 |
| 334 | 0 | | 1:45 PM | 08/17/2019 | 4:09 PM | 06/25/2019 | 0.1 | 11:06 AM | 08/15/2019 | 16040162 |
| 335 | 0 | | 1:45 PM | 08/17/2019 | 4:09 PM | 06/25/2019 | 0.1 | 11:06 AM | 08/15/2019 | 16040162 |
| 336 | 0.171 | | 3:40 AM | 08/25/2019 | 4:09 PM | 06/25/2019 | 0.1 | 11:06 AM | 08/15/2019 | 16040162 |
| 337 | 0.163 | | 3:46 AM | 08/25/2019 | 4:09 PM | 06/25/2019 | 0.1 | 11:06 AM | 08/15/2019 | 16040162 |
| 402 | 0.178 | | 6:19 PM | 08/17/2019 | 7:30 AM | 04/12/2019 | 0.098 | 11:33 AM | 08/15/2019 | 16040163 |
| 403 | 0.183 | | 6:21 PM | 08/17/2019 | 7:30 AM | 04/12/2019 | 0.098 | 11:33 AM | 08/15/2019 | 16040163 |
| 78 | NEG | | 1:04 AM | 08/17/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 79 | 0.02 | | 1:05 AM | 08/17/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 80 | 0.171 | | 4:14 PM | 08/17/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |

| Test Number | Result | ID | Test Time | Test Date | Cal Time | Cal Date | Check Result | Check Time | Check Date | Unit Serial Number |
|-------------|--------|----|-----------|------------|----------|------------|--------------|------------|------------|--------------------|
| 81 | 0.185 | | 4:16 PM | 08/17/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 82 | 0.121 | | 2:46 AM | 08/18/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 83 | 0.12 | | 2:49 AM | 08/18/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 84 | 0.235 | | 8:51 PM | 08/18/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 85 | 0.263 | | 8:54 PM | 08/18/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 86 | 0.241 | | 8:57 PM | 08/18/2019 | 3:58 PM | 03/28/2019 | 0.104 | 11:44 AM | 08/15/2019 | 16080027 |
| 38 | 0.247 | | 10:46 PM | 08/17/2019 | 4:05 PM | 03/28/2019 | 0.103 | 11:48 AM | 08/15/2019 | 16080028 |
| 39 | 0.204 | | 10:53 PM | 08/17/2019 | 4:05 PM | 03/28/2019 | 0.103 | 11:48 AM | 08/15/2019 | 16080028 |
| 40 | 0.154 | | 6:40 PM | 08/20/2019 | 4:05 PM | 03/28/2019 | 0.103 | 11:48 AM | 08/15/2019 | 16080028 |
| 41 | 0.157 | | 6:41 PM | 08/20/2019 | 4:05 PM | 03/28/2019 | 0.103 | 11:48 AM | 08/15/2019 | 16080028 |
| 37 | 0.238 | | 1:06 AM | 08/18/2019 | 4:07 PM | 03/28/2019 | 0.103 | 11:48 AM | 08/15/2019 | 16080030 |
| 38 | 0.213 | | 1:08 AM | 08/18/2019 | 4:07 PM | 03/28/2019 | 0.103 | 11:48 AM | 08/15/2019 | 16080030 |
| 51 | 0.163 | | 11:03 PM | 08/16/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 52 | 0.165 | | 11:06 PM | 08/16/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 53 | 0 | | 2:22 AM | 08/18/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 54 | 0.07 | | 2:45 AM | 08/18/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 55 | 0.069 | | 2:47 AM | 08/18/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 56 | 0.109 | | 3:27 AM | 08/18/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 57 | 0.108 | | 3:29 AM | 08/18/2019 | 4:12 PM | 03/28/2019 | 0.101 | 11:51 AM | 08/15/2019 | 16080064 |
| 269 | 0.158 | | 10:25 PM | 08/24/2019 | 1:54 PM | 12/19/2017 | 0.104 | 11:53 AM | 08/15/2019 | 16080065 |
| 270 | 0.057 | | 2:47 AM | 08/25/2019 | 1:54 PM | 12/19/2017 | 0.104 | 11:53 AM | 08/15/2019 | 16080065 |
| 271 | 0.157 | | 3:06 AM | 08/25/2019 | 1:54 PM | 12/19/2017 | 0.104 | 11:53 AM | 08/15/2019 | 16080065 |
| 272 | 0.16 | | 3:08 AM | 08/25/2019 | 1:54 PM | 12/19/2017 | 0.104 | 11:53 AM | 08/15/2019 | 16080065 |
| 121 | 0.016 | | 12:59 AM | 08/17/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 122 | 0.119 | | 4:11 AM | 08/17/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 123 | 0.121 | | 4:13 AM | 08/17/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 124 | 0.333 | | 9:41 PM | 08/17/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 125 | 0.325 | | 9:42 PM | 08/17/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 126 | 0.014 | | 1:42 AM | 08/21/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 127 | 0.014 | | 1:45 AM | 08/21/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 128 | 0.013 | | 2:42 AM | 08/24/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 129 | 0.166 | | 2:01 AM | 08/25/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |
| 130 | 0.172 | | 2:04 AM | 08/25/2019 | 4:18 PM | 03/28/2019 | 0.098 | 11:57 AM | 08/15/2019 | 16080069 |

CHP SLA PAS Usage logs 8-15-19 to 8-26-19

| Test Number | Result | ID | Test Time | Test Date | Cal Time | Cal Date | Check Result | Check Time | Check Date | Unit Serial Number |
|-------------|--------|----|-----------|------------|----------|------------|--------------|------------|------------|--------------------|
| 409 | 0.087 | | 11:14 PM | 08/23/2019 | 10:41 AM | 08/14/2017 | 0.1 | 11:58 AM | 08/15/2019 | 16080071 |
| 14 | 0.108 | | 1:43 AM | 08/18/2019 | 4:24 PM | 03/28/2019 | 0.103 | 12:08 PM | 08/15/2019 | 16080075 |
| 15 | 0.114 | | 1:47 AM | 08/18/2019 | 4:24 PM | 03/28/2019 | 0.103 | 12:08 PM | 08/15/2019 | 16080075 |
| 16 | 0.039 | | 1:55 AM | 08/18/2019 | 4:24 PM | 03/28/2019 | 0.103 | 12:08 PM | 08/15/2019 | 16080075 |
| 106 | 0.017 | | 11:54 PM | 08/15/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 107 | 0 | | 1:13 AM | 08/17/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 108 | 0.138 | | 4:30 AM | 08/18/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 109 | 0.133 | | 4:32 AM | 08/18/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 110 | 0.17 | | 1:49 AM | 08/23/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 111 | 0.165 | | 1:54 AM | 08/23/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 112 | 0.176 | | 2:52 AM | 08/24/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 113 | 0.172 | | 2:55 AM | 08/24/2019 | 4:30 PM | 03/28/2019 | 0.104 | 12:10 PM | 08/15/2019 | 16080077 |
| 56 | 0.057 | | 9:44 PM | 08/22/2019 | 4:36 PM | 03/28/2019 | 0.108 | 12:10 PM | 08/15/2019 | 16120175 |
| 57 | 0.057 | | 9:46 PM | 08/22/2019 | 4:36 PM | 03/28/2019 | 0.108 | 12:10 PM | 08/15/2019 | 16120175 |
| 58 | 0.132 | | 8:08 PM | 08/25/2019 | 4:36 PM | 03/28/2019 | 0.108 | 12:10 PM | 08/15/2019 | 16120175 |
| 59 | 0.131 | | 8:10 PM | 08/25/2019 | 4:36 PM | 03/28/2019 | 0.108 | 12:10 PM | 08/15/2019 | 16120175 |
| 117 | 0 | | 12:31 AM | 08/18/2019 | 9:50 AM | 04/25/2017 | 0.104 | 12:19 PM | 08/15/2019 | 16120177 |
| 118 | 0.07 | | 11:58 PM | 08/22/2019 | 9:50 AM | 04/25/2017 | 0.104 | 12:19 PM | 08/15/2019 | 16120177 |
| 119 | 0.175 | | 1:18 AM | 08/24/2019 | 9:50 AM | 04/25/2017 | 0.104 | 12:19 PM | 08/15/2019 | 16120177 |
| 120 | 0.187 | | 1:20 AM | 08/24/2019 | 9:50 AM | 04/25/2017 | 0.104 | 12:19 PM | 08/15/2019 | 16120177 |
| 90 | 0.152 | | 12:57 AM | 08/24/2019 | 8:16 AM | 10/21/2016 | 0.101 | 12:18 PM | 08/15/2019 | 16120188 |
| 91 | 0.152 | | 1:00 AM | 08/24/2019 | 8:16 AM | 10/21/2016 | 0.101 | 12:18 PM | 08/15/2019 | 16120188 |
| 92 | 0 | | 2:28 AM | 08/24/2019 | 8:16 AM | 10/21/2016 | 0.101 | 12:18 PM | 08/15/2019 | 16120188 |
| 354 | 0.187 | | 2:08 AM | 08/23/2019 | 12:48 PM | 06/09/2019 | 0.104 | 12:29 PM | 08/15/2019 | 16120195 |
| 355 | 0.184 | | 2:10 AM | 08/23/2019 | 12:48 PM | 06/09/2019 | 0.104 | 12:29 PM | 08/15/2019 | 16120195 |

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-----------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1:54 | 54 | 181P41076 | | | | | | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|
| .100 | 4:09 PM | 06/25/2019 | .102 | .100 | 3:15 PM | 08/26/2019 | BAC | 16040162 | 8.7.24 |

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS | |
|------------|---------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|------------|---------------------------|--|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | | |
| 08/26/2019 | 1520 34 | 181P41076 | | 100 | 23 | | .102 | | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 3:25 PM | 03/28/2019 | .102 | .100 | 3:21 PM | 08/26/2019 | BAC | 16040119 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-----------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1513 | 59 | 181P41076 | 73 | 100 | 100 | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 07:46 AM | 04/12/2019 | .100 | .100 | 3:14 PM | 08/26/2019 | BAC | 16040140 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|---------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1579 34 | 181P41076 | .100 | 73 | | .104 | | | X | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 3:41 PM | 03/28/2019 | .104 | .100 | 3:19 PM | 08/26/2019 | BAC | 16040141 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------|-------|---------|--------------------|-----------------------------|--------------------|----------------------|--------------------|------------|----------|---------------------------|
| | | | | | | | TEST #1 (Optional) | TEST #2 (Optional) | | | | | |
| 08/26/2019 | 1526 | 74 | 181P41076 | 100 | 23 | | 106 | | NO | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifeloc Technologies |
|--------------|--|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|----------------------|
| .100 | | 08:56 AM | 04/25/2017 | .106 | .100 | 3:27 PM | 08/26/2019 | BAC | 16040161 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|-------|-----------|-----------------------------------|--------------------|-------|--------------------|--------------------|-----------------------------|-----|----------------------|---|--------------------|------------|----------|---------------------------|
| | | | LOT NO. | TEST #1 (Optional) | | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | | |
| 08/26/2019 | 15:21 | 34 | 181P41076 | | 100 | 24 | | -100 | | | X | | M. Tippett | 20137 | |

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date: 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 07:30 AM | 04/12/2019 | .100 | .100 | 3:22 PM | 08/26/2019 | BAC | 16040163 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 153634 | 181P41076 | | 100 | 24 | .103 | | | | X | M. Tippett | 20137 | |

| Solution prepared by: | | Cal Standard | | Last Cal | | Cal Standard | | Last Check | | UOM | | Serial No. | | FC20 | | Calibration / Check Printout | Lifeloc Technologies |
|---------------------------|--|--------------|----------|----------|---------|--------------|------------|------------|-------|-----|--|------------|--------|------|--|------------------------------|----------------------|
| | | Value | Time: | Result | Time: | Result | Time: | Result | Time: | | | | | | | | |
| Cesar A. Mejia | | .100 | 12:59 PM | .103 | 3:37 PM | .100 | 08/26/2019 | BAC | | | | 16080013 | 8.7.24 | | | | |
| Date: 09-10-18 | | | | | | | | | | | | | | | | | |
| Expiration date: 09-10-20 | | | | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | | OPERATOR | ID. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|-----------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|---------|------------|---------|-------------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | TEST #1 | TEST #2 | | | |
| 08/26/2019 | 1532 | 34 | 181P41076 | .100 | 24 | .103 | | | | | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifeloc Technologies |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|----------------------|
| .100 | 11:51 AM | 07/05/2019 | .103 | .100 | 3:33 PM | 08/26/2019 | BAC | 16080014 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|-------|-----------------------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|---------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | TEST #1 | TEST #2 | | | |
| 08/26/2019 | 15:32 | 34 | 181P41076 | .100 | 24 | | .105 | | | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| FC20 | Serial No. | UOM | Last Check | Time: | Cal Standard | Result | Last Cal | Time: | Cal Standard | Result | Last Cal | Time: | Cal Standard |
|--------|------------|-----|------------|---------|--------------|--------|------------|---------|--------------|--------|------------|---------|--------------|
| 8.7.24 | 16080015 | BAC | 08/26/2019 | 3:33 PM | .100 | .105 | 03/28/2019 | 3:58 PM | .100 | .105 | 03/28/2019 | 3:58 PM | .100 |

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|--------|--------------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 153239 | 181P41076 | | .100 | 23 | .104 | | | X | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 3:58 PM | 03/28/2019 | .104 | .100 | 3:32 PM | 08/26/2019 | BAC | 16080027 | 8.7.24 | Lifeloc Technologies |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 153634 | 181P41076 | | .100 | 27 | | | | | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifeloc Technologies |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|----------------------|
| .100 | 4:05 PM | 03/28/2019 | .102 | .100 | 3:37 PM | 08/26/2019 | BAC | 16080028 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------------------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 153634 | | 181P41076 | 100 | 23 | | 0/01 | | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifeloc Technologies |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|----------------------|
| .100 | 4:07 PM | 03/28/2019 | .101 | .100 | 3:37 PM | 08/26/2019 | BAC | 16080030 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#13

| | | | | | |
|-------------------|----------|------------------|-------------|----------------|----------------|
| AREA NAME | AREA NO. | COORDINATOR NAME | I.D. NUMBER | PAS MAKE/MODEL | PAS SERIAL NO. |
| South Los Angeles | 530 | Michael Tippett | 20137 | Lifoloc FC20 | 16080064 |

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|--------|--------------------------------------|---------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 153854 | 181P41076 | | .100 | | 25 | | .103 | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifoloc Technologies |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|----------------------|
| .100 | 4:12 PM | 03/28/2019 | .103 | .100 | 3:38 PM | 08/26/2019 | BAC | 16080064 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|----------|-----------------------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 15:34:34 | | 181P41076 | .100 | 24 | | .105 | | | | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | REPAIRS/PROBLEMS/COMMENTS |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 1:54 PM | 12/19/2017 | .105 | .100 | 3:40 PM | 08/26/2019 | BAC | 16080065 | 8.7.24 | Calibration / Check Printout |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1545 | 34 | 181P41076 | .100 | 24 | | | .102 | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifeloc Technologies |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|----------------------|
| .100 | 10:41 AM | 08/14/2017 | .102 | .100 | 3:45 PM | 08/26/2019 | BAC | 16080071 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Liteloc FC20
PAS SERIAL NO.: 160 80072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 160/34 | 181P41076 | .100 | 22 | | .107 | | | | | M. Tippett | 20137 | |

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 2:05 PM | 12/19/2017 | .107 | .100 | 4:02 PM | 08/26/2019 | BAC | 16080072 | 8.7.24 | |

Liteloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|---------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 5:34 PM | 181P41076 | .100 | 24 | | .103 | | | X | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | REPAIRS/PROBLEMS/COMMENTS |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|---------------------------|
| .100 | 4:24 PM | 03/28/2019 | .103 | .100 | 3:54 PM | 08/26/2019 | BAC | 16080075 | 8.7.24 | |

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 153434 | 181P41076 | | .100 | 24 | | | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 4:30 PM | 03/28/2019 | .104 | .100 | 3:55 PM | 08/26/2019 | BAC | 16080077 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|---------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1:55 PM | 181P41076 | | .100 | 23 | | 106 | | X | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | REPAIRS/PROBLEMS/COMMENTS |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|---------------------------|
| .100 | 4:36 PM | 03/28/2019 | .106 | .100 | 3:55 PM | 08/26/2019 | BAC | 16120175 | 8.7.24 | |

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | TEMP (°C) | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|---------|--------------------------------------|---------|-----------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1602 34 | 181P41076 | | .100 | 24 | .105 | | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 09:50 AM | 04/25/2017 | .105 | .100 | 4:03 PM | 08/26/2019 | BAC | 16120177 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------------------------------|---------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 160334 | 181P41076 | | .100 | 24 | | .103 | | | | X | M. Tippett | 20137 | |

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | REPAIRS/PROBLEMS/COMMENTS |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|---------------------------|
| .100 | 08:50 AM | 04/12/2019 | .103 | .100 | 4:03 PM | 08/26/2019 | BAC | 16120189 | 8.7.24 | |

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|---------|-----------------------------------|-----------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1:59:34 | | 181P41076 | | | .100 | | | | M. Tippett | 20137 | |

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Lifeloc Technologies |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 12:20 PM | 08/15/2019 | .100 | .100 | 4:00 PM | 08/26/2019 | BAC | 16120194 | 8.7.24 | |
| | | | | | | | | | | Calibration / Check Printout |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | TEST #1 | TEST #2 (Optional) | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|---------|--------------------------------------|-----------|-------|---------|-----------------------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | | | | | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1605 34 | | 181P41076 | .100 | 24 | | | | | | | X | M. Tippett | 20137 | |

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 08:16 AM | 10/21/2016 | .101 | .100 | 4:06 PM | 08/26/2019 | BAC | 16120188 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16/20197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|--------|--------------------------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 16/234 | | 181D41076 | .100 | 23 | | .104 | | | | X | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 3:25 PM | 06/25/2019 | .104 | .100 | 4:12 PM | 08/26/2019 | BAC | 16120197 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippet
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifloc FC20
PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS/ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|---------|--------------------------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|-----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1610 34 | | 181D41076 | .100 | | 24 | | .106 | | | X | M. Tippet | 20137 | |

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout | Lifloc Technologies |
|--------------|----------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|---------------------|
| .100 | 12:48 PM | 06/09/2019 | .106 | .100 | 4:10 PM | 08/26/2019 | BAC | 16120195 | 8.7.24 | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | TEST #1 | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|--------|-----------|------------|-------|---------|--------------------|--------------------|-----------------------------|-----|----------------------|--|--------------------|------------|----------|---------------------------|
| | | | | | | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | | |
| 08/26/2019 | 161734 | | 181DP41076 | .100 | 23 | | | | | | | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | Calibration / Check Printout |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|------------------------------|
| .100 | 4:42 PM | 03/28/2019 | .106 | .100 | 4:18 PM | 08/26/2019 | BAC | 17170066 | 8.7.24 | |

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|---------|--------------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1621 34 | 181P41076 | .100 | 72 | | 104 | | | X | | M. Tippett | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 | |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|--|
| .100 | 4:46 PM | 03/28/2019 | .104 | .100 | 4:22 PM | 08/26/2019 | BAC | 17170155 | 8.7.24 | |

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | LOT NO. | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|---------|-------|--------------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|------------|----------|---------------------------|
| | | | | | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | | 18 | 1P41076 | 100 | | | | | | | | M. Tippett | 20137 | 02 SERVICE |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 CHP202K_1108.pdf

#31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippet I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | TEMP (°C) | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-----------|-----------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|-----------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | | | TEST #1 (Optional) | TEST #2 (Optional) | YES | NO | | | | |
| 08/26/2019 | 1616 | 57 | 181P41076 | .100 | 22 | .109 | | | X | | M. Tippet | 20137 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

| Cal Standard | Time: | Last Cal | Result | Cal Standard | Time: | Last Check | UOM | Serial No. | FC20 |
|--------------|---------|------------|--------|--------------|---------|------------|-----|------------|--------|
| .100 | 5:05 PM | 03/28/2019 | .109 | .100 | 4:17 PM | 08/26/2019 | BAC | 17170064 | 8.7.24 |

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.