

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
47	0.079		3:08 AM	08/04/2019	3:25 PM	03/28/2019	0.096	11:48 AM	07/25/2019	16040119
48	0.077		3:12 AM	08/04/2019	3:25 PM	03/28/2019	0.096	11:48 AM	07/25/2019	16040119
49	0.071		3:18 AM	08/04/2019	3:25 PM	03/28/2019	0.096	11:48 AM	07/25/2019	16040119
50	0.228		5:04 AM	08/04/2019	3:25 PM	03/28/2019	0.096	11:48 AM	07/25/2019	16040119
51	0.219		5:06 AM	08/04/2019	3:25 PM	03/28/2019	0.096	11:48 AM	07/25/2019	16040119
36	0.049		1:02 AM	07/28/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
37	0.219		2:41 AM	07/28/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
38	0.213		2:45 AM	07/28/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
39	0.156		2:52 AM	07/28/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
40	0.156		2:55 AM	07/28/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
41	0.01		3:12 AM	07/28/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
42	0		12:46 PM	07/30/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
43	0		10:27 PM	07/30/2019	3:41 PM	03/28/2019	0.102	11:47 AM	07/25/2019	16040141
235	0.147		2:35 AM	07/28/2019	8:56 AM	04/25/2017	0.1	11:55 AM	07/25/2019	16040161
236	0.145		2:38 AM	07/28/2019	8:56 AM	04/25/2017	0.1	11:55 AM	07/25/2019	16040161
320	0		7:12 PM	07/28/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
321	0		7:15 PM	07/28/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
322	0.157		9:04 PM	07/28/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
323	0.172		9:08 PM	07/28/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
324	0.034		12:41 PM	08/02/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
325	0.014		12:45 PM	08/02/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
326	0.031		12:48 PM	08/02/2019	4:09 PM	06/25/2019	0.095	11:37 AM	07/25/2019	16040162
396	0.15		12:40 AM	07/27/2019	7:30 AM	04/12/2019	0.096	11:51 AM	07/25/2019	16040163
397	0.14		12:43 AM	07/27/2019	7:30 AM	04/12/2019	0.096	11:51 AM	07/25/2019	16040163
398	0.019		1:00 AM	07/28/2019	7:30 AM	04/12/2019	0.096	11:51 AM	07/25/2019	16040163
399	0.045		10:24 AM	07/28/2019	7:30 AM	04/12/2019	0.096	11:51 AM	07/25/2019	16040163
400	0.192		9:56 PM	08/02/2019	7:30 AM	04/12/2019	0.096	11:51 AM	07/25/2019	16040163
401	0.189		10:00 PM	08/02/2019	7:30 AM	04/12/2019	0.096	11:51 AM	07/25/2019	16040163
57	0.024		7:42 PM	07/28/2019	11:51 AM	07/05/2019	0.102	12:00 PM	07/25/2019	16080014
58	0.022		7:45 PM	07/28/2019	11:51 AM	07/05/2019	0.102	12:00 PM	07/25/2019	16080014
76	0.261		1:59 AM	07/27/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
77	0.266		2:02 AM	07/27/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
78	0.029		4:18 AM	07/27/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
79	0.026		4:20 AM	07/27/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
80	0.067		9:50 PM	07/28/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
81	0.07		9:53 PM	07/28/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
82	0.138		12:38 AM	07/29/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
83	0.134		12:41 AM	07/29/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
84	0.151		2:17 AM	07/29/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
85	0.143		2:19 AM	07/29/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
86	0.058		10:24 PM	07/30/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
87	0.177		1:43 AM	07/31/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
88	0.172		1:47 AM	07/31/2019	3:58 PM	03/28/2019	0.101	12:02 PM	07/25/2019	16080015
66	0.097		12:58 AM	07/27/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
67	0.098		1:00 AM	07/27/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
68	0.076		3:23 AM	07/28/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
69	0.077		3:26 AM	07/28/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
70	0.153		11:15 PM	08/04/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
71	0.164		11:17 PM	08/04/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
72	0		11:51 PM	08/04/2019	3:58 PM	03/28/2019	0.103	12:01 PM	07/25/2019	16080027
31	0		1:37 AM	07/28/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
32	0.167		4:14 AM	07/28/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
33	0.069		2:15 AM	07/30/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
34	0.133		11:21 PM	08/03/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
35	0.138		11:23 PM	08/03/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
36	0.098		12:39 AM	08/04/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
37	0.099		12:41 AM	08/04/2019	4:05 PM	03/28/2019	0.102	12:06 PM	07/25/2019	16080028
36	0.06		1:45 AM	08/04/2019	4:07 PM	03/28/2019	0.102	12:05 PM	07/25/2019	16080030
40	0.206		5:30 PM	07/28/2019	4:12 PM	03/28/2019	0.101	12:07 PM	07/25/2019	16080064
41	0.211		5:32 PM	07/28/2019	4:12 PM	03/28/2019	0.101	12:07 PM	07/25/2019	16080064
42	0		5:33 PM	08/03/2019	4:12 PM	03/28/2019	0.101	12:07 PM	07/25/2019	16080064
43	0.137		7:48 PM	08/03/2019	4:12 PM	03/28/2019	0.101	12:07 PM	07/25/2019	16080064
44	0.12		7:50 PM	08/03/2019	4:12 PM	03/28/2019	0.101	12:07 PM	07/25/2019	16080064
256	0.049		1:57 AM	07/27/2019	1:54 PM	12/19/2017	0.101	12:09 PM	07/25/2019	16080065
257	0.017		2:44 AM	07/29/2019	1:54 PM	12/19/2017	0.101	12:09 PM	07/25/2019	16080065
258	0.05		1:57 AM	08/03/2019	1:54 PM	12/19/2017	0.101	12:09 PM	07/25/2019	16080065

CHP South LA PAS usage logs 7-25-2019 to 8-5-2019

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
259	0.218		3:33 AM	08/03/2019	1:54 PM	12/19/2017	0.101	12:09 PM	07/25/2019	16080065
260	0.213		3:35 AM	08/03/2019	1:54 PM	12/19/2017	0.101	12:09 PM	07/25/2019	16080065
113	0.146		12:42 AM	08/04/2019	4:18 PM	03/28/2019	0.099	12:13 PM	07/25/2019	16080069
114	0.147		12:44 AM	08/04/2019	4:18 PM	03/28/2019	0.099	12:13 PM	07/25/2019	16080069
403	0.204		1:53 AM	07/29/2019	10:41 AM	08/14/2017	0.102	12:17 PM	07/25/2019	16080071
404	0.201		1:55 AM	07/29/2019	10:41 AM	08/14/2017	0.102	12:17 PM	07/25/2019	16080071
405	0.083		4:23 AM	08/04/2019	10:41 AM	08/14/2017	0.102	12:17 PM	07/25/2019	16080071
406	0.084		4:26 AM	08/04/2019	10:41 AM	08/14/2017	0.102	12:17 PM	07/25/2019	16080071
25	0.124		11:50 PM	08/02/2019	2:05 PM	12/19/2017	0.106	12:33 PM	07/25/2019	16080072
26	0.116		11:54 PM	08/02/2019	2:05 PM	12/19/2017	0.106	12:33 PM	07/25/2019	16080072
27	0.084		4:21 AM	08/03/2019	2:05 PM	12/19/2017	0.106	12:33 PM	07/25/2019	16080072
28	0.086		4:24 AM	08/03/2019	2:05 PM	12/19/2017	0.106	12:33 PM	07/25/2019	16080072
29	0.08		1:58 AM	08/04/2019	2:05 PM	12/19/2017	0.106	12:33 PM	07/25/2019	16080072
30	0.074		2:01 AM	08/04/2019	2:05 PM	12/19/2017	0.106	12:33 PM	07/25/2019	16080072
103	0		1:34 AM	07/26/2019	4:30 PM	03/28/2019	0.104	12:26 PM	07/25/2019	16080077
104	0		1:35 AM	07/26/2019	4:30 PM	03/28/2019	0.104	12:26 PM	07/25/2019	16080077
105	0		1:35 AM	07/26/2019	4:30 PM	03/28/2019	0.104	12:26 PM	07/25/2019	16080077
52	0		9:43 PM	07/27/2019	4:36 PM	03/28/2019	0.105	12:26 PM	07/25/2019	16120175
53	POS		9:43 PM	07/27/2019	4:36 PM	03/28/2019	0.105	12:26 PM	07/25/2019	16120175
54	0		9:47 PM	07/27/2019	4:36 PM	03/28/2019	0.105	12:26 PM	07/25/2019	16120175
322	NEG		12:33 PM	07/25/2019	6:53 AM	05/30/2019	0.104	11:04 AM	07/15/2019	16120194
323	0.064		12:54 AM	07/27/2019	6:53 AM	05/30/2019	0.109	12:33 PM	07/25/2019	16120194
343	0.12		1:08 AM	07/25/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
344	0.123		1:11 AM	07/25/2019	12:48 PM	06/09/2019	0.099	11:11 AM	07/15/2019	16120195
345	0		12:08 AM	08/03/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
346	0.149		1:50 AM	08/03/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
347	0.132		1:53 AM	08/03/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
348	0.122		2:47 AM	08/04/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
349	0.127		2:50 AM	08/04/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
350	0.103		4:53 AM	08/04/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
351	0.105		4:56 AM	08/04/2019	12:48 PM	06/09/2019	0.105	12:39 PM	07/25/2019	16120195
316	0.204		12:49 AM	08/05/2019	3:25 PM	06/25/2019	0.098	12:46 PM	07/25/2019	16120197
317	0.202		12:53 AM	08/05/2019	3:25 PM	06/25/2019	0.098	12:46 PM	07/25/2019	16120197

#2

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0810 34		181P41076	.100	23	.094			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:25 PM	03/28/2019	.094	.100	08:10 AM	08/05/2019	BAC	16040119	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lieloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	080434		181P41076							M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lieloc Technologies
.100	07:46 AM	04/12/2019	.096	.100	08:04 AM	08/05/2019	BAC	16040140	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
08/05/2019	0809	34	181P41076	100	23	101			X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:41 PM	03/28/2019	.101	.100	08:09 AM	08/05/2019	BAC	16040141	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0915 34		181P41076	.100	22		.101				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	08:56 AM	04/25/2017	.101	.100	09:16 AM	08/05/2019	BAC	16040161	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0911	34	181P41076	.100	24			.095				M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:30 AM	04/12/2019	.095	.100	09:12 AM	08/05/2019	BAC	16040163	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEM/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	082334	181P41076		.100	23	.100	.100	X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEM/COMMENTS
.100	12:59 PM	01/29/2018	.100	.100	08:23 AM	08/05/2019	BAC	16080013	8.7.24	

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

8

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0819	34	181P41076	100	23	.101		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	11:51 AM	07/05/2019	.101	.100	08:20 AM	08/05/2019	BAC	16080014	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				VALUE	TEST #1	TEST #2	YES				
08/05/2019	0820	34	181P41076	.100	23		.100			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	3:58 PM	03/28/2019	.100	.100	08:21 AM	08/05/2019	BAC	16080015	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	082034	181P41076		.100	25		.100				M. Tippett	20137	

Solution prepared by:													
Cesar A. Mejia													
Date: 09-10-18													
Expiration date: 09-10-20													
Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies			
.100	3:58 PM	03/28/2019	.100	.100	08:20 AM	08/05/2019	BAC	16080027	8.7.24	Calibration / Check Printout			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0825	34	181P41076	.100	22	100			X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:05 PM	03/28/2019	.100	.100	08:25 AM	08/05/2019	BAC	16080028	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0825	34	181P41076	.100	22	.101			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:07 PM	03/28/2019	.101	.100	08:25 AM	08/05/2019	BAC	16080030	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)				
08/05/2019	0835	34	181P41076	.100	.098				M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:18 PM	03/28/2019	.098	.100	08:35 AM	08/05/2019	BAC	16080069	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0835	34	181P41076	.100	23		.102					M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	10:41 AM	08/14/2017	.102	.100	08:35 AM	08/05/2019	BAC	16080071	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)				
08/05/2019	0850 34		181P41076		.100				M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.104	.100	08:51 AM	08/05/2019	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifoloc FC20
PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
08/05/2019	0843 34		181P41076	100	22	102			X	M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.102	.100	08:44 AM	08/05/2019	BAC	16080075	8.7.24	Lifoloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS	OPERATOR		I.D. NO.	
08/05/2019	0845	34	181P41076	23		103			X			M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:30 PM	03/28/2019	.103	.100	08:45 AM	08/05/2019	BAC	16080077	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
08/05/2019	0843 34		181P41076		.100	22				M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.107	.100	08:46 AM	08/05/2019	BAC	16120175	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

21

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
ID. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	0853	34	181P41076	100	25			X		M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.104	.100	08:53 AM	08/05/2019	BAC	16120177	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAC VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO						
08/05/2019	0851	34	181P41076	.100	22					X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	08:16 AM	10/21/2016	.100	.100	08:52 AM	08/05/2019	BAC	16120188	8.7.24	
Calibration / Check Printout										
Lifeloc Technologies										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES				
08/05/2019	0853	37	181P41076	.100	23	.102		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.102	.100	08:56 AM	08/05/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2	YES	NO				
08/05/2019	085234		181P41076									M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 09-10-18 Expiration date: 09-10-20														

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	06:53 AM	05/30/2019	.109	.100	08:53 AM	08/05/2019	BAC	16120194	8.7.24	Calibration / Check Printout
Lifeloc Technologies										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett ID. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)					YES
08/05/2019	0859	34	181P41076	.100	22					X			

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.105	.100	08:59 AM	08/05/2019	BAC	16120195	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 15120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
08/05/2019	1437	34	181P41076	.100	23		.097		X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 09-10-18
 Expiration date:
 09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	06/25/2019	.097	.100	2:37 PM	08/05/2019	BAC	16120197	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES				
08/05/2019	0905 34		181P41076	.100	21									M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:42 PM	03/28/2019	.103	.100	09:06 AM	08/05/2019	BAC	17170066	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/05/2019	090834	181P41076		.100	22	102					M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 09-10-18
Expiration date:
09-10-20

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:46 PM	03/28/2019	.102	.100	09:09 AM	08/05/2019	BAC	17170155	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					YES
08/05/2019	090834		181P41076	.100	20		.107				X				
Solution prepared by: Cesar A. Mejia Date: 09-10-18 Expiration date: 09-10-20															
Cal Standard															
Time: 5:05 PM															
Last Cal 03/28/2019															
Result .107															
Cal Standard .100															
Time: 09:09 AM															
Last Check 08/05/2019															
UOM BAC															
Serial No. 17170064															
FC20 8.7.24															
Lifeloc Technologies															
Calibration / Check Printout															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.