

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
62	0.109		2:55 AM	09/06/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
63	0.107		2:58 AM	09/06/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
77	0.027		1:42 AM	09/06/2019	7:46 AM	04/12/2019	0.1	3:14 PM	08/26/2019	16040140
78	0		5:32 PM	09/06/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
79	0.254		7:21 PM	09/06/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
80	0.25		7:24 PM	09/06/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
81	0		12:23 AM	09/08/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
82	0.122		12:43 AM	09/08/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
83	0.156		12:45 AM	09/08/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
84	0.148		12:47 AM	09/08/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
85	0.154		3:22 AM	09/08/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
86	0.138		3:26 AM	09/08/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
87	0		7:59 PM	09/15/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
88	0.152		1:22 AM	09/16/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
89	0.149		1:25 AM	09/16/2019	7:46 AM	04/12/2019	0.101	11:36 AM	09/06/2019	16040140
48	0.006		10:30 PM	09/06/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
49	0.156		1:22 AM	09/07/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
50	0.167		1:25 AM	09/07/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
51	0.126		12:19 AM	09/13/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
52	0.13		12:21 AM	09/13/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
53	0.12		2:37 AM	09/14/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
54	0.12		2:38 AM	09/14/2019	3:41 PM	03/28/2019	0.106	11:41 AM	09/06/2019	16040141
259	0		1:07 AM	09/07/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
260	0.018		1:50 AM	09/07/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
261	0.09		4:42 AM	09/07/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
262	0.092		4:44 AM	09/07/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
263	0.144		5:46 AM	09/07/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
264	0.113		12:16 AM	09/11/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
265	0.102		12:23 AM	09/11/2019	8:56 AM	04/25/2017	0.107	11:56 AM	09/06/2019	16040161
410	0.144		12:01 AM	09/08/2019	7:30 AM	04/12/2019	0.1	11:51 AM	09/06/2019	16040163
411	0.14		12:04 AM	09/08/2019	7:30 AM	04/12/2019	0.1	11:51 AM	09/06/2019	16040163
412	0.04		2:18 AM	09/08/2019	7:30 AM	04/12/2019	0.1	11:51 AM	09/06/2019	16040163
413	0.152		3:09 AM	09/17/2019	7:30 AM	04/12/2019	0.1	11:51 AM	09/06/2019	16040163

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
414	0.165		3:13 AM	09/17/2019	7:30 AM	04/12/2019	0.1	11:51 AM	09/06/2019	16040163
90	0.085		10:40 PM	09/08/2019	3:58 PM	03/28/2019	0.104	12:01 PM	09/06/2019	16080027
91	0.086		10:42 PM	09/08/2019	3:58 PM	03/28/2019	0.104	12:01 PM	09/06/2019	16080027
92	0.217		3:44 AM	09/14/2019	3:58 PM	03/28/2019	0.104	12:01 PM	09/06/2019	16080027
93	0.218		3:47 AM	09/14/2019	3:58 PM	03/28/2019	0.104	12:01 PM	09/06/2019	16080027
42	0.094		11:33 PM	09/07/2019	4:05 PM	03/28/2019	0.102	12:09 PM	09/06/2019	16080028
43	0.093		11:36 PM	09/07/2019	4:05 PM	03/28/2019	0.102	12:09 PM	09/06/2019	16080028
44	0.146		2:53 AM	09/10/2019	4:05 PM	03/28/2019	0.102	12:09 PM	09/06/2019	16080028
45	0.132		2:56 AM	09/10/2019	4:05 PM	03/28/2019	0.102	12:09 PM	09/06/2019	16080028
60	0.044		2:42 AM	09/15/2019	4:12 PM	03/28/2019	0.103	12:08 PM	09/06/2019	16080064
61	0.045		2:44 AM	09/15/2019	4:12 PM	03/28/2019	0.103	12:08 PM	09/06/2019	16080064
62	0.078		3:40 AM	09/15/2019	4:12 PM	03/28/2019	0.103	12:08 PM	09/06/2019	16080064
63	0.078		3:44 AM	09/15/2019	4:12 PM	03/28/2019	0.103	12:08 PM	09/06/2019	16080064
140	0.159		3:09 AM	09/06/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
141	0.161		3:13 AM	09/06/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
142	0.045		2:00 AM	09/07/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
143	0.041		2:01 AM	09/07/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
144	0.034		2:26 AM	09/07/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
145	0.098		2:24 AM	09/08/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
146	0.098		2:27 AM	09/08/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
147	0.008		1:06 AM	09/09/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
148	0		1:48 AM	09/09/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
149	0.018		11:27 PM	09/10/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
150	0.017		11:29 PM	09/10/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
151	0		11:07 PM	09/14/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
152	0.167		5:25 PM	09/15/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
153	0.151		5:27 PM	09/15/2019	4:18 PM	03/28/2019	0.099	12:15 PM	09/06/2019	16080069
416	0.156		4:46 PM	09/11/2019	10:41 AM	08/14/2017	0.102	12:15 PM	09/06/2019	16080071
417	0.154		4:48 PM	09/11/2019	10:41 AM	08/14/2017	0.102	12:15 PM	09/06/2019	16080071
418	0.132		12:07 AM	09/15/2019	10:41 AM	08/14/2017	0.102	12:15 PM	09/06/2019	16080071
419	0.136		12:11 AM	09/15/2019	10:41 AM	08/14/2017	0.102	12:15 PM	09/06/2019	16080071
17	0.015		3:38 AM	09/06/2019	4:24 PM	03/28/2019	0.103	3:54 PM	08/26/2019	16080075
18	0.016		3:41 AM	09/06/2019	4:24 PM	03/28/2019	0.103	3:54 PM	08/26/2019	16080075

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
19	0.114		10:30 PM	09/07/2019	4:24 PM	03/28/2019	0.104	12:25 PM	09/06/2019	16080075
20	0.116		10:33 PM	09/07/2019	4:24 PM	03/28/2019	0.104	12:25 PM	09/06/2019	16080075
21	0.123		2:31 AM	09/08/2019	4:24 PM	03/28/2019	0.104	12:25 PM	09/06/2019	16080075
22	0.113		2:34 AM	09/08/2019	4:24 PM	03/28/2019	0.104	12:25 PM	09/06/2019	16080075
66	0.161		10:43 PM	09/14/2019	4:36 PM	03/28/2019	0.109	12:26 PM	09/06/2019	16120175
67	0.166		10:46 PM	09/14/2019	4:36 PM	03/28/2019	0.109	12:26 PM	09/06/2019	16120175
123	0		10:07 PM	09/07/2019	9:50 AM	04/25/2017	0.106	12:34 PM	09/06/2019	16120177
94	0.109		10:44 PM	09/07/2019	8:16 AM	10/21/2016	0.101	12:33 PM	09/06/2019	16120188
95	0.113		10:46 PM	09/07/2019	8:16 AM	10/21/2016	0.101	12:33 PM	09/06/2019	16120188
96	0.035		9:14 PM	09/14/2019	8:16 AM	10/21/2016	0.101	12:33 PM	09/06/2019	16120188
364	0.146		2:37 AM	09/09/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
365	0.133		2:39 AM	09/09/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
366	0.082		12:46 AM	09/12/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
367	0.082		12:48 AM	09/12/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
368	0.077		12:52 AM	09/12/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
369	0.133		12:10 AM	09/13/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
370	0.132		12:13 AM	09/13/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
371	0.23		1:48 AM	09/14/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
372	0.229		1:50 AM	09/14/2019	12:48 PM	06/09/2019	0.107	12:41 PM	09/06/2019	16120195
321	0.205		1:31 AM	09/09/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
322	0.2		1:33 AM	09/09/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
323	0		10:35 PM	09/14/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
324	0.231		3:05 AM	09/15/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
325	0.234		3:07 AM	09/15/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
326	0.122		9:48 PM	09/15/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
327	0.119		9:50 PM	09/15/2019	3:25 PM	06/25/2019	0.104	12:48 PM	09/06/2019	16120197
74	0.191		10:16 PM	09/06/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
75	0.187		10:18 PM	09/06/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
76	0.091		10:49 PM	09/07/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
77	0.1		10:52 PM	09/07/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
78	0.229		11:01 PM	09/16/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
79	0.227		11:04 PM	09/16/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
80	0.224		12:55 AM	09/17/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064

CHP South Los Angeles PAS Usage Log 9/6/19-9/17/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
81	0.231		12:57 AM	09/17/2019	5:05 PM	03/28/2019	0.108	12:52 PM	09/06/2019	17170064
10	0.155		9:28 AM	09/14/2019	4:42 PM	03/28/2019	0.107	12:51 PM	09/06/2019	17170066
11	0.155		9:30 AM	09/14/2019	4:42 PM	03/28/2019	0.107	12:51 PM	09/06/2019	17170066

#1

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)					
09/17/2019	1316	34	19DP41092						M. Tippett	20137	
Solution prepared by Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21											

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:09 PM	06/25/2019	.104	.100	1:16 PM	09/17/2019	BAC	16040162	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

2

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
09/17/2019	1322	34	19DP41092	.100	22	.104			X			M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	03/28/2019	.104	.100	1:23 PM	09/17/2019	BAC	16040119	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1314 34	19DP41092	.100	24					X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:46 AM	04/12/2019	.102	.100	1:15 PM	09/17/2019	BAC	16040140	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	152034	19DP41092		.100	23	.108		X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:41 PM	03/28/2019	.108	.100	1:20 PM	09/17/2019	BAC	16040141	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

HS

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
09/17/2019	152834	19DP41092	.100	25		.107			X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:56 AM	04/25/2017	.107	.100	1:28 PM	09/17/2019	BAC	16040161	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1:23 PM	34	19DP41092	.100	24	.100	X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	07:30 AM	04/12/2019	.100	.100	1:23 PM	09/17/2019	BAC	16040163	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1:39 PM	34	19DP41092	.100	24		104				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	12:59 PM	01/29/2018	.104	.100	1:39 PM	09/17/2019	BAC	16080013	8.7.24	
Calibration / Check Printout										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

9

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1335	34	19DP41092	.100	24	.105				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.105	.100	1:35 PM	09/17/2019	BAC	16080015	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
09/17/2019	1574	34	19DP41092	.100	24	.104						M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.104	.100	1:35 PM	09/17/2019	BAC	16080027	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	153434	19DP41092	.100	24	.103			X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:05 PM	03/28/2019	.103	.100	1:40 PM	09/17/2019	BAC	16080028	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1354	34	19DP41092	.100		23		.104			M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:07 PM	03/28/2019	.104	.100	1:39 PM	09/17/2019	BAC	16080030	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1:34	34	19DP41092	.100	23						M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:12 PM	03/28/2019	.103	.100	1:45 PM	09/17/2019	BAC	16080064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#114

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1346	54	19DP41092	.100	23		.108					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	1:54 PM	12/19/2017	.108	.100	1:47 PM	09/17/2019	BAC	16080065	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifloc FC20 PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
09/17/2019	1:52 PM	19DP41092	.100	23	TEST #1	TEST #2	YES	NO		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:18 PM	03/28/2019	.102	.100	1:52 PM	09/17/2019	BAC	16080069	8.7.24	

Lifloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1:52 PM	34	19DP41092	.100	24		.104					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
10:41 AM	08/14/2017	.104	.100	1:52 PM	09/17/2019	BAC	16080071	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	140934	19DP41092	.100	23		.108				X	M. Tippett	20137	
Solution prepared by: Cesar Munoz													
Date: 04-22-19													
Expiration date: 04-22-21													

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	2:05 PM	12/19/2017	.108	.100	2:09 PM	09/17/2019	BAC	16080072	8.7.24	
Calibration / Check Printout										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1400	34	19DP41092	.100	23	.110			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:36 PM	03/28/2019	.110	.100	2:01 PM	09/17/2019	BAC	16120175	8.7.24	

Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lieloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
09/17/2019	1400 34		19DP41092							M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	09:50 AM	04/25/2017	.107	.100	2:09 PM	09/17/2019	BAC	16120177	8.7.24	Calibration / Check Printout

Lieloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1407 34	19DP41092	100	24	103			X		M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.103	.100	2:08 PM	09/17/2019	BAC	16120188	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

23

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 ID. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1411	34	19DP41092	100	24		106					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:50 AM	04/12/2019	.106	.100	2:12 PM	09/17/2019	BAC	16120189	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1408 34	19DP41092	.100	23		.102			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	12:20 PM	08/15/2019	.102	.100	2:08 PM	09/17/2019	BAC	16120194	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
09/17/2019	1414	34	19DP41092	.100	24		.108		X				M. Tippett	20137	

Solution prepared by:
 Cesar Minoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:48 PM	06/09/2019	.108	.100	2:15 PM	09/17/2019	BAC	16120195	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				VALUE	TEST #	TEST #2 (Optional)	YES				
09/17/2019	1422	34	19DP41092		23		.102			X		M. Tippett	20137	

Solution prepared by:
Cesar M/mnoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	3:25 PM	06/25/2019	.102	.100	2:22 PM	09/17/2019	BAC	16120197	8.7.24

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/17/2019	1425 34	19DP41092	.100	23		.108			X		M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:42 PM	03/28/2019	.108	.100	2:25 PM	09/17/2019	BAC	17170066	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/17/2019	1427 34	19DP41092		.100		23		.106						
Solution prepared by: Cesar Munoz														
Date: 04-22-19														
Expiration date: 04-22-21														

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:46 PM	03/28/2019	.106	.100	2:27 PM	09/17/2019	BAC	17170155	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#31

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippet
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2 (Optional)	YES	NO				
09/17/2019	1425	39	19DDP41092	.100	22		108			X		M. Tippet	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	5:05 PM	03/28/2019	.108	.100	2:26 PM	09/17/2019	BAC	17170064	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.