

CHP South LA PAS Usage Logs 8/26/19-9/6/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
58	0.104		3:34 AM	08/31/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
59	0.1		3:36 AM	08/31/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
60	0.108		2:34 AM	09/04/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
61	0.107		2:36 AM	09/04/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
62	0.109		2:55 AM	09/06/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
63	0.107		2:58 AM	09/06/2019	3:25 PM	03/28/2019	0.102	3:21 PM	08/26/2019	16040119
76	0		1:55 AM	08/30/2019	7:46 AM	04/12/2019	0.1	3:14 PM	08/26/2019	16040140
77	0.027		1:42 AM	09/06/2019	7:46 AM	04/12/2019	0.1	3:14 PM	08/26/2019	16040140
46	0.144		10:13 PM	09/05/2019	3:41 PM	03/28/2019	0.104	3:19 PM	08/26/2019	16040141
47	0.145		10:15 PM	09/05/2019	3:41 PM	03/28/2019	0.104	3:19 PM	08/26/2019	16040141
257	0.07		4:11 AM	08/28/2019	8:56 AM	04/25/2017	0.106	3:27 PM	08/26/2019	16040161
258	0.162		6:43 PM	09/01/2019	8:56 AM	04/25/2017	0.106	3:27 PM	08/26/2019	16040161
338	0.137		3:01 AM	08/30/2019	4:09 PM	06/25/2019	0.102	3:15 PM	08/26/2019	16040162
339	0.134		3:03 AM	08/30/2019	4:09 PM	06/25/2019	0.102	3:15 PM	08/26/2019	16040162
404	0.166		11:01 PM	08/30/2019	7:30 AM	04/12/2019	0.1	3:22 PM	08/26/2019	16040163
405	0.173		11:03 PM	08/30/2019	7:30 AM	04/12/2019	0.1	3:22 PM	08/26/2019	16040163
406	0.012		2:35 AM	08/31/2019	7:30 AM	04/12/2019	0.1	3:22 PM	08/26/2019	16040163
407	0.033		3:06 AM	08/31/2019	7:30 AM	04/12/2019	0.1	3:22 PM	08/26/2019	16040163
408	0.206		10:03 PM	08/31/2019	7:30 AM	04/12/2019	0.1	3:22 PM	08/26/2019	16040163
409	0.159		10:05 PM	08/31/2019	7:30 AM	04/12/2019	0.1	3:22 PM	08/26/2019	16040163
87	0		10:34 PM	08/28/2019	3:58 PM	03/28/2019	0.104	3:32 PM	08/26/2019	16080027
88	0.161		12:02 AM	08/29/2019	3:58 PM	03/28/2019	0.104	3:32 PM	08/26/2019	16080027
89	0.165		12:05 AM	08/29/2019	3:58 PM	03/28/2019	0.104	3:32 PM	08/26/2019	16080027
58	0		5:56 PM	08/27/2019	4:12 PM	03/28/2019	0.103	3:38 PM	08/26/2019	16080064
59	0		8:13 PM	08/27/2019	4:12 PM	03/28/2019	0.103	3:38 PM	08/26/2019	16080064
273	0.139		1:03 AM	08/31/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
274	0.136		1:04 AM	08/31/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
275	0.125		3:36 AM	08/31/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
276	0.129		3:38 AM	08/31/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
277	0.041		11:47 PM	08/31/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
278	0.172		12:54 AM	09/02/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
279	0.159		12:56 AM	09/02/2019	1:54 PM	12/19/2017	0.105	3:40 PM	08/26/2019	16080065
131	0.272		10:42 PM	08/29/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
132	0.287		10:45 PM	08/29/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
133	0.198		10:15 PM	08/30/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
134	0.191		10:18 PM	08/30/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
135	0.106		2:07 AM	08/31/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
136	0.125		2:10 AM	08/31/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
137	0.13		2:12 AM	08/31/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
138	0.164		1:28 AM	09/05/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
139	0.156		1:31 AM	09/05/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
140	0.159		3:09 AM	09/06/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
141	0.161		3:13 AM	09/06/2019	4:18 PM	03/28/2019	0.101	3:45 PM	08/26/2019	16080069
410	0.198		2:19 AM	09/02/2019	10:41 AM	08/14/2017	0.102	3:45 PM	08/26/2019	16080071
411	0.198		2:23 AM	09/02/2019	10:41 AM	08/14/2017	0.102	3:45 PM	08/26/2019	16080071
412	0.006		3:54 AM	09/03/2019	10:41 AM	08/14/2017	0.102	3:45 PM	08/26/2019	16080071
413	0		11:19 PM	09/03/2019	10:41 AM	08/14/2017	0.102	3:45 PM	08/26/2019	16080071
414	0.137		2:49 AM	09/04/2019	10:41 AM	08/14/2017	0.102	3:45 PM	08/26/2019	16080071
415	0.139		2:52 AM	09/04/2019	10:41 AM	08/14/2017	0.102	3:45 PM	08/26/2019	16080071
17	0.015		3:38 AM	09/06/2019	4:24 PM	03/28/2019	0.103	3:54 PM	08/26/2019	16080075
18	0.016		3:41 AM	09/06/2019	4:24 PM	03/28/2019	0.103	3:54 PM	08/26/2019	16080075
114	0		2:00 AM	08/29/2019	4:30 PM	03/28/2019	0.104	3:55 PM	08/26/2019	16080077
60	0.041		11:08 PM	08/30/2019	4:36 PM	03/28/2019	0.106	3:55 PM	08/26/2019	16120175
61	0.04		11:11 PM	08/30/2019	4:36 PM	03/28/2019	0.106	3:55 PM	08/26/2019	16120175
62	0.147		2:24 AM	08/31/2019	4:36 PM	03/28/2019	0.106	3:55 PM	08/26/2019	16120175
63	0.153		2:27 AM	08/31/2019	4:36 PM	03/28/2019	0.106	3:55 PM	08/26/2019	16120175
64	0.095		8:35 PM	09/01/2019	4:36 PM	03/28/2019	0.106	3:55 PM	08/26/2019	16120175
65	0.092		8:38 PM	09/01/2019	4:36 PM	03/28/2019	0.106	3:55 PM	08/26/2019	16120175
121	0.117		11:09 PM	08/30/2019	9:50 AM	04/25/2017	0.105	4:03 PM	08/26/2019	16120177
122	0.116		11:11 PM	08/30/2019	9:50 AM	04/25/2017	0.105	4:03 PM	08/26/2019	16120177
93	0		10:51 PM	08/30/2019	8:16 AM	10/21/2016	0.101	4:06 PM	08/26/2019	16120188
356	0.211		2:14 AM	08/31/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
357	0.201		2:16 AM	08/31/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
358	0.113		1:19 AM	09/01/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
359	0.105		1:21 AM	09/01/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
360	0.172		2:17 AM	09/02/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195

CHP South LA PAS Usage Logs 8/26/19-9/6/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
361	0.151		2:20 AM	09/02/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
362	0.123		4:21 AM	09/04/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
363	0.119		4:22 AM	09/04/2019	12:48 PM	06/09/2019	0.106	4:10 PM	08/26/2019	16120195
318	0.095		11:48 PM	08/31/2019	3:25 PM	06/25/2019	0.104	4:12 PM	08/26/2019	16120197
319	0.091		11:50 PM	08/31/2019	3:25 PM	06/25/2019	0.104	4:12 PM	08/26/2019	16120197
320	0.097		11:53 PM	08/31/2019	3:25 PM	06/25/2019	0.104	4:12 PM	08/26/2019	16120197

#1

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
I.D. NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					YES
09/06/2019	1447	34	19DP41092	.100	23		.103					X	M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:09 PM	06/25/2019	.103	.100	10:47 AM	09/06/2019	BAC	16040162	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

42

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: *16040119*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/06/2019	<i>1145 34</i>		19DP41092	.100		<i>23</i>		<i>.102</i>				<i>X</i>			
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21															
					Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
					.100	3:25 PM	03/28/2019	.102	.100	11:43 AM	09/06/2019	BAC	16040119	8.7.24	
Calibration / Check Printout															
Lifeloc Technologies															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)						
09/06/2019	11:35	34	19DP41092	.100	23				YES	NO	X			

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:46 AM	04/12/2019	.101	.100	11:36 AM	09/06/2019	BAC	16040140	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#4

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1141	54	19DDP41092	.100		22		.106			X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:41 PM	03/28/2019	.106	.100	11:41 AM	09/06/2019	BAC	16040141	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1153	34	19DP41092	.100	23		.107				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:56 AM	04/25/2017	.107	.100	11:56 AM	09/06/2019	BAC	16040161	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1150	34	19DP41092	100	23		100					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.100	.100	11:51 AM	09/06/2019	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2					
09/06/2019	1206	34	19DP41092	.100	22		.104		NO		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard		Time:		Last Cal		Result		Cal Standard		Time:		Last Check		UOM		Serial No.		FC20	
.100		12:59 PM		01/29/2018		.104		.100		12:06 PM		09/06/2019		BAC		16080013		8.7.24	

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	12:01	34	19DP41092	100	23		.104				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	11:51 AM	07/05/2019	.104	.100	12:02 PM	09/06/2019	BAC	16080014	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Michael Tippett	I.D. NUMBER 20137	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. <i>16080015</i>
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-----------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1 (Optional)		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/06/2019	120234	19DP41092			100	23							M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:58 PM	03/28/2019	.105	.100	12:02 PM	09/06/2019	BAC	16080015	8.7.24	

Lifoloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	12:01	34	19DP41092		100	23		104				X	M. Tippett	20137	

Solution prepared by:														
Cesar Munoz														
Date: 04-22-19														
Expiration date: 04-22-21														
Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout			
.100		3:58 PM	03/28/2019	.104	.100	12:01 PM	09/06/2019	BAC	16080027	8.7.24	Lifeloc Technologies			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
09/06/2019	1204	34	19DP41092	.100	22		.103			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:07 PM	03/28/2019	.103	.100	12:04 PM	09/06/2019	BAC	16080030	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/06/2019	1208	34	19DDP41092	100	25		102					X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:05 PM	03/28/2019	.102	.100	12:09 PM	09/06/2019	BAC	16080028	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1207	34	19DP41092	.100	22		.103		X		M. Tippett	20137		

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:12 PM	03/28/2019	.103	.100	12:08 PM	09/06/2019	BAC	16080064	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Michael Tippett	I.D. NUMBER 20137	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-----------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1:29	34	19DP41092	100	22			106					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	1:54 PM	12/19/2017	.106	.100	12:10 PM	09/06/2019	BAC	16080065	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16082069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	12:14	34		19DP41092	.100	23		.099				X	M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	4:18 PM	03/28/2019	.099	.100	12:15 PM	09/06/2019	BAC	16080069	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	12:15	34	19DDP41092	.100	23		.102				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	10:41 AM	08/14/2017	.102	.100	12:15 PM	09/06/2019	BAC	16080071	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Michael Tippett | I.D. NUMBER: 20137 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/06/2019	1232	34	19DDP41092	100	22		107					X	M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	2:05 PM	12/19/2017	.107	.100	12:32 PM	09/06/2019	BAC	16080072	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
09/06/2019	1224	34	19DP41092	.100	23	.104					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:24 PM	03/28/2019	.104	.100	12:25 PM	09/06/2019	BAC	16080075	8.7.24	

Lifoloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
09/06/2019	122534		19DP41092	.100		-23						X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard		Time:		Last Cal		Result		Cal Standard		Time:		Last Check		UOM		Serial No.		FC20		Calibration / Check Printout	Lifeloc Technologies
	.100		4:30 PM		03/28/2019		.106		.100		12:26 PM		09/06/2019		BAC		16080077		8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/06/2019	1225	34	19DP41092	.100										M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.109	.100	12:26 PM	09/06/2019	BAC	16120175	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Michael Tippett	I.D. NUMBER 20137	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120175
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES	NO					
09/06/2019	12:34	34	19DP41092	100	22			106					X		M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard			Last Cal			Cal Standard			Last Check			UOM		Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
Time:	Result	Time:	Result	Time:	Result	Time:	Result	UOM	Serial No.	FC20	Calibration / Check Printout		Lifeloc Technologies				
4:36 PM	.100	03/28/2019	.109	12:26 PM	.100	09/06/2019	BAC	16120175	8.7.24		Calibration / Check Printout		Lifeloc Technologies				

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet path simulator.

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
09/06/2019	1233	54	19DP41092	.100	22		.101					X	M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:16 AM	10/21/2016	.101	.100	12:33 PM	09/06/2019	BAC	16120188	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1237 34	19DDP41092	.100	22	.103					X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:50 AM	04/12/2019	.103	.100	12:38 PM	09/06/2019	BAC	16120189	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1234/34		19DP41092	.100	22							M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:20 PM	08/15/2019	.101	.100	12:35 PM	09/06/2019	BAC	16120194	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO.: 161 20195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	124034		19DDP41092	.100	22		.107					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.107	.100	12:41 PM	09/06/2019	BAC	16120195	8.7.24	

Lifoloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Michael Tippett
ID NUMBER: 20137
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	124834	19DP41092	.100	22	104	X					M. Tippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date: 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:25 PM	06/25/2019	.104	.100	12:48 PM	09/06/2019	BAC	16120197	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
										YES	NO				
09/06/2019	1251	34	19DP41092	.100	21		107			X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:42 PM	03/28/2019	.107	.100	12:51 PM	09/06/2019	BAC	17170066	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
								TEST #1	TEST #2 (Optional)	YES	NO				
09/06/2019	125334		19DP41092	.100	71			.105			X		M. Tippett	20137	

Solution prepared by:
 Cesar Minoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	4:46 PM	03/28/2019	.105	.100	12:53 PM	09/06/2019	BAC	17170155	8.7.24	Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Michael Tippett
 I.D. NUMBER: 20137
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019	1252	34	19DP41092	.100	21	.108					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	5:05 PM	03/28/2019	.108	.100	12:52 PM	09/06/2019	BAC	17170064	8.7.24

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/06/2019			19DP41092	.100							M. Tippett	20137	NO SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Michael Tippett I.D. NUMBER: 20137 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
09/06/2019			19DP41092	.100							M. Tippett	20137	00 SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions CHP202K_1108.pdf