

CHP South LA PAS Usage logs 11/21/19-12/23/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
84	0.223		7:04 PM	11/24/2019	3:25 PM	03/28/2019	0.101	3:11 PM	11/21/2019	16040119
86	0.127		3:46 AM	11/28/2019	3:25 PM	03/28/2019	0.101	3:11 PM	11/21/2019	16040119
89	0.17		2:27 AM	12/14/2019	3:25 PM	03/28/2019	0.1	2:10 PM	12/12/2019	16040119
90	0.151		2:29 AM	12/14/2019	3:25 PM	03/28/2019	0.1	2:10 PM	12/12/2019	16040119
91	0.159		2:24 AM	12/22/2019	3:25 PM	03/28/2019	0.1	2:10 PM	12/12/2019	16040119
92	0.153		2:27 AM	12/22/2019	3:25 PM	03/28/2019	0.1	2:10 PM	12/12/2019	16040119
118	0		4:35 AM	12/20/2019	7:46 AM	04/12/2019	0.097	1:59 PM	12/12/2019	16040140
119	0.063		4:50 AM	12/20/2019	7:46 AM	04/12/2019	0.097	1:59 PM	12/12/2019	16040140
120	0.067		4:53 AM	12/20/2019	7:46 AM	04/12/2019	0.097	1:59 PM	12/12/2019	16040140
74	0.05		10:36 AM	12/01/2019	3:41 PM	03/28/2019	0.101	3:30 PM	11/21/2019	16040141
75	0.05		10:39 AM	12/01/2019	3:41 PM	03/28/2019	0.101	3:30 PM	11/21/2019	16040141
84	0.101		1:42 AM	12/18/2019	3:41 PM	03/28/2019	0.101	3:10 PM	12/12/2019	16040141
294	0.05		2:39 AM	12/01/2019	8:56 AM	04/25/2017	0.102	3:55 PM	11/21/2019	16040161
301	0.155		11:23 PM	12/16/2019	8:56 AM	04/25/2017	0.102	3:22 PM	12/12/2019	16040161
302	0.143		11:26 PM	12/16/2019	8:56 AM	04/25/2017	0.102	3:22 PM	12/12/2019	16040161
369	0		7:02 PM	11/21/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
370	0		7:03 PM	11/21/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
371	0.222		8:51 PM	11/28/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
372	0.186		8:53 PM	11/28/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
373	0.206		8:56 PM	11/28/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
385	0.152		3:51 AM	12/14/2019	4:09 PM	06/25/2019	0.105	1:59 PM	12/12/2019	16040162
386	0.144		3:54 AM	12/14/2019	4:09 PM	06/25/2019	0.105	1:59 PM	12/12/2019	16040162
446	0.157		1:47 AM	12/01/2019	7:30 AM	04/12/2019	0.1	3:51 PM	11/21/2019	16040163
447	0.158		1:50 AM	12/01/2019	7:30 AM	04/12/2019	0.1	3:51 PM	11/21/2019	16040163
448	0		11:03 AM	12/15/2019	7:30 AM	04/12/2019	0.097	3:15 PM	12/12/2019	16040163
106	0		3:52 AM	11/23/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
110	0.051		8:36 PM	11/23/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
111	0.094		3:34 AM	11/24/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
112	0.089		3:38 AM	11/24/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
116	0.185		2:29 AM	12/15/2019	12:59 PM	01/29/2018	0.099	3:33 PM	12/12/2019	16080013
117	0.177		2:32 AM	12/15/2019	12:59 PM	01/29/2018	0.099	3:33 PM	12/12/2019	16080013
118	0.013		3:35 AM	12/15/2019	12:59 PM	01/29/2018	0.099	3:33 PM	12/12/2019	16080013
119	0.032		2:24 AM	12/22/2019	12:59 PM	01/29/2018	0.099	3:33 PM	12/12/2019	16080013

CHP South LA PAS Usage logs 11/21/19-12/23/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
120	0.153		2:59 AM	12/22/2019	12:59 PM	01/29/2018	0.099	3:33 PM	12/12/2019	16080013
121	0.152		3:00 AM	12/22/2019	12:59 PM	01/29/2018	0.099	3:33 PM	12/12/2019	16080013
83	0.016		10:10 PM	12/14/2019	11:51 AM	07/05/2019	0.097	3:27 PM	12/12/2019	16080014
120	0.233		7:13 PM	11/27/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
121	0.214		7:17 PM	11/27/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
122	0.119		8:58 AM	11/28/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
123	0.116		9:01 AM	11/28/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
124	0.063		10:15 AM	12/01/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
125	0.059		10:17 AM	12/01/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
63	0.091		6:42 AM	11/21/2019	4:05 PM	03/28/2019	0.101	1:38 PM	11/13/2019	16080028
64	0.09		6:45 AM	11/21/2019	4:05 PM	03/28/2019	0.101	1:38 PM	11/13/2019	16080028
65	0.154		1:23 AM	12/14/2019	4:05 PM	03/28/2019	0.096	3:33 PM	12/12/2019	16080028
66	0.172		1:26 AM	12/14/2019	4:05 PM	03/28/2019	0.096	3:33 PM	12/12/2019	16080028
67	0.111		1:37 AM	12/15/2019	4:05 PM	03/28/2019	0.096	3:33 PM	12/12/2019	16080028
68	0.112		1:39 AM	12/15/2019	4:05 PM	03/28/2019	0.096	3:33 PM	12/12/2019	16080028
69	0		10:44 PM	12/22/2019	4:05 PM	03/28/2019	0.096	3:33 PM	12/12/2019	16080028
59	0.197		1:12 AM	11/24/2019	4:07 PM	03/28/2019	0.093	4:43 PM	11/21/2019	16080030
60	0.2		1:14 AM	11/24/2019	4:07 PM	03/28/2019	0.093	4:43 PM	11/21/2019	16080030
101	0.077		10:47 PM	11/22/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
102	0.042		11:22 PM	12/01/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
103	0.04		11:23 PM	12/01/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
104	0.079		2:53 AM	12/02/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
105	0.085		2:55 AM	12/02/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
106	0.123		7:46 PM	12/21/2019	4:12 PM	03/28/2019	0.101	3:33 PM	12/12/2019	16080064
107	0.123		7:49 PM	12/21/2019	4:12 PM	03/28/2019	0.101	3:33 PM	12/12/2019	16080064
287	0.088		10:01 PM	11/21/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
288	0.083		10:05 PM	11/21/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
289	0		5:34 PM	11/22/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
290	0.201		10:55 PM	11/29/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
291	0.213		10:58 PM	11/29/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
292	0.113		4:05 AM	11/30/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
293	0.11		4:08 AM	11/30/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
294	0.085		12:32 AM	12/18/2019	1:54 PM	12/19/2017	0.1	3:34 PM	12/12/2019	16080065

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
295	0.086		12:35 AM	12/18/2019	1:54 PM	12/19/2017	0.1	3:34 PM	12/12/2019	16080065
296	0.235		8:57 AM	12/22/2019	1:54 PM	12/19/2017	0.1	3:34 PM	12/12/2019	16080065
297	0.235		9:00 AM	12/22/2019	1:54 PM	12/19/2017	0.1	3:34 PM	12/12/2019	16080065
442	0.116		1:34 AM	11/22/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
443	0.111		1:37 AM	11/22/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
444	0.103		3:02 AM	11/23/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
445	0.106		3:04 AM	11/23/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
446	0.019		3:10 AM	11/24/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
447	0.018		3:13 AM	11/24/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
448	0		12:44 AM	11/30/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
449	0.139		4:44 AM	11/30/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
450	0.133		4:47 AM	11/30/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
451	0.154		12:23 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
452	0.157		12:25 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
453	0.124		2:57 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
454	0.126		3:00 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
455	0.055		12:05 AM	12/02/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
458	0.118		11:17 PM	12/12/2019	10:41 AM	08/14/2017	0.094	3:38 PM	12/12/2019	16080071
459	0.122		11:19 PM	12/12/2019	10:41 AM	08/14/2017	0.094	3:38 PM	12/12/2019	16080071
462	0.215		3:45 AM	12/14/2019	10:41 AM	08/14/2017	0.094	3:38 PM	12/12/2019	16080071
463	0.217		3:47 AM	12/14/2019	10:41 AM	08/14/2017	0.094	3:38 PM	12/12/2019	16080071
38	0.119		1:02 AM	11/24/2019	2:05 PM	12/19/2017	0.094	5:46 PM	11/21/2019	16080072
39	0.124		1:05 AM	11/24/2019	2:05 PM	12/19/2017	0.094	5:46 PM	11/21/2019	16080072
115	0.157		1:59 AM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
116	0.161		2:01 AM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
117	0.231		5:19 PM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
118	0.226		5:21 PM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
128	0.096		3:38 AM	12/15/2019	4:30 PM	03/28/2019	0.097	3:50 PM	12/12/2019	16080077
129	0.094		1:03 AM	12/16/2019	4:30 PM	03/28/2019	0.097	3:50 PM	12/12/2019	16080077
130	0.09		1:06 AM	12/16/2019	4:30 PM	03/28/2019	0.097	3:50 PM	12/12/2019	16080077
131	0.028		12:19 PM	12/17/2019	4:30 PM	03/28/2019	0.097	3:50 PM	12/12/2019	16080077
132	0.027		12:21 PM	12/17/2019	4:30 PM	03/28/2019	0.097	3:50 PM	12/12/2019	16080077
86	0.248		11:51 AM	11/23/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
87	0.252		11:53 AM	11/23/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
88	0.053		2:33 AM	11/24/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
89	0.182		7:31 PM	11/24/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
90	0.178		7:33 PM	11/24/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
96	0		1:17 AM	12/14/2019	4:36 PM	03/28/2019	0.098	3:47 PM	12/12/2019	16120175
97	0.095		3:21 AM	12/14/2019	4:36 PM	03/28/2019	0.098	3:47 PM	12/12/2019	16120175
98	0.095		3:25 AM	12/14/2019	4:36 PM	03/28/2019	0.098	3:47 PM	12/12/2019	16120175
99	0.091		3:28 AM	12/14/2019	4:36 PM	03/28/2019	0.098	3:47 PM	12/12/2019	16120175
147	0.171		11:57 PM	11/21/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
148	0.179		11:59 PM	11/21/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
149	0.044		12:06 AM	11/22/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
150	0.075		2:44 AM	11/23/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
151	0.072		2:46 AM	11/23/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
152	0.194		11:12 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
153	0.191		11:14 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
154	0.158		11:20 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
155	0.192		11:22 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
156	0.236		5:09 PM	11/25/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
157	0.246		5:12 PM	11/25/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
158	0.201		3:41 AM	11/28/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
159	0.195		3:44 AM	11/28/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
166	0.205		3:19 AM	12/17/2019	9:50 AM	04/25/2017	0.1	3:57 PM	12/12/2019	16120177
167	0.209		3:21 AM	12/17/2019	9:50 AM	04/25/2017	0.1	3:57 PM	12/12/2019	16120177
112	0		9:39 PM	11/23/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
113	0.077		3:14 AM	11/28/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
114	0.078		3:17 AM	11/28/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
115	0		7:58 PM	11/29/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
117	0.104		3:29 AM	12/14/2019	8:16 AM	10/21/2016	0.096	3:55 PM	12/12/2019	16120188
118	0.11		3:32 AM	12/14/2019	8:16 AM	10/21/2016	0.096	3:55 PM	12/12/2019	16120188
390	0.053		2:13 AM	11/23/2019	12:48 PM	06/09/2019	0.103	5:56 PM	11/21/2019	16120195
395	0.103		3:50 AM	12/15/2019	12:48 PM	06/09/2019	0.101	4:01 PM	12/12/2019	16120195
396	0.098		3:52 AM	12/15/2019	12:48 PM	06/09/2019	0.101	4:01 PM	12/12/2019	16120195
397	0		10:18 AM	12/19/2019	12:48 PM	06/09/2019	0.101	4:01 PM	12/12/2019	16120195

CHP South LA PAS Usage logs 11/21/19-12/23/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
398	0.185		3:25 AM	12/22/2019	12:48 PM	06/09/2019	0.101	4:01 PM	12/12/2019	16120195
399	0.182		3:27 AM	12/22/2019	12:48 PM	06/09/2019	0.101	4:01 PM	12/12/2019	16120195
348	0.099		6:34 AM	11/29/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
349	0.101		6:36 AM	11/29/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
350	0.133		5:31 AM	11/30/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
351	0.143		5:33 AM	11/30/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
355	0.182		9:52 PM	12/17/2019	3:25 PM	06/25/2019	0.095	4:09 PM	12/12/2019	16120197
356	0.179		9:54 PM	12/17/2019	3:25 PM	06/25/2019	0.095	4:09 PM	12/12/2019	16120197
357	0.213		1:04 AM	12/23/2019	3:25 PM	06/25/2019	0.095	4:09 PM	12/12/2019	16120197
358	0.197		1:07 AM	12/23/2019	3:25 PM	06/25/2019	0.095	4:09 PM	12/12/2019	16120197
139	0		1:20 AM	12/13/2019	5:05 PM	03/28/2019	0.104	4:14 PM	12/12/2019	17170064
140	0.098		1:26 AM	12/15/2019	5:05 PM	03/28/2019	0.104	4:14 PM	12/12/2019	17170064
141	0.095		1:29 AM	12/15/2019	5:05 PM	03/28/2019	0.104	4:14 PM	12/12/2019	17170064
142	0		12:41 AM	12/20/2019	5:05 PM	03/28/2019	0.104	4:14 PM	12/12/2019	17170064
143	0.137		2:57 AM	12/23/2019	5:05 PM	03/28/2019	0.104	4:14 PM	12/12/2019	17170064
144	0.136		2:59 AM	12/23/2019	5:05 PM	03/28/2019	0.104	4:14 PM	12/12/2019	17170064
16	0.157		8:42 PM	12/21/2019	4:42 PM	03/28/2019	0.096	4:13 PM	12/12/2019	17170066
17	0.168		8:45 PM	12/21/2019	4:42 PM	03/28/2019	0.096	4:13 PM	12/12/2019	17170066

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward MeElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	11:39 34	19DP41092	.100	21	.102		X		M. Tippett	20137		
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21												
Cal Standard												
Time: 4:09 PM												
Last Cal												
06/25/2019												
Result: .102												
Cal Standard												
Time: 11:40 AM												
Last Check												
12/23/2019												
UOM: BAC												
Serial No.: 16040162												
FC20: 8.7.24												
Lifeloc Technologies												
Calibration / Check Printout												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS WAKEMODEL: Lifeloc FC20 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	153834	19DP41092	.100	22		.095					M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	03/28/2019	.095	.100	1:39 PM	12/23/2019	BAC	16040119	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward McElroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	11:36	34	19DPR41092	.100	21	.097				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	07:46 AM	04/12/2019	.097	.100	11:37 AM	12/23/2019	BAC	16040140	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. <i>16040141</i>
--------------------------------	-----------------	------------------------------------	----------------------	--------------------------------	-----------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/23/2019	1142 34	19DP41092	.100	21	.100					X	M. Tippet	20137	
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21													
Cal Standard													
Time: 3:41 PM													
Last Cal													
03/28/2019													
Result													
.100													
Cal Standard													
.100													
Time: 11:42 AM													
Last Check													
12/23/2019													
UOM													
BAC													
Serial No. 16040141													
FC20													
8.7.24													
Calibration / Check Printout													
Lifeloc Technologies													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/23/2019	1154	34	19DB41092	.100	20	.100			X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.100	.100	11:54 AM	12/23/2019	BAC	16040161	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/23/2019	1146	34	19DPA1092	.100	21		.096				X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	07:30 AM	04/12/2019	.096	.100	11:46 AM	12/23/2019	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

7

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	1205 34		19DB41092		.100	21	.099		X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies Calibration / Check Printout
.100	12:59 PM	01/29/2018	.099	.100	12:04 PM	12/23/2019	BAC	16080013	8.7.24	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	15:34		19DP41092	.100								M. Trippett	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:51 AM	07/05/2019	.100	.100	11:58 AM	12/23/2019	BAC	16080014	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

9

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/23/2019	11:56 AM	19DP41092	.100	21	.099			X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.099	.095	11:56 AM	12/23/2019	BAC	16080015	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	1:30:54	19DP41092	.100	23		.097			NO	X	M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	3:58 PM	03/28/2019	.097	.100	1:30 PM	12/23/2019	BAC	16080027	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

11

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Edward McElroy
 I.C. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO: 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REPEAT RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	1200	34	19DP41092	.100	20	.098			X	M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:05 PM	03/28/2019	.098	.100	12:01 PM	12/23/2019	BAC	16080028	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: 16080630

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	153834		19DP41092	.100								M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:07 PM	03/28/2019	.097	.100	1:33 PM	12/23/2019	BAC	16080030	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	11:58 AM		19JDP41092	.100		20		.100			X	M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	4:12 PM	03/28/2019	.100	.100	11:58 AM	12/23/2019	BAC	16080064	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#14

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward MeElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
12/23/2019	1:24:54		19DP41092		.100	19	.096		X	M. Tippett	20137	

Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21												
Cal Standard												
Result	Time:	Last Cal	Result	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout			
.100	1:54 PM	12/19/2017	.096	1:25 PM	12/23/2019	BAC	16080065	8.7.24	Lifeloc Technologies			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS					
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO				
12/23/2019	1205 34		19DR41092		.100		20		.096		X		M. Tippet	20137			
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21																	
		Cal Standard			Time:	Last Cal		Result	Cal Standard		Time:	Last Check		UOM	Serial No.	FC20	Lifeloc Technologies Calibration / Check Printout
			.100		4:18 PM		03/28/2019	.096		.100	12:03 PM		12/23/2019	BAC	16080069	8.7.24	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	1200	34	19DP41092	.100	21	.098			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	10:41 AM	08/14/2017	.098	.090	12:01 PM	12/23/2019	BAC	16080071	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES				
12/23/2019	122034	19DP41092		.100	70	.102		X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.102	.100	12:21 PM	12/23/2019	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO						
12/23/2019	1217	74	19DP41092		.100					X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	4:24 PM	03/28/2019	.099	.100	12:13 PM	12/23/2019	BAC	16080075	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO	
12/23/2019	1:26 PM	19DP41092		.100	22		.099				X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Call Standard	Time:	Last Cal	Result	Call Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
	4:30 PM	03/28/2019	.099		1:26 PM	12/23/2019	BAC	16080077	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/23/2019	1:30	34	19DR41092	.100	72	.097					X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	4:36 PM	03/28/2019	.097	.100	1:19 PM	12/23/2019	BAC	16120175	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/23/2019	1245 34		19DR41092							M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	09:50 AM	04/25/2017	.102	.100	12:16 PM	12/23/2019	BAC	16120177	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO	
12/23/2019	1213 54	19DP41092		.100	20		.095				X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.095	.100	12:14 PM	12/23/2019	BAC	16120188	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16/20/89

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	12:19	34	19DP41092	.100	21	.099	X			M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	04/12/2019	.099	.100	12:19 PM	12/23/2019	BAC	16120189	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO						
12/23/2019	12/13 34		19DP41092		.100	20	.094			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:20 PM	08/15/2019	.094	.100	12:14 PM	12/23/2019	BAC	16120194	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

25

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Medley
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifloc FC20
 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
12/23/2019	122054	19DP41092	.100	70	103			X		M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:48 PM	06/09/2019	.103	.100	12:20 PM	12/23/2019	BAC	16120195	8.7.24	

Lifloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward McElroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	1228 54	19DP41092	.100	20		.096					M. Tippett	20137	
Solution prepared by: Cesar Munoz													
Date: 04-22-19													
Expiration date: 04-22-21													

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	3:25 PM	06/25/2019	.096	.100	12:29 PM	12/23/2019	BAC	16120197	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
12/23/2019	1253H		19DP41092	.100		19						X	M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	4:42 PM	03/28/2019	.099	.100	12:33 PM	12/23/2019	BAC	17170066	8.7.24	

Calibration / Check Printout
Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward McElroy	16193	Lifeloc FC20	

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/23/2019			19DP41092	.100							M. Tippett	20137	00 SERVICE	
		Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019	12:34:34		19DP41092	.100	19	.097			X		M. Tippett	20137	

Solution prepared by:
 Cesar Munoz
 Date: 04-22-19
 Expiration date:
 04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	4:46 PM	03/28/2019	.097	.100	12:34 PM	12/23/2019	BAC	17170155	8.7.24	
Calibration / Check Printout										

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

30

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/23/2019		19DP41092	.100								M. Tippett	20137	00 SERVICE
		Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration date: 04-22-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

31

AREA NAME: South Los Angeles
AREA NO: 530
COORDINATOR NAME: Edward McElroy
I.D. NUMBER: 16193
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS	OPERATOR		I.D. NO.	
12/23/2019	1231	34	19DP41092	100	18	099				X		M. Tippet	20137	

Solution prepared by:
Cesar Munoz
Date: 04-22-19
Expiration date:
04-22-21

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	5:05 PM	03/28/2019	.099	.100	12:31 PM	12/23/2019	BAC	17170064	8.7.24

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.