

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
99	0.09		2:34 PM	01/11/2020	0.102	5:49 PM	01/03/2020	16040119
100	0.091		2:37 PM	01/11/2020	0.102	5:49 PM	01/03/2020	16040119
101	0.094		2:41 PM	01/11/2020	0.102	5:49 PM	01/03/2020	16040119
102	0.275		6:06 PM	01/11/2020	0.102	5:49 PM	01/03/2020	16040119
103	0.279		6:09 PM	01/11/2020	0.102	5:49 PM	01/03/2020	16040119
128	0.123		4:11 AM	01/03/2020	0.097	11:37 AM	12/23/2019	16040140
129	0.129		4:15 AM	01/03/2020	0.097	11:37 AM	12/23/2019	16040140
90	0.162		1:40 AM	01/05/2020	0.104	5:42 PM	01/03/2020	16040141
91	0.159		1:43 AM	01/05/2020	0.104	5:42 PM	01/03/2020	16040141
92	0.046		2:07 AM	01/05/2020	0.104	5:42 PM	01/03/2020	16040141
305	0.007		10:53 PM	01/05/2020	0.103	5:55 PM	01/03/2020	16040161
306	0.008		10:54 PM	01/05/2020	0.103	5:55 PM	01/03/2020	16040161
307	0.008		10:54 PM	01/05/2020	0.103	5:55 PM	01/03/2020	16040161
391	0.042		12:49 AM	01/03/2020	0.102	11:40 AM	12/23/2019	16040162
392	0		1:45 AM	01/03/2020	0.102	11:40 AM	12/23/2019	16040162
393	0		1:48 AM	01/03/2020	0.102	11:40 AM	12/23/2019	16040162
394	0.037		4:04 PM	01/06/2020	0.104	5:33 PM	01/03/2020	16040162
395	0.04		4:07 PM	01/06/2020	0.104	5:33 PM	01/03/2020	16040162
396	0		10:39 PM	01/06/2020	0.104	5:33 PM	01/03/2020	16040162
397	0.178		11:13 PM	01/08/2020	0.104	5:33 PM	01/03/2020	16040162
398	0.17		11:16 PM	01/08/2020	0.104	5:33 PM	01/03/2020	16040162
449	0.046		2:19 AM	01/04/2020	0.09	5:46 PM	01/03/2020	16040163
450	0.162		1:20 AM	01/05/2020	0.09	5:46 PM	01/03/2020	16040163
451	0.16		1:23 AM	01/05/2020	0.09	5:46 PM	01/03/2020	16040163
452	0.079		11:57 PM	01/08/2020	0.09	5:46 PM	01/03/2020	16040163
453	0.082		12:00 AM	01/09/2020	0.09	5:46 PM	01/03/2020	16040163
454	0		2:01 AM	01/12/2020	0.09	5:46 PM	01/03/2020	16040163
123	0		3:50 PM	01/06/2020	0.096	6:04 PM	01/03/2020	16080013
124	0		3:53 PM	01/06/2020	0.096	6:04 PM	01/03/2020	16080013
125	0.121		9:57 PM	01/08/2020	0.096	6:04 PM	01/03/2020	16080013
126	0.109		10:01 PM	01/08/2020	0.096	6:04 PM	01/03/2020	16080013
127	0.106		2:18 AM	01/13/2020	0.096	6:04 PM	01/03/2020	16080013
128	0.102		2:21 AM	01/13/2020	0.096	6:04 PM	01/03/2020	16080013

84	0	11:39 PM	01/10/2020	0.096	5:58 PM	01/03/2020	16080014
130	0	12:57 AM	01/06/2020	0.093	5:57 PM	01/03/2020	16080027
131	0	4:17 AM	01/12/2020	0.093	5:57 PM	01/03/2020	16080027
132	0	4:20 AM	01/12/2020	0.093	5:57 PM	01/03/2020	16080027
73	0.055	1:33 AM	01/03/2020	0.098	12:01 PM	12/23/2019	16080028
74	0.063	2:12 AM	01/03/2020	0.098	12:01 PM	12/23/2019	16080028
75	0.066	11:48 PM	01/03/2020	0.097	6:04 PM	01/03/2020	16080028
76	0.066	11:51 PM	01/03/2020	0.097	6:04 PM	01/03/2020	16080028
77	0.078	3:06 AM	01/04/2020	0.097	6:04 PM	01/03/2020	16080028
78	0.08	3:08 AM	01/04/2020	0.097	6:04 PM	01/03/2020	16080028
79	0	11:58 PM	01/04/2020	0.097	6:04 PM	01/03/2020	16080028
80	0.143	12:55 AM	01/05/2020	0.097	6:04 PM	01/03/2020	16080028
81	0.13	12:58 AM	01/05/2020	0.097	6:04 PM	01/03/2020	16080028
82	0.146	2:33 AM	01/12/2020	0.097	6:04 PM	01/03/2020	16080028
83	0.135	2:37 AM	01/12/2020	0.097	6:04 PM	01/03/2020	16080028
61	0.109	10:50 PM	01/04/2020	0.093	6:04 PM	01/03/2020	16080030
62	0.101	10:53 PM	01/04/2020	0.093	6:04 PM	01/03/2020	16080030
63	0	1:38 AM	01/12/2020	0.093	6:04 PM	01/03/2020	16080030
298	0.064	2:47 AM	01/08/2020	0.096	6:07 PM	01/03/2020	16080065
299	0.062	2:50 AM	01/08/2020	0.096	6:07 PM	01/03/2020	16080065
300	0.11	4:04 AM	01/08/2020	0.096	6:07 PM	01/03/2020	16080065
301	0.112	4:07 AM	01/08/2020	0.096	6:07 PM	01/03/2020	16080065
302	0.041	2:31 AM	01/11/2020	0.096	6:07 PM	01/03/2020	16080065
303	0.043	2:33 AM	01/11/2020	0.096	6:07 PM	01/03/2020	16080065
171	0.053	1:26 AM	01/05/2020	0.092	6:12 PM	01/03/2020	16080069
172	0.052	1:30 AM	01/05/2020	0.092	6:12 PM	01/03/2020	16080069
173	0.06	2:47 AM	01/05/2020	0.092	6:12 PM	01/03/2020	16080069
174	0.066	2:50 AM	01/05/2020	0.092	6:12 PM	01/03/2020	16080069
175	0.153	3:58 AM	01/08/2020	0.092	6:12 PM	01/03/2020	16080069
176	0.163	4:01 AM	01/08/2020	0.092	6:12 PM	01/03/2020	16080069
464	0.078	12:29 AM	01/05/2020	0.097	6:11 PM	01/03/2020	16080071
465	0.074	12:31 AM	01/05/2020	0.097	6:11 PM	01/03/2020	16080071
466	0.119	11:08 PM	01/05/2020	0.097	6:11 PM	01/03/2020	16080071
467	0.112	11:10 PM	01/05/2020	0.097	6:11 PM	01/03/2020	16080071

468	0.051	12:50 AM	01/07/2020	0.097	6:11 PM	01/03/2020	16080071
469	0.109	12:04 AM	01/11/2020	0.097	6:11 PM	01/03/2020	16080071
470	0.107	12:06 AM	01/11/2020	0.097	6:11 PM	01/03/2020	16080071
471	0.113	10:27 PM	01/11/2020	0.097	6:11 PM	01/03/2020	16080071
472	0.107	10:29 PM	01/11/2020	0.097	6:11 PM	01/03/2020	16080071
473	0.179	10:35 PM	01/11/2020	0.097	6:11 PM	01/03/2020	16080071
40	0	11:45 PM	01/11/2020	0.099	6:30 PM	01/03/2020	16080072
41	0.2	12:58 AM	01/12/2020	0.099	6:30 PM	01/03/2020	16080072
42	0.211	1:01 AM	01/12/2020	0.099	6:30 PM	01/03/2020	16080072
43	0.136	3:57 AM	01/12/2020	0.099	6:30 PM	01/03/2020	16080072
44	0.129	3:59 AM	01/12/2020	0.099	6:30 PM	01/03/2020	16080072
32	0.197	1:02 AM	01/04/2020	0.098	6:23 PM	01/03/2020	16080075
33	0.177	1:04 AM	01/04/2020	0.098	6:23 PM	01/03/2020	16080075
108	0.148	10:02 PM	01/06/2020	0.099	6:23 PM	01/03/2020	16120175
109	0.152	10:05 PM	01/06/2020	0.099	6:23 PM	01/03/2020	16120175
110	0	9:08 PM	01/07/2020	0.099	6:23 PM	01/03/2020	16120175
174	0.163	3:22 AM	01/05/2020	0.099	6:33 PM	01/03/2020	16120177
175	0.155	3:25 AM	01/05/2020	0.099	6:33 PM	01/03/2020	16120177
176	0	3:31 AM	01/05/2020	0.099	6:33 PM	01/03/2020	16120177
207	0.169	2:48 AM	01/13/2020	0.095	6:36 PM	01/03/2020	16120189
336	0.163	1:33 AM	01/05/2020	0.095	6:31 PM	01/03/2020	16120194
337	0.163	1:36 AM	01/05/2020	0.095	6:31 PM	01/03/2020	16120194
338	0.144	1:29 AM	01/11/2020	0.095	6:31 PM	01/03/2020	16120194
339	0.124	1:32 AM	01/11/2020	0.095	6:31 PM	01/03/2020	16120194
340	0.125	3:03 AM	01/12/2020	0.095	6:31 PM	01/03/2020	16120194
341	0.132	3:07 AM	01/12/2020	0.095	6:31 PM	01/03/2020	16120194
145	0.039	12:51 AM	01/12/2020	0.101	6:56 PM	01/03/2020	17170064
146	0.137	1:51 AM	01/12/2020	0.101	6:56 PM	01/03/2020	17170064
147	0.129	1:53 AM	01/12/2020	0.101	6:56 PM	01/03/2020	17170064
148	0.157	3:57 AM	01/12/2020	0.101	6:56 PM	01/03/2020	17170064
149	0.156	4:00 AM	01/12/2020	0.101	6:56 PM	01/03/2020	17170064
150	0	10:30 AM	01/13/2020	0.101	6:56 PM	01/03/2020	17170064

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 1640162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/13/2020	1501	34.0°	19DP41092	.100	24.1		.104			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
01/13/2020	5:12	34.0°	19DP41092	.100	27.6	.100				X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: #16040490

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
01/13/2020		34.0°	19DP41092								E. Mc Elroy	16193	OUT OF SERVICE

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO: 1604941

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
01/13/2020	1617	34.0°	19DP41092	100	23.3		1098			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202K\_1108.pdf

#5

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16040161

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
01/13/2020	1513	34.0°	19DP41092	.100	21.8		104			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf



























































