

Test Number	Result	ID	Test Time	Test Date	Unit Serial Number
88	0.103		2:02 PM	12/12/2019	16040119
117	0.058		1:45 AM	12/09/2019	16040140
76	0		11:44 PM	12/05/2019	16040141
77	0.107		2:18 AM	12/08/2019	16040141
78	0.124		2:22 AM	12/08/2019	16040141
79	0.075		3:32 AM	12/08/2019	16040141
80	0.091		3:34 AM	12/08/2019	16040141
81	0.085		3:37 AM	12/08/2019	16040141
82	0.109		11:31 PM	12/08/2019	16040141
83	0.105		11:34 PM	12/08/2019	16040141
295	0.195		12:19 AM	12/05/2019	16040161
296	0.198		12:23 AM	12/05/2019	16040161
297	0.103		2:29 AM	12/05/2019	16040161
298	0.097		2:30 AM	12/05/2019	16040161
299	0.16		6:23 AM	12/08/2019	16040161
300	0.158		6:25 AM	12/08/2019	16040161
374	0.201		10:03 AM	12/06/2019	16040162
375	0.213		10:05 AM	12/06/2019	16040162
376	0.108		8:50 AM	12/07/2019	16040162
377	0.106		8:52 AM	12/07/2019	16040162
378	0		3:13 AM	12/08/2019	16040162
379	0		4:57 AM	12/08/2019	16040162
380	NEG		2:06 AM	12/09/2019	16040162
381	NEG		2:07 AM	12/09/2019	16040162
382	0.052		2:23 AM	12/09/2019	16040162
383	0.053		2:25 AM	12/09/2019	16040162
384	0.105		1:32 PM	12/12/2019	16040162
113	0.02		3:48 AM	12/07/2019	16080013
114	0.119		12:01 AM	12/12/2019	16080013
115	0.127		12:03 AM	12/12/2019	16080013
81	0.177		1:44 AM	12/07/2019	16080014
82	0.178		1:46 AM	12/07/2019	16080014
126	0.091		12:23 AM	12/07/2019	16080027

chp life loc usage log

Test Number	Result	ID	Test Time	Test Date	Unit Serial Number
127	0.084		12:25 AM	12/07/2019	16080027
456	0.113		3:19 AM	12/06/2019	16080071
457	0.108		3:21 AM	12/06/2019	16080071
91	0.119		12:04 PM	12/04/2019	16120175
92	0.103		12:14 PM	12/04/2019	16120175
93	0		12:19 PM	12/06/2019	16120175
94	0.254		2:57 AM	12/07/2019	16120175
95	0.244		3:01 AM	12/07/2019	16120175
160	0.193		12:42 AM	12/07/2019	16120177
161	0.189		12:43 AM	12/07/2019	16120177
162	0.131		11:45 PM	12/07/2019	16120177
163	0.127		11:48 PM	12/07/2019	16120177
164	0.122		4:16 AM	12/11/2019	16120177
165	0.12		4:19 AM	12/11/2019	16120177
116	0		5:39 PM	12/06/2019	16120188
391	0.152		1:01 AM	12/09/2019	16120195
392	0.136		1:04 AM	12/09/2019	16120195
393	0.28		3:03 AM	12/09/2019	16120195
394	0.277		3:08 AM	12/09/2019	16120195
352	NEG		11:38 PM	12/02/2019	16120197
353	0.347		9:23 AM	12/06/2019	16120197
354	0.346		9:26 AM	12/06/2019	16120197



#2

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16090119

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/12/2019	1409	34.0°	19DP41092	.100	23					X	E. Mc Elroy	16193	
<b>Lifeloc Technologies</b> Calibration / Check Printout													
	FC20		8.7.24										
	Serial No.		16040119										
	UOM		BAC										
	Last Check		12/12/2019										
	Time:		2:10 PM										
	Cal Standard		.100										
	Result		.100										
	Last Cal		03/28/2019										
	Time:		3:25 PM										
	Cal Standard		.100										

\* If the reading is within plus or minus (±) 0.010 per cent  
\*\* If using dry gas for accuracy checks, note the canistic







#6

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16040163

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/12/2019	1515	34.0°	19DDP41092	.100	25	.097		X			E. Mc Elroy	16193	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24
Serial No.	16040163
UOM	BAC
Last Check	12/12/2019
Time:	3:15 PM
Cal Standard	.100
Result	.097
Last Cal	04/12/2019
Time:	07:30 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 per cent  
 \*\* If using dry gas for accuracy checks, note the canistic



#8

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16080014

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/12/2019	15:26	34.0°	19DDP41092	.100	23		0.097			X		E. Mc Elroy	16193	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24
Serial No.	16080014
UOM	BAC
Last Check	12/12/2019
Time:	3:27 PM
Cal Standard	.100
Result	.097
Last Cal	07/05/2019
Time:	11:51 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 perc  
 \*\* If using dry gas for accuracy checks, note the canist

#9

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward Mc Elroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080015

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #1	TEST #2		YES	NO						
12/12/2019	1526	34.0°	19DDP41092	.100	22				094			X		E. Mc Elroy	16193	

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24
Serial No.	16080015
UOM	BAC
Last Check	12/12/2019
Time:	3:26 PM
Cal Standard	.095
Result	.094
Last Cal	03/28/2019
Time:	3:58 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 per cent  
 \*\* If using dry gas for accuracy checks, note the canister

#10

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080027

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
12/12/2019	15:26	34.0°	19DDP41092	.100	22								E. Mc Elroy	16193	

SIMULATOR SOLUTION (OR DRY GAS)\*\*  
 Solution prepared by: Cesar Munoz  
 Date: 04-22-19  
 Expiration date: 04-22-21

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24	Serial No.	16080027	UOM	BAC	Last Check	12/12/2019	Time:	3:27 PM	Cal Standard	.100	Result	.097	Last Cal	03/28/2019	Time:	3:58 PM	Cal Standard	.100
------	--------	------------	----------	-----	-----	------------	------------	-------	---------	--------------	------	--------	------	----------	------------	-------	---------	--------------	------

\* If the reading is within plus or minus (±) 0.010 percent  
 \*\* If using dry gas for accuracy checks, note the canister



























#24

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES	NO				
12/12/2019	1554	34.0°	19DDP41092	.100	21			1094				X		E. Mc Elroy	16193	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24
Serial No.	16120194
UOM	BAC
Last Check	12/12/2019
Time:	3:54 PM
Cal Standard	.100
Result	.094
Last Cal	08/15/2019
Time:	12:20 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent  
 \*\* If using dry gas for accuracy checks, note the canister

#25

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/12/2019	1600	34.0°	19DP41092	.100	21	101	101	X			E. Mc Elroy	16193	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24
Serial No.	16120195
UOM	BAC
Last Check	12/12/2019
Time:	4:01 PM
Cal Standard	.100
Result	.101
Last Cal	06/09/2019
Time:	12:48 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent  
 \*\* If using dry gas for accuracy checks, note the canister

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1	TEST #2	YES	NO					
12/12/2019	1:09	34.0°	19DDP41092	.100	21		095				X	NO	E. Mc Elroy	16193	
<b>Lifeloc Technologies</b> Calibration / Check Printout															
				FC20	8.7.24										
				Serial No.	16120197										
				UOM	BAC										
				Last Check	12/12/2019										
				Time:	4:09 PM										
				Cal Standard	.100										
				Result	.095										
				Last Cal	06/25/2019										
				Time:	3:25 PM										
				Cal Standard	.100										

\* If the reading is within plus or minus (±) 0.010 percent  
 \*\* If using dry gas for accuracy checks, note the canister

#27

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 17170066

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
12/12/2019	1:13	34.0°	19DP41092	.100	20			.096				X	E. Mc Elroy	16193	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

## Lifeloc Technologies

### Calibration / Check Printout

FC20	8.7.24
Serial No.	17170066
UOM	BAC
Last Check	12/12/2019
Time:	4:13 PM
Cal Standard	.100
Result	.096
Last Cal	03/28/2019
Time:	4:42 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 perce  
 \*\* If using dry gas for accuracy checks, note the canistic

#28

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 17170150

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
12/12/2019	1610	34.0°	19DP41092	.100	21		.094				X		E. Mc Elroy	16193	

Solution prepared by: Cesar Munoz  
 Date: 04-22-19  
 Expiration date: 04-22-21

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	17170150
UOM	BAC
Last Check	12/12/2019
Time:	4:11 PM
Cal Standard	.100
Result	.094
Last Cal	08/24/2018
Time:	11:22 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percc  
 \*\* If using dry gas for accuracy checks, note the caniste





#31

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/12/2019	1613	34.0°	19DP41092	.100	20		.104			X		E. Mc Elroy	16193	

Solution prepared by:  
 Cesar Munoz  
 Date: 04-22-19  
 Expiration date:  
 04-22-21

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	17170064
UOM	BAC
Last Check	12/12/2019
Time:	4:14 PM
Cal Standard	.100
Result	.104
Last Cal	03/28/2019
Time:	5:05 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent  
 \*\* If using dry gas for accuracy checks, note the canister