

CHP SLA PAS USAGE LOGS
11/21/19-12/2/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
84	0.223		7:04 PM	11/24/2019	3:25 PM	03/28/2019	0.101	3:11 PM	11/21/2019	16040119
86	0.127		3:46 AM	11/28/2019	3:25 PM	03/28/2019	0.101	3:11 PM	11/21/2019	16040119
74	0.05		10:36 AM	12/01/2019	3:41 PM	03/28/2019	0.101	3:30 PM	11/21/2019	16040141
75	0.05		10:39 AM	12/01/2019	3:41 PM	03/28/2019	0.101	3:30 PM	11/21/2019	16040141
294	0.05		2:39 AM	12/01/2019	8:56 AM	04/25/2017	0.102	3:55 PM	11/21/2019	16040161
369	0		7:02 PM	11/21/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
370	0		7:03 PM	11/21/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
371	0.222		8:51 PM	11/28/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
372	0.186		8:53 PM	11/28/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
373	0.206		8:56 PM	11/28/2019	4:09 PM	06/25/2019	0.102	2:20 PM	11/21/2019	16040162
446	0.157		1:47 AM	12/01/2019	7:30 AM	04/12/2019	0.1	3:51 PM	11/21/2019	16040163
447	0.158		1:50 AM	12/01/2019	7:30 AM	04/12/2019	0.1	3:51 PM	11/21/2019	16040163
106	0		3:52 AM	11/23/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
110	0.051		8:36 PM	11/23/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
111	0.094		3:34 AM	11/24/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
112	0.089		3:38 AM	11/24/2019	12:59 PM	01/29/2018	0.1	4:16 PM	11/21/2019	16080013
120	0.233		7:13 PM	11/27/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
121	0.214		7:17 PM	11/27/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
122	0.119		8:58 AM	11/28/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
123	0.116		9:01 AM	11/28/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
124	0.063		10:15 AM	12/01/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
125	0.059		10:17 AM	12/01/2019	3:58 PM	03/28/2019	0.097	4:30 PM	11/21/2019	16080027
63	0.091		6:42 AM	11/21/2019	4:05 PM	03/28/2019	0.101	1:38 PM	11/13/2019	16080028
64	0.09		6:45 AM	11/21/2019	4:05 PM	03/28/2019	0.101	1:38 PM	11/13/2019	16080028
59	0.197		1:12 AM	11/24/2019	4:07 PM	03/28/2019	0.093	4:43 PM	11/21/2019	16080030
60	0.2		1:14 AM	11/24/2019	4:07 PM	03/28/2019	0.093	4:43 PM	11/21/2019	16080030
101	0.077		10:47 PM	11/22/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
102	0.042		11:22 PM	12/01/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
103	0.04		11:23 PM	12/01/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
104	0.079		2:53 AM	12/02/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
105	0.085		2:55 AM	12/02/2019	4:12 PM	03/28/2019	0.099	4:45 PM	11/21/2019	16080064
287	0.088		10:01 PM	11/21/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
288	0.083		10:05 PM	11/21/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065

CHP SLA PAS USAGE LOGS

11/21/19-12/2/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
289	0		5:34 PM	11/22/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
290	0.201		10:55 PM	11/29/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
291	0.213		10:58 PM	11/29/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
292	0.113		4:05 AM	11/30/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
293	0.11		4:08 AM	11/30/2019	1:54 PM	12/19/2017	0.097	4:49 PM	11/21/2019	16080065
442	0.116		1:34 AM	11/22/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
443	0.111		1:37 AM	11/22/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
444	0.103		3:02 AM	11/23/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
445	0.106		3:04 AM	11/23/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
446	0.019		3:10 AM	11/24/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
447	0.018		3:13 AM	11/24/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
448	0		12:44 AM	11/30/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
449	0.139		4:44 AM	11/30/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
450	0.133		4:47 AM	11/30/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
451	0.154		12:23 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
452	0.157		12:25 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
453	0.124		2:57 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
454	0.126		3:00 AM	12/01/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
455	0.055		12:05 AM	12/02/2019	10:41 AM	08/14/2017	0.092	4:58 PM	11/21/2019	16080071
38	0.119		1:02 AM	11/24/2019	2:05 PM	12/19/2017	0.094	5:46 PM	11/21/2019	16080072
39	0.124		1:05 AM	11/24/2019	2:05 PM	12/19/2017	0.094	5:46 PM	11/21/2019	16080072
115	0.157		1:59 AM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
116	0.161		2:01 AM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
117	0.231		5:19 PM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
118	0.226		5:21 PM	11/23/2019	4:30 PM	03/28/2019	0.102	5:42 PM	11/21/2019	16080077
86	0.248		11:51 AM	11/23/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
87	0.252		11:53 AM	11/23/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
88	0.053		2:33 AM	11/24/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
89	0.182		7:31 PM	11/24/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
90	0.178		7:33 PM	11/24/2019	4:36 PM	03/28/2019	0.104	5:40 PM	11/21/2019	16120175
147	0.171		11:57 PM	11/21/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
148	0.179		11:59 PM	11/21/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
149	0.044		12:06 AM	11/22/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177

CHP SLA PAS USAGE LOGS
11/21/19-12/2/19

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
150	0.075		2:44 AM	11/23/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
151	0.072		2:46 AM	11/23/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
152	0.194		11:12 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
153	0.191		11:14 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
154	0.158		11:20 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
155	0.192		11:22 PM	11/24/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
156	0.236		5:09 PM	11/25/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
157	0.246		5:12 PM	11/25/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
158	0.201		3:41 AM	11/28/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
159	0.195		3:44 AM	11/28/2019	9:50 AM	04/25/2017	0.098	5:50 PM	11/21/2019	16120177
112	0		9:39 PM	11/23/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
113	0.077		3:14 AM	11/28/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
114	0.078		3:17 AM	11/28/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
115	0		7:58 PM	11/29/2019	8:16 AM	10/21/2016	0.089	5:48 PM	11/21/2019	16120188
390	0.053		2:13 AM	11/23/2019	12:48 PM	06/09/2019	0.103	5:56 PM	11/21/2019	16120195
348	0.099		6:34 AM	11/29/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
349	0.101		6:36 AM	11/29/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
350	0.133		5:31 AM	11/30/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197
351	0.143		5:33 AM	11/30/2019	3:25 PM	06/25/2019	0.096	6:05 PM	11/21/2019	16120197

Test Number	Result	ID	Test Time	Test Date	Unit Serial Number
73	0		11:54 PM	11/27/2019	16080014
74	0.087		2:16 AM	11/28/2019	16080014
75	0.087		2:17 AM	11/28/2019	16080014
76	0.05		1:38 AM	11/30/2019	16080014
77	0.077		3:36 AM	11/30/2019	16080014
78	0.078		3:38 AM	11/30/2019	16080014
79	0.109		3:24 AM	12/01/2019	16080014
80	0.105		3:28 AM	12/01/2019	16080014
352	NEG		11:38 PM	12/02/2019	16120197

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lieloc FC20 PAS SERIAL NO: 16046162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1719	34	19DP41092	.100	23	.100			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 04-22-19										
			Expiration Date: 04-22-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

2

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1754	34	19DP41092	.100	25	.099				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-09) OPI 051

#3

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/02/2019	1720	34	19DP41092	.100	22		.096		X			E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz														
Date: 04-22-19														
Expiration Date: 04-22-21														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS/ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.			TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/02/2019	1745	34	19DP41092		.100	21		.101				X		E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz																
Date: 04-22-19																
Expiration Date: 04-22-21																

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/02/2019	1757	34	19DDP41092	.100	21	.099					E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

#6

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/02/2019	1748	34	19DP41092	.100	23		.093			X		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 04-22-19											
			Expiration Date:											
			04-22-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/02/2019	1805	34		19DRP41092	.100	22		096		X			E. Mc Elroy	16193	
			Solution prepared by:												
			Cesar Munoz												
			Date: 04-22-19												
			Expiration Date:												
			04-22-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-09) OPI 051

8

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy ID. NUMBER: 16193 PAS MAKE/MODEL: Lifelog FC20 PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019			19DP41092							E. Mc Elroy	16193	DEVICE NOT IN LOCKER FOR TESTING

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1106.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#9

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO	
12/02/2019	1757	34	19DP41092	.100		22		.095			X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz												
			Date: 04-22-19												
			Expiration Date: 04-22-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO.: 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/02/2019	1803 34		19DP41092	.100	22		1.095			X		E. Mc Elroy	16193	
		Solution prepared by:												
		Cesar Munoz												
		Date: 04-22-19												
		Expiration Date:												
		04-22-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifloc FC20	16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1808	34	19DP41092	.100	21	.093				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifefloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1808	34	19DP41092	.100	21	.095					E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 04-22-19										
			Expiration Date: 04-22-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles
AREA NO: 530
COORDINATOR NAME: Edward Mc Elroy
I.D. NUMBER: 16193
PAS MAKE/MODEL: Lifeloc FC20
PAS SERIAL NO: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1809	34	19DP41092	.100	21						E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Edward Mc Elroy
I.D. NUMBER: 16193
PAS MAKE/MODEL: Lifeloc PC20
PAS SERIAL NO.: 1608065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/02/2019	1810	34	19DP41092	.100	21		.092		X		E. Mc Elroy	16193		
Solution prepared by: Cesar Munoz Date: 04-22-19 Expiration Date: 04-22-21														

* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1811	34	19DP41092	.100	22		.092		X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 04-22-19										
			Expiration Date: 04-22-21										

* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward Mc Elroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifefloc FC20 | PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO	
12/02/2019	18:29	34	19DP41092	.100	21	.097				X			E. Mc Elroy	16193	
			Solution prepared by:												
			Cesar Munoz												
			Date: 04-22-19												
			Expiration Date: 04-22-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	18:27	34	19DP41092	.100	21	093			X		E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz													
Date: 04-22-19													
Expiration Date: 04-22-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051 #19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080077
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/02/2019	18:26	34	19DP41092	.100	22		.097		X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. <i>16120175</i>
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-----------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1823	34	19DP41092	.100	21		100		X		E. Mc Elroy	16193	
Solution prepared by:													
Cesar Munoz													
Date: 04-22-19													
Expiration Date: 04-22-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/02/2019	1833	34	19DP41092	.100	22		099		X		E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz													
Date: 04-22-19													
Expiration Date: 04-22-21													

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-09) OP1 051

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1832	34	19DP41092	.100	22	.092		X			E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 04-22-19										
			Expiration Date: 04-22-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Eroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1844 34		19DP41092	.100	22		095			X	E. Mc Eroy	16193	
		Solution prepared by:											
		Cesar Munoz											
		Date: 04-22-19											
		Expiration Date: 04-22-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifelog FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1908	34	19DP41092	.100	22		.091				E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 04-22-19										
			Expiration Date:										
			04-22-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifelog FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	2:55	34	19DP41092	.100	25		.097		X		E. Mc Elroy	16193	DEVICE NOT IN LOCKER FOR TESTING

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1930	34	19DP41092	.100	22		096			X	E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 04-22-19										
			Expiration Date: 04-22-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
AREA NO: 530
COORDINATOR NAME: Edward Mc Elroy
ID NUMBER: 16193
PAS MAKE/MODEL: Lifloc FC20
PAS SERIAL NO: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/02/2019	1926	34	19DP41092	.100	21		.099			X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170150

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1924	34	19DP41092	.100	21	098			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Edward Mc Elroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Intelec FC20 | PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/02/2019	1929	34	19DP41092	.100	21		.097			X		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 04-22-19											
			Expiration Date:											
			04-22-21											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME Edward Mc Elroy I.D. NUMBER 16193 PAS MAKE/MODEL Lifeloc FC20 PAS SERIAL NO. 90102844-82

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	34		19DP41092	.100							E. Mc Elroy	16193	OUT OF SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/02/2019	1933	34	19DP41092	.100	21	.097			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.