

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
119	0.158		9:48 PM	03/02/2020	0.1	9:55 AM	02/24/2020	16040119
120	0.148		9:51 PM	03/02/2020	0.1	9:55 AM	02/24/2020	16040119
147	0	0	8:29 PM	02/26/2020	0.095	9:42 AM	02/24/2020	16040140
148	0	0	8:30 PM	02/26/2020	0.095	9:42 AM	02/24/2020	16040140
114	0.178		2:58 AM	02/29/2020	0.092	9:53 AM	02/24/2020	16040141
115	0.175		3:00 AM	02/29/2020	0.092	9:53 AM	02/24/2020	16040141
116	0.157		2:16 AM	03/03/2020	0.092	9:53 AM	02/24/2020	16040141
117	0.15		2:18 AM	03/03/2020	0.092	9:53 AM	02/24/2020	16040141
323	0.097		1:35 AM	03/01/2020	0.1	10:10 AM	02/24/2020	16040161
324	0.097		1:38 AM	03/01/2020	0.1	10:10 AM	02/24/2020	16040161
419	0.093		8:00 AM	02/29/2020	0.102	9:38 AM	02/24/2020	16040162
420	0.095		8:02 AM	02/29/2020	0.102	9:38 AM	02/24/2020	16040162
421	0.089		8:04 AM	02/29/2020	0.102	9:38 AM	02/24/2020	16040162
422	0.139		1:59 AM	03/01/2020	0.102	9:38 AM	02/24/2020	16040162
423	0.128		2:03 AM	03/01/2020	0.102	9:38 AM	02/24/2020	16040162
424	0	0	2:14 AM	03/01/2020	0.102	9:38 AM	02/24/2020	16040162
463	0.119		11:59 PM	02/26/2020	0.096	10:00 AM	02/24/2020	16040163
464	0.115		12:01 AM	02/27/2020	0.096	10:00 AM	02/24/2020	16040163
465	0.108		10:00 PM	02/28/2020	0.096	10:00 AM	02/24/2020	16040163
466	0	0	5:01 AM	02/29/2020	0.096	10:00 AM	02/24/2020	16040163
467	0.196		11:21 PM	03/03/2020	0.096	10:00 AM	02/24/2020	16040163
468	0.199		11:24 PM	03/03/2020	0.096	10:00 AM	02/24/2020	16040163
143	0.029		2:53 AM	03/01/2020	0.103	10:20 AM	02/24/2020	16080013
144	0.026		2:55 AM	03/01/2020	0.103	10:20 AM	02/24/2020	16080013
148	0	0	9:29 AM	02/29/2020	0.098	10:18 AM	02/24/2020	16080027
98	0.177		3:12 AM	02/24/2020	0.091	12:10 PM	02/13/2020	16080028
99	0.174		3:14 AM	02/24/2020	0.091	12:10 PM	02/13/2020	16080028
100	0.174		4:27 AM	02/29/2020	0.093	10:25 AM	02/24/2020	16080028
101	0.176		4:30 AM	02/29/2020	0.093	10:25 AM	02/24/2020	16080028
102	0.06		2:18 AM	03/01/2020	0.093	10:25 AM	02/24/2020	16080028
103	0.191		11:15 PM	03/04/2020	0.093	10:25 AM	02/24/2020	16080028
104	0.186		11:19 PM	03/04/2020	0.093	10:25 AM	02/24/2020	16080028
84	0.041		2:18 AM	02/25/2020	0.096	10:24 AM	02/24/2020	16080030

chp life loc usage log

pas logs 2 24 20 to 3 5 20

85	0.164	2:47 AM	02/25/2020	0.096	10:24 AM	02/24/2020	16080030
86	0.168	2:50 AM	02/25/2020	0.096	10:24 AM	02/24/2020	16080030
123	0.127	10:39 PM	02/28/2020	0.097	10:26 AM	02/24/2020	16080064
124	0.129	10:42 PM	02/28/2020	0.097	10:26 AM	02/24/2020	16080064
125	0.12	1:14 AM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
126	0.118	1:16 AM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
127	0.101	3:58 AM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
128	0.101	4:01 AM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
129	0.148	7:21 PM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
130	0.144	7:24 PM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
131	0.064	7:31 PM	02/29/2020	0.097	10:26 AM	02/24/2020	16080064
132	0.079	1:35 AM	03/03/2020	0.097	10:26 AM	02/24/2020	16080064
133	0.096	1:37 AM	03/03/2020	0.097	10:26 AM	02/24/2020	16080064
319	0	8:08 PM	02/26/2020	0.098	10:28 AM	02/24/2020	16080065
320	0.046	10:39 PM	02/28/2020	0.098	10:28 AM	02/24/2020	16080065
321	0.064	1:47 AM	02/29/2020	0.098	10:28 AM	02/24/2020	16080065
322	0.066	3:14 AM	02/29/2020	0.098	10:28 AM	02/24/2020	16080065
38	0.111	5:45 PM	02/25/2020	0.09	10:47 AM	02/24/2020	16080075
39	0.106	5:49 PM	02/25/2020	0.09	10:47 AM	02/24/2020	16080075
40	0.149	3:31 AM	03/01/2020	0.09	10:47 AM	02/24/2020	16080075
41	0.142	3:33 AM	03/01/2020	0.09	10:47 AM	02/24/2020	16080075
141	0	5:11 PM	03/04/2020	0.103	10:52 AM	02/24/2020	16080077
142	0	5:14 PM	03/04/2020	0.103	10:52 AM	02/24/2020	16080077
143	0	7:22 PM	03/04/2020	0.103	10:52 AM	02/24/2020	16080077
144	0	7:24 PM	03/04/2020	0.103	10:52 AM	02/24/2020	16080077
145	0	7:56 PM	03/04/2020	0.103	10:52 AM	02/24/2020	16080077
146	0	7:59 PM	03/04/2020	0.103	10:52 AM	02/24/2020	16080077
133	0.234	4:14 PM	03/03/2020	0.099	10:50 AM	02/24/2020	16120175
134	0.238	4:16 PM	03/03/2020	0.099	10:50 AM	02/24/2020	16120175
192	0.162	3:21 AM	03/01/2020	0.095	11:06 AM	02/24/2020	16120177
193	0.139	3:23 AM	03/01/2020	0.095	11:06 AM	02/24/2020	16120177
194	0	5:19 PM	03/04/2020	0.095	11:06 AM	02/24/2020	16120177
143	0.171	3:04 AM	02/24/2020	0.099	12:31 PM	02/13/2020	16120188
144	0.184	3:07 AM	02/24/2020	0.099	12:31 PM	02/13/2020	16120188

chp life loc usage log

pas logs 2 24 20 to 3 5 20

145	0.233	1:12 AM	02/29/2020	0.094	11:04 AM	02/24/2020	16120188
146	0.208	1:14 AM	02/29/2020	0.094	11:04 AM	02/24/2020	16120188
422	0.161	2:35 AM	02/28/2020	0.098	11:16 AM	02/24/2020	16120195
423	0.157	2:37 AM	02/28/2020	0.098	11:16 AM	02/24/2020	16120195
424	0.147	9:28 PM	02/28/2020	0.098	11:16 AM	02/24/2020	16120195
425	0.153	9:30 PM	02/28/2020	0.098	11:16 AM	02/24/2020	16120195































































AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. # 17170155
--------------------------------	-----------------	---------------------------------	---------------------	--------------------------------	------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/05/2020	1536	34.0°	1P41101	.100	203		1095		X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





