

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/06/2020 08:45

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|------|--------------------|------|--------|------------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Fail | | |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass | | |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 2859 | | |
| Check sum | 69730 | Download date/time: | 02/06/2020 | 09:28 | | Record ID | AZN3 | | |
| Record imported (Y/N) | | | | N | | | | Source | Diagnostic |

Instrument serial No: 68-012338 Accuracy check date/time: 02/06/2020 08:46

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|--|--------------------|------|--------|-----------|
| Eeprom | | Serial No. | | Temp | | Internal standards | | | |
| Prom | | Real Time clock | | Motor | | Printer | | | |
| Ram | | Auto calibration | | Range | | Digital voltmeter | | | |
| Check sum | 55305 | Download date/time: | 02/06/2020 | 09:28 | | Record ID | AZR0 | | |
| Record imported (Y/N) | | | | N | | | | Source | Exception |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/10/2020 16:59

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|-------------------|--------------------|------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Fail |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 2865 |
| Check sum | 14662 | Download date/time: | 02/27/2020 | 16:19 | | Record ID | B074 |
| Record imported (Y/N) | | | | N | Source Diagnostic | | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/10/2020 20:38

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 011

Operator _____

Diagnostics check

| | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|--------------------|--------------------|------|
| Eeprom | | Serial No. | | Temp | | Internal standards | |
| Prom | | Real Time clock | | Motor | | Printer | |
| Ram | | Auto calibration | | Range | | Digital voltmeter | |
| Check sum | 22775 | Download date/time: | 02/27/2020 | 16:19 | | Record ID | AZX2 |
| Record imported (Y/N) | | | | N | Source Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/13/2020 08:38

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|------|--------------------|------|-------------------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass | |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass | |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 2884 | |
| Check sum | 60698 | Download date/time: | 02/27/2020 | 16:19 | | Record ID | B085 | |
| Record imported (Y/N) | | | | N | | | | Source Diagnostic |

Instrument serial No: 68-012338 Accuracy check date/time: 02/13/2020 08:40

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator _____

Diagnostics check

| | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|--|--------------------|------|--------------------|
| Eeprom | | Serial No. | | Temp | | Internal standards | | |
| Prom | | Real Time clock | | Motor | | Printer | | |
| Ram | | Auto calibration | | Range | | Digital voltmeter | | |
| Check sum | 35139 | Download date/time: | 02/27/2020 | 16:19 | | Record ID | AZY1 | |
| Record imported (Y/N) | | | | N | | | | Source Calibration |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/13/2020 08:42

First calibration check: .141 Interferent detected: No
Second calibration check: .143 Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 014

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 53692 | Download date/time: 02/27/2020 | 16:19 | Record ID AZZZ |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/14/2020 19:04

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 60059 | Download date/time: 02/27/2020 | 16:19 | Record ID B0B8 |
| Record imported (Y/N) N | Source Exception | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/14/2020 19:10

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Range exceeded

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-----------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 67800 | Download date/time: 02/27/2020 | 16:19 | Record ID B0C7 |
| Record imported (Y/N) N | Source | Exception | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/20/2020 08:41

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | |
|-------------------------|--------------------------------|------------------|-----------|-------|--------------------|-------------------|------|
| Eeprom | Pass | Serial No. Pass | Temp | Pass | Internal standards | Pass | |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 2890 |
| Check sum 20423 | Download date/time: 02/27/2020 | 16:19 | Record ID | B090 | | | |
| Record imported (Y/N) N | Source | Diagnostic | | | | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/20/2020 08:43

First calibration check: .STD Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 57547 | Download date/time: 02/27/2020 | 16:19 | Record ID B001 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/20/2020 08:43

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 63949 | Download date/time: 02/27/2020 | 16:19 | Record ID B0D0 |
| Record imported (Y/N) N | Source Exception | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/20/2020 08:45

First calibration check: Unstable reference Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 20675 | Download date/time: 02/27/2020 | 16:19 | Record ID B016 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/20/2020 08:45

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 51741 | Download date/time: 02/27/2020 | 16:19 | Record ID B0E9 |
| Record imported (Y/N) N | Source Exception | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/20/2020 08:47

First calibration check: .146 Interferent detected: No
Second calibration check: .146 Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 65320 | Download date/time: 02/27/2020 | 16:19 | Record ID B025 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/27/2020 08:00

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------------------|-------------------------|
| Eeprom Pass | Serial No. Pass | Temp Pass | Internal standards Fail |
| Prom Pass | Real Time clock Pass | Motor Pass | Printer Pass |
| Ram Pass | Auto calibration Pass | Range Pass | Digital voltmeter 2514 |
| Check sum 34508 | Download date/time: 02/27/2020 | 16:19 | Record ID B0A3 |
| Record imported (Y/N) N | | Source Diagnostic | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/27/2020 08:02

First calibration check: .STD Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 013

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 34329 | Download date/time: 02/27/2020 | 16:19 | Record ID B030 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/27/2020 08:02

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 69920 | Download date/time: 02/27/2020 | 16:19 | Record ID B0F4 |
| Record imported (Y/N) N | Source Exception | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/27/2020 08:05

First calibration check: Aborted Interferent detected:
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 013

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 55226 | Download date/time: 02/27/2020 | 16:19 | Record ID B045 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012338 Accuracy check date/time: 02/27/2020 08:07

First calibration check: Aborted Interferent detected:
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 46134 | Download date/time: 02/27/2020 | 16:19 | Record ID B050 |
| Record imported (Y/N) N | Source Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 02/27/2020 08:09

First calibration check: .137 Interferent detected: No
Second calibration check: Aborted Interferent detected:

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 59554 | Download date/time: 02/27/2020 | 16:19 | Record ID B069 |
| Record imported (Y/N) N | Source Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.