

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/05/2020 10:30

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|------|--------------------|------|-------------------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass | |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass | |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3222 | |
| Check sum | 50885 | Download date/time: | 03/05/2020 | 11:16 | | Record ID | B1N7 | |
| Record imported (Y/N) | | | | N | | | | Source Diagnostic |

Instrument serial No: 68-012333 Accuracy check date/time: 03/05/2020 10:35

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .147

Solution lot#: IRS011620

Soln temp C: 34.00

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

| | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|--|--------------------|------|--------------------|
| Eeprom | | Serial No. | | Temp | | Internal standards | | |
| Prom | | Real Time clock | | Motor | | Printer | | |
| Ram | | Auto calibration | | Range | | Digital voltmeter | | |
| Check sum | 52930 | Download date/time: | 03/05/2020 | 11:16 | | Record ID | B1J1 | |
| Record imported (Y/N) | | | | N | | | | Source Calibration |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/05/2020 10:37

First calibration check: .150 Interferent detected: No
Second calibration check: .152 Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.03
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 18694 | Download date/time: 03/05/2020 | 11:16 | Record ID B1K2 |
| Record imported (Y/N) N | Source Calibration | | |

Instrument serial No: 68-012333 Accuracy check date/time: 03/08/2020 03:25

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

CHANGED TIME FROM 0225 TO 0325

Number of people tested since last periodic determination of accuracy:

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 28330 | Download date/time: 04/02/2020 | 10:38 | Record ID B3Q9 |
| Record imported (Y/N) N | Source Memo | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/12/2020 06:44

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | |
|-----------|-----------------------|---------------------|------------|--------|------------|--------------------|------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3225 |
| Check sum | 50594 | Download date/time: | 04/02/2020 | 10:38 | | Record ID | B3M0 |
| | Record imported (Y/N) | N | | Source | Diagnostic | | |

Instrument serial No: 68-012333 Accuracy check date/time: 03/12/2020 06:45

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .147

Solution lot#: IRS011620

Soln temp C: 34.03

Simulator Ser No. MP5021

CHANGED TIME FROM 0225 TO 0325

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

| | | | | | | | |
|-----------|-----------------------|---------------------|------------|--------|-------------|--------------------|------|
| Eeprom | | Serial No. | | Temp | | Internal standards | |
| Prom | | Real Time clock | | Motor | | Printer | |
| Ram | | Auto calibration | | Range | | Digital voltmeter | |
| Check sum | 59968 | Download date/time: | 04/02/2020 | 10:38 | | Record ID | B3E4 |
| | Record imported (Y/N) | N | | Source | Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/12/2020 06:49

First calibration check: .144 Interferent detected: No
Second calibration check: .148 Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 37232 | Download date/time: 04/02/2020 | 10:38 | Record ID B3F9 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012333 Accuracy check date/time: 03/19/2020 10:18

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------------------|-------------------------|
| Eeprom Pass | Serial No. Pass | Temp Pass | Internal standards Pass |
| Prom Pass | Real Time clock Pass | Motor Pass | Printer Pass |
| Ram Pass | Auto calibration Pass | Range Pass | Digital voltmeter 3219 |
| Check sum 31787 | Download date/time: 04/02/2020 | 10:38 | Record ID B3N5 |
| Record imported (Y/N) N | | Source Diagnostic | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/19/2020 10:19

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 47401 | Download date/time: 04/02/2020 | 10:38 | Record ID B3G4 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012333 Accuracy check date/time: 03/19/2020 10:21

First calibration check: .145 Interferent detected: No
Second calibration check: .150 Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 33.99
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 32813 | Download date/time: 04/02/2020 | 10:38 | Record ID B3H7 |
| Record imported (Y/N) N | | Source Calibration | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/26/2020 10:38

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|------|--------------------|------|-------------------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass | |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass | |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3201 | |
| Check sum | 21446 | Download date/time: | 04/02/2020 | 10:38 | | Record ID | B300 | |
| Record imported (Y/N) | | | | N | | | | Source Diagnostic |

Instrument serial No: 68-012333 Accuracy check date/time: 03/26/2020 10:39

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .147

Solution lot#: IRS011620

Soln temp C: 33.99

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

| | | | | | | | | |
|-----------------------|-------|---------------------|------------|-------|--|--------------------|------|--------------------|
| Eeprom | | Serial No. | | Temp | | Internal standards | | |
| Prom | | Real Time clock | | Motor | | Printer | | |
| Ram | | Auto calibration | | Range | | Digital voltmeter | | |
| Check sum | 25005 | Download date/time: | 04/02/2020 | 10:38 | | Record ID | B312 | |
| Record imported (Y/N) | | | | N | | | | Source Calibration |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 03/26/2020 10:41

First calibration check: .142 Interferent detected: No
Second calibration check: .152 Interferent detected: No

Concentration (%W/V) .147 Solution lot#: IRS011620 Soln temp C: 33.64
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 42592 | Download date/time: 04/02/2020 | 10:38 | Record ID B3J7 |
| Record imported (Y/N) N | Source Calibration | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.