

CHP SLA PAS Usage Logs 4/3/2020

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
153	0.278		7:57 PM	03/28/2020	0.095	12:59 PM	03/24/2020	16040140
154	0.277		7:59 PM	03/28/2020	0.095	12:59 PM	03/24/2020	16040140
155	0.174		7:39 AM	03/29/2020	0.095	12:59 PM	03/24/2020	16040140
156	0.175		7:42 AM	03/29/2020	0.095	12:59 PM	03/24/2020	16040140
122	0.149		10:41 PM	03/24/2020	0.095	1:01 PM	03/24/2020	16040141
123	0.142		10:44 PM	03/24/2020	0.095	1:01 PM	03/24/2020	16040141
326	0.242		10:19 PM	03/29/2020	0.098	1:03 PM	03/24/2020	16040161
327	0.276		10:22 PM	03/29/2020	0.098	1:03 PM	03/24/2020	16040161
442	0.137		6:08 PM	03/27/2020	0.098	12:56 PM	03/24/2020	16040162
443	0.139		6:11 PM	03/27/2020	0.098	12:56 PM	03/24/2020	16040162
444	0.192		3:15 AM	03/29/2020	0.098	12:56 PM	03/24/2020	16040162
445	0.182		3:18 AM	03/29/2020	0.098	12:56 PM	03/24/2020	16040162
446	0.248		2:59 PM	03/30/2020	0.098	12:56 PM	03/24/2020	16040162
447	0.243		3:01 PM	03/30/2020	0.098	12:56 PM	03/24/2020	16040162
448	0		3:17 PM	04/01/2020	0.098	12:56 PM	03/24/2020	16040162
473	0.017		1:34 AM	03/25/2020	0.097	1:04 PM	03/24/2020	16040163
150	0		5:03 PM	03/26/2020	0.093	3:48 PM	03/24/2020	16080013
151	0.146		12:50 AM	03/29/2020	0.093	3:48 PM	03/24/2020	16080013
152	0.151		12:53 AM	03/29/2020	0.093	3:48 PM	03/24/2020	16080013
106	0.107		1:25 AM	03/25/2020	0.099	1:11 PM	03/24/2020	16080015
107	0.1		1:27 AM	03/25/2020	0.099	1:11 PM	03/24/2020	16080015
152	0.066		1:15 AM	03/29/2020	0.098	1:13 PM	03/24/2020	16080027
153	0.064		1:18 AM	03/29/2020	0.098	1:13 PM	03/24/2020	16080027
323	0		7:55 AM	03/28/2020	0.106	1:19 PM	03/24/2020	16080065
324	0.168		3:57 AM	03/29/2020	0.106	1:19 PM	03/24/2020	16080065
325	0.17		4:01 AM	03/29/2020	0.106	1:19 PM	03/24/2020	16080065
326	0.049		9:36 AM	04/01/2020	0.106	1:19 PM	03/24/2020	16080065
327	0.044		9:38 AM	04/01/2020	0.106	1:19 PM	03/24/2020	16080065
46	0		10:12 PM	03/28/2020	0.099	1:26 PM	03/24/2020	16080075
47	0.12		12:33 AM	03/29/2020	0.099	1:26 PM	03/24/2020	16080075
48	0.116		12:35 AM	03/29/2020	0.099	1:26 PM	03/24/2020	16080075
142	0		1:52 AM	04/03/2020	0.1	3:51 PM	03/24/2020	16120175
143	0		1:52 AM	04/03/2020	0.1	3:51 PM	03/24/2020	16120175

CHP SLA PAS Usage Logs 4/3/2020

144	0	1:53 AM	04/03/2020	0.1	3:51 PM	03/24/2020	16120175
197	0.205	11:33 PM	03/26/2020	0.102	1:31 PM	03/24/2020	16120177
198	0.115	11:36 PM	03/26/2020	0.102	1:31 PM	03/24/2020	16120177
158	0	8:12 PM	03/29/2020	0.106	1:37 PM	03/24/2020	16120188
347	0.148	2:09 AM	03/29/2020	0.095	1:41 PM	03/24/2020	16120194
348	0.136	2:12 AM	03/29/2020	0.095	1:41 PM	03/24/2020	16120194
430	0	2:03 AM	04/02/2020	0.101	1:43 PM	03/24/2020	16120195

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16080077
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/03/2020	1249	34.0	19IP41101	.100	23.3	.104				✓		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/03/2020	1258	34.0		19P41101	.100	23.8	.099					✓	E. Mc Elroy	16193	
			Solution prepared by:												
			Cesar Munoz												
			Date: 09-10-19												
			Expiration date:												
			09-10-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/03/2020	1300	34.0	19IP41101	.100	23.6	1094				✓	E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 09-10-19										
			Expiration date: 09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/03/2020			19IP41101	.100							E. Mc Elroy	16193	OUT OF SERVICE
		Solution prepared by:											
		Cesar Munoz											
		Date: 09-10-19											
		Expiration date:											
		09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 17170064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
04/03/2020	1316	34.0	19IP41101	.100	22.3	.100						E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.