

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
149	0.217		7:41 PM	03/07/2020	0.097	2:01 PM	03/05/2020	16040140
150	0.208		7:44 PM	03/07/2020	0.097	2:01 PM	03/05/2020	16040140
151	0.098		5:48 PM	03/08/2020	0.097	2:01 PM	03/05/2020	16040140
152	0.099		5:51 PM	03/08/2020	0.097	2:01 PM	03/05/2020	16040140
118	0.176		6:02 PM	03/15/2020	0.099	2:06 PM	03/05/2020	16040141
119	0.159		6:09 PM	03/15/2020	0.099	2:06 PM	03/05/2020	16040141
325	0.021		3:43 PM	03/06/2020	0.099	2:21 PM	03/05/2020	16040161
425	0		9:45 AM	03/07/2020	0.096	2:02 PM	03/05/2020	16040162
426	0.117		1:05 AM	03/08/2020	0.096	2:02 PM	03/05/2020	16040162
427	0.102		1:08 AM	03/08/2020	0.096	2:02 PM	03/05/2020	16040162
428	0		1:18 AM	03/08/2020	0.096	2:02 PM	03/05/2020	16040162
429	0.106		2:46 AM	03/08/2020	0.096	2:02 PM	03/05/2020	16040162
430	0.106		2:48 AM	03/08/2020	0.096	2:02 PM	03/05/2020	16040162
431	0.276		9:17 PM	03/10/2020	0.096	2:02 PM	03/05/2020	16040162
432	0.292		9:21 PM	03/10/2020	0.096	2:02 PM	03/05/2020	16040162
433	0.046		9:51 PM	03/10/2020	0.096	2:02 PM	03/05/2020	16040162
434	0.065		5:02 PM	03/11/2020	0.096	2:02 PM	03/05/2020	16040162
435	0.064		5:04 PM	03/11/2020	0.096	2:02 PM	03/05/2020	16040162
436	0.069		2:19 AM	03/12/2020	0.096	2:02 PM	03/05/2020	16040162
469	0.109		1:41 AM	03/14/2020	0.095	2:11 PM	03/05/2020	16040163
470	0.108		1:44 AM	03/14/2020	0.095	2:11 PM	03/05/2020	16040163
471	0.138		1:13 AM	03/15/2020	0.095	2:11 PM	03/05/2020	16040163
472	0.139		1:15 AM	03/15/2020	0.095	2:11 PM	03/05/2020	16040163
145	0		9:00 PM	03/13/2020	0.102	2:31 PM	03/05/2020	16080013
146	0.152		2:36 AM	03/14/2020	0.102	2:31 PM	03/05/2020	16080013
147	0.146		2:39 AM	03/14/2020	0.102	2:31 PM	03/05/2020	16080013
148	0.104		3:41 AM	03/14/2020	0.102	2:31 PM	03/05/2020	16080013
149	0.096		3:44 AM	03/14/2020	0.102	2:31 PM	03/05/2020	16080013
99	0.104		1:27 AM	03/06/2020	0.098	2:24 PM	03/05/2020	16080014
100	0.112		1:30 AM	03/06/2020	0.098	2:24 PM	03/05/2020	16080014
101	0.09		12:49 AM	03/12/2020	0.098	2:24 PM	03/05/2020	16080014
102	0.09		12:51 AM	03/12/2020	0.098	2:24 PM	03/05/2020	16080014
103	0.129		10:28 PM	03/15/2020	0.098	2:24 PM	03/05/2020	16080014

104	0.129	10:31 PM	03/15/2020	0.098	2:24 PM	03/05/2020	16080014
104	0.157	1:04 AM	03/12/2020	0.098	2:21 PM	03/05/2020	16080015
105	0.154	1:08 AM	03/12/2020	0.098	2:21 PM	03/05/2020	16080015
149	0.012	8:52 PM	03/05/2020	0.096	2:22 PM	03/05/2020	16080027
105	0.022	2:43 AM	03/08/2020	0.097	2:29 PM	03/05/2020	16080028
106	0.086	3:34 AM	03/08/2020	0.097	2:29 PM	03/05/2020	16080028
107	0.07	3:37 AM	03/08/2020	0.097	2:29 PM	03/05/2020	16080028
108	0.08	3:39 AM	03/08/2020	0.097	2:29 PM	03/05/2020	16080028
109	0.1	12:00 AM	03/14/2020	0.097	2:29 PM	03/05/2020	16080028
110	0.104	12:03 AM	03/14/2020	0.097	2:29 PM	03/05/2020	16080028
111	0.05	11:31 AM	03/15/2020	0.097	2:29 PM	03/05/2020	16080028
87	0.084	8:45 PM	03/08/2020	0.096	2:28 PM	03/05/2020	16080030
88	0.088	8:48 PM	03/08/2020	0.096	2:28 PM	03/05/2020	16080030
89	0.161	1:11 AM	03/09/2020	0.096	2:28 PM	03/05/2020	16080030
90	0.166	1:14 AM	03/09/2020	0.096	2:28 PM	03/05/2020	16080030
91	0.111	12:51 AM	03/14/2020	0.096	2:28 PM	03/05/2020	16080030
92	0.112	12:53 AM	03/14/2020	0.096	2:28 PM	03/05/2020	16080030
93	0.015	11:08 PM	03/14/2020	0.096	2:28 PM	03/05/2020	16080030
94	0.224	1:35 AM	03/15/2020	0.096	2:28 PM	03/05/2020	16080030
95	0.227	1:38 AM	03/15/2020	0.096	2:28 PM	03/05/2020	16080030
134	0.151	1:47 AM	03/15/2020	0.097	2:29 PM	03/05/2020	16080064
135	0.138	1:50 AM	03/15/2020	0.097	2:29 PM	03/05/2020	16080064
479	0.033	10:08 PM	03/08/2020	0.095	2:38 PM	03/05/2020	16080071
480	0.095	1:13 AM	03/09/2020	0.095	2:38 PM	03/05/2020	16080071
481	0.106	1:15 AM	03/09/2020	0.095	2:38 PM	03/05/2020	16080071
42	0	2:59 AM	03/08/2020	0.091	3:01 PM	03/05/2020	16080075
43	0.083	1:04 AM	03/11/2020	0.091	3:01 PM	03/05/2020	16080075
44	0.085	1:07 AM	03/11/2020	0.091	3:01 PM	03/05/2020	16080075
147	0.103	2:18 PM	03/10/2020	0.102	3:04 PM	03/05/2020	16080077
148	0.092	2:19 PM	03/10/2020	0.102	3:04 PM	03/05/2020	16080077
135	0	7:38 PM	03/07/2020	0.099	3:02 PM	03/05/2020	16120175
136	0.014	6:04 PM	03/08/2020	0.099	3:02 PM	03/05/2020	16120175
137	0.014	6:06 PM	03/08/2020	0.099	3:02 PM	03/05/2020	16120175
138	0	3:58 AM	03/12/2020	0.099	3:02 PM	03/05/2020	16120175

chnp life loc usage log

pas logs 3 5 20 to 3 16 20

139	0.105	6:13 PM	03/15/2020	0.099	3:02 PM	03/05/2020	16120175
140	0.095	6:16 PM	03/15/2020	0.099	3:02 PM	03/05/2020	16120175
147	0.248	9:39 PM	03/10/2020	0.105	3:12 PM	03/05/2020	16120188
148	0.238	9:41 PM	03/10/2020	0.105	3:12 PM	03/05/2020	16120188
149	0.093	3:39 AM	03/14/2020	0.105	3:12 PM	03/05/2020	16120188
150	0.092	3:43 AM	03/14/2020	0.105	3:12 PM	03/05/2020	16120188
151	0.12	12:29 AM	03/15/2020	0.105	3:12 PM	03/05/2020	16120188
152	0.145	12:32 AM	03/15/2020	0.105	3:12 PM	03/05/2020	16120188
153	0.207	10:25 PM	03/15/2020	0.105	3:12 PM	03/05/2020	16120188
154	0.206	10:28 PM	03/15/2020	0.105	3:12 PM	03/05/2020	16120188
426	0.117	11:05 PM	03/06/2020	0.101	3:22 PM	03/05/2020	16120195
427	0.096	11:07 PM	03/06/2020	0.101	3:22 PM	03/05/2020	16120195
428	0.094	3:58 AM	03/07/2020	0.101	3:22 PM	03/05/2020	16120195
429	0.095	4:01 AM	03/07/2020	0.101	3:22 PM	03/05/2020	16120195
166	0	9:55 PM	03/07/2020	0.097	3:34 PM	03/05/2020	17170064
167	0.163	2:39 AM	03/08/2020	0.097	3:34 PM	03/05/2020	17170064
168	0.175	2:42 AM	03/08/2020	0.097	3:34 PM	03/05/2020	17170064
169	0	11:21 PM	03/15/2020	0.097	3:34 PM	03/05/2020	17170064

#1

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc PC20 PAS SERIAL NO: # 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020	134	34.0°	19IP41101	.100	23.1	.101			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME
 South Los Angeles

AREA NO. 530
 COORDINATOR NAME
 E. Mc Elroy

I.D. NUMBER
 16193

PAS MAKE/MODEL
 LifaLoc PC20

PAS SERIAL NO.
 # 16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020	154	34.0°	19IP41101	.100	21.6	100				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: # 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/16/2020	1141	34.0°	19IP41101	.100	22.1	.099				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

#4

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: E. Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifenoc FC20
 PAS SERIAL NO.: # 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #1	TEST #2 (Optional)					
03/16/2020	1115	34.0°	19IP41101	.100	20.8	.096			X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: E. Mc Elroy
I.D. NUMBER: 16193
PAS MAKE/MODEL: Liteloc FC20
PAS SERIAL NO.: # 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1434	34.0°	191P41101	.100	23.1	.092		X			E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions
Chp202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FCC20 PAS SERIAL NO.: # 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP		SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		(°C)		LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1158	34.0°		19IP41101	.100	23.1		0.97				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: E. Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lieloc PC20
 PAS SERIAL NO.: # 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/16/2020	1202	34.0°	191P41101	.100	22.13	.103		X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the caustic serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. # 16080014
--------------------------------	-----------------	---------------------------------	----------------------	-------------------------------	------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1205	34.0°		19IP41101	.100	22.3		098			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1109.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: # 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1208	34.0°	19IP41101	.100	22.3		.097		X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

#10

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: # 16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1212	34.0°	19IP41101	.100	22.1	.096			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME E. Mc Elroy I.D. NUMBER 16193 PAS MAKE/MODEL Lifeloc FC20 PAS SERIAL NO. # 16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020	1215	34.0°	19IP41101	.100	12.3	.097					E. Mc Elroy	16193	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. # 16080030
--------------------------------	-----------------	---------------------------------	----------------------	--------------------------------	------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020	12:18	34.0°	19IP41101	.100	21.8	.096			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifenoc PC20 PAS SERIAL NO.: # 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	TEMP #1	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
03/16/2020	1222	34.0°	19DP41101			21.8		.097		X		E. Mc Elroy	16193	
				Solution prepared by:										
				Cesar Munoz										
				Date: 09-10-19										
				Expiration Date:										
				09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifelog FC20 PAS SERIAL NO.: # 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1226	34.0°	191P41101	.100	20.6		100					E. Mc Elroy	16193	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy ID. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc PC20 PAS SERIAL NO.: # 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1230	34.0°	191P41101	.100	21.8		1.094			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions CHP202K_1108.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: # 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020	1233	34.0°	19IP41101	.100	22.1		.096			X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO: #16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1 (Optional)	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1236	34.0°	19TP41101								E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME E. Mc Elroy ID. NUMBER 16193 PAS MAKE/MODEL Lifeloc PC20 PAS SERIAL NO. #16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1242	34.0°	191P41101	.100	22.6		0.94			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions
 Chip202k_1108.pdf

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: E. Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloq FCC20
 PAS SERIAL NO.: # 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS			
03/16/2020	1245	34.0°	19IP41101	.100	21.8	104			X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc PC20 PAS SERIAL NO.: # 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020	1249	34.0°	19IP41101	.100	21.6		.098				E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: # 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1252	34.0°	191P41101	.100	22.6		.100					E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	E. Mc Elroy	16193	Lifeloc FC20	# 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1255	34.0°	19DP41101	.100	22.3	.105					E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifenoc FC20 PAS SERIAL NO.: # 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1301	34.0°	19IP41101	22.8		1.095			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

CHP202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. # 16120194
--------------------------------	-----------------	---------------------------------	----------------------	--------------------------------	------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	13:24	34.0°		19DP41101	.100		21.8		.092		X		E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 CHP202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: E. Mc Elroy ID. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc PC20 PAS SERIAL NO.: # 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1308	34.0°	19IP41101		.100	21.3		.099		X			E. Mc Elroy	16193	
			Solution prepared by:												
			Cesar Munoz												
			Date: 09-10-19												
			Expiration Date:												
			09-10-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc PC20 PAS SERIAL NO.: # 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020		34.0°	19IP41101		.100								E. Mc Elroy	16193	<i>OUT of SERVICE</i>

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. # 17170066
--------------------------------	-----------------	---------------------------------	----------------------	--------------------------------	------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
03/16/2020	1312	34.0°	191P41101		.100	21.1		1.07				X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz													
			Date: 09-10-19													
			Expiration Date: 09-10-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	E. Mc Elroy	16193	Lifeloc FC20	# 1770150

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020		34.0°	19IP41101								E. Mc Elroy	16193	OUT OF SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc PC20	PAS SERIAL NO. # 17170155
--------------------------------	-----------------	---------------------------------	---------------------	--------------------------------	------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/16/2020	1335	34.0°	19IP41101	.100	21.2		.092					E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: #90102844-82

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/16/2020		34.0°	19IP41101	.100							E. Mc Elroy	16193	OUT OF SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifoloc FC20 PAS SERIAL NO. # 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1 (Optional)	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				

03/16/2020	1354	34.0°	19IP41101	.100	21.6	1.097			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 CHP202K_1108.pdf