

CHP SLA PAS Usage Logs 07-06-2020

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
152	0.041		3:20 PM	06/27/2020	0.099	7:50 AM	06/25/2020	16040119
153	0.04		3:22 PM	06/27/2020	0.099	7:50 AM	06/25/2020	16040119
175	0.183		2:14 AM	07/02/2020	0.098	7:49 AM	06/25/2020	16040140
176	0.18		2:16 AM	07/02/2020	0.098	7:49 AM	06/25/2020	16040140
177	0.144		11:37 PM	07/02/2020	0.098	7:49 AM	06/25/2020	16040140
178	0.148		11:40 PM	07/02/2020	0.098	7:49 AM	06/25/2020	16040140
138	0.218		4:00 PM	06/27/2020	0.104	7:53 AM	06/25/2020	16040141
139	0.215		4:02 PM	06/27/2020	0.104	7:53 AM	06/25/2020	16040141
140	0.082		12:44 AM	07/03/2020	0.104	7:53 AM	06/25/2020	16040141
141	0.091		12:47 AM	07/03/2020	0.104	7:53 AM	06/25/2020	16040141
349	0.128		11:55 PM	07/05/2020	0.099	8:00 AM	06/25/2020	16040161
350	0.119		11:58 PM	07/05/2020	0.099	8:00 AM	06/25/2020	16040161
480	0.13		10:57 PM	07/05/2020	0.102	7:43 AM	06/25/2020	16040162
481	0.143		10:59 PM	07/05/2020	0.102	7:43 AM	06/25/2020	16040162
188	0.245		10:58 PM	06/25/2020	0.102	8:05 AM	06/25/2020	16080013
189	0.238		11:01 PM	06/25/2020	0.102	8:05 AM	06/25/2020	16080013
190	0		1:43 AM	07/05/2020	0.102	8:05 AM	06/25/2020	16080013
191	0		1:44 AM	07/05/2020	0.102	8:05 AM	06/25/2020	16080013
125	0.069		12:37 AM	07/05/2020	0.099	8:11 AM	06/25/2020	16080015
126	0.069		12:40 AM	07/05/2020	0.099	8:11 AM	06/25/2020	16080015
121	0.194		1:21 AM	06/29/2020	0.1	8:30 AM	06/25/2020	16080028
122	0.189		1:24 AM	06/29/2020	0.1	8:30 AM	06/25/2020	16080028
108	0		1:30 AM	06/27/2020	0.096	8:32 AM	06/25/2020	16080030
339	0.159		10:42 PM	07/02/2020	0.101	8:41 AM	06/25/2020	16080065
340	0.162		10:44 PM	07/02/2020	0.101	8:41 AM	06/25/2020	16080065
341	0.183		12:22 AM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
342	0.189		12:25 AM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
343	0		12:27 AM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
344	0.147		2:54 AM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
345	0.075		2:55 AM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
346	0.253		9:02 PM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
347	0.251		9:05 PM	07/05/2020	0.101	8:41 AM	06/25/2020	16080065
182	0.173		11:11 PM	06/27/2020	0.094	8:45 AM	06/25/2020	16080069

CHP SLA PAS Usage Logs 07-06-2020

183	0.164	11:14 PM	06/27/2020	0.094	8:45 AM	06/25/2020	16080069
184	NEG	2:42 AM	07/06/2020	0.094	8:45 AM	06/25/2020	16080069
185	0.104	2:55 AM	07/06/2020	0.094	8:45 AM	06/25/2020	16080069
186	0.111	2:57 AM	07/06/2020	0.094	8:45 AM	06/25/2020	16080069
512	0.076	1:10 AM	06/26/2020	0.099	8:46 AM	06/25/2020	16080071
513	0.138	3:29 AM	06/26/2020	0.099	8:46 AM	06/25/2020	16080071
514	0.144	3:30 AM	06/26/2020	0.099	8:46 AM	06/25/2020	16080071
75	0.196	4:55 PM	06/27/2020	0.1	9:27 AM	06/25/2020	16080072
76	0.212	4:57 PM	06/27/2020	0.1	9:27 AM	06/25/2020	16080072
158	0.222	6:08 AM	07/05/2020	0.104	10:09 AM	06/25/2020	16120175
159	0.213	6:12 AM	07/05/2020	0.104	10:09 AM	06/25/2020	16120175
160	0.244	10:16 PM	07/05/2020	0.104	10:09 AM	06/25/2020	16120175
161	0.249	10:18 PM	07/05/2020	0.104	10:09 AM	06/25/2020	16120175
214	0.113	2:40 AM	06/27/2020	0.102	10:17 AM	06/25/2020	16120177
215	0.132	2:42 AM	06/27/2020	0.102	10:17 AM	06/25/2020	16120177
216	0.144	1:48 AM	07/04/2020	0.102	10:17 AM	06/25/2020	16120177
217	0.143	1:49 AM	07/04/2020	0.102	10:17 AM	06/25/2020	16120177
218	0.186	10:21 PM	07/04/2020	0.102	10:17 AM	06/25/2020	16120177
219	0.191	10:23 PM	07/04/2020	0.102	10:17 AM	06/25/2020	16120177
220	0.098	1:02 AM	07/05/2020	0.102	10:17 AM	06/25/2020	16120177
221	0.101	1:04 AM	07/05/2020	0.102	10:17 AM	06/25/2020	16120177
185	0.075	9:18 PM	06/29/2020	0.103	11:57 AM	06/25/2020	17170064
186	0.08	9:21 PM	06/29/2020	0.103	11:57 AM	06/25/2020	17170064

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)				
07/06/2020	1254	34.0°	19IP41102	.100	33.3	.104			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09/17/19										
			Expiration Date:										
			09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2020	1143	34.0°	19IP41102	.100	27.6		102		X			E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16040141
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2020	1259	34.0°	19IP41102	23.1		105		X			E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09/17/19										
			Expiration Date:										
			09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. 1640161
--------------------------------	-----------------	-------------------------------------	----------------------	-------------------------------	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2020	1305	34.0°	19IP41102	.100	23.3	.100			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09/17/19										
			Expiration Date:										
			09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

6

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16090163
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
07/06/2020		34.0°	19IP41102	.100							E. Mc Elroy	16193	OUT OF SERVICE	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
07/06/2020	1307	34.0°		19IP41102	.100	23.8	1104			X	E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09/17/19										
			Expiration Date:										
			09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME		AREA NO.	COORDINATOR NAME		I.D. NUMBER	PAS MAKE/MODEL		PAS SERIAL NO.						
South Los Angeles		530	Edward Mc Elroy		16193	LifeLoc FC20		16080014						
DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	TEST #1	TEST #2	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2020	1315	34.0°	19IP41102		24.6			097		X		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

9

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.	PAS ACCURACY CHECK RESULTS*			CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEMP (°C)	YES	NO				
South Los Angeles	530	Edward Mc Elroy	16193	Lifloc FC20	16080015									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16080027
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2020	1318	34.0°	19IP41102	.100	24.1			0.095			X	E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 160800 28
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)		SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
		TEMP #1	TEMP #2	LOT NO	VALUE	TEST #1	TEST #2	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2020	1321	34.0°	19IP41102	.100	27.1	.101	X						E. Mc Elroy	16193	
			Solution prepared by:												
			Cesar Munoz												
			Date: 09/17/19												
			Expiration Date:												
			09/17/21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080030
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
07/06/2020	1314	34.0°		19IP41102	.100	23.6			0.95	X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2020	1326	34.0°	19IP41102	23.8	.101			X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)		LOT NO.		SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	YES	NO				
07/06/2020	1335	34.0°		19IP41102				23.6	100		.100	X			E. Mc Elroy	16193	
							Solution prepared by:										
							Cesar Munoz										
							Date: 09/17/19										
							Expiration Date:										
							09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080069
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2020	1339	34.0°	19IP41102	.100	24.1	.096			X			E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.				
South Los Angeles		530	Edward Mc Elroy	16193	Lifoloc FC20	16080071				
PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.										
DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)				
07/06/2020	1340	34.0°	19IP41102	TEST #1 23.8	TEST #2 (Optional)	TEST #1 .096	TEST #2 X	E. Mc Elroy	16193	
			Solution prepared by:							
			Cesar Munoz							
			Date: 09/17/19							
			Expiration Date:							
			09/17/21							

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. 1608007a
--------------------------------	-----------------	-------------------------------------	----------------------	-------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2020	1345	34.0°	19IP41102	.100	23.8	.101				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080075
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2	YES				
07/06/2020	1347	34.0°		19IP41102									E. Mc Elroy	16193	
					.100	23.6	100				X				
			Solution prepared by:												
			Cesar Munoz												
			Date: 09/17/19												
			Expiration Date:												
			09/17/21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080077
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
07/06/2020	1143	34.0°		19IP41102	29.6	.110			X			E. Mc Elroy	16193		
			Solution prepared by: Cesar Munoz Date: 09/17/19 Expiration Date: 09/17/21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16120175
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
07/06/2020	1347	34.0°		19IP41102	27.1	.102			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09/17/19										
			Expiration Date:										
			09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2020	1401	34.0°	19IP41102		.100	24.1	.101			X		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120188
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2020	1402	34.0°	19IP41102		TEST #1 24.1	TEST #2 :109				X		E. Mc Elroy	16193	
				Solution prepared by:										
				Cesar Munoz										
				Date: 09/17/19										
				Expiration Date:										
				09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS			
			TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)					TEST #2 (Optional)	YES	NO
07/06/2020	1405	34.0	19IP41102		24.3	101	X				E. Mc Elroy	16193				

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2020	1405	34.0	19IP41102	.100	13.3	1096		X			E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09/17/19										
			Expiration Date:										
			09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2020	1408	34.0°	19IP41102	.100	23.8	102				X		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09/17/19											
			Expiration Date:											
			09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
07/06/2020		34.0°	19IP41102		.100							E. Mc Elroy	16193	OUT OF SERVICE
				Solution prepared by:										
				Cesar Munoz										
				Date: 09/17/19										
				Expiration Date:										
				09/17/21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170066
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2020	1420	34.0°	19IP41102	.100	23.1	099				X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz											
			Date: 09/17/19											
			Expiration Date: 09/17/21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
07/06/2020		34.0°	19IP41102							E. Mc Elroy	16193	OUT OF SERVICE
			Solution prepared by:									
			Cesar Munoz									
			Date: 09/17/19									
			Expiration Date:									
			09/17/21									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

29

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170155
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
07/06/2020	1422	34.0°	19IP41102	.100	23.3	.100			X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 90102844-82
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)		LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2020		34.0°		19IP41102	.100								E. Mc Elroy	16193	OUT of SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2020	1431	34.0°	19IP41102	.100	23.8	.103	X				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.