

CHP SLA PAS Usage Logs 08-27-2020

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
172	0.399		10:56 AM	08/20/2020	0.099	12:02 PM	08/17/2020	16040119
173	0.378		10:59 AM	08/20/2020	0.099	12:02 PM	08/17/2020	16040119
174	0.158		2:30 AM	08/25/2020	0.099	12:02 PM	08/17/2020	16040119
175	0.152		2:33 AM	08/25/2020	0.099	12:02 PM	08/17/2020	16040119
188	0		3:27 AM	08/19/2020	0.096	11:59 AM	08/17/2020	16040140
189	0.201		3:13 AM	08/23/2020	0.096	11:59 AM	08/17/2020	16040140
190	0.193		3:17 AM	08/23/2020	0.096	11:59 AM	08/17/2020	16040140
191	0.017		12:56 AM	08/26/2020	0.096	11:59 AM	08/17/2020	16040140
147	0.043		4:31 PM	08/18/2020	0.103	12:01 PM	08/17/2020	16040141
148	0.043		4:34 PM	08/18/2020	0.103	12:01 PM	08/17/2020	16040141
149	0.182		10:02 PM	08/22/2020	0.103	12:01 PM	08/17/2020	16040141
150	0.177		10:04 PM	08/22/2020	0.103	12:01 PM	08/17/2020	16040141
151	0		10:09 PM	08/25/2020	0.103	12:01 PM	08/17/2020	16040141
361	0		1:30 PM	08/18/2020	0.097	12:10 PM	08/17/2020	16040161
362	0		1:30 PM	08/18/2020	0.097	12:10 PM	08/17/2020	16040161
488	0.14		1:40 AM	08/17/2020	0.099	1:51 PM	08/06/2020	16040162
489	0.141		1:43 AM	08/17/2020	0.099	1:51 PM	08/06/2020	16040162
490	0		7:05 PM	08/17/2020	0.101	11:54 AM	08/17/2020	16040162
491	0.083		6:10 PM	08/23/2020	0.101	11:54 AM	08/17/2020	16040162
492	0.081		6:13 PM	08/23/2020	0.101	11:54 AM	08/17/2020	16040162
493	0.11		1:35 AM	08/26/2020	0.101	11:54 AM	08/17/2020	16040162
494	0.114		1:37 AM	08/26/2020	0.101	11:54 AM	08/17/2020	16040162
204	0.148		10:34 PM	08/19/2020	0.102	12:13 PM	08/17/2020	16080013
142	0.035		10:23 PM	08/20/2020	0.099	12:12 PM	08/17/2020	16080015
143	0.138		4:22 PM	08/23/2020	0.099	12:12 PM	08/17/2020	16080015
144	0.125		4:25 PM	08/23/2020	0.099	12:12 PM	08/17/2020	16080015
186	0.077		10:20 PM	08/21/2020	0.093	12:19 PM	08/17/2020	16080027
187	0.074		10:22 PM	08/21/2020	0.093	12:19 PM	08/17/2020	16080027
130	0.13		12:09 AM	08/26/2020	0.098	12:23 PM	08/17/2020	16080028
131	0.139		12:12 AM	08/26/2020	0.098	12:23 PM	08/17/2020	16080028
112	0		2:49 PM	08/20/2020	0.094	12:24 PM	08/17/2020	16080030
113	0		2:49 PM	08/20/2020	0.094	12:24 PM	08/17/2020	16080030
531	0		1:55 AM	08/17/2020	0.098	2:21 PM	08/06/2020	16080071

CHP SLA PAS Usage Logs 08-27-2020

532	0	1:57 AM	08/17/2020	0.098	2:21 PM	08/06/2020	16080071
77	0.202	4:52 PM	08/19/2020	0.097	12:39 PM	08/17/2020	16080072
78	0.216	4:54 PM	08/19/2020	0.097	12:39 PM	08/17/2020	16080072
79	0.262	8:44 PM	08/22/2020	0.097	12:39 PM	08/17/2020	16080072
80	0.116	9:05 AM	08/23/2020	0.097	12:39 PM	08/17/2020	16080072
81	0.109	9:06 AM	08/23/2020	0.097	12:39 PM	08/17/2020	16080072
175	NEG	3:22 PM	08/24/2020	0.102	12:42 PM	08/17/2020	16080077
176	NEG	3:24 PM	08/24/2020	0.102	12:42 PM	08/17/2020	16080077
172	0.152	2:06 AM	08/25/2020	0.1	12:42 PM	08/17/2020	16120175
173	0.153	2:09 AM	08/25/2020	0.1	12:42 PM	08/17/2020	16120175
235	0.147	12:49 AM	08/22/2020	0.099	12:47 PM	08/17/2020	16120177
236	0.149	12:51 AM	08/22/2020	0.099	12:47 PM	08/17/2020	16120177

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Edward Mc Elroy
 I.D. NUMBER: 16193
 PAS MAKE/MODEL: Lifeloc FC20
 PAS SERIAL NO: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO	
08/27/2020	0915	34.0°	19IP41102	.100	23.3		098				X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifaloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
08/27/2020	0917	34.0°	19JP41102	.100	23.3	.098					X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Ifaloc FC20 PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
08/27/2020	0918	34.0°	19IP41102	.100	22.6		101		Y		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

AREA NAME	AREA NO.	COORDINATOR NAME	ID NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/27/2020	0936	34.0°	19P41102	.100	21.8	.100					E. Mc Elroy	16193	
Solution prepared by: Cesar Munoz													
Date: 09-17-19													
Expiration date: 09-17-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/27/2020	0928	34.0°	19TP41102	100	21.6		1095			X	E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration date: 09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

