CHP SLA PAS Usage Logs 11/16/20

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
202	0.209		12:54 AM	11/05/2020	0.097	10:05 AM	10/27/2020	16040119
203	0.197		12:57 AM	11/05/2020	0.097	10:05 AM	10/27/2020	16040119
204	0.134		8:00 AM	11/07/2020	0.099	1:41 PM	11/05/2020	16040119
205	0.138		8:02 AM	11/07/2020	0.099	1:41 PM	11/05/2020	16040119
204	0		11:35 PM	11/10/2020	0.098	1:34 PM	11/05/2020	16040140
205	0.036		5:59 PM	11/14/2020	0.098	1:34 PM	11/05/2020	16040140
206	0.036		6:01 PM	11/14/2020	0.098	1:34 PM	11/05/2020	16040140
379	0.164		11:58 PM	11/13/2020	0.099	1:46 PM	11/05/2020	16040161
380	0.099		3:12 AM	11/15/2020	0.099	1:46 PM	11/05/2020	16040161
381	0.097		3:15 AM	11/15/2020	0.099	1:46 PM	11/05/2020	16040161
382	0.143		4:07 AM	11/16/2020	0.099	1:46 PM	11/05/2020	16040161
383	0.137		4:09 AM	11/16/2020	0.099	1:46 PM	11/05/2020	16040161
504	0.081		11:27 PM	11/07/2020	0.102	1:29 PM	11/05/2020	16040162
505	0.077		11:30 PM	11/07/2020	0.102	1:29 PM	11/05/2020	16040162
506	0.202		1:50 AM	11/10/2020	0.102	1:29 PM	11/05/2020	16040162
507	0.192		1:53 AM	11/10/2020	0.102	1:29 PM	11/05/2020	16040162
508	0.247		3:18 AM	11/15/2020	0.102	1:29 PM	11/05/2020	16040162
509	0.236		3:20 AM	11/15/2020	0.102	1:29 PM	11/05/2020	16040162
200	0.098		11:50 PM	11/12/2020	0.096	2:52 PM	11/05/2020	16080027
201	0.097		11:52 PM	11/12/2020	0.096	2:52 PM	11/05/2020	16080027
207	0.158		11:52 PM	11/13/2020	0.099	3:07 PM	11/05/2020	16080069
208	0.164		11:55 PM	11/13/2020	0.099	3:07 PM	11/05/2020	16080069
548	0		5:02 AM	11/06/2020	0.102	3:10 PM	11/05/2020	16080071
549	0.205		11:46 PM	11/06/2020	0.102	3:10 PM	11/05/2020	16080071
550	0.194		11:48 PM	11/06/2020	0.102	3:10 PM	11/05/2020	16080071
551	0.194		10:42 PM	11/08/2020	0.102	3:10 PM	11/05/2020	16080071
552	0.191		10:44 PM	11/08/2020	0.102	3:10 PM	11/05/2020	16080071
553	0.158		12:16 AM	11/16/2020	0.102	3:10 PM	11/05/2020	16080071
554	0.157		12:18 AM	11/16/2020	0.102	3:10 PM	11/05/2020	16080071
259	0.143		3:55 AM	11/06/2020	0.103	3:21 PM	11/05/2020	16120177
260	0.132		3:58 AM	11/06/2020	0.103	3:21 PM	11/05/2020	16120177
261	0.076		5:21 PM	11/12/2020	0.103	3:21 PM	11/05/2020	16120177
262	0.072		5:24 PM	11/12/2020	0.103	3:21 PM	11/05/2020	16120177

CHP SLA PAS Usage Logs 11/16/20

263	0.169	10:28 PM	11/13/2020	0.103	3:21 PM	11/05/2020	16120177
264	0.176	10:30 PM	11/13/2020	0.103	3:21 PM	11/05/2020	16120177
197	0.164	6:36 AM	11/08/2020	0.104	3:24 PM	11/05/2020	16120188
198	0.186	6:39 AM	11/08/2020	0.104	3:24 PM	11/05/2020	16120188
199	0.199	6:41 AM	11/08/2020	0.104	3:24 PM	11/05/2020	16120188
200	0.135	11:50 PM	11/12/2020	0.104	3:24 PM	11/05/2020	16120188
201	0.131	11:52 PM	11/12/2020	0.104	3:24 PM	11/05/2020	16120188

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0807	34.0°	19IP41102	.100	22.1		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16040119

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0823	34.0°	19IP41102	.100	25.6		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0811	34.0°	19IP41102	.100	32.1		.097			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16040191

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0813	34.0°	19IP41102	.100	21.1		.103			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0827	34.0°	19IP41102	.100	23.6		.098			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

11 6

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020		34.0°	19IP41102	.100								E. Mc Elroy	16193	OUT OF SERVICE
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0815	34.0°	19IP41102	.100	21.3		.096			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0829	34.0°	19IP41102	.100	23.1		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0836	34.0°	19IP41102	.100	22.6		.101					E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080027

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0839	34.0°	19IP41102	.100	22.3		.094					E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080028

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	1157	34.0°	19IP41102	.100	29.3		.099			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0841	34.0°	19IP41102	.100	21.6		.094			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0844	34.0°	19IP41102	.100	22.1		.101			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0846	34.0°	19IP41102	.100	21.8		.105			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0849	34.0°	19IP41102	.100	21.8		.097			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0851	34.0°	19IP41102	.100	22.6		.097			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0853	34.0°	19IP41102	.100	21.3		.101			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080075

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0856	34.0°	19IP41102	.100	21.6		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	1608077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0901	34.0°	19IP41102	.100	22.6		.108			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	1423	34.0°	19IP41102	.100	26.6		.103			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0915	34.0°	19IP41102	.100	22.6		.097			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0910	34.0°	19IP41102	.100	22.8		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0923	34.0°	19IP41102	.100	22.3		.096			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0924	34.0°	19IP41102	.100	21.8		.096			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0919	34.0°	19IP41102	.100	23.6		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0927	34.0°	19IP41102	.100	21.8		.096			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/16/2020	0929	34.0°	19IP41102	.100	20.6		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	17170150

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0930	34.0°	19IP41102	.100	22.1		.105			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0932	34.0°	19IP41102	.100	21.3		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	90102844-82

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020		34.0°	19IP41102	.100								E. Mc Elroy	16193	OUT OF SERVICE
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
11/16/2020	0935	34.0°	19IP41102	.100	21.1		.100			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.