

CHP SLA PAS Usage Logs 12/04/20

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
210	0.198		1:45 AM	11/25/2020	0.1	8:23 AM	11/16/2020	16040119
211	0.185		1:47 AM	11/25/2020	0.1	8:23 AM	11/16/2020	16040119
212	0.332		12:23 AM	11/29/2020	0.095	7:29 AM	11/25/2020	16040119
213	0.302		12:25 AM	11/29/2020	0.095	7:29 AM	11/25/2020	16040119
214	0.295		12:27 AM	11/29/2020	0.095	7:29 AM	11/25/2020	16040119
215	0.11		4:45 AM	11/29/2020	0.095	7:29 AM	11/25/2020	16040119
216	0.01		11:03 PM	12/02/2020	0.095	7:29 AM	11/25/2020	16040119
208	0.14		2:54 AM	12/02/2020	0.096	7:32 AM	11/25/2020	16040140
209	0.139		2:57 AM	12/02/2020	0.096	7:32 AM	11/25/2020	16040140
386	0.081		4:29 AM	11/29/2020	0.096	7:36 AM	11/25/2020	16040161
387	0.08		4:33 AM	11/29/2020	0.096	7:36 AM	11/25/2020	16040161
230	0.086		8:44 AM	11/28/2020	0.095	7:39 AM	11/25/2020	16080013
231	0.077		8:49 AM	11/28/2020	0.095	7:39 AM	11/25/2020	16080013
232	0		10:16 PM	11/30/2020	0.095	7:39 AM	11/25/2020	16080013
233	0		10:17 PM	11/30/2020	0.095	7:39 AM	11/25/2020	16080013
234	0.083		8:49 PM	12/02/2020	0.095	7:39 AM	11/25/2020	16080013
235	0.073		8:52 PM	12/02/2020	0.095	7:39 AM	11/25/2020	16080013
236	0.104		4:00 PM	12/03/2020	0.095	7:39 AM	11/25/2020	16080013
237	0.085		4:04 PM	12/03/2020	0.095	7:39 AM	11/25/2020	16080013
238	0		7:44 PM	12/03/2020	0.095	7:39 AM	11/25/2020	16080013
138	0.164		1:04 AM	11/28/2020	0.097	8:35 AM	11/25/2020	16080028
139	0.156		1:06 AM	11/28/2020	0.097	8:35 AM	11/25/2020	16080028
140	0.118		1:22 AM	11/30/2020	0.097	8:35 AM	11/25/2020	16080028
141	0.115		1:24 AM	11/30/2020	0.097	8:35 AM	11/25/2020	16080028
218	0.14		4:02 PM	11/29/2020	0.098	8:02 AM	11/25/2020	16080069
219	0.126		4:05 PM	11/29/2020	0.098	8:02 AM	11/25/2020	16080069
557	0		1:26 AM	12/04/2020	0.095	8:04 AM	11/25/2020	16080071
58	0.244		1:13 AM	12/02/2020	0.098	8:08 AM	11/25/2020	16080075
59	0.212		1:16 AM	12/02/2020	0.098	8:08 AM	11/25/2020	16080075
60	0.228		1:18 AM	12/02/2020	0.098	8:08 AM	11/25/2020	16080075
191	0.266		5:02 PM	11/29/2020	0.1	8:11 AM	11/25/2020	16120175
192	0.272		5:05 PM	11/29/2020	0.1	8:11 AM	11/25/2020	16120175
193	0.063		3:23 AM	12/01/2020	0.1	8:11 AM	11/25/2020	16120175

CHP SLA PAS Usage Logs 12/04/20

194	0.063	3:25 AM	12/01/2020	0.1	8:11 AM	11/25/2020	16120175
208	0.297	1:05 AM	11/25/2020	0.1	9:18 AM	11/16/2020	16120188
209	0.296	1:07 AM	11/25/2020	0.1	9:18 AM	11/16/2020	16120188

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	1149	34.0°	19IP41101	.100	19.8	1097			X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040141
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			TEMP	LOT NO.	TEST #1	TEST #2	TEST #1	TEST #2	TEST #1	TEST #2	YES					NO
			(°C)		(Optional)	(Optional)	(Optional)	(Optional)	YES	NO	YES					NO
12/04/2020	1153	34.0°	19IP41101		19.1)			102			X		E. Mc Elroy	16193		

* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	-------------------------------------	----------------------	-------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2020	1157	34.0°	19IP41101		19.1	1097			X		E. Mc Elroy	16193		
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2020		34.0°	19IP41101		.100						E. Mc Elroy	16193	OUT OF SERVICE
			Prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifec FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	-------------------------------------	---------------------	------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	12:00	34.0°	19IP41101	.100	19.8	.097			X		E. Mc Elroy	16193	
			Prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080014
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
12/04/2020	12:02	34.0	19IP41101	19IP41101	21.3	1.00	1.099		X		E. Mc Elroy	16193	
			Prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080015
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2020	1203	34.0°	191P41101		.100	20.6	.101		X			E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080027
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAC CHECK RESULTS*		CALIBRATION REQUIRED		PAC RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2020	12:05	34.0°	19IP41101	20.1		096			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

||

AREA NAME South Los Angeles		AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080028
--------------------------------	--	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	12:09	34.0°	19IP41101	.100	20.8	094		X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. 16080030
--------------------------------	-----------------	-------------------------------------	----------------------	-------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	12:10	34.0	19IP41101	.100	19.6	.094	X				E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16080064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/04/2020	1213	34.0°	19IP41101		.100	20.3		.099				X	16193	E. Mc Elroy	
				Prepared by:											
				Cesar Munoz											
				Date: 09-10-19											
				Expiration Date:											
				09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
12/04/2020	12:15	34.0°	19P41101		.100	19.8		104		X		E. Mc Elroy	16193	
				Prepared by:										
				Cesar Munoz										
				Date: 09-10-19										
				Expiration Date:										
				09-10-21										

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/04/2020	1:30	34.0°	19TP41101		22.3	.100	.095		X		E. Mc Elroy	16193			
			Prepared by:												
			Cesar Munoz												
			Date: 09-10-19												
			Expiration Date:												
			09-10-21												

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME Edward Mc Elroy I.D. NUMBER 16193 PAS MAKE/MODEL LifeLoc FC20 PAS SERIAL NO. 1608007

#16

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2020	1243	34.0°	19IP41101	.100	TEST #1 21.8	TEST #2 10.95				X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.							
South Los Angeles		530	Edward Mc Elroy	16193	Lifloc FC20	16080072							
PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.													
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/04/2020	1245	34.0°	19IP41101	TEST #1	TEST #2	TEST #1	TEST #2	YES	NO		E. Mc Elroy	16193	
		Prepared by:											
		Cesar Munoz											
		Date: 09-10-19											
		Expiration Date:											
		09-10-21											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. 16080075
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2020	1246	34.0	19IP41101	.100	20.3	.098	X					E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080077
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	1248	34.0°	19IP41101		.100	20.8	108				X	E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifec FC20	PAS SERIAL NO. 16120175
--------------------------------	-----------------	-------------------------------------	----------------------	------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	12:49	34.0°	19IP41101	.100	29.3	.101				X	E. Mc Elroy	16193	
			Prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. 16120177
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	PAS					
										YES					NO
12/04/2020	12:51	34.0°	19IP41101		.100	21.1		101		X		E. Mc Elroy	16193		
				Prepared by:											
				Cesar Munoz											
				Date: 09-10-19											
				Expiration Date:											
				09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc FC20	16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	12:54	34.0°	19IP41101	.100	21.1	101		X			E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
12/04/2020	12:56	34.0°	191P41101		.100	20.3	098		X		E. Mc Elroy	16193	
			Prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	1258	34.0°	19IP41101	.100	20.6	20.6	.096			X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	1300	34.0°	19IP41101	.100	20.0	100				X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME		AREA NO.	COORDINATOR NAME		I.D. NUMBER	PAS MAKE/MODEL		PAS SERIAL NO.					
South Los Angeles		530	Edward Mc Elroy		16193	Lifeloc FC20		16120197					
PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.													
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	1303	34.0°	19IP41101	.100	20.1	1.098			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifloc FC20	17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2020	1305	34.0	19P41101		.100	19.1	.101			X		E. Mc Elroy	16193	
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2020		34.0°	19IP41101		.100						E. Mc Elroy	16193	OUT OF SERVICE
			Prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 17170155
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/04/2020	1308	34.0°	19IP41101	.100	20.6	.098		X			E. Mc Elroy	16193	
			Prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 90102844-82
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2020		34.0°	19IP41101		.100							E. Mc Elroy	16193	OUT OF SERVICE
			Prepared by:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date:											
			09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifetec FC20	PAS SERIAL NO. 17170064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/04/2020	1310	34.0°	19IP41101	.100	20.3	101				X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.