

CHP SLA PAS Usage Logs 12/14/2020

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
217	0.114		3:13 AM	12/09/2020	0.097	11:49 AM	12/04/2020	16040119
218	0.114		3:16 AM	12/09/2020	0.097	11:49 AM	12/04/2020	16040119
210	0.122		6:25 PM	12/05/2020	0.096	11:52 AM	12/04/2020	16040140
211	0.125		6:26 PM	12/05/2020	0.096	11:52 AM	12/04/2020	16040140
212	0		4:13 PM	12/06/2020	0.096	11:52 AM	12/04/2020	16040140
213	0		10:29 PM	12/09/2020	0.096	11:52 AM	12/04/2020	16040140
388	0.191		10:11 PM	12/05/2020	0.097	11:57 AM	12/04/2020	16040161
389	0.174		10:14 PM	12/05/2020	0.097	11:57 AM	12/04/2020	16040161
390	0.081		12:46 AM	12/06/2020	0.097	11:57 AM	12/04/2020	16040161
391	0.09		12:49 AM	12/06/2020	0.097	11:57 AM	12/04/2020	16040161
392	0.316		3:31 AM	12/12/2020	0.097	11:57 AM	12/04/2020	16040161
393	0.312		3:34 AM	12/12/2020	0.097	11:57 AM	12/04/2020	16040161
239	0.05		1:42 AM	12/07/2020	0.097	12:00 PM	12/04/2020	16080013
240	0.21		3:00 AM	12/13/2020	0.097	12:00 PM	12/04/2020	16080013
241	0.223		3:03 AM	12/13/2020	0.097	12:00 PM	12/04/2020	16080013
142	0.146		10:50 PM	12/05/2020	0.094	12:10 PM	12/04/2020	16080028
143	0.156		10:53 PM	12/05/2020	0.094	12:10 PM	12/04/2020	16080028
144	0.092		12:41 AM	12/12/2020	0.094	12:10 PM	12/04/2020	16080028
145	0.089		12:43 AM	12/12/2020	0.094	12:10 PM	12/04/2020	16080028
120	0.157		12:22 AM	12/13/2020	0.094	12:11 PM	12/04/2020	16080030
121	0.128		12:26 AM	12/13/2020	0.094	12:11 PM	12/04/2020	16080030
122	0.13		12:28 AM	12/13/2020	0.094	12:11 PM	12/04/2020	16080030
378	0		10:44 PM	12/04/2020	0.104	12:15 PM	12/04/2020	16080065
379	0.132		2:10 AM	12/06/2020	0.104	12:15 PM	12/04/2020	16080065
380	0.127		2:14 AM	12/06/2020	0.104	12:15 PM	12/04/2020	16080065
220	0.186		10:48 AM	12/05/2020	0.095	12:31 PM	12/04/2020	16080069
221	0.196		10:50 AM	12/05/2020	0.095	12:31 PM	12/04/2020	16080069
222	0.032		9:36 PM	12/09/2020	0.095	12:31 PM	12/04/2020	16080069
223	0.055		1:03 AM	12/10/2020	0.095	12:31 PM	12/04/2020	16080069
224	0.055		1:06 AM	12/10/2020	0.095	12:31 PM	12/04/2020	16080069
557	0		1:26 AM	12/04/2020	0.095	8:04 AM	11/25/2020	16080071
558	0.066		11:43 PM	12/04/2020	0.095	12:43 PM	12/04/2020	16080071
559	0.068		11:47 PM	12/04/2020	0.095	12:43 PM	12/04/2020	16080071

CHP SLA PAS Usage Logs 12/14/2020

560	0.24	1:33 AM	12/12/2020	0.095	12:43 PM	12/04/2020	16080071
561	0.22	1:37 AM	12/12/2020	0.095	12:43 PM	12/04/2020	16080071
562	0.018	9:44 PM	12/12/2020	0.095	12:43 PM	12/04/2020	16080071
563	0.113	11:48 PM	12/12/2020	0.095	12:43 PM	12/04/2020	16080071
564	0.108	11:50 PM	12/12/2020	0.095	12:43 PM	12/04/2020	16080071
565	0.01	11:54 PM	12/12/2020	0.095	12:43 PM	12/04/2020	16080071
566	0.139	2:17 AM	12/13/2020	0.095	12:43 PM	12/04/2020	16080071
567	0.13	2:20 AM	12/13/2020	0.095	12:43 PM	12/04/2020	16080071
568	0	4:57 AM	12/13/2020	0.095	12:43 PM	12/04/2020	16080071
569	0	4:58 AM	12/13/2020	0.095	12:43 PM	12/04/2020	16080071
570	0.169	7:41 PM	12/13/2020	0.095	12:43 PM	12/04/2020	16080071
571	0.162	7:44 PM	12/13/2020	0.095	12:43 PM	12/04/2020	16080071
265	0.193	12:05 AM	12/06/2020	0.101	12:52 PM	12/04/2020	16120177
266	0.201	12:07 AM	12/06/2020	0.101	12:52 PM	12/04/2020	16120177
267	0	3:15 PM	12/12/2020	0.101	12:52 PM	12/04/2020	16120177
210	0.068	3:40 AM	12/05/2020	0.101	12:55 PM	12/04/2020	16120188
211	0.07	3:43 AM	12/05/2020	0.101	12:55 PM	12/04/2020	16120188
212	0	3:27 AM	12/06/2020	0.101	12:55 PM	12/04/2020	16120188
213	0.138	3:57 AM	12/06/2020	0.101	12:55 PM	12/04/2020	16120188
214	0.132	3:58 AM	12/06/2020	0.101	12:55 PM	12/04/2020	16120188
353	0.15	1:08 AM	12/12/2020	0.096	12:59 PM	12/04/2020	16120194
354	0.161	1:10 AM	12/12/2020	0.096	12:59 PM	12/04/2020	16120194
355	0.207	2:23 AM	12/14/2020	0.096	12:59 PM	12/04/2020	16120194
356	0.213	2:26 AM	12/14/2020	0.096	12:59 PM	12/04/2020	16120194
368	0.138	1:29 AM	12/10/2020	0.098	1:03 PM	12/04/2020	16120197
369	0.148	1:32 AM	12/10/2020	0.098	1:03 PM	12/04/2020	16120197
370	0.083	2:44 AM	12/10/2020	0.098	1:03 PM	12/04/2020	16120197
371	0.084	2:47 AM	12/10/2020	0.098	1:03 PM	12/04/2020	16120197
372	0.039	11:43 PM	12/12/2020	0.098	1:03 PM	12/04/2020	16120197
373	0	5:01 AM	12/13/2020	0.098	1:03 PM	12/04/2020	16120197
374	0.132	5:28 PM	12/13/2020	0.098	1:03 PM	12/04/2020	16120197
375	0.142	5:31 PM	12/13/2020	0.098	1:03 PM	12/04/2020	16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

#1

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0737	34.0°	19IP41101	.100	20.6		098			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.0990 to 0.1100).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
								TEST #1 (Optional)	TEST #2 (Optional)					
12/14/2020	0741	34.0°		191P41101	.100	21.1				X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16040190

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO						
12/14/2020	0743	34.0°		19IP41101	.100	20.8		0.94				X	NO		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
16040141

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/14/2020	0745	34.0°	191P41101	.100	20.1		101			X		E. Mc Elroy	16193	
Prepared by:														
Cesar Munoz														
Date: 09-10-19														
Expiration Date: 09-10-21														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/14/2020	0747	34.0°	19IP41101	.100	19, 8	1.094				X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020		34.0°	191P41101	.100								E. Mc Elroy	16193	OUT OF SERVICE

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

CHP202K\_1108.pdf



AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Liteloc FC20    PAS SERIAL NO.: 16080013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)		LOT NO	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO			TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
12/14/2020	0751	34.0°	191P41101		.100	20.6		091				X		E. Mc Elroy	16193	

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

#8

AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Edward Mc Elroy

I.D. NUMBER  
 16193

PAS MAKE/MODEL  
 Lifloc PC20

PAS SERIAL NO.  
 16080014

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0808	34.0°		19IP41101	.100	22.1		.097			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				

Prepared by:  
Cesar Munoz

Date: 09-10-19

Expiration Date:  
09-10-21

12/14/2020 0810

34.0° 191P41101

.100

21.1

.099

X

E. Mc Elroy

16193

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080027

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS			
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO							
12/14/2020	0812	34.0°		19TP41101	.100		2018			1095			X		E. Mc Elroy	16193		

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080030
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	YES	NO				
12/14/2020	0830	34.0°	191P41101		.100	30.3			091					E. Mc Elroy	16193	
Prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21																

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202K\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0823	34.0°		19IP41101	.100	20.8		0.97		X			E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 CHP202K\_1108.pdf

AREA NAME South Los Angeles      AREA NO. 530      COORDINATOR NAME Edward Mc Elroy      I.D. NUMBER 16193      PAS MAKE/MODEL Lifeloc FC20      PAS SERIAL NO. 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/14/2020	0825	34.0°	191P41101	.100	20.3		101		X		E. Mc Elroy	16193	
Prepared by: Cesar Munoz													
Date: 09-10-19													
Expiration Date: 09-10-21													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



AREA NAME  
South Los Angeles

AREA NO. 530  
COORDINATOR NAME Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifoloc FC20

PAS SERIAL NO.  
16086069

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/14/2020	0937	34.0°	19IP41101	.100	20.8		093				E. Mc Elroy	16193	

\* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Edward Mc Elroy

I.D. NUMBER  
 16193

PAS MAKE/MODEL  
 Lifeloc FC20

PAS SERIAL NO.  
 16080071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/14/2020	0828	34.0°	191P41101	.100	20.8		093			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME Edward Mc Elroy I.D. NUMBER 16193 PAS MAKE/MODEL Lifoloc FC20 PAS SERIAL NO. 16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0831	34.0°	19IP41101	.100	20.1		0.097			X		E. Mc Elroy	16193	
Prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#18

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    ID. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc PC20    PAS SERIAL NO.: 16080075

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0833	34.0°	19IP41101	.100	21.1	1096			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPl 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080077
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0835	34.0°	19IP41101	.100	20.8		105			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 CHP202K\_1108.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FCC20 PAS SERIAL NO.: 16120175

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0836	34.0°	19TP41101	.100	30.8		099			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf

AREA NAME **South Los Angeles** AREA NO. **530** COORDINATOR NAME **Edward Mc Elroy** I.D. NUMBER **16193** PAS MAKE/MODEL **Lifeloc FC20** PAS SERIAL NO. **16120177**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/14/2020	0839	34.0°	19IP41101	.100	21.1		1.00				X	E. Mc Elroy	16193	
Prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21														

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPl 051

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120188

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#22

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0843	34.0°		19IP41101	.100	33.1		0.96			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions      CHP202K\_1108.pdf



CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifeloc PC20	16120189

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020	0845	34.0°	19TP41101		1.00	21.1		1.095		Y	N		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16126194
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/14/2020	0852	34.0°	191P41101	.100	20.8		1.091			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#25

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO
12/14/2020	0854	34.0°	191P41101	.100	21.1		0.97			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/14/2020	0757	34.0°	19IP41101	.100	30.1		1.094			NO	X	E. Mc Elroy	16193	

Prepared by:  
Cesar Munoz  
Date: 09-10-19  
Expiration Date:  
09-10-21

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 CHP202K\_1.08.pdf

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc PC20  
 PAS SERIAL NO.: 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
12/14/2020	0857	34.0°	19IP41101	.100	19.8		099		X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
12/14/2020		34.0°	19IP41101	.100								E. Mc Elroy	16193	OUT OF SERVICE
Prepared by: Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: **South Los Angeles**    AREA NO.: **530**    COORDINATOR NAME: **Edward Mc Elroy**    ID. NUMBER: **16193**    PAS MAKE/MODEL: **Lifeloc FC20**    PAS SERIAL NO.: **17170155**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
									TEST #1	TEST #2					
12/14/2020	0900	34.0°	191P41101	100	20.3		1096				X		E. Mc Elroy	16193	

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FCC20      PAS SERIAL NO.: 90102844-82

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS			
12/14/2020		34.0°	19P41101	.100								E. Mc Elroy	16193	OUT OF SERVICE

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FCC20	PAS SERIAL NO. 17170064
--------------------------------	-----------------	-------------------------------------	----------------------	---------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	TEST #1 (Optional)		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
12/14/2020	0903	34.0°	19P41101		.100	19.6		1.098				X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.