

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
261	0.325		6:28 PM	05/28/2021	0.095	12:09 PM	05/28/2021	16040119
262	0.324		6:29 PM	05/28/2021	0.095	12:09 PM	05/28/2021	16040119
276	0.191		2:19 AM	05/29/2021	0.095	12:04 PM	05/28/2021	16040140
277	0.192		2:21 AM	05/29/2021	0.095	12:04 PM	05/28/2021	16040140
278	0.116		3:21 AM	05/30/2021	0.095	12:04 PM	05/28/2021	16040140
279	0.116		3:24 AM	05/30/2021	0.095	12:04 PM	05/28/2021	16040140
280	0.081		10:42 PM	06/05/2021	0.095	12:04 PM	05/28/2021	16040140
281	0.078		10:45 PM	06/05/2021	0.095	12:04 PM	05/28/2021	16040140
282	0.061		12:52 AM	06/06/2021	0.095	12:04 PM	05/28/2021	16040140
432	0.145		9:59 PM	05/30/2021	0.096	12:13 PM	05/28/2021	16040161
433	0.139		10:02 PM	05/30/2021	0.096	12:13 PM	05/28/2021	16040161
434	0.023		10:46 PM	05/31/2021	0.096	12:13 PM	05/28/2021	16040161
435	0.229		3:54 AM	06/05/2021	0.096	12:13 PM	05/28/2021	16040161
436	0.202		3:57 AM	06/05/2021	0.096	12:13 PM	05/28/2021	16040161
437	0.128		7:27 AM	06/06/2021	0.096	12:13 PM	05/28/2021	16040161
438	0.125		7:31 AM	06/06/2021	0.096	12:13 PM	05/28/2021	16040161
523	0.039		9:38 PM	05/30/2021	0.096	11:59 AM	05/28/2021	16040162
524	0.023		10:27 PM	05/30/2021	0.096	11:59 AM	05/28/2021	16040162
185	0.122		3:03 AM	05/30/2021	0.096	12:17 PM	05/28/2021	16080015
186	0.127		3:06 AM	05/30/2021	0.096	12:17 PM	05/28/2021	16080015
173	0.122		4:13 AM	05/29/2021	0.101	1:23 PM	05/28/2021	16080028
174	0.12		4:16 AM	05/29/2021	0.101	1:23 PM	05/28/2021	16080028
175	0.01		1:32 AM	06/06/2021	0.101	1:23 PM	05/28/2021	16080028
176	0		12:00 PM	06/07/2021	0.101	1:23 PM	05/28/2021	16080028
145	0.173		1:48 PM	05/30/2021	0.103	12:23 PM	05/28/2021	16080030
146	0.176		1:50 PM	05/30/2021	0.103	12:23 PM	05/28/2021	16080030
170	0.114		12:44 AM	06/02/2021	0.094	12:24 PM	05/28/2021	16080064
171	0.119		12:46 AM	06/02/2021	0.094	12:24 PM	05/28/2021	16080064
287	0.122		10:07 PM	05/31/2021	0.103	12:36 PM	05/28/2021	16080069
288	0.165		1:09 AM	06/01/2021	0.103	12:36 PM	05/28/2021	16080069
289	0.153		1:12 AM	06/01/2021	0.103	12:36 PM	05/28/2021	16080069
290	0.174		8:27 PM	06/05/2021	0.103	12:36 PM	05/28/2021	16080069
291	0.171		8:29 PM	06/05/2021	0.103	12:36 PM	05/28/2021	16080069

CHP SLA Usage Logs 06/07/2021

617	0.089	12:51 AM	05/30/2021	0.102	12:46 PM	05/28/2021	16080071
618	0.083	12:54 AM	05/30/2021	0.102	12:46 PM	05/28/2021	16080071
619	0.066	12:58 AM	05/30/2021	0.102	12:46 PM	05/28/2021	16080071
620	0.324	4:39 PM	05/31/2021	0.102	12:46 PM	05/28/2021	16080071
621	0.317	4:40 PM	05/31/2021	0.102	12:46 PM	05/28/2021	16080071
622	0.16	7:34 PM	06/03/2021	0.102	12:46 PM	05/28/2021	16080071
623	0.163	7:35 PM	06/03/2021	0.102	12:46 PM	05/28/2021	16080071
624	0.219	6:19 PM	06/04/2021	0.102	12:46 PM	05/28/2021	16080071
625	0.226	6:22 PM	06/04/2021	0.102	12:46 PM	05/28/2021	16080071
626	0	10:35 PM	06/06/2021	0.102	12:46 PM	05/28/2021	16080071
124	0.071	12:09 AM	06/05/2021	0.096	12:53 PM	05/28/2021	16080072
125	0.066	12:12 AM	06/05/2021	0.096	12:53 PM	05/28/2021	16080072
105	0.121	12:31 AM	05/31/2021	0.097	12:57 PM	05/28/2021	16080075
106	0.124	12:34 AM	05/31/2021	0.097	12:57 PM	05/28/2021	16080075
107	0.158	2:25 AM	05/31/2021	0.097	12:57 PM	05/28/2021	16080075
108	0.155	2:27 AM	05/31/2021	0.097	12:57 PM	05/28/2021	16080075
109	0.019	3:11 PM	05/31/2021	0.097	12:57 PM	05/28/2021	16080075
181	0.225	12:57 AM	05/30/2021	0.104	12:58 PM	05/28/2021	16080077
182	0.208	1:00 AM	05/30/2021	0.104	12:58 PM	05/28/2021	16080077
183	0.158	1:41 AM	06/04/2021	0.104	12:58 PM	05/28/2021	16080077
184	0.156	1:43 AM	06/04/2021	0.104	12:58 PM	05/28/2021	16080077
185	0.022	1:49 AM	06/04/2021	0.104	12:58 PM	05/28/2021	16080077
230	0.228	8:26 PM	05/31/2021	0.096	12:58 PM	05/28/2021	16120175
231	0.214	8:28 PM	05/31/2021	0.096	12:58 PM	05/28/2021	16120175
232	0	9:01 PM	06/05/2021	0.096	12:58 PM	05/28/2021	16120175
233	0.013	12:30 AM	06/07/2021	0.096	12:58 PM	05/28/2021	16120175
249	0.168	1:58 AM	05/29/2021	0.095	1:04 PM	05/28/2021	16120188
250	0.165	2:00 AM	05/29/2021	0.095	1:04 PM	05/28/2021	16120188
251	0.168	4:32 AM	05/29/2021	0.095	1:04 PM	05/28/2021	16120188
252	0.178	4:34 AM	05/29/2021	0.095	1:04 PM	05/28/2021	16120188
253	0	9:04 PM	05/30/2021	0.095	1:04 PM	05/28/2021	16120188
254	0.255	2:20 AM	06/01/2021	0.095	1:04 PM	05/28/2021	16120188
255	0.271	2:22 AM	06/01/2021	0.095	1:04 PM	05/28/2021	16120188
256	0	6:33 AM	06/03/2021	0.095	1:04 PM	05/28/2021	16120188

CHP SLA Usage Logs 06/07/2021

479	0.12	1:24 AM	05/29/2021	0.098	1:09 PM	05/28/2021	16120195
480	0.123	1:26 AM	05/29/2021	0.098	1:09 PM	05/28/2021	16120195
481	0	4:08 AM	05/31/2021	0.098	1:09 PM	05/28/2021	16120195
482	0	11:12 PM	05/31/2021	0.098	1:09 PM	05/28/2021	16120195
483	0.209	10:17 PM	06/02/2021	0.098	1:09 PM	05/28/2021	16120195
484	0.204	10:20 PM	06/02/2021	0.098	1:09 PM	05/28/2021	16120195
485	0.122	11:29 PM	06/02/2021	0.098	1:09 PM	05/28/2021	16120195
486	0.122	11:31 PM	06/02/2021	0.098	1:09 PM	05/28/2021	16120195
487	0.136	10:22 PM	06/04/2021	0.098	1:09 PM	05/28/2021	16120195
488	0.145	10:25 PM	06/04/2021	0.098	1:09 PM	05/28/2021	16120195
489	0.172	12:20 AM	06/05/2021	0.098	1:09 PM	05/28/2021	16120195
490	0.174	12:22 AM	06/05/2021	0.098	1:09 PM	05/28/2021	16120195
491	0.058	12:07 AM	06/06/2021	0.098	1:09 PM	05/28/2021	16120195
492	0.059	12:09 AM	06/06/2021	0.098	1:09 PM	05/28/2021	16120195
192	0.057	8:26 PM	05/30/2021	0.098	1:21 PM	05/28/2021	17170064

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO					
06/07/2021	1526	34.0°	19IP41102	.100	TEST #1 21.8	TEST #2 (Optional)	TEST #1 .099	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040141
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES					NO
06/07/2021	1536	34.0°	19IP41102	.100	TEST #1 21.1	TEST #2 (Optional)	TEST #1 1.097	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021	1543	34.0°	19IP41102	.100	22.6	100			X		E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080027
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021	1550	34.0°	19IP41102	.100	21.8	104				X	E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080028
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1					
06/07/2021	1640	34.0°	19IP41102	.100	24.1	102		X		E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz									
			Date: 09-17-19									
			Expiration Date: 09-17-21									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080030
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/07/2021	1552	34.0°	19IP41102		TEST #1 20.6	TEST #2 (Optional) 10.6				E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz									
			Date: 09-17-19									
			Expiration Date: 09-17-21									

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)									
06/07/2021	1554	34.0°	19IP41102	.100	21.1		1.097			X	E. Mc Elroy	16193		
			Solution Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			09-17-21											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021	1555	34.0°	19IP41102	.100			1099			X	E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#15

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080069
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021	1559	34.0°	19IP41102	.100	24.1	103			X		E. Mc Elroy	16193	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME		AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.									
South Los Angeles		530	Edward Mc Elroy	16193	Lifeloc FC20	16080072									
PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.															
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2021	1605	34.0°	19IP41102	.100	21.1		1096				X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifelog FC20	16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021	1624	34.0°	19IP41102	.100	21.1	1096				X	E. Mc Elroy	16193		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 17170066
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021	1629	34.0°	19IP41102	.100	2011		.101			X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#28

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Edward Mc Elroy	16193	Lifoloc FC20	17170150

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/07/2021		34.0°	19IP41102	.100							E. Mc Elroy	16193	OUT OF SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170155
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO					
06/07/2021	1631	34.0°	19IP41102	TEST #1 20.8	TEST #2 (Optional)	TEST #1 1.097	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	
		Solution Prepared by:											
		Cesar Munoz											
		Date: 09-17-19											
		Expiration Date:											
		09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 90102844-82
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/07/2021		34.0°	19IP41102	.100							E. Mc Elroy	16193	OUT OF SERVICE
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO					
06/07/2021	1632	34.0°	19IP41102	.100	TEST #1 20.6	TEST #2 (Optional)	TEST #1 1098	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.