

CHP SLA Usage Logs 03/31/2021

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
250	0.172		12:15 AM	03/31/2021	0.109	12:18 PM	03/21/2021	16040119
251	0.163		12:18 AM	03/31/2021	0.109	12:18 PM	03/21/2021	16040119
252	0.101		4:53 AM	03/24/2021	0.091	11:36 AM	03/21/2021	16040140
253	0.11		4:55 AM	03/24/2021	0.091	11:36 AM	03/21/2021	16040140
254	0.142		11:17 PM	03/25/2021	0.091	11:36 AM	03/21/2021	16040140
255	0.158		11:19 PM	03/25/2021	0.091	11:36 AM	03/21/2021	16040140
256	0.133		3:51 AM	03/26/2021	0.091	11:36 AM	03/21/2021	16040140
257	0.133		3:54 AM	03/26/2021	0.091	11:36 AM	03/21/2021	16040140
411	0.165		11:26 PM	03/21/2021	0.093	11:43 AM	03/21/2021	16040161
412	0.154		11:29 PM	03/21/2021	0.093	11:43 AM	03/21/2021	16040161
261	0		5:43 PM	03/21/2021	0.108	11:46 AM	03/21/2021	16080013
262	0		5:45 PM	03/21/2021	0.108	11:46 AM	03/21/2021	16080013
219	0		7:49 AM	03/22/2021	0.103	2:31 PM	03/21/2021	16080027
220	0.089		9:59 AM	03/26/2021	0.103	2:31 PM	03/21/2021	16080027
221	0.079		10:01 AM	03/26/2021	0.103	2:31 PM	03/21/2021	16080027
142	0.061		3:26 AM	03/27/2021	0.103	12:00 PM	03/21/2021	16080030
402	0.179		3:09 AM	03/21/2021	0.094	1:35 PM	03/11/2021	16080065
403	0.179		3:11 AM	03/21/2021	0.094	1:35 PM	03/11/2021	16080065
254	0.169		10:38 PM	03/25/2021	0.099	12:10 PM	03/21/2021	16080069
255	0.155		10:41 PM	03/25/2021	0.099	12:10 PM	03/21/2021	16080069
80	0		2:07 AM	03/21/2021	0.092	1:45 PM	03/11/2021	16080075
81	0.069		3:23 AM	03/27/2021	0.095	12:22 PM	03/21/2021	16080075
82	0.067		3:27 AM	03/27/2021	0.095	12:22 PM	03/21/2021	16080075
83	0.173		11:22 PM	03/28/2021	0.095	12:22 PM	03/21/2021	16080075
84	0.18		11:24 PM	03/28/2021	0.095	12:22 PM	03/21/2021	16080075
85	0		2:52 AM	03/29/2021	0.095	12:22 PM	03/21/2021	16080075
86	0.026		1:41 AM	03/31/2021	0.095	12:22 PM	03/21/2021	16080075
216	0		6:58 PM	03/24/2021	0.097	12:24 PM	03/21/2021	16120175
217	0.012		4:40 AM	03/28/2021	0.097	12:24 PM	03/21/2021	16120175
218	0.014		4:43 AM	03/28/2021	0.097	12:24 PM	03/21/2021	16120175
295	0.17		5:21 PM	03/22/2021	0.098	12:28 PM	03/21/2021	16120177
296	0.159		5:24 PM	03/22/2021	0.098	12:28 PM	03/21/2021	16120177
297	0		4:55 AM	03/31/2021	0.098	12:28 PM	03/21/2021	16120177

CHP SLA Usage Logs 03/31/2021

229	0.11	11:17 PM	03/25/2021	0.096	12:30 PM	03/21/2021	16120188
230	0.108	11:19 PM	03/25/2021	0.096	12:30 PM	03/21/2021	16120188
231	0.189	10:45 PM	03/26/2021	0.096	12:30 PM	03/21/2021	16120188
232	0.174	10:47 PM	03/26/2021	0.096	12:30 PM	03/21/2021	16120188
233	0.106	4:09 AM	03/27/2021	0.096	12:30 PM	03/21/2021	16120188
234	0.111	4:11 AM	03/27/2021	0.096	12:30 PM	03/21/2021	16120188
235	0.06	10:27 PM	03/30/2021	0.096	12:30 PM	03/21/2021	16120188
236	0.1	11:41 PM	03/30/2021	0.096	12:30 PM	03/21/2021	16120188
237	0.105	11:44 PM	03/30/2021	0.096	12:30 PM	03/21/2021	16120188

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEMP (°C)	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/31/2021		34.0°	19IP41101		.100								E. Mc Elroy	16193	OUT OF SERVICE
				Solution prepared by: Cesar Munoz											
				Date: 09-10-19											
				Expiration Date: 09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/31/2021	1850	34.0°	19IP41101	.100	23.8	.102					E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/31/2021	1855	34.0°	19IP41101	25.1		.101		X			E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		PAS CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/31/2021	1858	34.0°	19IP41101	.100	24.3	.100			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-10-19										
			Expiration Date:										
			09-10-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 17170066
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
	TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
03/31/2021	34.0°	19IP41101							E. Mc Elroy	16193	out of service

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/31/2021	1645	34.0°	19IP41101	.100	TEST #1 25.8	TEST #2 (Optional)	TEST #1 .114	TEST #2 (Optional)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	.099	E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz											
			Date: 09-10-19											
			Expiration Date: 09-10-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OP1 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170155
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
				VALUE	TEST #2 (Optional)	YES	NO					
03/31/2021	1902	34.0°	19IP41101	23.3	.113	X		.098	E. Mc Elroy	16193		
			Solution prepared by:									
			Cesar Munoz									
			Date: 09-10-19									
			Expiration Date:									
			09-10-21									

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions
Chp202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 90102844-82
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/31/2021		34.0°	19IP41101	.100							E. Mc Elroy	16193	OUT OF SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

