

CHP SLA Usage Logs 04/09/2021

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
250	0.172		12:15 AM	03/31/2021	0.109	12:18 PM	03/21/2021	16040119
251	0.163		12:18 AM	03/31/2021	0.109	12:18 PM	03/21/2021	16040119
252	0.169		12:56 AM	04/03/2021	0.096	6:03 PM	03/31/2021	16040119
253	0.153		12:59 AM	04/03/2021	0.096	6:03 PM	03/31/2021	16040119
254	0.082		5:51 PM	04/04/2021	0.096	6:03 PM	03/31/2021	16040119
255	0.082		5:53 PM	04/04/2021	0.096	6:03 PM	03/31/2021	16040119
258	0.109		12:00 PM	04/02/2021	0.095	6:53 PM	03/31/2021	16040140
259	0.102		12:03 PM	04/02/2021	0.095	6:53 PM	03/31/2021	16040140
222	0.079		12:59 AM	04/04/2021	0.107	6:26 PM	03/31/2021	16080027
223	0.078		1:02 AM	04/04/2021	0.107	6:26 PM	03/31/2021	16080027
404	0.094		3:33 AM	04/04/2021	0.104	6:34 PM	03/31/2021	16080065
405	0.09		3:36 AM	04/04/2021	0.104	6:34 PM	03/31/2021	16080065
256	0.305		10:00 PM	04/01/2021	0.108	6:37 PM	03/31/2021	16080069
257	0.313		10:02 PM	04/01/2021	0.108	6:37 PM	03/31/2021	16080069
258	0.109		2:11 AM	04/04/2021	0.108	6:37 PM	03/31/2021	16080069
259	0.104		2:14 AM	04/04/2021	0.108	6:37 PM	03/31/2021	16080069
260	0.195		2:30 AM	04/05/2021	0.108	6:37 PM	03/31/2021	16080069
261	0.208		2:33 AM	04/05/2021	0.108	6:37 PM	03/31/2021	16080069
262	0		1:33 AM	04/09/2021	0.108	6:37 PM	03/31/2021	16080069
601	0.104		1:21 AM	04/02/2021	0.11	6:38 PM	03/31/2021	16080071
602	0.114		1:23 AM	04/02/2021	0.11	6:38 PM	03/31/2021	16080071
603	0.176		12:06 AM	04/04/2021	0.11	6:38 PM	03/31/2021	16080071
604	0.166		12:08 AM	04/04/2021	0.11	6:38 PM	03/31/2021	16080071
605	0.148		3:26 AM	04/04/2021	0.11	6:38 PM	03/31/2021	16080071
606	0.144		3:28 AM	04/04/2021	0.11	6:38 PM	03/31/2021	16080071
115	0.184		8:23 PM	04/03/2021	0.099	6:42 PM	03/31/2021	16080072
116	0.184		8:26 PM	04/03/2021	0.099	6:42 PM	03/31/2021	16080072
86	0.026		1:41 AM	03/31/2021	0.095	12:22 PM	03/21/2021	16080075
87	0		1:34 AM	04/07/2021	0.098	6:44 PM	03/31/2021	16080075
219	0		10:07 PM	04/03/2021	0.102	6:46 PM	03/31/2021	16120175
220	0		12:57 AM	04/04/2021	0.102	6:46 PM	03/31/2021	16120175
221	0		12:02 AM	04/07/2021	0.102	6:46 PM	03/31/2021	16120175
297	0		4:55 AM	03/31/2021	0.098	12:28 PM	03/21/2021	16120177

CHP SLA Usage Logs 04/09/2021

298	0	6:30 AM	04/02/2021	0.101	4:51 PM	03/31/2021	16120177
299	0	2:22 AM	04/03/2021	0.101	4:51 PM	03/31/2021	16120177
300	0	2:22 PM	04/08/2021	0.101	4:51 PM	03/31/2021	16120177
301	0.201	12:36 AM	04/09/2021	0.101	4:51 PM	03/31/2021	16120177
238	0.08	2:54 AM	04/01/2021	0.098	6:50 PM	03/31/2021	16120188
239	0.083	2:56 AM	04/01/2021	0.098	6:50 PM	03/31/2021	16120188
240	0.023	10:09 PM	04/03/2021	0.098	6:50 PM	03/31/2021	16120188
241	0.157	10:52 PM	04/03/2021	0.098	6:50 PM	03/31/2021	16120188
242	0.145	10:55 PM	04/03/2021	0.098	6:50 PM	03/31/2021	16120188
243	0.081	4:44 AM	04/09/2021	0.098	6:50 PM	03/31/2021	16120188
244	0.091	4:45 AM	04/09/2021	0.098	6:50 PM	03/31/2021	16120188















AREA NAME  
 South Los Angeles

AREA NO. 530  
 COORDINATOR NAME Edward Mc Elroy

I.D. NUMBER 16193

PAS MAKE/MODEL  
 Lifeloc FC20

PAS SERIAL NO. 16080013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1155	34.0°	191P41101	.100	22.6		.097			X		E. Mc Elroy	16193	
			Solution Prepared By:											
			Cesar Munoz											
			Date: 09-10-19											
			Expiration Date: 09-10-21											

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080014

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1205	34.0°	19IP41101	.100	23.8	.095		X			E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080015
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1205	34.0°	19IP41101	.100	22.8	.094			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf

#10

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080027

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)		LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP				TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1208	34.0°		19P41101	.100	22.3		.102			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 ID NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc PC20  
 PAS SERIAL NO.: 16080028

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1217	34.0°	19IP41101	.100	22.8		.93				X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf

AREA NAME: South Los Angeles     AREA NO.: 530     COORDINATOR NAME: Edward Mc Elroy     I.D. NUMBER: 16193     PAS MAKE/MODEL: Lifeloc FC20     PAS SERIAL NO.: 16080030

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1230	34.0°	191P41101	.100	22.3		1.02				X	E. Mc Elroy	16193	
Solution Prepared By: Cesar Munoz														
Date: 09-10-19														
Expiration Date: 09-10-21														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16080064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1222	34.0°	19IP41101	.100	22.6		.095			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 CHP202K\_1108.pdf

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME  
South Los Angeles

AREA NO. 530  
COORDINATOR NAME Edward Mc Elroy

I.D. NUMBER 16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1234	34.0°	191P41101	.100	22.6	.097			X		E. Mc Elroy	16193	

Solution Prepared By:  
Cesar Munoz

Date: 09-10-19  
Expiration Date:  
09-10-21

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#15

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
**16082069**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/09/2021	1951	34.0°	19IP41101		.100	31.8		.103		X	NO		E. Mc Elroy	16193	
			Solution Prepared By:												
			Cesar Munoz												
			Date: 09-10-19												
			Expiration Date:												
			09-10-21												

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS WAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16080071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1319	34.0°	19IP41101		.100	23.3		.103		X			E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

I.D. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1325	34.0°	191P41101	.100	22.6					X	E. Mc Elroy	16193	
Solution Prepared By: _____ Cesar Munoz Date: 09-10-19 Expiration Date: 09-10-21													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16080075

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
04/09/2021	1326	34.0°		19JP41101	.100	22.3		.092			X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Cnp202k\_1106.pdf

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifloc PC20      PAS SERIAL NO.: 16080077

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1329	34.0°	19IP41101	.100	23.1		.098			X		E. Mc Elroy	16193	
Solution Prepared By: Cesar Munoz														
Date: 09-10-19														
Expiration Date: 09-10-21														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 20

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120175

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1356	34.0°	19IP41101	.100	22.6		.097				E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
04/09/2021	1900	34.0°	191P41101		.100	23.3		.098			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chip202K\_1108.pdf

AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Edward Mc Elroy

I.D. NUMBER  
 16193

PAS MAKE/MODEL  
 Lifelog FC20

PAS SERIAL NO.  
 16120188

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
					TEST #1 (Optional)	TEST #2	TEST #1 (Optional)	TEST #2	YES	NO					
04/09/2021	1404	34.0°	19IP41101	.100		23.1		0.94			X		E. Mc Elroy	16193	
Solution Prepared By:															
Cesar Munoz															
Date: 09-10-19															
Expiration Date: 09-10-21															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf



AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FCC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	---------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1406	34.0°	19IP41101	.100		22.6		0.93			X		16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1407	34.0°		191P41101	.100	22.3							X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
 PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG  
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME **South Los Angeles** AREA NO. **530** COORDINATOR NAME **Edward Mc Elroy** I.D. NUMBER **16193** PAS MAKE/MODEL **Lifeloc FC20** PAS SERIAL NO. **16120195**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/09/2021	1409	34.0°	19IP41101	.100	22.3		.094					E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1412	34.0°	191P41101	.100	22.3	.092				X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.990 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Edward Mc Elroy    I.D. NUMBER: 16193    PAS MAKE/MODEL: Lifeloc FC20    PAS SERIAL NO.: 1170066

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1417	34.0°	191P41101	.100	21.1	.093			X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Edward Mc Elroy      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc PC20      PAS SERIAL NO.: 17170150

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1202	34.0°	19IP41101	.100	22.8	.089		X		.099	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Edward Mc Elroy  
 I.D. NUMBER: 16193  
 PAS MAKE/MODEL: Lifeloc FC20  
 PAS SERIAL NO.: 17170155

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1427	34.0°	191P41101	.100	21.8		.093		X		E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Edward Mc Elroy

ID. NUMBER  
16193

PAS MAKE/MODEL  
Lifeloc FC20

PAS SERIAL NO.  
**90102844-82**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/09/2021		34.0°	191P41101								E. Mc Elroy	16193	<b>OUT OF SERVICE</b>

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions



AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Edward Mc Elroy

I.D. NUMBER  
 16193

PAS MAKE/MODEL  
 Lifeloc FC20

PAS SERIAL NO.  
 17170064

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/09/2021	1428	34.0°	19IP41101	.100	21.6		.093			X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.