

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
288	0.055		4:57 AM	06/20/2021	0.094	8:26 AM	06/17/2021	16040140
289	0.056		4:59 AM	06/20/2021	0.094	8:26 AM	06/17/2021	16040140
444	0.142		12:13 AM	06/18/2021	0.095	8:41 AM	06/17/2021	16040161
445	0.127		12:15 AM	06/18/2021	0.095	8:41 AM	06/17/2021	16040161
446	0.095		2:09 AM	06/18/2021	0.095	8:41 AM	06/17/2021	16040161
447	0.096		2:11 AM	06/18/2021	0.095	8:41 AM	06/17/2021	16040161
448	0.132		2:53 AM	06/19/2021	0.095	8:41 AM	06/17/2021	16040161
449	0.136		2:54 AM	06/19/2021	0.095	8:41 AM	06/17/2021	16040161
450	0.152		6:36 PM	06/19/2021	0.095	8:41 AM	06/17/2021	16040161
451	0.131		6:40 PM	06/19/2021	0.095	8:41 AM	06/17/2021	16040161
452	0.145		6:43 PM	06/19/2021	0.095	8:41 AM	06/17/2021	16040161
527	0.042		3:53 PM	06/24/2021	0.095	8:20 AM	06/17/2021	16040162
528	0.045		3:55 PM	06/24/2021	0.095	8:20 AM	06/17/2021	16040162
174	0.044		10:12 PM	06/20/2021	0.094	9:46 AM	06/17/2021	16080064
425	0.139		4:42 AM	06/19/2021	0.097	9:48 AM	06/17/2021	16080065
426	0.137		4:45 AM	06/19/2021	0.097	9:48 AM	06/17/2021	16080065
427	0.098		2:48 AM	06/20/2021	0.097	9:48 AM	06/17/2021	16080065
428	0.093		2:51 AM	06/20/2021	0.097	9:48 AM	06/17/2021	16080065
429	0.176		3:25 AM	06/24/2021	0.097	9:48 AM	06/17/2021	16080065
430	0.166		3:28 AM	06/24/2021	0.097	9:48 AM	06/17/2021	16080065
299	0.107		7:22 AM	06/22/2021	0.101	8:50 AM	06/17/2021	16080069
300	0.102		7:25 AM	06/22/2021	0.101	8:50 AM	06/17/2021	16080069
301	0.185		10:48 PM	06/23/2021	0.101	8:50 AM	06/17/2021	16080069
302	0.172		10:50 PM	06/23/2021	0.101	8:50 AM	06/17/2021	16080069
627	0		12:06 AM	06/22/2021	0.101	9:51 AM	06/17/2021	16080071
112	0.156		1:34 AM	06/17/2021	0.094	4:07 PM	06/07/2021	16080075
113	0.145		1:36 AM	06/17/2021	0.094	4:07 PM	06/07/2021	16080075
186	0.222		1:33 AM	06/18/2021	0.102	10:00 AM	06/17/2021	16080077
187	0.22		1:35 AM	06/18/2021	0.102	10:00 AM	06/17/2021	16080077
188	0.099		3:06 AM	06/18/2021	0.102	10:00 AM	06/17/2021	16080077
189	0.104		3:09 AM	06/18/2021	0.102	10:00 AM	06/17/2021	16080077
234	0.202		2:22 AM	06/23/2021	0.098	10:00 AM	06/17/2021	16120175
235	0.214		2:24 AM	06/23/2021	0.098	10:00 AM	06/17/2021	16120175

CHP SLA Usage Logs 06/25/2021

257	0	12:42 AM	06/20/2021	0.096	10:07 AM	06/17/2021	16120188
258	0.118	3:45 AM	06/20/2021	0.096	10:07 AM	06/17/2021	16120188
259	0.108	3:47 AM	06/20/2021	0.096	10:07 AM	06/17/2021	16120188
243	0.105	6:47 AM	06/20/2021	0.092	10:08 AM	06/17/2021	16120189
244	0.106	6:50 AM	06/20/2021	0.092	10:08 AM	06/17/2021	16120189
499	0.09	5:03 AM	06/21/2021	0.094	10:11 AM	06/17/2021	16120195
500	0.094	5:06 AM	06/21/2021	0.094	10:11 AM	06/17/2021	16120195
195	0.195	1:26 AM	06/21/2021	0.097	10:24 AM	06/17/2021	17170064

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2021	0841	34.0°	19IP41102	.100	TEST #1 21.8	TEST #2 (Optional) .095	TEST #1	TEST #2 (Optional)	YES	NO	E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
06/25/2021	0856	34.0°	19IP41102	.100	22.1		.096				E. Mc Elroy	16193		

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16040140
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1				
06/25/2021	0849	34.0°	19IP41102		21.3	1093			X		E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16040141
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	0852	34.0°	19IP41102	.100	TEST #1 21.5	TEST #2 (Optional)	TEST #1 .097	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	0900	34.0°	19IP41102	.100	21.1	.094					E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	0858	34.0°	19IP41102	.100	TEST #1 22.6	TEST #2 (Optional)	TEST #1 .102	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#7

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL		PAS SERIAL NO.						
South Los Angeles		530	Edward Mc Elroy	16193	Lifeloc FC20		16080013						
<b>PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.</b>													
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	0952	34.0°	19IP41102	.100	21.6	116				X	E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles			AREA NO. 530		COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193		PAS MAKE/MODEL Lifeloc FC20		PAS SERIAL NO. 16080014	
--------------------------------	--	--	-----------------	--	-------------------------------------	----------------------	--	--------------------------------	--	----------------------------	--

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	0952	34.0°	19IP41102	.100	TEST #1 22.3	TEST #2 (Optional)	TEST #1 0.977	TEST #2 (Optional)	X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080015
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
		06/25/2021	0956	34.0°	19IP41102	.100	21.8	1.093						
			Solution Prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16080027
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	0958	34.0°	19IP41102	.100	22.1	.099				X	E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080028
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/25/2021	1007	34.0°	19IP41102	.100	TEST #1 20.8	TEST #2 (Optional)	TEST #1 .097	TEST #2 (Optional)			E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#12

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080030
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

### PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1007	34.0°	19IP41102		.100	20.8		103		X		E. Mc Elroy	16193	

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080064
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1009	34.0°	19IP41102	.100	2018		095			X	E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1011	34.0°	19JP41102	.100	21.1	1095			X		E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1				
06/25/2021	1014	34.0°	19IP41102		21.1	.100	100		X		E. Mc Elroy	16193	
				Solution Prepared by: Cesar Munoz									
				Date: 09-17-19									
				Expiration Date: 09-17-21									

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080072
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2021	1021	34.0°	19IP41102	.100	2018		1093				X		E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz Date: 09-17-19 Expiration Date: 09-17-21												

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles      I.D. NUMBER: 16193      PAS MAKE/MODEL: Lifeloc FC20      PAS SERIAL NO: 16080075  
 COORDINATOR NAME: Edward Mc Elroy

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
06/25/2021	1023	34.0°	19IP41102	TEST #1: 20.6	TEST #2 (Optional):	TEST #1: 1092	TEST #2 (Optional):	YES:	NO: X	E. Mc Elroy	16193	
			Solution Prepared by:									
			Cesar Munoz									
			Date: 09-17-19									
			Expiration Date:									
			09-17-21									

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#19

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080077
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1023	34.0°	19IP41102	.100	TEST #1 2018	TEST #2 (Optional)	TEST #1 .097	TEST #2 (Optional)			X	E. Mc Elroy	16193	
			Solution Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			09-17-21											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#20

CHP 202K (Rev. 8-08) OPI 051

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.				
South Los Angeles		530	Edward Mc Elroy	16193	Lifeloc FC20	16120175				
<b>PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.</b>										
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1					
06/25/2021	10:23	34.0°	19IP41102	.100	2018	1092		E. Mc Elroy	16193	
			Solution Prepared by:							
			Cesar Munoz							
			Date: 09-17-19							
			Expiration Date:							
			09-17-21							

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#21

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. <span style="color: blue; font-size: large;">16120177</span>
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	--

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1029	34.0°	19IP41102	.100	21.1		TEST #1	TEST #2				E. Mc Elroy	16193	
			Solution Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			09-17-21											

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120188
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1				
06/25/2021	1031	34.0°	19IP41102		21.6	TEST #1	TEST #2				E. Mc Elroy	16193	
			Solution Prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	-------------------------------	----------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1033	34.0°	19IP41102	.100	21.6	10.92		X			E. Mc Elroy	16193		
			Solution Prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1034	34.0°	19IP41102	.100	TEST #1 20.3	TEST #2 (Optional)	TEST #1 1097	TEST #2 (Optional)	YES	NO	X	E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021	1039	34.0°	19IP41102	.100	TEST #1 20.6	TEST #2 (Optional)	TEST #1 .092	TEST #2 (Optional)			X	E. Mc Elroy	16193	
			Solution Prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			09-17-21											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193
		PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170066

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1				
06/25/2021	1045	34.0°	19IP41102		.100	20.1	1099		X	E. Mc Elroy	16193	
				Solution Prepared by: Cesar Munoz								
				Date: 09-17-19								
				Expiration Date: 09-17-21								

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2021		34.0°	19IP41102	.100								E. Mc Elroy	16193	OUT OF SERVICE

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170155
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2021	1047	34.0°	19IP41102	TEST #1: 20.3	TEST #2 (Optional): .093				X	E. Mc Elroy	16193	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 90102844-82
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	-------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2021		34.0°	19IP41102	.100							E. Mc Elroy	16193	OUT OF SERVICE	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#31

CHP 202K (Rev. 8-08) OPT 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170064
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	OR DRY GAS**		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2021	1048	34.0°	19IP41102		.100	TEST #1 20.1	TEST #2 (Optional)	TEST #1 1095	TEST #2 (Optional)			X	E. Mc Elroy	16193	
			Solution Prepared by: Cesar Munoz												
			Date: 09-17-19												
			Expiration Date: 09-17-21												

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.