

CHP SLA PAS Usage Logs

Test Number	Result	ID	Test Time	Test Date	Check Result	Check Time	Check Date	Unit Serial Number
268	0.1		2:26 AM	08/07/2021	0.098	9:52 AM	08/05/2021	16040119
269	0.098		2:28 AM	08/07/2021	0.098	9:52 AM	08/05/2021	16040119
270	0.185		12:04 AM	08/15/2021	0.098	9:52 AM	08/05/2021	16040119
271	0.184		12:07 AM	08/15/2021	0.098	9:52 AM	08/05/2021	16040119
302	0.065		12:54 PM	08/07/2021	0.094	9:43 AM	08/05/2021	16040140
303	0.066		12:57 PM	08/07/2021	0.094	9:43 AM	08/05/2021	16040140
304	0		3:38 AM	08/08/2021	0.094	9:43 AM	08/05/2021	16040140
305	0		3:48 AM	08/10/2021	0.094	9:43 AM	08/05/2021	16040140
306	0.202		1:17 PM	08/15/2021	0.094	9:43 AM	08/05/2021	16040140
307	0.205		1:20 PM	08/15/2021	0.094	9:43 AM	08/05/2021	16040140
214	0.07		5:58 PM	08/05/2021	0.097	9:45 AM	08/05/2021	16040141
215	0.066		6:01 PM	08/05/2021	0.097	9:45 AM	08/05/2021	16040141
462	0.134		3:33 AM	08/15/2021	0.093	9:54 AM	08/05/2021	16040161
463	0.132		3:36 AM	08/15/2021	0.093	9:54 AM	08/05/2021	16040161
533	0		12:39 PM	08/14/2021	0.099	9:38 AM	08/05/2021	16040162
534	0.146		12:02 AM	08/15/2021	0.099	9:38 AM	08/05/2021	16040162
535	0.148		12:05 AM	08/15/2021	0.099	9:38 AM	08/05/2021	16040162
276	0		8:15 AM	08/08/2021	0.1	9:57 AM	08/05/2021	16080013
277	0		8:16 AM	08/08/2021	0.1	9:57 AM	08/05/2021	16080013
278	0.023		1:52 AM	08/13/2021	0.1	9:57 AM	08/05/2021	16080013
122	0		2:31 AM	08/12/2021	0.096	9:56 AM	08/05/2021	16080014
162	0.144		2:37 AM	08/07/2021	0.1	10:02 AM	08/05/2021	16080030
163	0.142		2:41 AM	08/07/2021	0.1	10:02 AM	08/05/2021	16080030
164	0.148		3:54 AM	08/07/2021	0.1	10:02 AM	08/05/2021	16080030
165	0.143		3:56 AM	08/07/2021	0.1	10:02 AM	08/05/2021	16080030
166	0.187		12:25 AM	08/12/2021	0.1	10:02 AM	08/05/2021	16080030
167	0.182		12:27 AM	08/12/2021	0.1	10:02 AM	08/05/2021	16080030
168	0.04		4:10 AM	08/15/2021	0.1	10:02 AM	08/05/2021	16080030
169	0.037		4:14 AM	08/15/2021	0.1	10:02 AM	08/05/2021	16080030
184	0.186		8:01 PM	08/07/2021	0.09	10:03 AM	08/05/2021	16080064
185	0.18		8:04 PM	08/07/2021	0.09	10:03 AM	08/05/2021	16080064
186	0.083		8:29 PM	08/07/2021	0.09	10:03 AM	08/05/2021	16080064
187	0.027		3:41 PM	08/10/2021	0.09	10:03 AM	08/05/2021	16080064

CHP SLA PAS Usage Logs

444	0.279	10:28 PM	08/08/2021	0.094	10:04 AM	08/05/2021	16080065
445	0.278	10:31 PM	08/08/2021	0.094	10:04 AM	08/05/2021	16080065
446	0.187	4:07 AM	08/14/2021	0.094	10:04 AM	08/05/2021	16080065
447	0.186	4:08 AM	08/14/2021	0.094	10:04 AM	08/05/2021	16080065
318	0	10:04 AM	08/07/2021	0.097	10:09 AM	08/05/2021	16080069
319	0	11:33 AM	08/07/2021	0.097	10:09 AM	08/05/2021	16080069
633	0	9:53 PM	08/07/2021	0.1	10:08 AM	08/05/2021	16080071
634	0.027	12:44 AM	08/08/2021	0.1	10:08 AM	08/05/2021	16080071
635	0.028	12:46 AM	08/08/2021	0.1	10:08 AM	08/05/2021	16080071
133	0.037	6:38 PM	08/13/2021	0.092	10:38 AM	08/05/2021	16080075
134	0.041	6:40 PM	08/13/2021	0.092	10:38 AM	08/05/2021	16080075
196	0.168	4:16 AM	08/08/2021	0.099	10:40 AM	08/05/2021	16080077
197	0.173	4:18 AM	08/08/2021	0.099	10:40 AM	08/05/2021	16080077
245	0.157	9:11 PM	08/13/2021	0.09	10:47 AM	08/05/2021	16120189
246	0.164	9:14 PM	08/13/2021	0.09	10:47 AM	08/05/2021	16120189
377	0.151	4:39 AM	08/08/2021	0.095	10:54 AM	08/05/2021	16120197
378	0.15	4:41 AM	08/08/2021	0.095	10:54 AM	08/05/2021	16120197
71	0	2:04 PM	08/08/2021	0.098	11:01 AM	08/05/2021	17170066
72	0	1:27 PM	08/11/2021	0.098	11:01 AM	08/05/2021	17170066
73	0	2:07 PM	08/14/2021	0.098	11:01 AM	08/05/2021	17170066

South Los Angeles

1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/15/2021	2201	34.0°	191P41102	.100	22.1	0.95					E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	2214	34.0°	19IP41102	.100	25.8	.097			X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040140
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	22:06	34.0°	19IP41102	.100	22.6	.092			X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 1604014
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	22:09	34.0°	19IP41102	.100	TEST #1 22.1	TEST #2 (Optional)	TEST #1 .099	TEST #2 (Optional)	X			E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	2113	34.0°	19IP41102	.100	21.6	21.6	.096	.096	X			E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/15/2021	2215	34.0°	19IP41102		.100	22.8		.106		X			E. Mc Elroy	16193	
			Solution prepared by:												
			Cesar Munoz												
			Date: 09-17-19												
			Expiration Date:												
			09-17-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
08/15/2021	2222	34.0°	19P41102	.100	22.3	.093		X			E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifloc FC20	PAS SERIAL NO. 16080027
--------------------------------	-----------------	-------------------------------------	----------------------	-------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/15/2021	2221	34.0°	19IP41102	.100	TEST #1 24.8	TEST #2 (Optional) 104	TEST #1	TEST #2 (Optional)		X		E. Mc Elroy	16193	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080028
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	22:39	34.0°	19IP41102	.100	23.3	095	095			X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. 16080066
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	2253	34.0°	19IP41102	.100	22.3	1097		X			E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120188
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
08/15/2021	1322	34.0°	19IP41102		.100	22.1			.093			X		E. Mc Elroy	16193	
			Solution prepared by:													
			Cesar Munoz													
			Date: 09-17-19													
			Expiration Date:													
			09-17-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/15/2021	2323	34.0°	19IP41102		23.6	.092			X		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz										
			Date: 09-17-19										
			Expiration Date: 09-17-21										

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#24

AREA NAME		AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.							
South Los Angeles		530	Edward Mc Elroy	16193	Lifeloc FC20	16120194							
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	2323	34.0°	19IP41102	.100	21.6	.098			X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)				
08/15/2021	23:26	34.0°		19IP41102	21.8	.100	096				E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions
Chp202K_1106.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Edward Mc Elroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifelog FC20 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/15/2021	2330	34.0°	19IP41102	.100	TEST #1: 21.3	TEST #2 (Optional): .095	TEST #1: .095	TEST #2 (Optional):	YES: X	NO:		E. Mc Elroy	16193	
			Solution prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 17170066
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/15/2021	2337	34.0°	19IP41102		.100	21.1		.098		X		E. Mc Elroy	16193	
			Solution prepared by:											
			Cesar Munoz											
			Date: 09-17-19											
			Expiration Date:											
			09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifelog FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/15/2021		34.0°	19IP41102	.100								E. Mc Elroy	16193	OUT of SERVICE
			Solution prepared by: Cesar Munoz											
			Date: 09-17-19											
			Expiration Date: 09-17-21											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#29

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	I.D. NUMBER 16193	PAS MAKE/MODEL LifeLoc FC20	PAS SERIAL NO. 17170155
--------------------------------	-----------------	-------------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021	2338	34.0°	19IP41102	.100	21.1		0.010		X		E. Mc Elroy	16193	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Edward Mc Elroy	ID. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 90102844-82
--------------------------------	-----------------	-------------------------------------	---------------------	--------------------------------	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/15/2021		34.0°	19P41102	.100							E. Mc Elroy	16193	OUT OF SERVICE
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration Date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

