

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
263	0		9:15 AM	06/26/2021	6:02 PM	03/31/2021	0.096	8:56 AM	06/25/2021	16040119
264	0.291		1:22 PM	06/27/2021	6:02 PM	03/31/2021	0.096	8:56 AM	06/25/2021	16040119
265	0.295		1:25 PM	06/27/2021	6:02 PM	03/31/2021	0.096	8:56 AM	06/25/2021	16040119
290	0.19		12:00 AM	07/05/2021	7:46 AM	04/12/2019	0.093	8:49 AM	06/25/2021	16040140
291	0.187		12:02 AM	07/05/2021	7:46 AM	04/12/2019	0.093	8:49 AM	06/25/2021	16040140
292	0.178		4:22 AM	07/06/2021	7:46 AM	04/12/2019	0.093	8:49 AM	06/25/2021	16040140
293	0.178		4:24 AM	07/06/2021	7:46 AM	04/12/2019	0.093	8:49 AM	06/25/2021	16040140
203	0.045		7:47 AM	06/27/2021	2:17 PM	04/03/2020	0.097	8:53 AM	06/25/2021	16040141
204	0.109		9:19 AM	07/05/2021	2:17 PM	04/03/2020	0.097	8:53 AM	06/25/2021	16040141
205	0.109		9:21 AM	07/05/2021	2:17 PM	04/03/2020	0.097	8:53 AM	06/25/2021	16040141
453	0.127		3:23 AM	07/02/2021	1:45 PM	02/03/2020	0.094	9:01 AM	06/25/2021	16040161
454	0.125		3:25 AM	07/02/2021	1:45 PM	02/03/2020	0.094	9:01 AM	06/25/2021	16040161
529	0.141		2:25 AM	07/03/2021	3:25 PM	02/03/2020	0.095	8:42 AM	06/25/2021	16040162
530	0.138		2:28 AM	07/03/2021	3:25 PM	02/03/2020	0.095	8:42 AM	06/25/2021	16040162
269	0.114		4:45 AM	07/04/2021	6:11 PM	03/31/2021	0.11	9:53 AM	06/25/2021	16080013
270	0.115		4:47 AM	07/04/2021	6:11 PM	03/31/2021	0.11	9:53 AM	06/25/2021	16080013
147	0		7:27 PM	06/27/2021	8:37 AM	01/13/2021	0.103	10:08 AM	06/25/2021	16080030
431	NEG		3:21 AM	06/29/2021	2:33 PM	03/05/2020	0.095	10:12 AM	06/25/2021	16080065
432	0		5:35 PM	07/04/2021	2:33 PM	03/05/2020	0.095	10:12 AM	06/25/2021	16080065
433	0		10:58 AM	07/05/2021	2:33 PM	03/05/2020	0.095	10:12 AM	06/25/2021	16080065
303	0.153		5:03 AM	06/26/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
304	0.155		5:04 AM	06/26/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
305	0.173		6:34 AM	06/27/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
306	0.184		6:36 AM	06/27/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
307	0		9:23 AM	06/28/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
308	0.097		4:44 AM	07/03/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
309	0.098		4:46 AM	07/03/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
310	0.18		4:20 AM	07/05/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
311	0.167		4:23 AM	07/05/2021	6:19 AM	01/22/2021	0.098	10:55 AM	06/25/2021	16080069
628	0.034		4:36 AM	07/04/2021	8:38 AM	03/03/2021	0.1	10:15 AM	06/25/2021	16080071
629	0.032		4:39 AM	07/04/2021	8:38 AM	03/03/2021	0.1	10:15 AM	06/25/2021	16080071
630	0		10:52 PM	07/05/2021	8:38 AM	03/03/2021	0.1	10:15 AM	06/25/2021	16080071
114	0.119		10:04 PM	06/26/2021	4:24 PM	03/28/2019	0.092	10:23 AM	06/25/2021	16080075

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
115	0.109		10:07 PM	06/26/2021	4:24 PM	03/28/2019	0.092	10:23 AM	06/25/2021	16080075
116	0.129		1:47 AM	06/28/2021	4:24 PM	03/28/2019	0.092	10:23 AM	06/25/2021	16080075
117	0.126		1:51 AM	06/28/2021	4:24 PM	03/28/2019	0.092	10:23 AM	06/25/2021	16080075
118	0.152		10:06 PM	07/02/2021	4:24 PM	03/28/2019	0.092	10:23 AM	06/25/2021	16080075
119	0.166		10:09 PM	07/02/2021	4:24 PM	03/28/2019	0.092	10:23 AM	06/25/2021	16080075
304	0.154		11:32 PM	07/03/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
305	0.135		11:34 PM	07/03/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
306	0		1:07 AM	07/04/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
307	0.169		1:45 AM	07/04/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
308	0.172		1:48 AM	07/04/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
309	0.086		2:33 AM	07/04/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
310	0.086		2:35 AM	07/04/2021	9:50 AM	04/25/2017	0.094	10:30 AM	06/25/2021	16120177
260	0.174		4:58 AM	06/28/2021	2:36 PM	08/06/2020	0.094	10:32 AM	06/25/2021	16120188
261	0.168		5:01 AM	06/28/2021	2:36 PM	08/06/2020	0.094	10:32 AM	06/25/2021	16120188
262	0.172		12:43 PM	06/29/2021	2:36 PM	08/06/2020	0.094	10:32 AM	06/25/2021	16120188
263	0.163		12:46 PM	06/29/2021	2:36 PM	08/06/2020	0.094	10:32 AM	06/25/2021	16120188
264	0.145		1:10 PM	06/29/2021	2:36 PM	08/06/2020	0.094	10:32 AM	06/25/2021	16120188
265	0.132		1:12 PM	06/29/2021	2:36 PM	08/06/2020	0.094	10:32 AM	06/25/2021	16120188
501	0.018		12:26 AM	06/26/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
502	0.017		12:28 AM	06/26/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
503	0.148		2:57 AM	06/26/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
504	0.15		2:59 AM	06/26/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
505	0.143		4:02 AM	06/28/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
506	0.138		4:04 AM	06/28/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
507	0.1		4:38 AM	07/03/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
508	0.101		4:41 AM	07/03/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
509	0.052		2:49 AM	07/04/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
510	0.206		10:44 PM	07/04/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
511	0.208		10:47 PM	07/04/2021	12:48 PM	06/09/2019	0.105	12:22 PM	06/25/2021	16120195
196	0.166		3:41 AM	06/28/2021	7:08 PM	03/31/2021	0.095	10:48 AM	06/25/2021	17170064
197	0.156		3:45 AM	06/28/2021	7:08 PM	03/31/2021	0.095	10:48 AM	06/25/2021	17170064
198	0.179		10:32 PM	07/04/2021	7:08 PM	03/31/2021	0.095	10:48 AM	06/25/2021	17170064
199	0.19		10:34 PM	07/04/2021	7:08 PM	03/31/2021	0.095	10:48 AM	06/25/2021	17170064

PAS USage Logs CHP SLA 7/6/21

Test Number	Result	ID	Test Time	Test Date	Cal Time	Cal Date	Check Result	Check Time	Check Date	Unit Serial Number
200	0.158		11:36 PM	07/04/2021	7:08 PM	03/31/2021	0.095	10:48 AM	06/25/2021	17170064
201	0.158		11:38 PM	07/04/2021	7:08 PM	03/31/2021	0.095	10:48 AM	06/25/2021	17170064

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Liteloc FC20	PAS SERIAL NO. 16040119
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1415	34	191P41102	.100	22		0.099					M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: E. McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1408	34	191P41102		.100	22		.094			X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifenoc FC20 PAS SERIAL NO.: 16040141

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1411	34	191P41102	.100	20		.096			X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1420	34	191P41102	.100	22		.098					M. Tippet	20137	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions Chp202K_1108.pdf

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1425	34	191P41102	.100	23		106		X			M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021	1532	34	191P41102	.100	33						M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#8

CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080014

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1	TEMP #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2021	144934	191P41102		.100	23				X		M. Tippet	20137	
			Solution prepared by:										
			Cesar Munoz										
			Date: 09-17-19										
			Expiration date:										
			09-17-21										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: E. McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Liteloc FC20 | PAS SERIAL NO.: 16080015

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021	1445	34	191P41102	.100	22		.097				M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080027
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	144734		191P41102	.100	27		108			X		M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 09-17-19
 Expiration date:
 09-17-21

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Liteloc FC20	PAS SERIAL NO. 16080028
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	-----------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1450	34	19IP41102	.100	22		104				X	M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Liteloc FC20 PAS SERIAL NO.: 16080030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1450	34	191P41102	.100	21		106					M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Liteloc FC20 PAS SERIAL NO.: 16080064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021			191P41102	.100							M. Tippett	20137	DEVICE WTS NOT IN WORK FOR TESTING. THE DEVICE IS OUT OF SERVICE UNTIL IT CAN BE CHECKED.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/06/2021	1454	34	191P41102	.100	21		.099				X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#15

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: E. McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 16080069

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1459	34	191P41102	.100	23		103			X		M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 09-17-19
 Expiration date:
 09-17-21

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021	1459	34	191P41102	.100	22		106				M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

#117

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1506	34	191P41102	.100	21		098			X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k_1108.pdf

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16080075
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/06/2021	1507	34	191P41102	.100	22		094				X		M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions
Chp202K_1108.pdf

#19

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: E. McElroy
I.D. NUMBER: 16193
PAS MAKE/MODEL: Liteloc FC20
PAS SERIAL NO.: 16080077

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1507	34	191P41102	.100	22		102			X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. <i>16120175</i>
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	-----------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021	<i>151234</i>		191P41102	.100	<i>22</i>		<i>093</i>				M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120177
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021	1/5/18	34	191P41102	.100	22		.094				M. Tippet	20137	

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	E. McElroy	16193	Lifeloc FC20	16/20188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1520	34		191P41102	.100	22		.093			X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1521	34	191P41102	.100	22		093		X			M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2021	1521	24	191P41102	.100	21	.097			X		M. Tippet	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

CHP202K_1108.pdf

25

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	E. McElroy	16193	Lifeloc FC20	16/20195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1545	34	19IP41102		.100	21		.092		X	NO		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	151	34	19P41102	.100	21		093			X		M. Tippet	20137	

Solution prepared by:
Cesar Munoz
Date: 09-17-19
Expiration date:
09-17-21

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

#27

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME E. McElroy I.D. NUMBER 16193 PAS MAKE/MODEL Lifeloc FC20 PAS SERIAL NO. 17170066

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1557	34	19IP41102	.100	20		105			X		M. Tippet	20137	

Solution prepared by:
 Cesar Munoz
 Date: 09-17-19
 Expiration date:
 09-17-21

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME E. McElroy	I.D. NUMBER 16193	PAS MAKE/MODEL Lifoloc FC20	PAS SERIAL NO. 17170150
--------------------------------	-----------------	--------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021		19	P41102	.100								M. Tippett	20137	DEVICE IS OUT OF SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#29

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: E. McElroy
I.D. NUMBER: 16193
PAS MAKE/MODEL: Lifelog FC20
PAS SERIAL NO.: 17170155

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1559	34	191P41102		.100	20		0.096					M. Tippet	20137	
			Solution prepared by:												
			Cesar Munoz												
			Date: 09-17-19												
			Expiration date:												
			09-17-21												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: E. McElroy | I.D. NUMBER: 16193 | PAS MAKE/MODEL: Lifeloc FC20 | PAS SERIAL NO.: 90102844.82

#30

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2021		191P41102									M. Tippet	20137	DEVICE IS OUT OF SERVICE
Solution prepared by: Cesar Munoz Date: 09-17-19 Expiration date: 09-17-21													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

#31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: E. McElroy I.D. NUMBER: 16193 PAS MAKE/MODEL: Lifeloc FC20 PAS SERIAL NO.: 17170064

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2021	1558 /558	34	191P41102	20		094			X		M. Tippett	20137	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.