

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/05/2021 11:11

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3049	
Check sum	15981	Download date/time:	09/02/2021	13:36		Record ID	BO28	
Record imported (Y/N)				N	Source Diagnostic			

Instrument serial No: 68-012337 Accuracy check date/time: 08/05/2021 11:13

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V) .131

Solution lot#: IRS0122920

Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	34092	Download date/time:	09/02/2021	13:36		Record ID	BNP7	
Record imported (Y/N)				N	Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/05/2021 11:15

First calibration check: .134 Interferent detected: No
Second calibration check: .135 Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS0122920 Soln temp C: 34.02
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 58479	Download date/time: 09/02/2021	13:36	Record ID BNQ8
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012337 Accuracy check date/time: 08/11/2021 11:07

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3050
Check sum 77260	Download date/time: 09/02/2021	13:36	Record ID BO33
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/11/2021 11:08

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS0122920 Soln temp C: 34.02
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 44639	Download date/time: 09/02/2021	13:36	Record ID BNR3
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 08/11/2021 11:10

First calibration check: .132 Interferent detected: No
Second calibration check: .136 Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS0122920 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 10579	Download date/time: 09/02/2021	13:36	Record ID BNS4
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/11/2021 11:11

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3045
Check sum	72660	Download date/time:	09/02/2021 13:36	Record ID	BO48		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012337 Accuracy check date/time: 08/19/2021 09:50

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3047
Check sum	44865	Download date/time:	09/02/2021 13:36	Record ID	BO53		
Record imported (Y/N) N				Source Diagnostic			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/19/2021 09:51

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS0122920 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57180	Download date/time: 09/02/2021	13:36	Record ID BNT5
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 08/19/2021 09:53

First calibration check: .132 Interferent detected: No
Second calibration check: .133 Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS0122920 Soln temp C: 34.00
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 43392	Download date/time: 09/02/2021	13:36	Record ID BNU6
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/26/2021 15:48

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3032	
Check sum	33747	Download date/time:	09/02/2021	13:36		Record ID	BO60	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012337 Accuracy check date/time: 08/26/2021 15:50

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .131

Solution lot#: IRS0122920

Soln temp C: 34.00

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	59434	Download date/time:	09/02/2021	13:36		Record ID	BNV1	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 08/26/2021 15:52

First calibration check: .133 Interferent detected: No
Second calibration check: .135 Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS0122920 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 58747	Download date/time: 09/02/2021	13:36	Record ID BNW2
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.