

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/02/2021 09:11

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2552
Check sum	26106	Download date/time:	12/02/2021	10:17		Record ID	BSD1
Record imported (Y/N)				N	Source Diagnostic		

Instrument serial No: 68-012323 Accuracy check date/time: 12/02/2021 09:13

First calibration check: .STD
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 011

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	10195	Download date/time:	12/02/2021	10:17		Record ID	BS73
Record imported (Y/N)				N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/02/2021 09:13

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 15795	Download date/time: 12/02/2021	10:17	Record ID BSJ7
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012323 Accuracy check date/time: 12/02/2021 09:18

First calibration check: .130 Interferent detected: No
Second calibration check: .130 Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS122920 Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 011

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 31639	Download date/time: 12/02/2021	10:17	Record ID BS88
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/09/2021 08:19

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2550	
Check sum	36312	Download date/time:	12/10/2021	08:47		Record ID	BSQ7	
Record imported (Y/N)				N	Source Diagnostic			

Instrument serial No: 68-012323 Accuracy check date/time: 12/09/2021 08:21

First calibration check: .STD Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	19947	Download date/time:	12/10/2021	08:47		Record ID	BSO6	
Record imported (Y/N)				N	Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/09/2021 08:21

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 31605	Download date/time: 12/10/2021	08:47	Record ID BSR4
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012323 Accuracy check date/time: 12/09/2021 08:31

First calibration check: .127

Interferent detected: No

Second calibration check: .126

Interferent detected: No

Concentration (%W/V) .131

Solution lot#: IRS122920

Soln temp C: 34.00

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 014

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 30702	Download date/time: 12/10/2021	08:47	Record ID BSP8
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/16/2021 07:18

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2533
Check sum	59878	Download date/time:	12/23/2021	08:11		Record ID	BT67
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012323 Accuracy check date/time: 12/16/2021 07:20

First calibration check: .STD
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 004

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	49326	Download date/time:	12/23/2021	08:11		Record ID	BT21
Record imported (Y/N) N				Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/16/2021 07:20

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 61880	Download date/time: 12/23/2021	08:11	Record ID BT85
Record imported (Y/N) N	Source Exception		

Instrument serial No: 68-012323 Accuracy check date/time: 12/16/2021 07:24

First calibration check: .129 Interferent detected: No
Second calibration check: .129 Interferent detected: No

Concentration (%W/V) .131 Solution lot#: IRS122920 Soln temp C: 33.99
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 004

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 56645	Download date/time: 12/23/2021	08:11	Record ID BT32
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/23/2021 07:08

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2546
Check sum	48996	Download date/time:	12/23/2021	08:11		Record ID	BT78
Record imported (Y/N)				N	Source Diagnostic		

Instrument serial No: 68-012323 Accuracy check date/time: 12/23/2021 07:10

First calibration check: .STD
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 010

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	59120	Download date/time:	12/23/2021	08:11		Record ID	BT45
Record imported (Y/N)				N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/23/2021 07:10

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 39138	Download date/time: 12/23/2021	08:11	Record ID BT96
Record imported (Y/N) N	Source Exception		

Instrument serial No: 68-012323 Accuracy check date/time: 12/23/2021 07:14

First calibration check: .134

Interferent detected: No

Second calibration check: .134

Interferent detected: No

Concentration (%W/V) .140

Solution lot#: IRS113021

Soln temp C: 34.04

Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 010

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 56399	Download date/time: 12/23/2021	08:11	Record ID BT56
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/30/2021 08:35

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2537	
Check sum	39754	Download date/time:	01/04/2022	11:39		Record ID	BTC1	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012323 Accuracy check date/time: 12/30/2021 08:37

First calibration check: .STD
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 005

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	53053	Download date/time:	01/04/2022	11:39		Record ID	BTA1	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 12/30/2021 08:37

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 33001	Download date/time: 01/04/2022	11:39	Record ID BTD0
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012323 Accuracy check date/time: 12/30/2021 08:41

First calibration check: .132
Second calibration check: .132

Interferent detected: No
Interferent detected: No

Concentration (%W/V) .140

Solution lot#: IRS113021

Soln temp C: 34.01
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 005

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 56482	Download date/time: 01/04/2022	11:39	Record ID BTB6
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.