

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/03/2022 10:31

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:  
Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass  
Prom Pass Real Time clock Pass Motor Pass Printer Pass  
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 2974  
Check sum 13698 Download date/time: 02/03/2022 11:37 Record ID BVC1  
Record imported (Y/N) N Source Diagnostic

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/03/2022 10:34

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No  
Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.02  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom Serial No. Temp Internal standards  
Prom Real Time clock Motor Printer  
Ram Auto calibration Range Digital voltmeter  
Check sum 51395 Download date/time: 02/03/2022 11:37 Record ID BV72  
Record imported (Y/N) N Source Calibration

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/03/2022 10:37

First calibration check: .139 Interferent detected: No  
Second calibration check: .140 Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.00  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51433	Download date/time: 02/03/2022	11:37	Record ID BV87
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/10/2022 06:36

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 2980
Check sum 18039	Download date/time: 03/10/2022	10:40	Record ID BWJ9
Record imported (Y/N) N	Source Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/10/2022 06:37

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.00  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51287	Download date/time: 03/10/2022	10:40	Record ID BWA2
Record imported (Y/N) N		Source Calibration	

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/10/2022 06:39

First calibration check: .136 Interferent detected: No  
Second calibration check: .139 Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.02  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 30812	Download date/time: 03/10/2022	10:40	Record ID BWB5
Record imported (Y/N) N		Source Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/17/2022 10:32

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2919
Check sum	37390	Download date/time:	03/10/2022	10:40		Record ID	BWK0
	Record imported (Y/N)	N		Source	Diagnostic		

Instrument serial No: 68-012337 Accuracy check date/time: 02/17/2022 10:34

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2917
Check sum	38791	Download date/time:	03/10/2022	10:40		Record ID	BWL7
	Record imported (Y/N)	N		Source	Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/17/2022 10:37

First calibration check: .141 Interferent detected: No  
Second calibration check: .141 Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.00  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator T TYLER 11036

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 33696	Download date/time: 03/10/2022	10:40	Record ID BWC6
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/24/2022 14:33

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 2957
Check sum 11901	Download date/time: 03/10/2022	10:40	Record ID BWM8
Record imported (Y/N) N	Source Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/24/2022 14:35

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.00  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator T TYLER 11036

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51535	Download date/time: 03/10/2022	10:40	Record ID BWD3
Record imported (Y/N) N		Source Calibration	

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/24/2022 14:37

First calibration check: .138 Interferent detected: No  
Second calibration check: .140 Interferent detected: No

Concentration (%W/V) .140 Solution lot#: IRS113021 Soln temp C: 34.00  
Simulator Ser No. MP5021

Number of people tested since last periodic determination of accuracy: 000

Operator T TYLER 11036

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57648	Download date/time: 03/10/2022	10:40	Record ID BWE4
Record imported (Y/N) N		Source Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.