

INTOXILYZER 8000
Instrument Initialization
19:28 08/27/2021


INTOXILYZER 8000
Instrument Initialization
12:13 09/02/2021

Long Beach Police Department
DUI CHECKPOINT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000434
09/02/2021 13:04

Sim Conc (%W/U) = .131
Sim Temp (C) = 34.00
Lot = IRS122920
Soln Exp Date = 12/29/2021
Sim Serial # = MP5021
Name/Photo Num =
G GOSSAGE 7885
Remarks =

Sub Tests Since Last Cal Check = 000

Test	%BAC	Time
Air Blank	0.000	13:06
Cal Check	0.136	13:06
Air Blank	0.000	13:07
Cal Check	0.137	13:08
Air Blank	0.000	13:08



Operator's Signature